

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:39

Date of Inspection: 12/16/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2402328118 Exp: 01/20/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.078 | 0.198 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.199 | 0.080 |
| 0.000 | 0.050 | 0.079 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

12/16/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 20:16

Date of Inspection: 11/22/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2402328118 Exp: 01/20/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.077 | 0.191 | 0.080 |
| 0.000 | 0.050 | 0.078 | 0.197 | 0.079 |
| 0.000 | 0.050 | 0.078 | 0.197 | 0.079 |

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Eric D Abraham #1727/5028 _____ ERIC D ABRAHAM
Signature and Printed Name

11/22/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:36


Date of Inspection: 10/23/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2402328118 Exp: 01/20/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.046 | 0.078 | 0.188 / 0.195 | 0.080 |
| 0.000 | 0.048 | 0.079 | 0.196 / 0.197 | 0.081 |
| 0.000 | 0.049 | 0.080 | 0.199 / 0.197 | 0.080 |

Number of Simulators Used: 5

Remarks:

 20: Control Outside TEST WAS REPEATED AFTER SIM WAS AT TEMP

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 #1722

ERIC D ABRAHAM

Signature and Printed Name

10/23/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:47

Date of Inspection: 09/27/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2402328118 Exp: 01/20/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.048 | 0.077 | 0.194 | 0.080 |
| 0.000 | 0.049 | 0.078 | 0.197 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.197 | 0.080 |

Number of Simulators Used: 5

Remarks: N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

NEIL F RUBEL

Signature and Printed Name

09/27/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 10:24

Date of Inspection: 08/13/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:2402328118 Exp: 01/20/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.080 |

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

NEIL F RUBEL

08/13/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:00

Date of Inspection: 06/20/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.078 | 0.200 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name NICHOLAS M LITTLE

06/20/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:57

Date of Inspection: 05/01/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.050 | 0.082 | 0.204 | 0.081 |
| 0.000 | 0.050 | 0.083 | 0.204 | 0.081 |
| 0.000 | 0.050 | 0.082 | 0.203 | 0.081 |


Number of Simulators Used: 5

Remarks:

N/A 

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NEIL F RUBEL

Signature and Printed Name

05/01/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:47

Date of Inspection: 04/15/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | Yes | |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | | No |
| Interferent Detect Test: Interferent Detect | | No |
| Diagnostic Check (Post-Inspection): OK | | No |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.20g/210L Test (g/210L) Lot#: _____ Exp: _____ | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____ |
|----------------------------|---|---|---|--|
| | | | | |
| | | | | |
| | | | | |

Number of Simulators Used: 5

Remarks:

Time-Date changed. A F / M A: . NO MOUTH ALCOHOL.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

NICHOLAS M LITTLE

04/15/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:27

Date of Inspection: 04/15/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.079 | 0.197 | 0.081 |
| 0.000 | 0.049 | 0.080 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.200 | 0.081 |

Number of Simulators Used: 5

Remarks:
N/A

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NICHOLAS M LITTLE

Signature and Printed Name

04/15/2023
Date

After Annual Insp

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 13:10

Date of Inspection: 01/12/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023 | 0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.082 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SUSAN M BARGE

Signature and Printed Name

01/12/2023
Date

Before Annual Insp

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:47

Date of Inspection: 01/04/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023 | 0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.048 | 0.079 | 0.198 | 0.080 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.080 |

Number of Simulators Used: 5


Remarks:

N/A



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

NEIL F RUBEL

01/04/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:04

Date of Inspection: 02/15/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |


| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023 | 0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.048 | 0.079 | 0.198 | 0.081 |
| 0.000 | 0.049 | 0.079 | 0.199 | 0.081 |
| 0.000 | 0.048 | 0.078 | 0.199 | 0.080 |

Number of Simulators Used: 5

Remarks:
NA

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name SUSAN M BARGE

02/15/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 09:31


Date of Inspection: 03/06/2023

Serial Number: 80-001111
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1482048 Exp: 04/21/2024 |
|----------------------------|---|---|---|--|
| 0.000 | 0.046 | 0.077 | 0.198 | 0.079 |
| 0.000 | 0.047 | 0.077 | 0.197 | 0.079 |
| 0.000 | 0.047 | 0.077 | 0.198 | 0.080 |

Number of Simulators Used: 5

Remarks:
N/A 

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

NEIL F RUBEL

03/06/2023
Date