

Alcohol Testing Program - Instrument Processing Sheet

Agency: Largo PD

Instrument Serial Number: 80-000791

Date In: 6/26/2026

DI Completion Date: 7/22/26

☒ Ship

☐ P/U

☐ H/D

☒ CMI

☐ EE

Intake By: <u>IAS</u> Date: <u>6/26/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>7/14/2026</u>		Flow Adjustment By: <u>KTS</u>																															
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Came with 2 boxes. IAS 6/26/26		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>95</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.101</u> (.139-.169) 36 mm <u>0.109</u> (.156-.190) 53 mm <u>0.175</u> (.228-.278) 103 mm <u>0.464</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1017</u> ☒ Stability Checks		Flow Column #: <u>ATP105</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>96</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.156</u> (.156-.190) 53 mm <u>0.230</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: Date:</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td>☐ Other:</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td></td> </tr> <tr> <td>10/5/2027</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: Date:	0.050	MP5088	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP5089	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP5090	202603D	☐ Other:	3/4/2028		0.080 DGS	N/A	23825080A1		10/5/2027	
						Simulator	Serial #	Lot#/Exp	Maintenance By: Date:																										
						0.050	MP5088	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP5089	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP5090	202603D	☐ Other:																																
		3/4/2028																																	
0.080 DGS	N/A	23825080A1																																	
		10/5/2027																																	
Optical Bench Adjustment By: <u>WKP</u>		Department Inspection By: <u>KTS</u>																																	
Barometric Pressure Gauge: <u>1010</u> ID#: <u>28662</u>		Barometric Pressure ID#: <u>28427</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1009</u> Instrument: <u>1010</u>																															
0.000	MP6294	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u>																															
0.040	MP6295	25090	3/11/2027	Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																															
0.100	MP6296	24390	10/29/2026	Simulator	Serial Number																														
0.200	MP6297	25020	1/14/2027	0.000	MP5086																														
0.300	MP6298	24430	12/10/2026	Interferent	MP5087																														
0.080 DGS	N/A	AG526603	9/23/2027	0.050	MP5088																														
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																														
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																														
0.050	MP5088	202603A	3/3/2028	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☒ Flow Adjustment ☐ Form 40 ☒ Other: Return Material Authorization																															
0.080	MP5089	202603B	3/3/2028																																
0.200	MP5090	202603D	3/4/2028																																
0.080 DGS	N/A	23825080A1	10/5/2027																																
Gauge ID #: <u>34420</u> Gauge: <u>1010</u> Instrument: <u>1009</u>																																			
Notes/Suggested Service: Stability Checks conducted on 7/15/26. KTS Tech Review: added checkmark for CMI. KTS 8/3/26				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.08.03 12:01:47 -04'00'				Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.04 10:14:19 -04'00'																															
Tech Review				Admin Review																															

Flow Cal. Adj.

80-000791

~~Katz~~ 7/14/26

LARCO PD
Intoxilyzer - Alcohol Analyzer
Model 8100 SN 80-000791
17/14/2026
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 3.871

2: Rate (Liters/min) = 15
SQRT(Diff) = 6.363

3: Rate (Liters/min) = 30
SQRT(Diff) = 18.785

Dependent Data Scale Factor = 100000 L/min



Independent Data Scale Factor = 256

Rounded Slope = 639

Rounded Intercept = -23454

Correlation = 0.99388

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																
LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	12:58	Control Test	0.049	12:58	Air Blank	0.000	12:59	Control Test	0.050	12:59	Air Blank	0.000	13:00	Control Test	0.049	13:01	Air Blank	0.000	13:01	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0763</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:03	Control Test	0.077	13:03	Air Blank	0.000	13:04	Control Test	0.078	13:05	Air Blank	0.000	13:05	Control Test	0.076	13:06	Air Blank	0.000	13:06	Control Test Stats			Average	0.0763		Std Dev	0.0006		Rel Std Dev(%)	0.7564		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.193</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.193</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.193</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1930</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	13:08	Control Test	0.193	13:09	Air Blank	0.000	13:10	Control Test	0.193	13:10	Air Blank	0.000	13:11	Control Test	0.193	13:12	Air Blank	0.000	13:12	Control Test Stats			Average	0.1930		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/15/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: 	Test	g/210L	Time	Air Blank	0.000	12:54	Control Test	0.080	12:54	Air Blank	0.000	12:55	Control Test	0.080	12:55	Air Blank	0.000	12:55	Control Test	0.080	12:56	Air Blank	0.000	12:56	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.049	12:58																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.050	12:59																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test	0.049	13:01																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.077	13:03																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.078	13:05																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.076	13:06																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test	0.193	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.193	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.193	13:12																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1930																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.080	12:54																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.080	12:55																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.080	12:56																																																																																																																																																	
Air Blank	0.000	12:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

LEAD RD
Intoxilyzer - Alcohol Analyzer
Model 8000
07/21/2026

WKP

12:17:16

Auto Calibration
Max Power Res Value = 33
Auto Range Res value = 22

Sol Value = 0.100 g/210L ***
Fit Value = 0.1000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12598, Sum Io = 14088
***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1140	(-0.0270)
Sample #2 =	0.1360	(0.0010)
Sample #3 =	0.1690	(-0.0140)
Sample #4 =	0.1610	(0.0020)
Avg % Abs =	0.1620	(-0.0037)
STD DEV =	0.0166	(0.0090)
REL STD DEV =	10.577	(244.442)

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.1240	(-0.0190)
Sample #2 =	0.0940	(0.0000)
Sample #3 =	0.1000	(-0.0010)
Sample #4 =	0.1030	(-0.0110)
Avg % Abs =	0.0990	(-0.0040)
STD DEV =	0.0046	(0.0061)
REL STD DEV =	4.629	(152.069)

Sol Value = 0.140 g/210L ***
Fit Value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12594, Sum Io = 14089

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	0.8420	(-0.0040)
Sample #2 =	0.8530	(0.0120)
Sample #3 =	0.8440	(-0.0180)
Sample #4 =	0.8140	(0.0120)
Avg % Abs =	0.8370	(0.0020)
STD DEV =	0.0224	(0.0173)
REL STD DEV =	2.440	(866.025)

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.5210	(-0.0030)
Sample #2 =	1.5120	(0.0140)
Sample #3 =	1.5110	(0.0070)
Sample #4 =	1.4960	(0.0150)
Avg % Abs =	1.5060	(0.0120)
STD DEV =	0.0190	(0.0044)
REL STD DEV =	0.595	(36.324)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12590, Sum Io = 14088

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	1.9900	(-0.0290)
Sample #2 =	1.9860	(-0.0150)
Sample #3 =	1.9720	(0.0130)
Sample #4 =	1.9910	(-0.0090)
Avg % Abs =	1.9830	(-0.0037)
STD DEV =	0.0198	(0.0147)
REL STD DEV =	0.497	(402.061)

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.5720	(-0.0090)
Sample #2 =	3.5620	(0.0020)
Sample #3 =	3.5540	(0.0180)
Sample #4 =	3.5660	(0.0130)
Avg % Abs =	3.5660	(0.0110)
STD DEV =	0.0140	(0.0180)
REL STD DEV =	0.392	(74.412)

Sol Value = 0.200 g/210L ***
Fit Value = 1.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12599, Sum Io = 14086

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.8150	(-0.0110)
Sample #2 =	3.8110	(-0.0300)
Sample #3 =	3.8150	(-0.0030)
Sample #4 =	3.8060	(-0.0250)
Avg % Abs =	3.8107	(-0.0193)
STD DEV =	0.0045	(0.0144)
REL STD DEV =	0.118	(74.266)

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	6.8480	(-0.0150)
Sample #2 =	6.8430	(-0.0070)
Sample #3 =	6.8290	(0.0160)
Sample #4 =	6.8410	(0.0020)
Avg % Abs =	6.8377	(0.0037)
STD DEV =	0.0076	(0.0116)
REL STD DEV =	0.111	(316.097)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12593, Sum Io = 14085

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	5.6280	(-0.0300)
Sample #2 =	5.5970	(0.0270)
Sample #3 =	5.5690	(0.0370)
Sample #4 =	5.5940	(0.0360)
Avg % Abs =	5.5867	(0.0333)
STD DEV =	0.1154	(0.0055)
REL STD DEV =	0.275	(16.523)

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	9.9970	(-0.0270)
Sample #2 =	9.9770	(0.0190)
Sample #3 =	9.9450	(0.0340)
Sample #4 =	9.9690	(0.0340)
Avg % Abs =	9.9637	(0.0290)
STD DEV =	0.0167	(0.0187)
REL STD DEV =	0.167	(29.863)

Sol Value = 0.200 g/210L ***
Fit Value = 1.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12599, Sum Io = 14086

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 =	3.8150	(-0.0110)
Sample #2 =	3.8110	(-0.0300)
Sample #3 =	3.8150	(-0.0030)
Sample #4 =	3.8060	(-0.0250)
Avg % Abs =	3.8107	(-0.0193)
STD DEV =	0.0045	(0.0144)
REL STD DEV =	0.118	(74.266)

***** AUTO CAL DATA *****

***** CHANNEL 1 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.062
Std Dev = 0.01 Rel Std Dev = 10.58
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.937
Std Dev = 0.02 Rel Std Dev = 12.44
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.983
Std Dev = 0.01 Rel Std Dev = 0.51
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.811
Std Dev = 0.00 Rel Std Dev = 0.12
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.587
Std Dev = 0.02 Rel Std Dev = 0.28
Zero Order Coef = -152.73
First Order Coef = 2430.58
Second Order Coef = 27.67
Standard Deviation = 9.858246

***** CHANNEL 2 *****

Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.099
Std Dev = 0.00 Rel Std Dev = 4.63
Sol Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 1.506
Std Dev = 0.01 Rel Std Dev = 0.60
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.069
Std Dev = 0.01 Rel Std Dev = 0.39
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.838
Std Dev = 0.01 Rel Std Dev = 0.11
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.964
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -132.13
First Order Coef = 1332.70
Second Order Coef = 11.49
Standard Deviation = 5.108758

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	0.0001
0.100	0.100	-0.0003
0.200	0.200	0.0003
0.300	0.300	-0.0001

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	0.0001
0.100	0.100	-0.0002
0.200	0.200	0.0001
0.300	0.300	-0.0000

Sol Value = 0.060 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****

Sample #1 =	3166.00
Sample #2 =	3164.00
Sample #3 =	3169.00
Sample #4 =	3202.00
Average Result =	3191.6667
STD DEV =	9.2916
REL STD DEV =	0.291

***** CHANNEL 2 *****

Sample #1 =	3345.00
Sample #2 =	3347.00
Sample #3 =	3339.00
Sample #4 =	3347.00
Average Result =	3344.3333
STD DEV =	4.6168
REL STD DEV =	0.138

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1019
3 in H2O Adjust (mg/l*10,000) = 618
9 in H2O Adjust (mg/l*10,000) = 465
***** AUTO CAL PASS *****

80-000791
Optical Bench
calibration
adjustment
WKP 7/21/2026

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053 ✓	0.08g/210L 0.077 to 0.083 ✓	0.20g/210L 0.194 to 0.206 ✓	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:09	Control Test	0.049	14:10	Air Blank	0.000	14:10	Control Test	0.050	14:11	Air Blank	0.000	14:11	Control Test	0.049	14:12	Air Blank	0.000	14:13	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:14	Control Test	0.078	14:15	Air Blank	0.000	14:15	Control Test	0.078	14:16	Air Blank	0.000	14:17	Control Test	0.078	14:17	Air Blank	0.000	14:18	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	14:26	Control Test	0.198	14:26	Air Blank	0.000	14:27	Control Test	0.198	14:28	Air Blank	0.000	14:28	Control Test	0.198	14:29	Air Blank	0.000	14:29	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LARGO PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000791 07/21/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>menpi</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:17	Control Test	0.081	13:18	Air Blank	0.000	13:18	Control Test	0.081	13:18	Air Blank	0.000	13:19	Control Test	0.081	13:19	Air Blank	0.000	13:20	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.049	14:10																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test	0.050	14:11																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.049	14:12																																																																																																																																																	
Air Blank	0.000	14:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test	0.078	14:15																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.078	14:16																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.078	14:17																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:26																																																																																																																																																	
Control Test	0.198	14:26																																																																																																																																																	
Air Blank	0.000	14:27																																																																																																																																																	
Control Test	0.198	14:28																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test	0.198	14:29																																																																																																																																																	
Air Blank	0.000	14:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.081	13:18																																																																																																																																																	
Air Blank	0.000	13:18																																																																																																																																																	
Control Test	0.081	13:18																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.081	13:19																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0810																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LARGO PD
Time of Inspection: 11:57

Date of Inspection: 07/22/2026

Serial Number: 80-000791
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.198	0.082
0.000	0.050	0.078	0.198	0.082
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.199	0.082
0.000	0.050	0.078	0.199	0.081
0.000	0.049	0.079	0.199	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.078	0.198	0.081
0.000	0.050	0.079	0.198	0.081
Standard Deviations	0.0004	0.0005	0.0004	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 _____
Signature and Printed Name

KATIE T SPEARIN

07/22/2026
Date

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Matthew Gosier on 7/14/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-000791

Bill To Address:

Matthew Gosier
Largo PD

201 North Highland Avenue, Largo, FL 33770

Ship to Address:

FL Dept of Law Enforcement

813 B Lake Bradford Rd, Tallahassee, FL 32304

Reason for Return:

Flow sensor R-value is 95.

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Matthew Gosier Phone #: 315-427-1173Email: mgosier@largo.com

ATP Contact Name: Katie Spearin

ATP Email: KaitlynSpearin@FDLE.State.FL.US



ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000791, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number: 80-000791
Owning Agency: LARGO PD
Calibration Date: 07/22/2026
Calibration Time: 11:57

UNCERTAINTY* ±	
0.050 g/210 L	0.004
0.080 g/210 L	0.004
0.200 g/210 L	0.008
0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Date

07/22/2026


KATIE T SPEARIN,
Department Inspector