

Alcohol Testing Program - Instrument Processing Sheet

Agency: FFWCC

Instrument Serial Number: 80-007168

Date In: 6/16/2026

DI Completion Date: 6/30/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>6/23/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/24/2026</u>		Flow Adjustment By: _____																				
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☒ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: DGS cylinder.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>230</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1019</u> Instrument: <u>1016</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																				
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A 3/3/2028	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202603D 3/4/2028	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: <u>TDG</u> Date: <u>6/23/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Added screw to printer port.	
						Simulator	Serial #	Lot#/Exp																
						0.050	MP6286	202603A 3/3/2028																
						0.080	MP6287	202603B 3/3/2028																
0.200	MP6288	202603D 3/4/2028																						
0.080 DGS	N/A	AG526603 9/23/2027																						
Optical Bench Adjustment By: <u>TDG</u>		Department Inspection By: <u>TDG</u>																						
Barometric Pressure Gauge: <u>1019</u> ID#: <u>33364</u>		Barometric Pressure ID#: <u>33363</u>																						
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1017</u>	Instrument: <u>1018</u>																			
0.000	MP5097	N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/2027</u>																			
0.040	MP5098	25090	3/11/2027	Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>9/22/2027</u>																			
0.100	MP5099	24390	10/29/2026	Simulator	Serial Number																			
0.200	MP5100	25020	1/14/2027	0.000	MP6284																			
0.300	MP5101	24430	12/10/2026	Interferent	MP6285																			
0.080 DGS	N/A	28424080A3	11/5/2026	0.050	MP6286																			
☒ Post Optical Bench Adjustment Stability Checks				0.080	MP6287																			
Simulator	Serial #	Lot #	Expiration	0.200	MP6288																			
0.050	MP6286	202603A	3/3/2028	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment ☒ Post-Stability Checks ☒ Flow Adjustment ☒ Form 40 ☐ Other:																				
0.080	MP6287	202603B	3/3/2028																					
0.200	MP6288	202603D	3/4/2028																					
0.080 DGS	N/A	AG526603	9/23/2027																					
Gauge ID #: <u>33363</u>				Gauge: <u>1017</u> Instrument: <u>1017</u>																				
Notes/Suggested Service: Tech Review: Added checkmark to Post-Stability Checks and strikeout to Flow Adjustment in Attachments. (TDG 7/1/26)				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																				
Digitally signed by Wen-Chi Pierson Date: 2026.07.01 16:07:27 -04'00'				Digitally signed by Kaitlyn Spearin Date: 2026.07.09 10:11:59 -04'00'																				
Tech Review				Admin Review																				

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 10:38

Date of Inspection: 06/24/2026

Serial Number: 80-007168

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/24/2026
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																											
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																											
☒	☒	☒	☒																																																																																																											
FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:37	Control Test	0.048	11:38	Air Blank	0.000	11:38	Control Test	0.049	11:39	Air Blank	0.000	11:39	Control Test	0.048	11:40	Air Blank	0.000	11:41	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:44	Control Test	0.076	11:44	Air Blank	0.000	11:45	Control Test	0.076	11:46	Air Blank	0.000	11:46	Control Test	0.076	11:47	Air Blank	0.000	11:47	Control Test Stats			Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000		Printed to external printer. Will attach. ML 6/24/26 FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:49	Control Test	0.078	11:50	Air Blank	0.000	11:50	Control Test	0.078	11:50	Air Blank	0.000	11:51	Control Test	0.078	11:51	Air Blank	0.000	11:52	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																												
Air Blank	0.000	11:37																																																																																																												
Control Test	0.048	11:38																																																																																																												
Air Blank	0.000	11:38																																																																																																												
Control Test	0.049	11:39																																																																																																												
Air Blank	0.000	11:39																																																																																																												
Control Test	0.048	11:40																																																																																																												
Air Blank	0.000	11:41																																																																																																												
Control Test Stats																																																																																																														
Average	0.0483																																																																																																													
Std Dev	0.0006																																																																																																													
Rel Std Dev(%)	1.1945																																																																																																													
Test	g/210L	Time																																																																																																												
Air Blank	0.000	11:44																																																																																																												
Control Test	0.076	11:44																																																																																																												
Air Blank	0.000	11:45																																																																																																												
Control Test	0.076	11:46																																																																																																												
Air Blank	0.000	11:46																																																																																																												
Control Test	0.076	11:47																																																																																																												
Air Blank	0.000	11:47																																																																																																												
Control Test Stats																																																																																																														
Average	0.0760																																																																																																													
Std Dev	0.0000																																																																																																													
Rel Std Dev(%)	0.0000																																																																																																													
Test	g/210L	Time																																																																																																												
Air Blank	0.000	11:49																																																																																																												
Control Test	0.078	11:50																																																																																																												
Air Blank	0.000	11:50																																																																																																												
Control Test	0.078	11:50																																																																																																												
Air Blank	0.000	11:51																																																																																																												
Control Test	0.078	11:51																																																																																																												
Air Blank	0.000	11:52																																																																																																												
Control Test Stats																																																																																																														
Average	0.0780																																																																																																													
Std Dev	0.0000																																																																																																													
Rel Std Dev(%)	0.0000																																																																																																													

FFWCC

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-007168

06/24/2026

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.194	11:31
Air Blank	0.000	11:31
Control Test	0.193	11:32
Air Blank	0.000	11:33
Control Test	0.193	11:33
Air Blank	0.000	11:34
Control Test Stats		
Average	0.1933	
Std Dev	0.0006	
Rel Std Dev(%)	0.2986	



Operator's Signature

#1

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007168
06/29/2026 13:47:14

Auto Calibration
Max Power Res Value = 70
Auto Range Res Value = 50

**** AUTO CAL FAIL

Mistyped one of the concentrations at the beginning of the adjustment process. Aborted prior to beginning the first adjustment and restarted a second adjustment.

NL
6/29/26

Optical Calibration Adjustment	
By:	TDG

#2

FIUCC
Intoxilyzer - Alcohol Analyzer
Model 8000
06/29/2026
SN 80-007168
13:51:31

Auto Calibration
Max Power Res Value = 73
Auto Range Res Value = 50

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12494, 9um Io = 12733
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0850 (-0.0110)
Sample #2 = 0.0640 (-0.0000)
Sample #3 = 0.0760 (-0.0120)
Sample #4 = 0.1080 (-0.0040)
Avg % Abs = 0.0827 (-0.0027)
STD DEV = 0.0227 (0.0083)
REL STD DEV = 27.514 (312.250)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1120 (-0.0100)
Sample #2 = 0.1030 (-0.0040)
Sample #3 = 0.1230 (-0.0290)
Sample #4 = 0.1310 (-0.0140)
Avg % Abs = 0.1190 (-0.0157)
STD DEV = 0.0144 (0.0126)
REL STD DEV = 12.119 (80.317)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12479, 9um Io = 12727
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7770 (-0.0000)
Sample #2 = 0.7800 (-0.0050)
Sample #3 = 0.8280 (-0.0090)
Sample #4 = 0.8080 (-0.0250)
Avg % Abs = 0.8053 (-0.0070)
STD DEV = 0.0241 (0.0171)
REL STD DEV = 2.994 (244.114)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4290 (-0.0010)
Sample #2 = 1.4380 (-0.0010)
Sample #3 = 1.4380 (-0.0070)
Sample #4 = 1.4400 (-0.0100)
Avg % Abs = 1.4387 (-0.0053)
STD DEV = 0.0012 (0.0057)
REL STD DEV = 0.080 (106.617)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12474, 9um Io = 12725
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8460 (-0.0030)
Sample #2 = 1.9040 (-0.0190)
Sample #3 = 1.8600 (-0.0150)
Sample #4 = 1.9030 (-0.0010)
Avg % Abs = 1.8890 (-0.0010)
STD DEV = 0.0251 (0.0171)
REL STD DEV = 1.330 (1708.800)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.3750 (-0.0010)
Sample #2 = 3.4180 (-0.0110)
Sample #3 = 3.4040 (-0.0060)
Sample #4 = 3.4210 (-0.0020)
Avg % Abs = 3.4143 (-0.0010)
STD DEV = 0.0091 (0.0089)
REL STD DEV = 0.266 (888.820)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12467, 9um Io = 12720
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5920 (-0.0070)
Sample #2 = 3.5770 (-0.0080)
Sample #3 = 3.5850 (-0.0160)
Sample #4 = 3.6220 (-0.0050)
Avg % Abs = 3.5947 (-0.0063)
STD DEV = 0.0240 (0.0106)
REL STD DEV = 0.668 (167.349)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.5220 (-0.0160)
Sample #2 = 6.5060 (-0.0050)
Sample #3 = 6.4920 (-0.0010)
Sample #4 = 6.5130 (-0.0180)
Avg % Abs = 6.5037 (-0.0073)
STD DEV = 0.0107 (0.0097)
REL STD DEV = 0.164 (132.444)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12465, 9um Io = 12719
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3230 (-0.0210)
Sample #2 = 5.2550 (-0.0230)
Sample #3 = 5.3010 (-0.0090)
Sample #4 = 5.2540 (-0.0320)
Avg % Abs = 5.2700 (-0.0213)
STD DEV = 0.0269 (0.0116)
REL STD DEV = 0.510 (54.329)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.4740 (-0.0160)
Sample #2 = 9.4320 (-0.0150)
Sample #3 = 9.4610 (-0.0010)
Sample #4 = 9.4310 (-0.0160)
Avg % Abs = 9.4413 (-0.0100)
STD DEV = 0.0170 (0.0095)
REL STD DEV = 0.180 (95.394)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.083
Std Dev = 0.02 Rel Std Dev = 27.51
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.805
Std Dev = 0.02 Rel Std Dev = 2.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.889
Std Dev = 0.03 Rel Std Dev = 1.33
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.595
Std Dev = 0.02 Rel Std Dev = 0.67
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.270
Std Dev = 0.03 Rel Std Dev = 0.51
Zero Order Coef = -215.95
First Order Coef = 2594.08
Second Order Coef = 30.20
Standard Deviation = 20.764763

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.119
Std Dev = 0.01 Rel Std Dev = 12.12
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.439
Std Dev = 0.00 Rel Std Dev = 0.08
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.414
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.504
Std Dev = 0.01 Rel Std Dev = 0.16
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.441
Std Dev = 0.02 Rel Std Dev = 0.18
Zero Order Coef = -155.31
First Order Coef = 1395.91
Second Order Coef = 14.15
Standard Deviation = 14.361886

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0002
0.040 0.040 0.0005
0.100 0.100 -0.0003
0.200 0.200 0.0000
0.300 0.300 0.0000
Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3220.00
Sample #2 = 3220.00
Sample #3 = 3185.00
Sample #4 = 3175.00
Average Result = 3193.3333
STD DEV = 23.6291
REL STD DEV = 0.740
***** CHANNEL 2 *****
Sample #1 = 3402.00
Sample #2 = 3377.00
Sample #3 = 3390.00
Sample #4 = 3375.00
Average Result = 3381.6667
STD DEV = 8.1445
REL STD DEV = 0.241

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1017
3 um H2O Adjust (mg/l x 10.000) = 616
9 um H2O Adjust (mg/l x 10.000) = 429
***** AUTO CAL PRSS *****

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0000
0.040 0.040 0.0003
0.100 0.101 -0.0006
0.200 0.199 0.0005
0.300 0.300 -0.0002

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/29/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:53</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:53	Control Test	0.050	15:54	Air Blank	0.000	15:54	Control Test	0.049	15:55	Air Blank	0.000	15:55	Control Test	0.049	15:56	Air Blank	0.000	15:56	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/29/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:03</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	16:01	Control Test	0.078	16:02	Air Blank	0.000	16:02	Control Test	0.078	16:03	Air Blank	0.000	16:03	Control Test	0.077	16:04	Air Blank	0.000	16:05	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/29/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:08</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:10</td></tr> <tr><td>Control Test</td><td>0.197</td><td>16:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	16:08	Control Test	0.198	16:08	Air Blank	0.000	16:09	Control Test	0.198	16:10	Air Blank	0.000	16:10	Control Test	0.197	16:11	Air Blank	0.000	16:11	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/29/2026 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:57</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:57	Control Test	0.079	15:58	Air Blank	0.000	15:58	Control Test	0.079	15:59	Air Blank	0.000	15:59	Control Test	0.079	15:59	Air Blank	0.000	16:00	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:53																																																																																																																																																	
Control Test	0.050	15:54																																																																																																																																																	
Air Blank	0.000	15:54																																																																																																																																																	
Control Test	0.049	15:55																																																																																																																																																	
Air Blank	0.000	15:55																																																																																																																																																	
Control Test	0.049	15:56																																																																																																																																																	
Air Blank	0.000	15:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:01																																																																																																																																																	
Control Test	0.078	16:02																																																																																																																																																	
Air Blank	0.000	16:02																																																																																																																																																	
Control Test	0.078	16:03																																																																																																																																																	
Air Blank	0.000	16:03																																																																																																																																																	
Control Test	0.077	16:04																																																																																																																																																	
Air Blank	0.000	16:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:08																																																																																																																																																	
Control Test	0.198	16:08																																																																																																																																																	
Air Blank	0.000	16:09																																																																																																																																																	
Control Test	0.198	16:10																																																																																																																																																	
Air Blank	0.000	16:10																																																																																																																																																	
Control Test	0.197	16:11																																																																																																																																																	
Air Blank	0.000	16:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:57																																																																																																																																																	
Control Test	0.079	15:58																																																																																																																																																	
Air Blank	0.000	15:58																																																																																																																																																	
Control Test	0.079	15:59																																																																																																																																																	
Air Blank	0.000	15:59																																																																																																																																																	
Control Test	0.079	15:59																																																																																																																																																	
Air Blank	0.000	16:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC
Time of Inspection: 13:54

Date of Inspection: 06/30/2026

Serial Number: 80-007168
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.050	0.077	0.196	0.078
0.000	0.050	0.077	0.196	0.079
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.077	0.196	0.078
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.077	0.195	0.079
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.050	0.077	0.195	0.079


Standard Deviations	0.0004	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007168 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007168</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:54</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.06.30 14:36:16
-04'00'

06/30/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality