

Alcohol Testing Program - Instrument Processing Sheet

Agency: MARION COUNTY SO

Instrument Serial Number: 80-000827

Date In: 6/9/2026

DI Completion Date: 6/18/2027 *2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>SLH</u> Date: <u>6/9/2026</u>		Quality Checks By: <u>IAS</u> Date: <u>6/11/2026</u>		Flow Adjustment By: _____																													
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Note from Agency the external printer port loose.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>211</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP102</u> 32 mm .152 (.139-.169) 36 mm .164 (.156-.190) 53 mm .230 (.228-.278) 103 mm .527 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1012</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																													
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202603B</td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202603D</td> </tr> <tr> <td>3/4/2028</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> </tr> <tr> <td>10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A	3/3/2028	0.080	MP6292	202603B	3/3/2028	0.200	MP6293	202603D	3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____										
Simulator	Serial #	Lot#/Exp																															
0.050	MP5088	202603A																															
		3/3/2028																															
0.080	MP6292	202603B																															
		3/3/2028																															
0.200	MP6293	202603D																															
		3/4/2028																															
0.080 DGS	N/A	23825080A1																															
		10/5/2027																															
Optical Bench Adjustment By: <u>SLH</u> Barometric Pressure Gauge: <u>1011</u> ID#: <u>34420</u>		Department Inspection By: <u>WKP</u> Barometric Pressure ID#: <u>34419 (x2)</u>																															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>03/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>08/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>01/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>05/06/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	25282	08/12/2027	0.200	MP6297	25020	01/14/2027	0.300	MP6298	25150	05/06/2027	0.080 DGS	N/A	AG510701	04/17/2027	Gauge: <u>1012 / 1012</u> Instrument: <u>1012 / 1011</u> Mouth Alcohol Solution Lot #: <u>2025-D (x2)</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294	N/A	N/A																														
0.040	MP6295	25090	03/11/2027																														
0.100	MP6296	25282	08/12/2027																														
0.200	MP6297	25020	01/14/2027																														
0.300	MP6298	25150	05/06/2027																														
0.080 DGS	N/A	AG510701	04/17/2027																														
☒ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A</td> <td>03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B</td> <td>03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D</td> <td>03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/2027</td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202603A	03/03/2028	0.080	MP6292	202603B	03/03/2028	0.200	MP6293	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027								
Simulator	Serial #	Lot #	Expiration																														
0.050	MP6291	202603A	03/03/2028																														
0.080	MP6292	202603B	03/03/2028																														
0.200	MP6293	202603D	03/04/2028																														
0.080 DGS	N/A	23825080A1	10/5/2027																														
Gauge ID #: <u>28427</u> Gauge: <u>1009</u> Instrument: <u>1010</u>		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																															
Notes/Suggested Service: Tech Review: Corrected DI completion date. WKP 6/30/2026		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct Agency Inspection Before Evidentiary Use Digitally signed by Shayla Platt Date: 2026.06.30 12:14:03 -04'00' Digitally signed by Shayla Platt Date: 2026.06.30 12:14:15 -04'00'																															

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet
MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:52 Control Test 0.048 12:53 Air Blank 0.000 12:54 Control Test 0.047 12:54 Air Blank 0.000 12:55 Control Test 0.047 12:55 Air Blank 0.000 12:56 Control Test Stats Average 0.0473 Std Dev 0.0005 Rel Std Dev(%) 1.2198	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 13:11 Control Test 0.077 13:11 Air Blank 0.000 13:12 Control Test 0.076 13:13 Air Blank 0.000 13:13 Control Test 0.076 13:14 Air Blank 0.000 13:14 Control Test Stats Average 0.0753 Std Dev 0.0016 Rel Std Dev(%) 0.7564	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:08 Control Test 0.197 14:09 Air Blank 0.000 14:09 Control Test 0.197 14:10 Air Blank 0.000 14:11 Control Test 0.197 14:11 Air Blank 0.000 14:12 Control Test Stats Average 0.1970 Std Dev 0.0000 Rel Std Dev(%) 0.0000	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:59 Control Test 0.080 11:59 Air Blank 0.000 12:00 Control Test 0.080 12:00 Air Blank 0.000 12:01 Control Test 0.080 12:01 Air Blank 0.000 12:02 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000
<i>Asheella Symons</i> Operator's Signature	<i>Asheella Symons</i> Operator's Signature	<i>Asheella Symons</i> Operator's Signature	<i>Asheella Symons</i> Operator's Signature

MARION COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

06/11/2026

SN 80-000827 ✓

15:21:46

sub

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0950	(-0.0170)	0.0950	(-0.0100)
Sample #2	0.1080	(0.0010)	0.1010	(-0.0120)
Sample #3	0.0870	(0.0110)	0.0850	(0.0000)
Sample #4	0.0620	(0.0330)	0.0930	(0.0010)
Avg % Abs	0.0857	(0.0150)	0.0930	(-0.0037)
STD DEV	0.0230	(0.0164)	0.0080	(0.0072)
REL STD DEV	26.882	(109.138)	8.602	(197.296)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8710	(-0.0240)	1.5460	(0.0000)
Sample #2	0.8690	(-0.0220)	1.5410	(-0.0030)
Sample #3	0.8500	(0.0110)	1.5450	(-0.0100)
Sample #4	0.8660	(0.0030)	1.5670	(-0.0040)
Avg % Abs	0.8617	(-0.0027)	1.5510	(-0.0057)
STD DEV	0.0102	(0.0172)	0.0140	(0.0038)
REL STD DEV	1.185	(645.538)	0.903	(66.811)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.0020	(-0.0190)	3.6640	(-0.0150)
Sample #2	2.0060	(-0.0010)	3.6880	(-0.0160)
Sample #3	1.9900	(0.0120)	3.6530	(0.0040)
Sample #4	2.0190	(0.0170)	3.6590	(0.0000)
Avg % Abs	2.0050	(0.0093)	3.6667	(-0.0040)
STD DEV	0.0145	(0.0093)	0.0187	(0.0106)
REL STD DEV	0.724	(99.553)	0.510	(264.575)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.8900	(-0.0350)	7.0990	(-0.0150)
Sample #2	3.8900	(-0.0270)	7.0680	(-0.0120)
Sample #3	3.8860	(0.0010)	7.0660	(-0.0050)
Sample #4	3.8820	(-0.0030)	7.0500	(0.0030)
Avg % Abs	3.8860	(-0.0097)	7.0613	(-0.0047)
STD DEV	0.0040	(0.0151)	0.0099	(0.0075)
REL STD DEV	0.103	(156.660)	0.140	(160.833)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.7120	(-0.0150)	10.3020	(-0.0150)
Sample #2	5.6820	(0.0150)	10.2680	(0.0190)
Sample #3	5.6680	(0.0220)	10.2450	(0.0190)
Sample #4	5.6800	(0.0220)	10.2650	(0.0150)
Avg % Abs	5.6767	(0.0197)	10.2593	(0.0177)
STD DEV	0.0076	(0.0040)	0.0125	(0.0023)
REL STD DEV	0.133	(20.550)	0.122	(13.072)

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
06/11/2026

SN 80-000827 ✓
15:21:46

SWA

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -198.07
First Order Coef 2416.18
Second Order Coef 23.57

<<<< 9um >>>>

-116.64
1285.64
11.48

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002
0.040	0.040	0.0001
0.100	0.100	0.0004
0.200	0.200	-0.0005
0.300	0.300	0.0002

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	-0.0000
0.100	0.100	0.0002
0.200	0.200	-0.0002
0.300	0.300	0.0001

<<<< 3um >>>>

<<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample

Sample #1	3326.00	3582.00
Sample #2	3302.00	3521.00
Sample #3	3339.00	3562.00
Sample #4	3282.00	3523.00
Avg	3307.6667	3535.3333
STD DEV	28.9194	23.1157
REL STD DEV	0.874	0.654
H2O adjust (mg/l*10k)	502	274

Barometric Pressure = 1011

*****CALIBRATION SUCCESSFUL***** ✓

Post-Stability Checks

80-00 0827

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet
MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000827 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:49 Control Test 0.050 16:49 Air Blank 0.000 16:50 Control Test 0.045 16:51 Air Blank 0.000 16:51 Control Test 0.050 16:52 Air Blank 0.000 16:52 Control Test Stats Average 0.0457 Std Dev 0.0006 Rel Std Dev(%) 1.1625 <i>M. Sigilob</i> Operator's Signature	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000827 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:29 Control Test 0.077 16:30 Air Blank 0.000 16:30 Control Test 0.078 16:31 Air Blank 0.000 16:31 Control Test 0.078 16:32 Air Blank 0.000 16:33 Control Test Stats Average 0.0777 Std Dev 0.0006 Rel Std Dev(%) 0.7434 <i>M. Sigilob</i> Operator's Signature	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000827 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:36 Control Test 0.196 16:37 Air Blank 0.000 16:37 Control Test 0.195 16:38 Air Blank 0.000 16:38 Control Test 0.196 16:39 Air Blank 0.000 16:40 Control Test Stats Average 0.1957 Std Dev 0.0006 Rel Std Dev(%) 0.2951 <i>M. Sigilob</i> Operator's Signature	MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000827 06/11/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 16:44 Control Test 0.080 16:44 Air Blank 0.000 16:45 Control Test 0.080 16:45 Air Blank 0.000 16:46 Control Test 0.079 16:46 Air Blank 0.000 16:47 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 <i>M. Sigilob</i> Operator's Signature

Verified instrument was
able to read gas pressure.

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 10:39

Date of Inspection: 06/18/2026

Serial Number: 80-000827
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

A F / M A: . Non-compliance: . Inadvertently provided mouth alcohol during the alcohol free subject test and did not repeat, thus ending the inspection. WKP 6/18/2026

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson

WEN-CHI K PIERSON
Signature and Printed Name

06/18/2026
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO

Serial Number: 80-000827

Time of Inspection: 12:37

Date of Inspection: 06/18/2026

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#: 202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#: 202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 23825080A1 Exp: 10/05/2027
0.000	0.000 / 0.050	0.078	0.199	0.080
0.000	0.000 / 0.051	0.079	0.200	0.080
0.000	0.000 / 0.051	0.078	0.199	0.080
0.000	0.000 / 0.050	0.078	0.200	0.079
0.000	0.000 / 0.050	0.079	0.199	0.079
0.000	0.000 / 0.051	0.079	0.199	0.079
0.000	0.000 / 0.051	0.079	0.199	0.080
0.000	0.000 / 0.051	0.079	0.200	0.079
0.000	0.000 / 0.051	0.079	0.200	0.079
0.000	0.000 / 0.051	0.079	0.199	0.080
Standard Deviations	0.0000 / 0.0004	0.0004	0.0005	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance.

0.05g/210L simulator was not fully attached to the instrument during the first test. Connected simulator completely and repeated. ulp

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

6/18/26

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

ulp

WEN-CHI K PIERSON

Signature and Printed Name

06/18/2026

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000827, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000827</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/18/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:37</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson

Digitally signed by Wen-Chi
Pierson
Date: 2026.06.18 13:34:11
-04'00'

06/18/2026

Date

WEN-CHI K PIERSON,

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality