

Instrument Processing Sheet

Agency: Cape Coral Police Department Instrument Serial Number: 80-001647
 Date In: 10/14/2025 DI Completion Date: 10/30/2025 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>10/24/25</u>	Quality Checks By: <u>TDG</u> Date: <u>10/28/25</u>	Flow Adjustment By: _____ Date: _____															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>194</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP104</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1009</u> Instrument: <u>1011</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)															
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
Simulator	Serial #	Lot#/Exp															
0.050	MP6286	202406K 6/19/2026															
0.080	MP6287	202406L 6/19/2026															
0.200	MP6288	202406N 6/20/2026															
0.080 DGS	N/A	AG526603 9/23/2027															

Optical Bench Adjustment By: _____	Department Inspection By: <u>TDG</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1013</u> Instrument: <u>1015</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other:																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								
Barometric Pressure Gauge: _____ ID#: _____	<input type="checkbox"/> Other:																																								

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Digitally signed by Shayla Platt Date: 2025.11.17 20:37:33 -05'00'	Digitally signed by LeAndra Higginbotham Date: 2025.11.18 19:55:02 -05'00'
Shayla Platt Tech Review	LeAndra Higginbotham Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																																
<p>#2</p> <p>CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model: 8000 10/28/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:41</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:42</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:43</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	15:41	Control Test	0.049	15:41	Air Blank	0.000	15:42	Control Test	0.049	15:43	Air Blank	0.000	15:43	Control Test	0.048	15:44	Air Blank	0.000	15:44	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel. Std Dev(%)	1.1863		<p>CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model: 8000 10/28/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:00</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:02</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2821</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	15:59	Control Test	0.079	16:00	Air Blank	0.000	16:00	Control Test	0.078	16:01	Air Blank	0.000	16:02	Control Test	0.077	16:02	Air Blank	0.000	16:03	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel. Std Dev(%)	1.2821		<p>CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model: 8000 10/28/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:23</td></tr> <tr><td>Control Test</td><td>0.196</td><td>16:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:24</td></tr> <tr><td>Control Test</td><td>0.196</td><td>16:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:26</td></tr> <tr><td>Control Test</td><td>0.195</td><td>16:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	16:23	Control Test	0.196	16:24	Air Blank	0.000	16:24	Control Test	0.196	16:25	Air Blank	0.000	16:26	Control Test	0.195	16:26	Air Blank	0.000	16:27	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel. Std Dev(%)	0.2951		<p>CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model: 8000 10/28/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>16:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>16:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	16:06	Control Test	0.080	16:06	Air Blank	0.000	16:06	Control Test	0.080	16:07	Air Blank	0.000	16:07	Control Test	0.079	16:08	Air Blank	0.000	16:08	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel. Std Dev(%)	0.7247		<p>CAPE CORAL PD Intoxilyzer - Alcohol Analyzer Model: 8000 10/28/2025 Software: 8100.27</p> <p>Operator's Signature </p>
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	15:41																																																																																																																																																		
Control Test	0.049	15:41																																																																																																																																																		
Air Blank	0.000	15:42																																																																																																																																																		
Control Test	0.049	15:43																																																																																																																																																		
Air Blank	0.000	15:43																																																																																																																																																		
Control Test	0.048	15:44																																																																																																																																																		
Air Blank	0.000	15:44																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0487																																																																																																																																																			
Std Dev	0.0006																																																																																																																																																			
Rel. Std Dev(%)	1.1863																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	15:59																																																																																																																																																		
Control Test	0.079	16:00																																																																																																																																																		
Air Blank	0.000	16:00																																																																																																																																																		
Control Test	0.078	16:01																																																																																																																																																		
Air Blank	0.000	16:02																																																																																																																																																		
Control Test	0.077	16:02																																																																																																																																																		
Air Blank	0.000	16:03																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0780																																																																																																																																																			
Std Dev	0.0010																																																																																																																																																			
Rel. Std Dev(%)	1.2821																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	16:23																																																																																																																																																		
Control Test	0.196	16:24																																																																																																																																																		
Air Blank	0.000	16:24																																																																																																																																																		
Control Test	0.196	16:25																																																																																																																																																		
Air Blank	0.000	16:26																																																																																																																																																		
Control Test	0.195	16:26																																																																																																																																																		
Air Blank	0.000	16:27																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.1957																																																																																																																																																			
Std Dev	0.0006																																																																																																																																																			
Rel. Std Dev(%)	0.2951																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	16:06																																																																																																																																																		
Control Test	0.080	16:06																																																																																																																																																		
Air Blank	0.000	16:06																																																																																																																																																		
Control Test	0.080	16:07																																																																																																																																																		
Air Blank	0.000	16:07																																																																																																																																																		
Control Test	0.079	16:08																																																																																																																																																		
Air Blank	0.000	16:08																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0797																																																																																																																																																			
Std Dev	0.0006																																																																																																																																																			
Rel. Std Dev(%)	0.7247																																																																																																																																																			

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CAPE CORAL PD
Time of Inspection: 13:49

Date of Inspection: 10/30/2025

Serial Number: 80-001647
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.197	0.079
0.000	0.049	0.080	0.197	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.080	0.198	0.077
0.000	0.049	0.080	0.198	0.078
0.000	0.050	0.080	0.198	0.077
0.000	0.050	0.080	0.197	0.078


Standard Deviations	0.0005	0.0000	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/30/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001647, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001647</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>CAPE CORAL PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:49</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Date: 2025.10.31 15:43:02
-04'00'

10/30/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality