

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 12:21

Date of Inspection: 12/03/2021

Serial Number: 80-006760

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1076482 Exp: 01/15/2022
/ 0.000	0.045	0.075	0.193	0.079
/ 0.000	0.046	0.076	0.193	0.079
/ 0.000	0.047	0.077	0.194	0.079

Number of Simulators Used: 5


Remarks:

00: Ambient Fail. *Cleared Room of Contaminants "Mouth Alcohol" from Room. 2<sup>ND</sup> Alc-free test Good.*

*(12)*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



NEIL F RUBEL

Signature and Printed Name

12/03/2021  
Date