

Instrument Processing Sheet

Agency: SUMTER COUNTY SO Instrument Serial Number: 80-001471
 Date In: 12/9/2025 DI Completion Date: 1/30/26 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>12/9/25</u> <input type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable	Quality Checks By: <u>KTS</u> Date: <u>1/28/26</u> <input checked="" type="checkbox"/> Breath Tube Screen * <input checked="" type="checkbox"/> Replace External O-Rings * <input checked="" type="checkbox"/> Instrument Set Up Verified * <input checked="" type="checkbox"/> R-Value: <u>198</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.140</u> (.139-.169) 36 mm <u>0.152</u> (.156-.190) 53 mm <u>0.222</u> (.228-.278) 103 mm <u>0.480</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1023</u> Instrument: <u>1019</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: <u>KTS</u> Date: <u>1/28/26</u> Flow Column #: <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>198</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.164</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.511</u> (.447-.547)
--	---	--

Notes: Instrument received with bottom left corner keyboard screw panel broken. WKP 12/9/2025
 Will not go into ready mode- needs breath hose. KTS 12/10/25

Simulator	Serial #	Lot#/Exp
0.050	MP5088	202406K 6/19/2026
0.080	MP5089	202406L 6/19/2026
0.200	MP5090	202406N 6/20/2026
0.080 DGS	N/A	AG510701 4/17/2027

Maintenance By: KTS Date: 1/27/26

Battery Replacement
 Dry Gas Regulator Replacement and Tank Sensor Tare
 Breath Tube Replacement
 Other:

Optical Bench Adjustment By: WKP

Barometric Pressure Gauge: 1023 ID#: 34419

Simulator	Serial #	Lot #	Expiration
0.000	MP5086	N/A	N/A
0.040	MP6295	25090	3/11/27
0.100	MP6296	25282	8/12/27
0.200	MP6297	25020	1/14/27
0.300	MP6298	25150	5/6/27
0.080 DGS	N/A	28424080A3	11/5/26

Post Optical Bench Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050	MP5088	202406K	6/19/26
0.080	MP5089	202406L	6/19/26
0.200	MP5090	202406N	6/20/26
0.080 DGS	N/A	AG510701	4/17/27

Barometric Pressure Gauge: 1023 ID#: 34420

Department Inspection By: KTS

Barometric Pressure ID#: 28427

Gauge: 1015 Instrument: 1016
 Mouth Alcohol Solution Lot #: 2025-D Exp: 9/25/27
 Acetone Stock Solution Lot #: 2025-B Exp: 9/22/27

Simulator	Serial Number
0.000	MP6294
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Attachments

Form 41
 Stability Checks
 Calibration Certificate
 Optical Bench Adjustment

Post-Stability Checks
 Flow Adjustment
 Form 40
 Other:

Notes/Suggested Service: Replaced breath hose. 1/27/26 KTS
 Barometric pressure of the instrument was 1023. WKP 1/28/26
 * performed on 1/27/26. KTS 1/30/26
 Tech Review: Added initials to note. KTS 1/30/26

Instrument Complies with Chapter 11D-8, FAC
 Instrument Does Not Comply with Chapter 11D-8, FAC

Return to/Place into Evidentiary Use
 Remain Out of Evidentiary Use

Conduct an Agency Inspection Before Evidentiary Use

Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2026.01.30 11:35:58 -05'00'</small>	Wen-Chi Pierson <small>Digitally signed by wen-chi Pierson Date: 2026.01.30 13:19:42 -05'00'</small>
Tech Review	Admin Review

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 14:17

Date of Inspection: 01/27/2026

Serial Number: 80-001471
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Katie T Spearin  _____
Signature and Printed Name

01/27/2026
Date

SUTTER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 60-001471
01/28/2026
Software: 8100.27



Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff) = 6.855
2: Rate (Liters/min) = 15
SQRT(Diff) = 11.574
3: Rate (Liters/min) = 30
SQRT(Diff) = 20.660
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 699
Rounded Intercept = -665010
Correlation = 0.99784

Flow Adjustment

Katry

1/28/26

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																																						
<p>SUMNER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000470 01/28/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:00</td></tr> <tr><td>Control Test</td><td>0.051</td><td>08:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:02</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:03</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:00	Control Test	0.051	08:01	Air Blank	0.000	08:02	Control Test	0.050	08:02	Air Blank	0.000	08:03	Control Test	0.049	08:04	Air Blank	0.000	08:04	Control Test Stats			Average	0.0500		Std Dev	0.0010		Rel Std Dev(%)	2.0000		<p>SUMNER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000470 01/28/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:06</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:07</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:08</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:09</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:06	Control Test	0.081	08:07	Air Blank	0.000	08:07	Control Test	0.081	08:08	Air Blank	0.000	08:08	Control Test	0.081	08:09	Air Blank	0.000	08:09	Control Test	0.081	08:10	Air Blank	0.000	08:10	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>SUMNER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000470 01/28/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:11</td></tr> <tr><td>Control Test</td><td>0.201</td><td>08:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:13</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:14</td></tr> <tr><td>Control Test</td><td>0.200</td><td>08:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	08:11	Control Test	0.201	08:12	Air Blank	0.000	08:13	Control Test	0.200	08:13	Air Blank	0.000	08:14	Control Test	0.200	08:15	Air Blank	0.000	08:15	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		<p>SUMNER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000470 01/28/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>07:55</td></tr> <tr><td>Control Test</td><td>0.084</td><td>07:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:57</td></tr> <tr><td>Control Test</td><td>0.085</td><td>07:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr> <tr><td>Control Test</td><td>0.085</td><td>07:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0847</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.6819</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	07:55	Control Test	0.084	07:56	Air Blank	0.000	07:57	Control Test	0.085	07:57	Air Blank	0.000	07:58	Control Test	0.085	07:58	Air Blank	0.000	07:58	Control Test Stats			Average	0.0847		Std Dev	0.0006		Rel Std Dev(%)	0.6819	
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	08:00																																																																																																																																																							
Control Test	0.051	08:01																																																																																																																																																							
Air Blank	0.000	08:02																																																																																																																																																							
Control Test	0.050	08:02																																																																																																																																																							
Air Blank	0.000	08:03																																																																																																																																																							
Control Test	0.049	08:04																																																																																																																																																							
Air Blank	0.000	08:04																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0500																																																																																																																																																								
Std Dev	0.0010																																																																																																																																																								
Rel Std Dev(%)	2.0000																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	08:06																																																																																																																																																							
Control Test	0.081	08:07																																																																																																																																																							
Air Blank	0.000	08:07																																																																																																																																																							
Control Test	0.081	08:08																																																																																																																																																							
Air Blank	0.000	08:08																																																																																																																																																							
Control Test	0.081	08:09																																																																																																																																																							
Air Blank	0.000	08:09																																																																																																																																																							
Control Test	0.081	08:10																																																																																																																																																							
Air Blank	0.000	08:10																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0810																																																																																																																																																								
Std Dev	0.0000																																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	08:11																																																																																																																																																							
Control Test	0.201	08:12																																																																																																																																																							
Air Blank	0.000	08:13																																																																																																																																																							
Control Test	0.200	08:13																																																																																																																																																							
Air Blank	0.000	08:14																																																																																																																																																							
Control Test	0.200	08:15																																																																																																																																																							
Air Blank	0.000	08:15																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.2003																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.2882																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	07:55																																																																																																																																																							
Control Test	0.084	07:56																																																																																																																																																							
Air Blank	0.000	07:57																																																																																																																																																							
Control Test	0.085	07:57																																																																																																																																																							
Air Blank	0.000	07:58																																																																																																																																																							
Control Test	0.085	07:58																																																																																																																																																							
Air Blank	0.000	07:58																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0847																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.6819																																																																																																																																																								
<p> Operator's Signature</p>	<p> Operator's Signature</p>	<p> Operator's Signature</p>	<p> Operator's Signature</p>																																																																																																																																																						

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																																
<p>SUMNER COUNTY SD IntoxiLizer - Alcohol Analyzer Model 8000 01/28/2026 SN: 80-010470 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0017</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>3.6034</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:44	Control Test	0.050	12:45	Air Blank	0.000	12:45	Control Test	0.047	12:46	Air Blank	0.000	12:47	Control Test	0.047	12:47	Air Blank	0.000	12:48	Control Test Stats			Average	0.0480		Std Dev	0.0017		Rel. Std Dev(%)	3.6034		<p>SUMNER COUNTY SD IntoxiLizer - Alcohol Analyzer Model 8000 01/28/2026 SN: 80-010470 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7355</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.079	13:02	Air Blank	0.000	13:02	Control Test	0.079	13:03	Air Blank	0.000	13:03	Control Test	0.078	13:04	Air Blank	0.000	13:05	Control Test Stats			Average	0.0787		Std Dev	0.0016		Rel. Std Dev(%)	0.7355		<p>SUMNER COUNTY SD IntoxiLizer - Alcohol Analyzer Model 8000 01/28/2026 SN: 80-010470 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2335</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	13:13	Control Test	0.197	13:13	Air Blank	0.000	13:14	Control Test	0.196	13:15	Air Blank	0.000	13:15	Control Test	0.197	13:16	Air Blank	0.000	13:16	Control Test Stats			Average	0.1957		Std Dev	0.0016		Rel. Std Dev(%)	0.2335		<p>SUMNER COUNTY SD IntoxiLizer - Alcohol Analyzer Model 8000 01/28/2026 SN: 80-010470 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:36</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:35	Control Test	0.080	12:35	Air Blank	0.000	12:36	Control Test	0.080	12:36	Air Blank	0.000	12:36	Control Test	0.080	12:37	Air Blank	0.000	12:37	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p><i>meny</i> Operator's Signature</p>
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	12:44																																																																																																																																																		
Control Test	0.050	12:45																																																																																																																																																		
Air Blank	0.000	12:45																																																																																																																																																		
Control Test	0.047	12:46																																																																																																																																																		
Air Blank	0.000	12:47																																																																																																																																																		
Control Test	0.047	12:47																																																																																																																																																		
Air Blank	0.000	12:48																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0480																																																																																																																																																			
Std Dev	0.0017																																																																																																																																																			
Rel. Std Dev(%)	3.6034																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	13:01																																																																																																																																																		
Control Test	0.079	13:02																																																																																																																																																		
Air Blank	0.000	13:02																																																																																																																																																		
Control Test	0.079	13:03																																																																																																																																																		
Air Blank	0.000	13:03																																																																																																																																																		
Control Test	0.078	13:04																																																																																																																																																		
Air Blank	0.000	13:05																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0787																																																																																																																																																			
Std Dev	0.0016																																																																																																																																																			
Rel. Std Dev(%)	0.7355																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	13:13																																																																																																																																																		
Control Test	0.197	13:13																																																																																																																																																		
Air Blank	0.000	13:14																																																																																																																																																		
Control Test	0.196	13:15																																																																																																																																																		
Air Blank	0.000	13:15																																																																																																																																																		
Control Test	0.197	13:16																																																																																																																																																		
Air Blank	0.000	13:16																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.1957																																																																																																																																																			
Std Dev	0.0016																																																																																																																																																			
Rel. Std Dev(%)	0.2335																																																																																																																																																			
Test	g/210L	Time																																																																																																																																																		
Air Blank	0.000	12:35																																																																																																																																																		
Control Test	0.080	12:35																																																																																																																																																		
Air Blank	0.000	12:36																																																																																																																																																		
Control Test	0.080	12:36																																																																																																																																																		
Air Blank	0.000	12:36																																																																																																																																																		
Control Test	0.080	12:37																																																																																																																																																		
Air Blank	0.000	12:37																																																																																																																																																		
Control Test Stats																																																																																																																																																				
Average	0.0800																																																																																																																																																			
Std Dev	0.0000																																																																																																																																																			
Rel. Std Dev(%)	0.0000																																																																																																																																																			
<p><i>meny</i> Operator's Signature</p>	<p><i>meny</i> Operator's Signature</p>	<p><i>meny</i> Operator's Signature</p>	<p><i>meny</i> Operator's Signature</p>																																																																																																																																																	

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUMTER COUNTY SO
Time of Inspection: 08:02

Date of Inspection: 01/30/2026

Serial Number: 80-001471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.080	0.199	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.049	0.081	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.078
0.000	0.050	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.078


Standard Deviations	0.0007	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

01/30/2026
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001471</u>	UNCERTAINTY* ±	
Owning Agency:	<u>SUMTER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:02</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/30/2026

 Date
KATIE T SPEARIN,
 Department Inspector

FDLE/ATP Form 69 January 2026
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality