

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 04/03/2026 02:24:00

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.082	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.156	02:17
Breath Volume	1.101	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.146	02:22
Breath Volume	1.671	02:22
Air Blank	0.000	02:23
Control Test	0.081	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:24

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MONROY

DATE/TIME OF TEST: 04/06/2026 21:37:49

Diagnostics Check	OK	21:34	
Air Blank	0.000	21:35	
Control Test	0.082	21:35	
Air Blank	0.000	21:36	
Subject Sample	# 1	21:36	Subject Test Refused
Breath Volume	0.000	21:36	
Air Blank	0.000	21:36	
Control Test	0.082	21:37	
Air Blank	0.000	21:37	
Diagnostics Check	OK	21:37	

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON GONZALEZ,KEVIN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/07/2026 23:24:51

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.082	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.219	23:18
Breath Volume	1.398	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.211	23:23
Breath Volume	1.429	23:23
Air Blank	0.000	23:23
Control Test	0.081	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DAVID,A
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: PARKER,DAVID
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 04/08/2026 02:56:05

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.081	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.117	02:51
Breath Volume	1.601	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.114	02:54
Breath Volume	1.746	02:54
Air Blank	0.000	02:55
Control Test	0.081	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:56

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POPE,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: POPE,NICHOLAS
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 04/08/2026 18:25:31

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:23	
Control Test	0.081	18:23	
Air Blank	0.000	18:23	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNA,WARDELL,P
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HANNA,WARDELL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: RICKSECKER

DATE/TIME OF TEST: 04/11/2026 02:12:21

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:04	No Sample Provided
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 2	0.180	02:07	
Breath Volume		1.656	02:07	
Air Blank		0.000	02:08	
Air Blank		0.000	02:10	
Subject Sample	# 3	0.187	02:10	
Breath Volume		1.710	02:10	
Air Blank		0.000	02:11	
Control Test		0.081	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON GONZALEZ,KEVIN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LEIVA ROJAS

DATE/TIME OF TEST: 04/12/2026 03:43:30

Diagnostics Check	OK	03:32	
Air Blank	0.000	03:32	
Control Test	0.081	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2	03:42	No Sample Provided
Breath Volume	0.000	03:42	
Air Blank	0.000	03:42	
Control Test	0.082	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON GONZALEZ,KEVIN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 03/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LEIVA ROJAS

DATE/TIME OF TEST: 04/12/2026 03:51:49

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.110	03:47
Breath Volume	2.050	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:49
Subject Sample # 2	0.102	03:50
Breath Volume	1.187	03:50
Air Blank	0.000	03:50
Control Test	0.081	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/14/2026 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MATTHEW,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,MATTHEW
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: DUFFY

DATE/TIME OF TEST: 04/18/2026 01:18:11

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.081	01:16	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/18/2026 21:16:31

Diagnostics Check	OK	21:13	
Air Blank	0.000	21:13	
Control Test	0.081	21:13	
Air Blank	0.000	21:14	
Subject Sample	# 1	21:15	Subject Test Refused
Breath Volume	0.000	21:15	
Air Blank	0.000	21:15	
Control Test	0.081	21:15	
Air Blank	0.000	21:16	
Diagnostics Check	OK	21:16	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: CASTRO,SHANE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BARAKAT

DATE/TIME OF TEST: 04/19/2026 03:48:40

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.081	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:44	No Sample Provided
Breath Volume	0.000	03:44	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMAD,JIHAD,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 04/19/2026 06:25:42

Diagnostics Check	OK	06:22	
Air Blank	0.000	06:23	
Control Test	0.081	06:23	
Air Blank	0.000	06:24	
Subject Sample	# 1	06:24	Subject Test Refused
Breath Volume	0.000	06:24	
Air Blank	0.000	06:24	
Control Test	0.081	06:25	
Air Blank	0.000	06:25	
Diagnostics Check	OK	06:25	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMAD,JIHAD,B
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: HAMMAD,JIHAD
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 04/19/2026 06:31:14

Diagnostics Check	OK	06:28	
Air Blank	0.000	06:28	
Control Test	0.080	06:29	
Air Blank	0.000	06:29	
Subject Sample	# 1	06:29	Subject Test Refused
Breath Volume	0.000	06:29	
Air Blank	0.000	06:30	
Control Test	0.080	06:30	
Air Blank	0.000	06:31	
Diagnostics Check	OK	06:31	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEILL,CONNOR,M
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: O'NEILL,CONNOR
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/22/2026 19:28:32

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.081	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.020	19:24
Breath Volume	1.636	19:24
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.020	19:27
Breath Volume	1.597	19:27
Air Blank	0.000	19:27
Control Test	0.082	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,DANIEL,J
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: VALDES,DANIEL
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 09:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 04/24/2026 10:48:05

Diagnostics Check	OK	10:45	
Air Blank	0.000	10:45	
Control Test	0.081	10:46	
Air Blank	0.000	10:46	
Subject Sample	# 1	10:46	Subject Test Refused
Breath Volume	0.000	10:46	
Air Blank	0.000	10:47	
Control Test	0.081	10:47	
Air Blank	0.000	10:48	
Diagnostics Check	OK	10:48	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON GONZALEZ,KEVIN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: TORRES NODARSE

DATE/TIME OF TEST: 04/25/2026 21:50:14

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.081	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.092	21:44
Breath Volume	1.550	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.091	21:48
Breath Volume	1.882	21:48
Air Blank	0.000	21:49
Control Test	0.082	21:49
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOSE,C
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: HERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RIVERY HERRERA

DATE/TIME OF TEST: 04/26/2026 02:23:39

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.081	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1	02:22	Subject Test Refused
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Control Test	0.080	02:23	
Air Blank	0.000	02:23	
Diagnostics Check	OK	02:23	

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON GONZALEZ,KEVIN,G
BREATH TEST OPERATOR AGENCY: MONROE COUNTY SO
ARRESTING OFFICER: LEON GONZALEZ,KEVIN
ARRESTING OFFICER AGENCY: MONROE COUNTY SO
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HANEY

DATE/TIME OF TEST: 04/30/2026 21:53:24

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.081	21:47
Air Blank	0.000	21:48
Subject Sample # 1	0.116	21:48
Breath Volume	3.140	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.110	21:51
Breath Volume	1.824	21:51
Air Blank	0.000	21:52
Control Test	0.081	21:52
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 05/20/2026 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXANDRE,BOBBY,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TYLER,LAKAEVIA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80000871
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/02/2026 00:14:09

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.080	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.080	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 04/15/2026 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: IBARRA,JORDAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PEREZ PINILLA

DATE/TIME OF TEST: 04/08/2026 18:02:35

Diagnostics Check		OK	17:50	
Air Blank		0.000	17:51	
Control Test		0.079	17:51	
Air Blank		0.000	17:51	
Subject Sample	# 1	0.225	17:55	Volume Not Met
Breath Volume		0.109	17:55	
Air Blank		0.000	17:55	
Air Blank		0.000	17:57	
Subject Sample	# 2	0.199	18:00	Volume Not Met
Breath Volume		0.015	18:00	
Air Blank		0.000	18:01	
Control Test		0.077	18:01	
Air Blank		0.000	18:02	
Diagnostics Check		OK	18:02	

Data Download Date/Time: 04/15/2026 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYA,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: IBARRA,JORDAN
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PEREZPINILLA

DATE/TIME OF TEST: 04/08/2026 18:10:28

Diagnostics Check	OK	18:05	
Air Blank	0.000	18:05	
Control Test	0.076	18:06	
Air Blank	0.000	18:06	
Subject Sample	# 1	18:08	Subject Test Refused
Breath Volume	0.707	18:08	
Air Blank	0.000	18:09	
Control Test	0.076	18:09	
Air Blank	0.000	18:10	
Diagnostics Check	OK	18:10	

Data Download Date/Time: 04/15/2026 14:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMSON,ROBERT,H
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: THOMSON,ROBERT
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 04/15/2026 23:49:51

Diagnostics Check	OK	23:49	
Air Blank		23:49	RFI Detect
Air Blank		23:49	RFI Detect

Data Download Date/Time: 05/12/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMSON,ROBERT,H
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: THOMSON,ROBERT
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: CONCEPCION

DATE/TIME OF TEST: 04/15/2026 23:55:51

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.080	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 1	23:54	Subject Test Refused
Breath Volume	0.000	23:54	
Air Blank	0.000	23:54	
Control Test	0.079	23:55	
Air Blank	0.000	23:55	
Diagnostics Check	OK	23:55	

Data Download Date/Time: 05/12/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAGES, BIANCA,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GONZALEZ, JULIO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 04/25/2026 01:08:43

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.197	01:04
Breath Volume	1.699	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.190	01:07
Breath Volume	1.152	01:07
Air Blank	0.000	01:07
Control Test	0.078	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 05/12/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OZETE-FONTE, BARBARA,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ, JORGE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25024080A3
DRY GAS EXPIRY DATE: 10/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: AGUERO

DATE/TIME OF TEST: 04/27/2026 11:07:29

Diagnostics Check	OK	10:56	
Air Blank	0.000	10:56	
Control Test	0.079	10:56	
Air Blank	0.000	10:57	
Subject Sample	# 1	11:00	No Sample Provided
Breath Volume	0.000	11:00	
Air Blank	0.000	11:00	
Air Blank	0.000	11:02	
Subject Sample	# 2	11:06	No Sample Provided
Breath Volume	0.000	11:06	
Air Blank	0.000	11:06	
Control Test	0.080	11:06	
Air Blank	0.000	11:07	
Diagnostics Check	OK	11:07	

Data Download Date/Time: 05/12/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SARMIENTO

DATE/TIME OF TEST: 04/10/2026 23:22:32

Diagnostics Check	OK	23:19	
Air Blank	0.000	23:20	
Control Test	0.081	23:20	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:21	Subject Test Refused
Breath Volume	0.000	23:21	
Air Blank	0.000	23:21	
Control Test	0.079	23:21	
Air Blank	0.000	23:22	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 04/24/2026 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SILVA BUITRAGO

DATE/TIME OF TEST: 04/11/2026 20:06:01

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.080	20:03	
Air Blank	0.000	20:04	
Subject Sample	# 1	20:04	Subject Test Refused
Breath Volume	0.000	20:04	
Air Blank	0.000	20:05	
Control Test	0.079	20:05	
Air Blank	0.000	20:05	
Diagnostics Check	OK	20:06	

Data Download Date/Time: 04/24/2026 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MARTINEZ MENDEZ

DATE/TIME OF TEST: 04/12/2026 01:41:41

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.160	01:36
Breath Volume	2.468	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.166	01:39
Breath Volume	1.917	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 04/24/2026 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SAAVEDRA CORTES

DATE/TIME OF TEST: 04/18/2026 22:32:32

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1 0.213	22:28	
Breath Volume	1.964	22:28	
Air Blank	0.000	22:29	
Air Blank	0.000	22:30	
Subject Sample	# 2 0.211	22:31	
Breath Volume	3.871	22:31	
Air Blank		22:31	RFI Detect
Air Blank	0.000	22:32	

Data Download Date/Time: 04/24/2026 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ARCIA,RAUL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SAAVEDRA CORTES

DATE/TIME OF TEST: 04/18/2026 22:45:24

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.078	22:38
Air Blank	0.000	22:38
Subject Sample # 1	0.197	22:39
Breath Volume	1.507	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.209	22:43
Breath Volume	2.800	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 04/24/2026 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARCIA,RAUL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MARIN ARROLIGA

DATE/TIME OF TEST: 04/19/2026 01:39:07

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 04/24/2026 10:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEREDIA RUBIO,ROBERTO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203318190
DRY GAS EXPIRY DATE: 04/11/2028
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TYNES

DATE/TIME OF TEST: 04/23/2026 02:17:00

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 04/24/2026 10:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,IVAN,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MENOCA, CESAR
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: NOA-ROMERO

DATE/TIME OF TEST: 04/11/2026 01:20:17

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.082	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.108	01:15
Breath Volume		1.710	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.101	01:18
Breath Volume		1.660	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 05/05/2026 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,CHRISTOPHER,F
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FIGUEROA,IVAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 04/11/2026 05:06:27

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.082	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.142	05:01
Breath Volume	2.210	05:01
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.145	05:04
Breath Volume	2.097	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 05/05/2026 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIALLO MORELL,RAUL,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FIGUEROA,IVAN
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 04/22/2026 03:18:59

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.082	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:18	
Control Test	0.082	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 05/05/2026 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,IVAN,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: RUIZ,GILBERT
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/24/2026 02:16:15

Diagnostics Check	OK	02:12	
Air Blank	0.000	02:13	
Control Test	0.082	02:13	
Air Blank	0.000	02:13	
Subject Sample	# 1	02:14	Subject Test Refused
Breath Volume	0.000	02:14	
Air Blank	0.000	02:15	
Control Test	0.082	02:15	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 05/05/2026 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,IVAN,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: RUIZ,GILBERT
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403255643
DRY GAS EXPIRY DATE: 01/29/2028
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/24/2026 02:29:54

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.081	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.267	02:20
Breath Volume		1.156	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.231	02:25
Breath Volume		2.027	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 3	0.242	02:28
Breath Volume		1.218	02:28
Air Blank		0.000	02:28
Control Test		0.079	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 05/05/2026 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: LABANINO PORTUONDO

DATE/TIME OF TEST: 04/05/2026 22:53:34

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:51	
Control Test	0.081	22:51	
Air Blank	0.000	22:52	
Subject Sample	# 1	22:52	Subject Test Refused
Breath Volume	0.000	22:52	
Air Blank	0.000	22:52	
Control Test	0.082	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 04/16/2026 04:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: BALDERA

DATE/TIME OF TEST: 04/09/2026 22:43:26

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.081	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1	0.096	22:34	Volume Not Met
Breath Volume		0.359	22:34	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 2	0.121	22:38	
Breath Volume		1.722	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 3	0.115	22:41	
Breath Volume		1.441	22:41	
Air Blank		0.000	22:42	
Control Test		0.080	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 04/16/2026 04:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: HIDALGO SILVA

DATE/TIME OF TEST: 04/27/2026 22:33:44

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.080	22:27
Air Blank	0.000	22:27
Subject Sample # 1	0.183	22:28
Breath Volume	2.156	22:28
Air Blank	0.000	22:29
Air Blank	0.000	22:31
Subject Sample # 2	0.180	22:32
Breath Volume	1.535	22:32
Air Blank	0.000	22:32
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 05/01/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAYRAMOV, AGAJAN,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: PIERRE, WENDLEY
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06023080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 04/27/2026 10:25:56

Diagnostics Check	OK	10:15	
Air Blank	0.000	10:16	
Control Test	0.076	10:16	
Air Blank	0.000	10:17	
Subject Sample # 1	0.092	10:19	
Breath Volume	1.882	10:19	
Air Blank	0.000	10:19	
Air Blank	0.000	10:21	
Subject Sample # 2	0.081	10:24	
Breath Volume	1.699	10:24	
Air Blank	0.000	10:25	
Control Test	0.073	10:25	Control Outside Tolerance
Air Blank	0.000	10:25	

Data Download Date/Time: 04/30/2026 18:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 15:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:29

SUBJECT'S LAST NAME: GELORMINO

DATE/TIME OF TEST: 04/05/2026 15:55:18

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.077	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:51
Breath Volume	2.808	15:51
Air Blank	0.000	15:51
Air Blank	0.000	15:53
Subject Sample # 2	0.000	15:53
Breath Volume	4.250	15:53
Air Blank	0.000	15:54
Control Test	0.077	15:54
Air Blank	0.000	15:55
Diagnostics Check	OK	15:55

Data Download Date/Time: 05/05/2026 06:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: STRINGER

DATE/TIME OF TEST: 04/06/2026 23:00:02

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.077	22:54
Air Blank	0.000	22:54
Subject Sample # 1	0.170	22:55
Breath Volume	1.937	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:57
Subject Sample # 2	0.180	22:58
Breath Volume	3.382	22:58
Air Blank	0.000	22:59
Control Test	0.077	22:59
Air Blank	0.000	22:59
Diagnostics Check	OK	23:00

Data Download Date/Time: 05/05/2026 06:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELTHROP, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: BELTHROP, MATTHEW
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: ISAA

DATE/TIME OF TEST: 04/12/2026 06:10:02

Diagnostics Check		OK	06:02
Air Blank		0.000	06:02
Control Test		0.077	06:02
Air Blank		0.000	06:03
Subject Sample	# 1	0.178	06:05
Breath Volume		1.726	06:05
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.172	06:08
Breath Volume		1.734	06:08
Air Blank		0.000	06:09
Control Test		0.077	06:09
Air Blank		0.000	06:09
Diagnostics Check		OK	06:10

Data Download Date/Time: 05/05/2026 06:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 04/13/2026 22:40:25

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.074	22:39	Control Outside Tolerance
Air Blank	0.000	22:40	

Data Download Date/Time: 05/05/2026 06:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 04/04/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403056709
DRY GAS EXPIRY DATE: 05/28/2027
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 04/13/2026 22:54:03

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.077	22:42	
Air Blank		0.000	22:43	
Subject Sample	# 1	0.133	22:46	Volume Not Met
Breath Volume		0.230	22:46	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.170	22:49	
Breath Volume		2.183	22:49	
Air Blank		0.000	22:50	
Air Blank		0.000	22:52	
Subject Sample	# 3	0.180	22:52	
Breath Volume		1.515	22:52	
Air Blank		0.000	22:53	
Control Test		0.076	22:53	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 05/05/2026 06:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/04/2026 18:29:33

Diagnostics Check	OK	18:22
Air Blank	0.000	18:22
Control Test	0.081	18:23
Air Blank	0.000	18:23
Subject Sample # 1	0.214	18:24
Breath Volume	1.191	18:24
Air Blank	0.000	18:24
Air Blank	0.000	18:26
Subject Sample # 2	0.232	18:27
Breath Volume	2.140	18:27
Air Blank	0.000	18:28
Control Test	0.081	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/05/2026 21:08:15

Diagnostics Check		OK	20:56	
Air Blank		0.000	20:57	
Control Test		0.082	20:57	
Air Blank		0.000	20:57	
Subject Sample	# 1	0.227	21:01	Volume Not Met
Breath Volume		0.031	21:01	
Air Blank		0.000	21:01	
Air Blank		0.000	21:03	
Subject Sample	# 2	0.214	21:06	Volume Not Met
Breath Volume		0.621	21:06	
Air Blank		0.000	21:07	
Control Test		0.080	21:07	
Air Blank		0.000	21:08	
Diagnostics Check		OK	21:08	

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIE,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 04/05/2026 21:14:49

Diagnostics Check	OK	21:09	
Air Blank	0.000	21:10	
Control Test	0.080	21:10	
Air Blank	0.000	21:10	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.035	21:13	
Air Blank	0.000	21:13	
Control Test	0.080	21:14	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: ROMANICK

DATE/TIME OF TEST: 04/14/2026 18:02:02

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.081	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.117	17:57
Breath Volume	2.332	17:57
Air Blank	0.000	17:58
Air Blank	0.000	18:00
Subject Sample # 2	0.120	18:00
Breath Volume	1.992	18:00
Air Blank	0.000	18:01
Control Test	0.081	18:01
Air Blank	0.000	18:01
Diagnostics Check	OK	18:02

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: VANNIEUWENHOVEN,SCOT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: BARCELOS

DATE/TIME OF TEST: 04/14/2026 23:45:03

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.082	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.146	23:40
Breath Volume	1.386	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.152	23:43
Breath Volume	1.433	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:45

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILLIPS,CHARLES,B
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PHILLIPS,CHARLES
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: MCNABB

DATE/TIME OF TEST: 04/15/2026 19:08:28

Diagnostics Check	OK	19:02
Air Blank	0.000	19:02
Control Test	0.082	19:02
Air Blank	0.000	19:03
Subject Sample # 1	0.090	19:03
Breath Volume	2.246	19:03
Air Blank	0.000	19:04
Air Blank	0.000	19:06
Subject Sample # 2	0.095	19:06
Breath Volume	2.230	19:06
Air Blank	0.000	19:07
Control Test	0.081	19:07
Air Blank	0.000	19:08
Diagnostics Check	OK	19:08

Data Download Date/Time: 04/23/2026 19:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVES,JAMES,F
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: REEVES,JAMES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/25/2026 19:51:14

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.082	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.237	19:46
Breath Volume		2.167	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.228	19:49
Breath Volume		2.402	19:49
Air Blank		0.000	19:50
Control Test		0.081	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 05/29/2026 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,DANIEL,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 04/27/2026 17:04:24

Diagnostics Check	OK	17:01	
Air Blank	0.000	17:01	
Control Test	0.082	17:01	
Air Blank	0.000	17:02	
Subject Sample	# 1	17:02	Interferent Detect
Breath Volume	2.171	17:02	
Air Blank	0.000	17:03	
Control Test	0.082	17:03	
Air Blank	0.000	17:04	
Diagnostics Check	OK	17:04	

Data Download Date/Time: 05/29/2026 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FFWCC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER, WILLIAM, H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: OLIVER, WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: BUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: TREMER

DATE/TIME OF TEST: 04/27/2026 18:04:24

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.081	18:02	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:03	
Control Test	0.082	18:03	
Air Blank	0.000	18:04	
Diagnostics Check	OK	18:04	

Data Download Date/Time: 05/29/2026 13:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDWAN,NADER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: REDWAN,NADER
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: OCON

DATE/TIME OF TEST: 04/25/2026 02:43:45

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.082	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.082	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 05/23/2026 18:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REDWAN,NADER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: REDWAN,NADER
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: OCON

DATE/TIME OF TEST: 04/25/2026 02:55:18

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:52	
Control Test	0.081	02:53	
Air Blank	0.000	02:53	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 05/23/2026 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAVO,GEORGE,A
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: BANDON,ERICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 04/26/2026 21:53:13

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:46	
Control Test	0.081	21:46	
Air Blank	0.000	21:46	
Subject Sample	# 1 0.206	21:49	
Breath Volume	1.468	21:49	
Air Blank	0.000	21:49	
Air Blank	0.000	21:51	
Subject Sample	# 2	21:51	Subject Test Refused
Breath Volume	0.000	21:51	
Air Blank	0.000	21:52	
Control Test	0.080	21:52	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 05/23/2026 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JAVIER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: ROSADO,CARLOS
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/27/2026 04:52:47

Diagnostics Check	OK	04:49	
Air Blank	0.000	04:49	
Control Test	0.081	04:50	
Air Blank	0.000	04:50	
Subject Sample	# 1	04:51	Subject Test Refused
Breath Volume	0.000	04:51	
Air Blank	0.000	04:51	
Control Test	0.081	04:52	
Air Blank	0.000	04:52	
Diagnostics Check	OK	04:52	

Data Download Date/Time: 05/23/2026 18:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GONZALEZ-GARCIA

DATE/TIME OF TEST: 04/01/2026 23:16:23

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.136	23:11
Breath Volume	3.515	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.134	23:14
Breath Volume	3.726	23:14
Air Blank	0.000	23:15
Control Test	0.081	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/16/2026 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURLEY,JORDAN,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CURLEY,JORDAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 04/02/2026 00:58:51

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.207	00:54
Breath Volume		2.406	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.213	00:57
Breath Volume		1.542	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULCZAK, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LEE, BRYCE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/03/2026 20:07:01

Diagnostics Check	OK	19:59
Air Blank	0.000	20:00
Control Test	0.082	20:00
Air Blank	0.000	20:01
Subject Sample # 1	0.161	20:01
Breath Volume	1.664	20:01
Air Blank	0.000	20:02
Air Blank	0.000	20:04
Subject Sample # 2	0.151	20:05
Breath Volume	3.222	20:05
Air Blank	0.000	20:06
Control Test	0.082	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:07

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/04/2026 00:15:46

Diagnostics Check	OK	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.094	00:11
Breath Volume	2.843	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.089	00:14
Breath Volume	2.773	00:14
Air Blank	0.000	00:14
Control Test	0.081	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: JEAN-MARY

DATE/TIME OF TEST: 04/04/2026 01:42:43

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.139	01:38
Breath Volume	2.339	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.138	01:41
Breath Volume	2.410	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GAVIN,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: CRISPIN ORREGO

DATE/TIME OF TEST: 04/04/2026 03:30:18

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.175	03:25
Breath Volume	1.976	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.191	03:28
Breath Volume	2.640	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULCZAK, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HOLLYWOOD, TRENT
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: VICTORES

DATE/TIME OF TEST: 04/04/2026 03:45:46

Diagnostics Check	OK	03:39
Air Blank	0.000	03:39
Control Test	0.080	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.028	03:41
Breath Volume	2.589	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.028	03:44
Breath Volume	2.019	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GAVIN,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: PEREZ,A
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: VETTRAINO

DATE/TIME OF TEST: 04/04/2026 21:14:23

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.081	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.000	21:09
Breath Volume	2.996	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:13
Breath Volume	2.910	21:13
Air Blank	0.000	21:13
Control Test	0.081	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: JORDAN,KEVIN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: GONZALEZ MONTALVO

DATE/TIME OF TEST: 04/05/2026 02:25:32

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.177	02:21
Breath Volume		1.722	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.175	02:23
Breath Volume		1.753	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELLERS,BRANDON,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SELLERS,BRANDON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BONACOLTA

DATE/TIME OF TEST: 04/05/2026 02:35:46

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.119	02:31
Breath Volume	1.894	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.122	02:34
Breath Volume	3.984	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 04/05/2026 17:49:20

Diagnostics Check	OK	17:43
Air Blank	0.000	17:43
Control Test	0.081	17:44
Air Blank	0.000	17:44
Subject Sample # 1	0.158	17:44
Breath Volume	1.496	17:44
Air Blank	0.000	17:45
Air Blank	0.000	17:47
Subject Sample # 2	0.169	17:47
Breath Volume	1.886	17:47
Air Blank	0.000	17:48
Control Test	0.081	17:48
Air Blank	0.000	17:49
Diagnostics Check	OK	17:49

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHOEMAKER,JACLYN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER,JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BOHL

DATE/TIME OF TEST: 04/06/2026 00:04:20

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:01	
Subject Sample # 1	0.178	00:03	
Breath Volume	1.906	00:03	
Air Blank		00:03	Purge Fail
Air Blank	0.000	00:04	

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHOEMAKER,JACLYN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER,JACLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: BOHL

DATE/TIME OF TEST: 04/06/2026 00:13:09

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.161	00:08
Breath Volume		1.523	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.180	00:11
Breath Volume		2.570	00:11
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GAVIN,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: UPSHAW

DATE/TIME OF TEST: 04/06/2026 01:22:05

Diagnostics Check	OK	01:15
Air Blank	0.000	01:16
Control Test	0.081	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.192	01:17
Breath Volume	2.265	01:17
Air Blank	0.000	01:18
Air Blank	0.000	01:19
Subject Sample # 2	0.193	01:20
Breath Volume	2.214	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHOEMAKER,JACLYN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KARSNICK,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: VILLALOBOS

DATE/TIME OF TEST: 04/06/2026 02:35:54

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.081	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.165	02:31
Breath Volume		2.296	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.155	02:34
Breath Volume		2.464	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/07/2026 12:50:11

Diagnostics Check	OK	12:43
Air Blank	0.000	12:43
Control Test	0.081	12:44
Air Blank	0.000	12:44
Subject Sample # 1	0.180	12:45
Breath Volume	4.929	12:45
Air Blank	0.000	12:46
Air Blank	0.000	12:47
Subject Sample # 2	0.178	12:48
Breath Volume	5.132	12:48
Air Blank	0.000	12:49
Control Test	0.079	12:49
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AIGNER,BRIAN,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: AIGNER,BRIAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/08/2026 23:58:08

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.236	23:53
Breath Volume	2.351	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.234	23:56
Breath Volume	2.527	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/16/2026 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULCZAK, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GREEN, ALPHONSO
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CHANT

DATE/TIME OF TEST: 04/09/2026 02:15:17

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.121	02:10
Breath Volume	1.570	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.135	02:13
Breath Volume	1.488	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,CHASE,U
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: COLWELL,CHASE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/09/2026 23:50:40

Diagnostics Check	OK	23:43
Air Blank	0.000	23:44
Control Test	0.081	23:44
Air Blank	0.000	23:45
Subject Sample # 1	0.098	23:45
Breath Volume	1.292	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:48
Subject Sample # 2	0.099	23:49
Breath Volume	1.625	23:49
Air Blank	0.000	23:49
Control Test	0.082	23:50
Air Blank	0.000	23:50
Diagnostics Check	OK	23:50

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: WHITUS,OCTAVIOUS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: THELEN

DATE/TIME OF TEST: 04/11/2026 01:50:02

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.082	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.108	01:45
Breath Volume	2.195	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.108	01:48
Breath Volume	2.199	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:50

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASH,CONNOR,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HASH,CONNOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080Q2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/12/2026 02:13:17

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.081	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.116	02:05	Volume Not Met
Breath Volume		0.234	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.099	02:11	Volume Not Met
Breath Volume		0.488	02:11	
Air Blank		0.000	02:12	
Control Test		0.082	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: CHIAPPY,CHRISTOFER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 04/13/2026 01:31:43

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.185	01:27
Breath Volume	3.988	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.176	01:30
Breath Volume	2.597	01:30
Air Blank	0.000	01:30
Control Test	0.078	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WITTBOLD

DATE/TIME OF TEST: 04/15/2026 21:35:08

Diagnostics Check	OK	21:33	
Air Blank	0.000	21:33	
Control Test	0.082	21:34	
Air Blank		21:34	RFI Detect
Air Blank	0.000	21:35	

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WITTBOLD

DATE/TIME OF TEST: 04/15/2026 21:41:16

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:37	
Control Test	0.080	21:37	
Air Blank	0.000	21:38	
Subject Sample	# 1 0.248	21:39	
Breath Volume	2.882	21:39	
Air Blank		21:40	Purge Fail
Air Blank		21:41	Purge Fail

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WITTBOLD

DATE/TIME OF TEST: 04/15/2026 21:44:35

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.072	21:43	Control Outside Tolerance
Air Blank	0.000	21:44	

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WITTBOLD

DATE/TIME OF TEST: 04/15/2026 21:49:35

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:46	
Control Test	0.075	21:47	
Air Blank	0.000	21:47	
Subject Sample # 1	0.237	21:48	
Breath Volume	2.097	21:48	
Air Blank		21:49	Purge Fail
Air Blank	0.000	21:49	

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DRESSEL,GREGORY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: WITTBOLD

DATE/TIME OF TEST: 04/15/2026 21:55:42

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.075	21:51	
Air Blank	0.000	21:52	
Subject Sample # 1	0.203	21:54	
Breath Volume	1.410	21:54	
Air Blank		21:54	Purge Fail
Air Blank	0.000	21:55	

Data Download Date/Time: 04/16/2026 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALCON,J,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GARLICK,R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: SANCHEZ GOMEZ

DATE/TIME OF TEST: 04/17/2026 01:49:54

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Subject Sample # 1	0.195	01:45
Breath Volume	2.386	01:45
Air Blank	0.000	01:45
Air Blank	0.000	01:47
Subject Sample # 2	0.205	01:48
Breath Volume	2.187	01:48
Air Blank	0.000	01:48
Control Test	0.078	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 05/26/2026 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BLAKE,DAVID
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: VARNADORE

DATE/TIME OF TEST: 04/17/2026 18:42:18

Diagnostics Check	OK	18:36
Air Blank	0.000	18:36
Control Test	0.080	18:37
Air Blank	0.000	18:37
Subject Sample # 1	0.000	18:38
Breath Volume	2.582	18:38
Air Blank	0.000	18:38
Air Blank	0.000	18:40
Subject Sample # 2	0.000	18:40
Breath Volume	2.574	18:40
Air Blank	0.000	18:41
Control Test	0.080	18:41
Air Blank	0.000	18:42
Diagnostics Check	OK	18:42

Data Download Date/Time: 05/26/2026 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEE,BRYCE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: FLORES ARGUETA

DATE/TIME OF TEST: 04/17/2026 22:04:22

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:51	
Control Test		0.079	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 1	0.269	21:53	No .020 agreement
Breath Volume		1.355	21:53	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 2	0.237	21:58	Volume Not Met
Breath Volume		0.000	21:58	
Air Blank		0.000	21:59	
Air Blank		0.000	22:01	
Subject Sample	# 3	0.247	22:02	No .020 agreement
Breath Volume		1.117	22:02	
Air Blank		0.000	22:03	
Control Test		0.076	22:03	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 05/26/2026 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REICHEL,T,H
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: ITZEP CASTRO

DATE/TIME OF TEST: 04/17/2026 23:06:51

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.181	23:02
Breath Volume	2.714	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.178	23:05
Breath Volume	2.339	23:05
Air Blank	0.000	23:05
Control Test	0.078	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 05/26/2026 09:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HERNANDEZ GRAJALES

DATE/TIME OF TEST: 04/18/2026 02:42:48

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.212	02:38
Breath Volume	2.378	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.219	02:41
Breath Volume	2.812	02:41
Air Blank	0.000	02:41
Control Test	0.078	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/26/2026 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GAVIN,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,GAVIN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PAULEON

DATE/TIME OF TEST: 04/18/2026 03:37:53

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.149	03:33
Breath Volume	1.996	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.149	03:36
Breath Volume	2.066	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/26/2026 09:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: LOPEZ ROBLERO

DATE/TIME OF TEST: 04/18/2026 03:48:00

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.078	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.231	03:43
Breath Volume	1.851	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.228	03:46
Breath Volume	3.468	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:48

Data Download Date/Time: 05/26/2026 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,DAVID,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ALFONSI,ANTHONY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 04/19/2026 01:37:48

Diagnostics Check		OK	01:31
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.201	01:33
Breath Volume		1.464	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.193	01:36
Breath Volume		1.464	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 05/26/2026 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,MIGUEL,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GUTIERREZ,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: AUDIAS MARTIN

DATE/TIME OF TEST: 04/19/2026 03:28:24

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.136	03:23
Breath Volume	1.644	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.132	03:26
Breath Volume	3.703	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 05/26/2026 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/19/2026 03:36:31

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.133	03:31
Breath Volume	2.800	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.133	03:34
Breath Volume	1.824	03:34
Air Blank	0.000	03:35
Control Test	0.079	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 05/26/2026 09:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELL,AARON,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ANGELL,AARON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: PILOT

DATE/TIME OF TEST: 04/22/2026 16:25:17

Diagnostics Check	OK	16:18
Air Blank	0.000	16:19
Control Test	0.078	16:19
Air Blank	0.000	16:20
Subject Sample # 1	0.166	16:20
Breath Volume	1.812	16:20
Air Blank	0.000	16:21
Air Blank	0.000	16:23
Subject Sample # 2	0.160	16:23
Breath Volume	1.722	16:23
Air Blank	0.000	16:24
Control Test	0.076	16:24
Air Blank	0.000	16:25
Diagnostics Check	OK	16:25

Data Download Date/Time: 05/26/2026 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHOEMAKER,JACLYN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SHOEMAKER,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: QUEIPO SABORIT

DATE/TIME OF TEST: 04/23/2026 01:52:43

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.101	01:48
Breath Volume	3.042	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:50
Subject Sample # 2	0.100	01:51
Breath Volume	3.000	01:51
Air Blank	0.000	01:51
Control Test	0.077	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/25/2026 00:49:29

Diagnostics Check	OK	00:42
Air Blank	0.000	00:43
Control Test	0.080	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.152	00:44
Breath Volume	3.410	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.144	00:47
Breath Volume	3.328	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:49
Diagnostics Check	OK	00:49

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WHITUS,OCTAVIOUS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/25/2026 01:32:33

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.079	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.000	01:28
Breath Volume	1.902	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.000	01:31
Breath Volume	1.671	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANTIBANEZ, D
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/25/2026 11:28:07

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.079	11:22
Air Blank	0.000	11:22
Subject Sample # 1	0.110	11:23
Breath Volume	1.269	11:23
Air Blank	0.000	11:24
Air Blank	0.000	11:25
Subject Sample # 2	0.113	11:26
Breath Volume	2.148	11:26
Air Blank	0.000	11:27
Control Test	0.079	11:27
Air Blank	0.000	11:28
Diagnostics Check	OK	11:28

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,C,S
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MORALES,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DUPUIS

DATE/TIME OF TEST: 04/26/2026 22:27:50

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1 0.214	22:19	Volume Not Met
Breath Volume	1.097	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2 0.202	22:23	
Breath Volume	1.503	22:23	
Air Blank	0.000	22:23	
Air Blank	0.000	22:25	
Subject Sample	# 3 0.205	22:26	
Breath Volume	2.476	22:26	
Air Blank	0.000	22:26	
Control Test	0.074	22:27	Control Outside Tolerance
Air Blank	0.000	22:27	

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,C,S
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MORALES,C
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DUPUIS

DATE/TIME OF TEST: 04/26/2026 22:39:06

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.077	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1 0.208	22:32	
Breath Volume	2.125	22:32	
Air Blank	0.000	22:32	
Air Blank	0.000	22:34	
Subject Sample	# 2 0.186	22:35	
Breath Volume	1.812	22:35	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 3	22:37	Subject Test Refused
Breath Volume	0.000	22:37	
Air Blank	0.000	22:38	
Control Test	0.077	22:38	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: CHIAPPY,CHRISTOFER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SLAJDA

DATE/TIME OF TEST: 04/27/2026 00:39:38

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:28	
Control Test		0.079	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1	0.197	00:32	Volume Not Met
Breath Volume		0.105	00:32	
Air Blank		0.000	00:33	
Air Blank		0.000	00:35	
Subject Sample	# 2	0.174	00:38	Volume Not Met
Breath Volume		0.683	00:38	
Air Blank			00:39	Purge Fail
Air Blank		0.000	00:39	

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: CHIAPPY,CHRISTOFER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SLAJDA

DATE/TIME OF TEST: 04/27/2026 00:43:31

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.073	00:42	Control Outside Tolerance
Air Blank	0.000	00:43	

Data Download Date/Time: 05/26/2026 09:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHIAPPY,CHRISTOFER,J
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: CHIAPPY,CHRISTOFER
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SLAJDA

DATE/TIME OF TEST: 04/27/2026 00:46:45

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:45	
Control Test	0.074	00:46	Control Outside Tolerance
Air Blank	0.000	00:46	

Data Download Date/Time: 05/26/2026 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULCZAK, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FEIST, CASEY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: BEIHOFF

DATE/TIME OF TEST: 04/28/2026 02:39:57

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.147	02:35
Breath Volume	1.468	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.157	02:38
Breath Volume	1.628	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/26/2026 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, TRAEVONNE, M
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: ROGERS, KAITLYN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29425080A2
DRY GAS EXPIRY DATE: 12/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ORTIZ MEJIA

DATE/TIME OF TEST: 04/30/2026 01:15:08

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.255	01:10
Breath Volume		1.304	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.269	01:13
Breath Volume		1.480	01:13
Air Blank		0.000	01:14
Control Test		0.077	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/26/2026 09:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTABLE,AVRY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: REIFF,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: AHLBORN

DATE/TIME OF TEST: 04/26/2026 03:11:35

Diagnostics Check	OK	03:03
Air Blank	0.000	03:04
Control Test	0.081	03:04
Air Blank	0.000	03:05
Subject Sample # 1	0.108	03:06
Breath Volume	1.656	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.114	03:09
Breath Volume	2.316	03:09
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 05/30/2026 05:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVACK,B
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: TARAZONA

DATE/TIME OF TEST: 04/27/2026 01:03:38

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.129	00:58
Breath Volume		1.492	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.129	01:01
Breath Volume		1.968	01:01
Air Blank		0.000	01:02
Control Test		0.081	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/30/2026 05:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDENAS,L,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KNODLE,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: ENDRES

DATE/TIME OF TEST: 04/29/2026 01:16:25

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.098	01:08
Breath Volume		1.578	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.101	01:14
Breath Volume		2.351	01:14
Air Blank		0.000	01:15
Control Test		0.081	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 05/30/2026 05:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,C,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMYRIOS,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 04/01/2026 01:10:33

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.000	01:06
Breath Volume	2.371	01:06
Air Blank	0.000	01:06
Air Blank	0.000	01:08
Subject Sample # 2	0.000	01:09
Breath Volume	2.261	01:09
Air Blank	0.000	01:09
Control Test	0.079	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,TINA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ARCILA

DATE/TIME OF TEST: 04/02/2026 03:37:04

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.096	03:32
Breath Volume		4.230	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 2	0.095	03:35
Breath Volume		4.289	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: EMILIUS,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: GODINEZ-PEREZ

DATE/TIME OF TEST: 04/04/2026 03:32:14

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.186	03:27
Breath Volume	1.738	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.186	03:30
Breath Volume	1.558	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDON,QUINLAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OGDON,QUINLAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CASTILLO-ESPINOSA

DATE/TIME OF TEST: 04/09/2026 00:29:32

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.042	00:25
Breath Volume	3.429	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.039	00:28
Breath Volume	2.632	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYMAN,P,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HEATON,JACKSON
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ANDRES

DATE/TIME OF TEST: 04/11/2026 22:43:38

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.236	22:37
Breath Volume	2.304	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.225	22:42
Breath Volume	1.183	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPIN,TINA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ESPIN,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MORALESHERNANDEZ

DATE/TIME OF TEST: 04/12/2026 04:39:10

Diagnostics Check	OK	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.036	04:34
Breath Volume	1.785	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:37
Subject Sample # 2	0.037	04:37
Breath Volume	2.152	04:37
Air Blank	0.000	04:38
Control Test	0.079	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICE,MARCUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PHILLIPS,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DIAZDOMINGO

DATE/TIME OF TEST: 04/20/2026 00:18:12

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.277	00:13
Breath Volume	2.207	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.290	00:16
Breath Volume	1.816	00:16
Air Blank	0.000	00:17
Control Test	0.078	00:17
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUNN,GATLIN,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAIL,MATTHEW
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MONTEJODIAZ

DATE/TIME OF TEST: 04/22/2026 00:56:57

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1 0.164	00:51	
Breath Volume	1.222	00:51	
Air Blank	0.000	00:51	
Air Blank	0.000	00:53	
Subject Sample	# 2	00:55	Subject Test Refused
Breath Volume	0.187	00:55	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYMAN,PATRICK,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARZICOLE,GABRIELLE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: LOPEZLAINES

DATE/TIME OF TEST: 04/26/2026 03:57:32

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.139	03:53
Breath Volume	2.507	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.137	03:56
Breath Volume	2.734	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERRE,KRENID,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TARR,LOGAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BONIFACIO

DATE/TIME OF TEST: 04/26/2026 04:13:02

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.189	04:08
Breath Volume	3.855	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.188	04:11
Breath Volume	3.074	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

Data Download Date/Time: 04/26/2026 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KEITER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: MANCUSO

DATE/TIME OF TEST: 04/20/2026 03:23:06

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.088	03:16
Breath Volume	1.324	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.088	03:21
Breath Volume	1.933	03:21
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MARISSA,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SHAFFER,H
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: VANDERPOOL

DATE/TIME OF TEST: 04/20/2026 21:51:56

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.119	21:47
Breath Volume	1.960	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.118	21:50
Breath Volume	1.335	21:50
Air Blank	0.000	21:51
Control Test	0.080	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PANNULLO,G
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 04/20/2026 23:25:44

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.278	23:20
Breath Volume		2.226	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.279	23:24
Breath Volume		3.167	23:24
Air Blank		0.000	23:24
Control Test		0.080	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KEITER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RIMES

DATE/TIME OF TEST: 04/23/2026 02:37:39

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.096	02:33
Breath Volume	2.039	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.097	02:36
Breath Volume	1.496	02:36
Air Blank	0.000	02:36
Control Test	0.079	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NAZZARESE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: BURROUGHS

DATE/TIME OF TEST: 04/23/2026 23:10:19

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.213	23:05
Breath Volume	2.195	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.208	23:08
Breath Volume	1.312	23:08
Air Blank	0.000	23:09
Control Test	0.080	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,JUAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROSADO,JUAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CHAJ XEC

DATE/TIME OF TEST: 04/24/2026 00:56:52

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.189	00:52
Breath Volume	2.414	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:54
Subject Sample # 2	0.187	00:55
Breath Volume	2.503	00:55
Air Blank	0.000	00:55
Control Test	0.078	00:56
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 06/08/2026 12:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONN,CLARENCE,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRADY,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: HEMINGER

DATE/TIME OF TEST: 04/24/2026 01:17:07

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.140	01:12
Breath Volume	1.593	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.129	01:15
Breath Volume	2.152	01:15
Air Blank	0.000	01:16
Control Test	0.078	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 06/08/2026 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MARISSA,A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DELHOMME,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 04/25/2026 20:58:00

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.000	20:53
Breath Volume	2.468	20:53
Air Blank	0.000	20:54
Air Blank	0.000	20:56
Subject Sample # 2	0.000	20:56
Breath Volume	1.410	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:57
Diagnostics Check	OK	20:58

Data Download Date/Time: 06/08/2026 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GONZALEZ,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: COPELAND

DATE/TIME OF TEST: 04/26/2026 01:02:07

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.153	00:57
Breath Volume	1.820	00:57
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.163	01:00
Breath Volume	3.132	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 06/08/2026 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILES,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEIHS,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DUFRENE

DATE/TIME OF TEST: 04/26/2026 04:17:51

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.097	04:13
Breath Volume	1.582	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.091	04:16
Breath Volume	1.156	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 06/08/2026 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDROP,P,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LABOUNTY,DFC
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: JAGNECKI

DATE/TIME OF TEST: 04/30/2026 02:07:26

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.197	02:02
Breath Volume	2.308	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.207	02:05
Breath Volume	2.191	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 06/08/2026 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALLADINO,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: GRIFFITH,NYKOLAS
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: TERNS

DATE/TIME OF TEST: 04/07/2026 13:25:28

Diagnostics Check	OK	13:19
Air Blank	0.000	13:19
Control Test	0.078	13:20
Air Blank	0.000	13:20
Subject Sample # 1	0.000	13:21
Breath Volume	2.339	13:21
Air Blank	0.000	13:21
Air Blank	0.000	13:23
Subject Sample # 2	0.000	13:24
Breath Volume	3.343	13:24
Air Blank	0.000	13:24
Control Test	0.079	13:24
Air Blank	0.000	13:25
Diagnostics Check	OK	13:25

Data Download Date/Time: 04/07/2026 16:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MARTINEZ,EDUARDO
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: ADMIN
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 04/08/2026 02:10:43

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.168	02:06
Breath Volume	2.742	02:06
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.178	02:09
Breath Volume	3.578	02:09
Air Blank	0.000	02:09
Control Test	0.078	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/16/2026 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: ADMIN
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:39

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 04/08/2026 10:24:48

Diagnostics Check	OK	10:18
Air Blank	0.000	10:19
Control Test	0.079	10:19
Air Blank	0.000	10:20
Subject Sample # 1	0.016	10:20
Breath Volume	2.644	10:20
Air Blank	0.000	10:21
Air Blank	0.000	10:22
Subject Sample # 2	0.016	10:23
Breath Volume	3.144	10:23
Air Blank	0.000	10:23
Control Test	0.077	10:24
Air Blank	0.000	10:24
Diagnostics Check	OK	10:24

Data Download Date/Time: 05/16/2026 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCALERA,MARIANA,Y
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MARTINEZ,EDUARDO
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: ADMIN
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CUELLAR MEDINA

DATE/TIME OF TEST: 04/13/2026 00:56:13

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.196	00:51
Breath Volume	2.378	00:51
Air Blank	0.000	00:52
Air Blank	0.000	00:53
Subject Sample # 2	0.187	00:54
Breath Volume	1.253	00:54
Air Blank	0.000	00:55
Control Test	0.077	00:55
Air Blank	0.000	00:56
Diagnostics Check	OK	00:56

Data Download Date/Time: 05/16/2026 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: ORAMAS

DATE/TIME OF TEST: 04/13/2026 07:33:05

Diagnostics Check	OK	07:26
Air Blank	0.000	07:27
Control Test	0.079	07:27
Air Blank	0.000	07:28
Subject Sample # 1	0.063	07:28
Breath Volume	1.742	07:28
Air Blank	0.000	07:29
Air Blank	0.000	07:30
Subject Sample # 2	0.067	07:31
Breath Volume	1.578	07:31
Air Blank	0.000	07:32
Control Test	0.077	07:32
Air Blank	0.000	07:32
Diagnostics Check	OK	07:33

Data Download Date/Time: 05/16/2026 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARNOLD,TRAVIS,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: BARTON,LORA
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/07/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03326080A1
DRY GAS EXPIRY DATE: 03/05/2028
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: TRUONG

DATE/TIME OF TEST: 04/14/2026 17:22:05

Diagnostics Check	OK	17:15
Air Blank	0.000	17:16
Control Test	0.078	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.000	17:17
Breath Volume	1.980	17:17
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.000	17:20
Breath Volume	1.839	17:20
Air Blank	0.000	17:21
Control Test	0.077	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 05/16/2026 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROHECKER,KYNDAL,C
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: THARPE,BLAKE
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 03/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/06/2026 16:56:36

Diagnostics Check	OK	16:50
Air Blank	0.000	16:50
Control Test	0.078	16:51
Air Blank	0.000	16:51
Subject Sample # 1	0.084	16:52
Breath Volume	1.578	16:52
Air Blank	0.000	16:52
Air Blank	0.000	16:54
Subject Sample # 2	0.084	16:55
Breath Volume	1.648	16:55
Air Blank	0.000	16:55
Control Test	0.078	16:56
Air Blank	0.000	16:56
Diagnostics Check	OK	16:56

Data Download Date/Time: 04/08/2026 09:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NASH,CAMERON,L
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: SHIVER,ROBERT
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/13/2026 04:06:31

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:57
Subject Sample	# 1	0.193	03:58
Breath Volume		1.824	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.219	04:01
Breath Volume		1.558	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 3	0.233	04:04
Breath Volume		2.757	04:04
Air Blank		0.000	04:05
Control Test		0.078	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 05/18/2026 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHEA, COLBY, R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BETHEA, COLBY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 04/10/2026 14:36:32

Diagnostics Check	OK	14:30
Air Blank	0.000	14:30
Control Test	0.079	14:31
Air Blank	0.000	14:31
Subject Sample # 1	0.000	14:32
Breath Volume	2.445	14:32
Air Blank	0.000	14:32
Air Blank	0.000	14:34
Subject Sample # 2	0.000	14:35
Breath Volume	2.296	14:35
Air Blank	0.000	14:35
Control Test	0.078	14:35
Air Blank	0.000	14:36
Diagnostics Check	OK	14:36

Data Download Date/Time: 05/05/2026 07:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS, JASON, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LONG, KATHLENE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/19/2026 08:50:55

Diagnostics Check	OK	08:49	
Air Blank		08:50	Ambient Fail
Air Blank	0.000	08:50	

Data Download Date/Time: 05/05/2026 07:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASS,JASON,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LONG,KATHLENE
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:28

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 04/19/2026 09:01:25

Diagnostics Check	OK	08:55
Air Blank	0.000	08:55
Control Test	0.079	08:55
Air Blank	0.000	08:56
Subject Sample # 1	0.210	08:56
Breath Volume	2.339	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.196	08:59
Breath Volume	1.597	08:59
Air Blank	0.000	09:00
Control Test	0.077	09:00
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 05/05/2026 07:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEXANDER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/05/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 04/01/2026 01:35:06

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.162	01:30
Breath Volume	2.636	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.160	01:33
Breath Volume	1.738	01:33
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/01/2026 07:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: MARQUES,MARIA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BALTIMORE

DATE/TIME OF TEST: 04/02/2026 03:33:14

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.292	03:28
Breath Volume	2.113	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.293	03:31
Breath Volume	1.621	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 05/04/2026 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHELPS,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COLLINS,A
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SEGURA-DIAZ

DATE/TIME OF TEST: 04/05/2026 01:26:05

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.145	01:21
Breath Volume	2.324	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.149	01:24
Breath Volume	2.363	01:24
Air Blank	0.000	01:25
Control Test	0.080	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/04/2026 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEX,M
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SMITH,ALEX
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 04/09/2026 02:35:25

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.238	02:30
Breath Volume		2.902	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.227	02:33
Breath Volume		1.558	02:33
Air Blank		0.000	02:34
Control Test		0.081	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 05/04/2026 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL, EDWARD,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DAWSON, A
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/12/2026 03:31:25

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.189	03:27
Breath Volume	2.351	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.181	03:29
Breath Volume	2.371	03:29
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/04/2026 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BABB,CALEB,W
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BABB,CALEB
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 04/13/2026 04:54:56

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.081	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.163	04:49
Breath Volume	2.246	04:49
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.154	04:53
Breath Volume	1.542	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: GRANT,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/13/2026 23:47:59

Diagnostics Check	OK	23:41
Air Blank	0.000	23:42
Control Test	0.082	23:42
Air Blank	0.000	23:43
Subject Sample # 1	0.125	23:43
Breath Volume	1.953	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.124	23:46
Breath Volume	2.792	23:46
Air Blank	0.000	23:47
Control Test	0.081	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,DANIEL,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BROWN,DANIEL
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 11:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/14/2026 13:23:36

Diagnostics Check	OK	13:17
Air Blank	0.000	13:17
Control Test	0.081	13:18
Air Blank	0.000	13:18
Subject Sample # 1	0.332	13:19
Breath Volume	1.761	13:19
Air Blank	0.000	13:19
Air Blank	0.000	13:21
Subject Sample # 2	0.334	13:22
Breath Volume	1.312	13:22
Air Blank	0.000	13:22
Control Test	0.080	13:23
Air Blank	0.000	13:23
Diagnostics Check	OK	13:23

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LEHMANN,SHANE
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: KISKADEN

DATE/TIME OF TEST: 04/16/2026 03:50:03

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.160	03:45
Breath Volume	1.226	03:45
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.155	03:48
Breath Volume	1.265	03:48
Air Blank	0.000	03:49
Control Test	0.081	03:49
Air Blank	0.000	03:49
Diagnostics Check	OK	03:50

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ROJO AGUILAR

DATE/TIME OF TEST: 04/20/2026 04:22:56

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.082	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.099	04:18
Breath Volume	1.179	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.098	04:21
Breath Volume	1.386	04:21
Air Blank	0.000	04:22
Control Test	0.082	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUSH, TYLER, C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: GODWIN, BOBBY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/21/2026 02:46:52

Diagnostics Check	OK	02:41	
Air Blank	0.000	02:41	
Control Test	0.082	02:42	
Air Blank	0.000	02:42	
Subject Sample	# 1 0.113	02:43	
Breath Volume	3.019	02:43	
Air Blank	0.000	02:43	
Air Blank	0.000	02:45	
Subject Sample	# 2 0.113	02:45	
Breath Volume	1.648	02:45	
Air Blank	0.000	02:46	
		02:46	RFI Detect
Air Blank		02:46	RFI Detect

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHRITENOUR,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DAWSON,ANDREW
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 04/21/2026 04:29:29

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.082	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.122	04:24
Breath Volume	4.410	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.122	04:27
Breath Volume	3.164	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GERGEN,PATRICK,S
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: GERGEN,PATRICK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: JEREZ

DATE/TIME OF TEST: 04/24/2026 23:58:43

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.081	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.151	23:54
Breath Volume	1.261	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.145	23:57
Breath Volume	1.886	23:57
Air Blank	0.000	23:57
Control Test	0.080	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HALL,NOAH
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: STROM

DATE/TIME OF TEST: 04/26/2026 04:45:29

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.116	04:40
Breath Volume	1.269	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:42
Subject Sample # 2	0.127	04:43
Breath Volume	1.214	04:43
Air Blank	0.000	04:44
Control Test	0.080	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,NOAH,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: DAWSON,ANDREW
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BUCHANAN

DATE/TIME OF TEST: 04/27/2026 04:14:28

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.058	04:10
Breath Volume		2.433	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.058	04:12
Breath Volume		2.460	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,AARON,C
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: GRANT,JAMES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 12:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/29/2026 01:53:37

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.082	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.215	01:49
Breath Volume		4.074	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.218	01:52
Breath Volume		3.160	01:52
Air Blank		0.000	01:52
Control Test		0.082	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 05/04/2026 08:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTON,JEREMY,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BARTON,JEREMY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GANN

DATE/TIME OF TEST: 04/29/2026 04:21:32

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.081	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.124	04:17
Breath Volume	3.886	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.123	04:19
Breath Volume	3.339	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 05/04/2026 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,NATHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SABA,DUSTIN
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: RASOOL

DATE/TIME OF TEST: 04/28/2026 00:01:26

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.109	23:56
Breath Volume	2.824	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.109	23:59
Breath Volume	3.082	23:59
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 05/06/2026 07:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/02/2026 03:01:32

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.221	02:56
Breath Volume	3.226	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.216	03:00
Breath Volume	2.640	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/04/2026 01:48:02

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Subject Sample # 1	0.122	01:43
Breath Volume	1.210	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.136	01:46
Breath Volume	1.312	01:46
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:48

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROCCHETTI,LEONEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GALEANOARGUETA

DATE/TIME OF TEST: 04/04/2026 03:37:36

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.195	03:33
Breath Volume	3.089	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.192	03:36
Breath Volume	3.097	03:36
Air Blank	0.000	03:36
Control Test	0.078	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/04/2026 04:04:37

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:59
Air Blank	0.000	03:59
Subject Sample # 1	0.155	04:00
Breath Volume	4.082	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.144	04:03
Breath Volume	2.167	04:03
Air Blank	0.000	04:03
Control Test	0.078	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 04/05/2026 02:49:31

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.076	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.180	02:45
Breath Volume	3.109	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.179	02:48
Breath Volume	2.515	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FRANCO,HENRY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/06/2026 22:40:21

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.146	22:36
Breath Volume	3.671	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.148	22:38
Breath Volume	3.468	22:38
Air Blank	0.000	22:39
Control Test	0.080	22:39
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,GIAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/07/2026 02:22:22

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.211	02:18
Breath Volume		1.476	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.215	02:20
Breath Volume		1.625	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/08/2026 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: OLEA BARAJAS

DATE/TIME OF TEST: 04/09/2026 00:33:31

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.180	00:29
Breath Volume	3.878	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.174	00:32
Breath Volume	3.445	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:32
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 05/14/2026 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: STAFFINE

DATE/TIME OF TEST: 04/09/2026 02:50:10

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:39	
Subject Sample	# 1	0.115	02:42	Volume Not Met
Breath Volume		0.183	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 2	0.099	02:48	Volume Not Met
Breath Volume		0.558	02:48	
Air Blank		0.000	02:49	
Control Test		0.079	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 05/14/2026 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: STAFFINE

DATE/TIME OF TEST: 04/09/2026 03:06:27

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:52	
Control Test		0.077	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.112	02:56	Volume Not Met
Breath Volume		0.441	02:56	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2	0.105	02:59	
Breath Volume		1.117	02:59	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 3	0.095	03:04	Volume Not Met
Breath Volume		0.199	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 05/14/2026 13:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,PEDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HILL III

DATE/TIME OF TEST: 04/09/2026 03:42:58

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.078	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:36	No Sample Provided
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Air Blank	0.000	03:38	
Subject Sample	# 2	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 05/14/2026 13:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: APONTE,MIGUEL
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LA RATONDO JR

DATE/TIME OF TEST: 04/09/2026 04:16:14

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:15	
Control Test	0.080	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DWYER,NICHOLAS
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: SORTO III

DATE/TIME OF TEST: 04/09/2026 23:52:11

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:45	No Sample Provided
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2	23:50	No Sample Provided
Breath Volume	0.000	23:50	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYES,JONATHAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FANA DISLA

DATE/TIME OF TEST: 04/10/2026 01:00:14

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.203	00:55
Breath Volume		2.035	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.196	00:58
Breath Volume		2.035	00:58
Air Blank		0.000	00:59
Control Test		0.078	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALTINERLIELMAS,ADEM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SOUCY

DATE/TIME OF TEST: 04/10/2026 03:55:28

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.110	03:51
Breath Volume		1.613	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.105	03:54
Breath Volume		2.488	03:54
Air Blank		0.000	03:54
Control Test		0.078	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: BEATH

DATE/TIME OF TEST: 04/10/2026 04:10:11

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.152	04:05
Breath Volume	4.308	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.155	04:08
Breath Volume	3.402	04:08
Air Blank	0.000	04:09
Control Test	0.078	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUIZ,ANDRES
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: FOLLMER

DATE/TIME OF TEST: 04/11/2026 02:02:20

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CARRASQUILLO,JOHN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FARGAS

DATE/TIME OF TEST: 04/11/2026 02:31:17

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:28	
Control Test	0.080	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:30	
Control Test	0.079	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: UZCATEGUI,ANDRES
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RAMJISINGH

DATE/TIME OF TEST: 04/11/2026 02:42:24

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.080	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:38
Breath Volume	2.667	02:38
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:41
Breath Volume	2.601	02:41
Air Blank	0.000	02:41
Control Test	0.080	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 04/11/2026 04:30:19

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.040	04:25
Breath Volume	4.777	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.041	04:28
Breath Volume	4.875	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 04/12/2026 01:55:39

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.178	01:50
Breath Volume		1.402	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.169	01:54
Breath Volume		1.113	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: STURDIFEN, FLOYED
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ANGLERO

DATE/TIME OF TEST: 04/12/2026 03:35:47

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.168	03:31
Breath Volume		3.980	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.167	03:34
Breath Volume		4.007	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELCHER,ALEX
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/12/2026 04:33:47

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.190	04:28
Breath Volume	2.164	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.193	04:32
Breath Volume	1.316	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: BATIE JR

DATE/TIME OF TEST: 04/12/2026 20:44:30

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.081	20:38
Air Blank	0.000	20:39
Subject Sample # 1	0.030	20:39
Breath Volume	1.738	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.029	20:43
Breath Volume	3.253	20:43
Air Blank	0.000	20:43
Control Test	0.081	20:43
Air Blank	0.000	20:44
Diagnostics Check	OK	20:44

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: JABOUIN

DATE/TIME OF TEST: 04/13/2026 02:58:43

Diagnostics Check		OK	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.018	02:54
Breath Volume		1.226	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.018	02:57
Breath Volume		1.464	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: PARKINS

DATE/TIME OF TEST: 04/13/2026 22:48:27

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:37	
Control Test	0.080	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:41	No Sample Provided
Breath Volume	0.000	22:41	
Air Blank	0.000	22:41	
Air Blank	0.000	22:43	
Subject Sample	# 2	22:47	No Sample Provided
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Control Test	0.081	22:47	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIERRA,EDGAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BIRMINGHAM

DATE/TIME OF TEST: 04/14/2026 21:47:02

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.132	21:42
Breath Volume	2.378	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.144	21:45
Breath Volume	2.570	21:45
Air Blank	0.000	21:46
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:47

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAMANNA,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/14/2026 22:26:07

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.081	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.227	22:21
Breath Volume	1.546	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.215	22:24
Breath Volume	2.500	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CASTILLO,WILFRED
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:46

SUBJECT'S LAST NAME: CONEY

DATE/TIME OF TEST: 04/15/2026 09:18:41

Diagnostics Check		OK	09:11
Air Blank		0.000	09:12
Control Test		0.081	09:12
Air Blank		0.000	09:13
Subject Sample	# 1	0.106	09:13
Breath Volume		1.300	09:13
Air Blank		0.000	09:14
Air Blank		0.000	09:15
Subject Sample	# 2	0.108	09:17
Breath Volume		1.433	09:17
Air Blank		0.000	09:17
Control Test		0.080	09:18
Air Blank		0.000	09:18
Diagnostics Check		OK	09:18

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 04/16/2026 04:31:16

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.000	04:26
Breath Volume	4.414	04:26
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.000	04:29
Breath Volume	4.496	04:29
Air Blank	0.000	04:30
Control Test	0.080	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, ANTONIO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: ELLER

DATE/TIME OF TEST: 04/17/2026 00:02:53

Diagnostics Check	OK	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.228	23:58
Breath Volume	2.363	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.231	00:01
Breath Volume	3.308	00:01
Air Blank	0.000	00:01
Control Test	0.079	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MAINE,VINCENT
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 04/17/2026 02:41:27

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Subject Sample # 1	0.150	02:36
Breath Volume	3.843	02:36
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.148	02:39
Breath Volume	3.890	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/17/2026 04:55:11

Diagnostics Check	OK	04:48
Air Blank	0.000	04:48
Control Test	0.081	04:49
Air Blank	0.000	04:49
Subject Sample # 1	0.133	04:50
Breath Volume	2.058	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.132	04:53
Breath Volume	2.371	04:53
Air Blank	0.000	04:54
Control Test	0.079	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 05/14/2026 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GONZALEZPEREZ

DATE/TIME OF TEST: 04/17/2026 23:09:44

Diagnostics Check		OK	23:03
Air Blank		0.000	23:04
Control Test		0.081	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.223	23:05
Breath Volume		3.269	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.218	23:08
Breath Volume		3.425	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLYTH, KYLE
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: REYES NICANOR

DATE/TIME OF TEST: 04/18/2026 00:34:13

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.211	00:29
Breath Volume		2.613	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.211	00:32
Breath Volume		2.128	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/18/2026 03:52:02

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:40	
Control Test		0.080	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 1	0.134	03:44	Volume Not Met
Breath Volume		0.820	03:44	
Air Blank		0.000	03:45	
Air Blank		0.000	03:46	
Subject Sample	# 2	0.132	03:47	
Breath Volume		1.531	03:47	
Air Blank		0.000	03:47	
Air Blank		0.000	03:49	
Subject Sample	# 3	0.128	03:50	
Breath Volume		1.628	03:50	
Air Blank		0.000	03:51	
Control Test		0.079	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:52	

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRILEY, AUSTIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 04/18/2026 04:42:45

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.095	04:35	Volume Not Met
Breath Volume		0.015	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 2	0.089	04:41	Volume Not Met
Breath Volume		0.589	04:41	
Air Blank		0.000	04:41	
Control Test		0.078	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/18/2026 05:17:20

Diagnostics Check		OK	05:10
Air Blank		0.000	05:11
Control Test		0.078	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.158	05:12
Breath Volume		1.410	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.168	05:15
Breath Volume		4.011	05:15
Air Blank		0.000	05:16
Control Test		0.079	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS, BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/18/2026 21:56:40

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:46	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:49	No Sample Provided
Breath Volume	0.000	21:49	
Air Blank	0.000	21:50	
Air Blank	0.000	21:52	
Subject Sample	# 2	21:55	No Sample Provided
Breath Volume	0.000	21:55	
Air Blank	0.000	21:55	
Control Test	0.081	21:56	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II, DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 04/19/2026 03:32:54

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Subject Sample # 1	0.147	03:28
Breath Volume	1.894	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.140	03:31
Breath Volume	1.757	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUINONES,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: RUIZGONZALEZ

DATE/TIME OF TEST: 04/19/2026 07:59:24

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:57	
Control Test	0.081	07:57	
Air Blank	0.000	07:57	
Subject Sample	# 1	07:58	Subject Test Refused
Breath Volume	0.000	07:58	
Air Blank	0.000	07:58	
Control Test	0.080	07:58	
Air Blank	0.000	07:59	
Diagnostics Check	OK	07:59	

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ,PAUL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: DOMINGO

DATE/TIME OF TEST: 04/19/2026 20:49:14

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.081	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.220	20:44
Breath Volume	3.636	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.215	20:47
Breath Volume	1.402	20:47
Air Blank	0.000	20:48
Control Test	0.080	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS, JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: REYES TINEO

DATE/TIME OF TEST: 04/19/2026 21:28:39

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.080	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.225	21:20
Breath Volume		1.398	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.200	21:24
Breath Volume		1.640	21:24
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 3	0.209	21:27
Breath Volume		1.593	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REYES,JONATHAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: TOMAS MATIAS

DATE/TIME OF TEST: 04/20/2026 04:29:06

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.242	04:24
Breath Volume		1.371	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.239	04:27
Breath Volume		1.941	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REICH, PARKER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: FELIX CASTANEDA

DATE/TIME OF TEST: 04/21/2026 04:11:14

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.129	04:06
Breath Volume	1.144	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.130	04:09
Breath Volume	2.246	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ, JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SOBIEN

DATE/TIME OF TEST: 04/22/2026 02:01:45

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.081	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.231	01:57
Breath Volume		3.921	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.229	02:00
Breath Volume		3.218	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LIEBKNECHT,KEVIN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: POLK

DATE/TIME OF TEST: 04/22/2026 03:37:47

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.000	03:33
Breath Volume	1.210	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.000	03:36
Breath Volume	2.273	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUINONES,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: EICHMANN

DATE/TIME OF TEST: 04/22/2026 12:53:22

Diagnostics Check	OK	12:47
Air Blank	0.000	12:47
Control Test	0.081	12:48
Air Blank	0.000	12:48
Subject Sample # 1	0.274	12:49
Breath Volume	1.695	12:49
Air Blank	0.000	12:49
Air Blank	0.000	12:51
Subject Sample # 2	0.267	12:51
Breath Volume	3.035	12:51
Air Blank	0.000	12:52
Control Test	0.080	12:52
Air Blank	0.000	12:53
Diagnostics Check	OK	12:53

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: VEST

DATE/TIME OF TEST: 04/23/2026 01:28:21

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.081	01:22
Air Blank	0.000	01:23
Subject Sample # 1	0.087	01:24
Breath Volume	1.781	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.079	01:26
Breath Volume	1.492	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MORISHITA

DATE/TIME OF TEST: 04/23/2026 03:46:21

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.137	03:41
Breath Volume	1.722	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.136	03:44
Breath Volume	3.058	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RAMIREZ,MATTHEW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: VALENCIA

DATE/TIME OF TEST: 04/23/2026 22:26:22

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:20
Subject Sample # 1	0.207	22:22
Breath Volume	3.164	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.192	22:24
Breath Volume	1.191	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AVILA,ARNOLD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: VAZQUEZ GONZALEZ

DATE/TIME OF TEST: 04/24/2026 01:23:52

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.081	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.198	01:19
Breath Volume	1.757	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.197	01:22
Breath Volume	1.222	01:22
Air Blank	0.000	01:22
Control Test	0.079	01:23
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: MELENDEZ PENA

DATE/TIME OF TEST: 04/24/2026 03:05:21

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.081	02:57
Air Blank	0.000	02:58
Subject Sample # 1	0.180	03:00
Breath Volume	1.246	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.180	03:03
Breath Volume	1.445	03:03
Air Blank	0.000	03:04
Control Test	0.078	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORTENSEN,KEVIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/24/2026 04:41:40

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.188	04:35
Breath Volume		1.660	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.178	04:40
Breath Volume		1.343	04:40
Air Blank		0.000	04:40
Control Test		0.080	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOSSAIN,ANGELI
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: TELXEIRA

DATE/TIME OF TEST: 04/25/2026 23:04:50

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.058	23:00
Breath Volume	3.371	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.053	23:03
Breath Volume	1.820	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: ALFRED

DATE/TIME OF TEST: 04/26/2026 04:03:30

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.213	03:58
Breath Volume	1.507	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.200	04:01
Breath Volume	2.218	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 05/14/2026 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAMANNA,MATTHEW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: TOLOKONNIKOVA

DATE/TIME OF TEST: 04/28/2026 23:38:39

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.081	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.263	23:34
Breath Volume	1.472	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.269	23:37
Breath Volume	1.175	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ACOSTA,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FREZIN

DATE/TIME OF TEST: 04/29/2026 05:27:43

Diagnostics Check		OK	05:14	
Air Blank		0.000	05:15	
Control Test		0.080	05:15	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.217	05:19	Volume Not Met
Breath Volume		1.000	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 2	0.219	05:22	No .020 agreement
Breath Volume		1.289	05:22	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 3	0.257	05:26	No .020 agreement
Breath Volume		1.351	05:26	
Air Blank		0.000	05:26	
Control Test		0.079	05:27	
Air Blank		0.000	05:27	
Diagnostics Check		OK	05:27	

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ACOSTA,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FREZIN

DATE/TIME OF TEST: 04/29/2026 05:41:54

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.078	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1	0.234	05:33	Volume Not Met
Breath Volume		0.164	05:33	
Air Blank		0.000	05:34	
Air Blank		0.000	05:36	
Subject Sample	# 2	0.204	05:36	No .020 agreement
Breath Volume		1.257	05:36	
Air Blank		0.000	05:37	
Air Blank		0.000	05:39	
Subject Sample	# 3	0.229	05:40	No .020 agreement
Breath Volume		1.257	05:40	
Air Blank		0.000	05:41	
Control Test		0.077	05:41	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ACOSTA,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: FREZIN

DATE/TIME OF TEST: 04/29/2026 05:53:55

Diagnostics Check		OK	05:43
Air Blank		0.000	05:43
Control Test		0.077	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.236	05:45
Breath Volume		1.335	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.209	05:49
Breath Volume		1.613	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 3	0.249	05:52
Breath Volume		2.234	05:52
Air Blank		0.000	05:53
Control Test		0.077	05:53
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,LAWRENCE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MENDEZ ACEITUNO

DATE/TIME OF TEST: 04/29/2026 23:17:59

Diagnostics Check	OK	23:09
Air Blank	0.000	23:09
Control Test	0.081	23:10
Air Blank	0.000	23:10
Subject Sample # 1	0.206	23:13
Breath Volume	1.984	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.200	23:16
Breath Volume	1.773	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BLYTH, KYLE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: WALTHER

DATE/TIME OF TEST: 04/30/2026 02:10:20

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.183	02:05
Breath Volume		1.992	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.181	02:08
Breath Volume		2.015	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 05/14/2026 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GARRISON,JUSTIN
ARRESTING OFFICER AGENCY: Gadsden County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: REDDICK

DATE/TIME OF TEST: 04/18/2026 10:43:02

Diagnostics Check	OK	10:36	
Air Blank	0.000	10:37	
Control Test	0.079	10:37	
Air Blank	0.000	10:38	
Subject Sample	# 1 0.335	10:38	
Breath Volume	2.031	10:38	
Air Blank	0.000	10:39	
Air Blank	0.000	10:41	
Subject Sample	# 2	10:41	Slope Not Met
Breath Volume	1.515	10:41	
Air Blank	0.000	10:42	
Control Test	0.079	10:42	
Air Blank	0.000	10:42	
Diagnostics Check	OK	10:43	

Data Download Date/Time: 05/06/2026 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFORD,STERLING,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GARRISON,JUSTIN
ARRESTING OFFICER AGENCY: Gadsden County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A3
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: REDDICK

DATE/TIME OF TEST: 04/18/2026 10:52:09

Diagnostics Check	OK	10:45
Air Blank	0.000	10:46
Control Test	0.078	10:46
Air Blank	0.000	10:47
Subject Sample # 1	0.340	10:47
Breath Volume	1.683	10:47
Air Blank	0.000	10:48
Air Blank	0.000	10:50
Subject Sample # 2	0.340	10:50
Breath Volume	2.058	10:50
Air Blank	0.000	10:51
Control Test	0.078	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 05/06/2026 07:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HYLAND

DATE/TIME OF TEST: 04/03/2026 03:46:12

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.121	03:41
Breath Volume	2.355	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.120	03:44
Breath Volume	2.476	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 04/19/2026 18:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAKE,PALMER,A
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: LAKE,PALMER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/08/2026 01:42:03

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.119	01:34	Volume Not Met
Breath Volume		0.148	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2	0.142	01:40	Volume Not Met
Breath Volume		0.878	01:40	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 04/19/2026 18:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADOCCHIO,CASEY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GREIS-OSPINA,DAVID
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/10/2026 22:00:44

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:52	
Control Test	0.079	21:52	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:56	No Sample Provided
Breath Volume	0.000	21:56	
Air Blank	0.000	21:56	
Air Blank	0.000	21:58	
Subject Sample	# 2	22:00	Sequence Aborted
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADOCCHIO,CASEY,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GREIS-OSPINA,DAVID
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 04/10/2026 22:11:59

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.026	22:06
Breath Volume	1.917	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.025	22:10
Breath Volume	2.390	22:10
Air Blank	0.000	22:11
Control Test	0.079	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 04/10/2026 23:10:49

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.120	23:06
Breath Volume	3.093	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.117	23:09
Breath Volume	2.765	23:09
Air Blank	0.000	23:09
Control Test	0.078	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCONNELL,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: MCCONNELL,JEFFREY
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: PITT

DATE/TIME OF TEST: 04/17/2026 07:09:44

Diagnostics Check		OK	07:02
Air Blank		0.000	07:03
Control Test		0.079	07:03
Air Blank		0.000	07:04
Subject Sample	# 1	0.113	07:04
Breath Volume		1.875	07:04
Air Blank		0.000	07:05
Air Blank		0.000	07:07
Subject Sample	# 2	0.119	07:08
Breath Volume		2.300	07:08
Air Blank		0.000	07:08
Control Test		0.079	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: WILLEY

DATE/TIME OF TEST: 04/18/2026 03:13:14

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.054	03:08
Breath Volume		1.960	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.055	03:11
Breath Volume		1.882	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SOWA,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/18/2026 03:54:22

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:52	RFI Detect
Breath Volume	1.605	03:52	
Air Blank	0.000	03:53	
Control Test	0.078	03:53	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRACCICA,CARMEN,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: SOWA,T
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 04/18/2026 04:03:23

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.138	03:58
Breath Volume		2.941	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.143	04:01
Breath Volume		2.578	04:01
Air Blank		0.000	04:02
Control Test		0.077	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/19/2026 02:31:03

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:25
Air Blank	0.000	02:26
Subject Sample # 1	0.000	02:26
Breath Volume	1.878	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.000	02:29
Breath Volume	1.167	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:31

Data Download Date/Time: 04/19/2026 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AKERS,DEPUTY,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: VILLA,N
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 04/19/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: RODRIGUE

DATE/TIME OF TEST: 04/22/2026 01:21:23

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.220	01:16
Breath Volume	1.882	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:18
Subject Sample # 2	0.218	01:19
Breath Volume	2.125	01:19
Air Blank	0.000	01:20
Control Test	0.076	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/27/2026 20:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,TYLER,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS,TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/19/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: OMEROVIC

DATE/TIME OF TEST: 04/24/2026 01:42:04

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.078	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.214	01:31	Volume Not Met
Breath Volume		0.250	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.227	01:35	
Breath Volume		1.523	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.211	01:40	
Breath Volume		1.960	01:40	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 05/27/2026 20:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TYLER, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS, TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:12

SUBJECT'S LAST NAME: HUNTON

DATE/TIME OF TEST: 04/25/2026 17:38:22

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.079	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.000	17:34
Breath Volume	2.199	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.000	17:37
Breath Volume	2.730	17:37
Air Blank	0.000	17:37
Control Test	0.079	17:37
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 05/27/2026 20:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TYLER, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS, TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GARCIA SANCHEZ

DATE/TIME OF TEST: 04/25/2026 23:55:41

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Subject Sample # 1	0.246	23:48
Breath Volume	1.855	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.243	23:53
Breath Volume	1.589	23:53
Air Blank	0.000	23:54
Control Test	0.077	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 05/27/2026 20:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, TYLER, R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: CLEMENTS, TYLER
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 04/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 04/26/2026 02:56:06

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.000	02:51
Breath Volume	1.800	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.000	02:54
Breath Volume	1.691	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 06/02/2026 18:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GUTIERREZ BOLIVAR

DATE/TIME OF TEST: 04/05/2026 00:07:49

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.200	00:03
Breath Volume		2.304	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.200	00:06
Breath Volume		2.640	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 04/20/2026 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: TIRADO,I
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: CASTILLO ORTEGA

DATE/TIME OF TEST: 04/05/2026 04:16:34

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.078	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.125	04:12
Breath Volume	1.777	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.128	04:14
Breath Volume	2.343	04:14
Air Blank	0.000	04:15
Control Test	0.077	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/20/2026 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/15/2026 02:01:26

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.079	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.000	01:56
Breath Volume	2.546	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.000	02:00
Breath Volume	1.167	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 04/20/2026 11:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALDERON,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CALDERON,RAFAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: NARVAEZ GONZALEZ

DATE/TIME OF TEST: 04/17/2026 23:58:44

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.079	23:53
Air Blank	0.000	23:53
Subject Sample # 1	0.122	23:54
Breath Volume	2.242	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.124	23:57
Breath Volume	2.535	23:57
Air Blank	0.000	23:57
Control Test	0.077	23:58
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 04/20/2026 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,JOSEPH,E
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: QUINONES,JOSEPH
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/23/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: GARZON

DATE/TIME OF TEST: 04/12/2026 18:07:02

Diagnostics Check	OK	18:00
Air Blank	0.000	18:01
Control Test	0.078	18:01
Air Blank	0.000	18:02
Subject Sample # 1	0.227	18:02
Breath Volume	3.488	18:02
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.232	18:05
Breath Volume	3.605	18:05
Air Blank	0.000	18:06
Control Test	0.077	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:07

Data Download Date/Time: 04/21/2026 16:44