

Instrument Processing Sheet

Agency: NAS Pensacola Instrument Serial Number: 80-001205
 Date In: 12/12/2025 DI Completion Date: N/A Ship P/U H/D CMI EE

Intake By: <u>SLH</u> Date: <u>12/12/25</u> <input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Dropped off by agency in box to assess power issue.	Quality Checks By: _____ Date: _____ <input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) <input type="checkbox"/> Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ <input type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Date: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
--	--	--

Simulator	Serial #	Lot#/Exp	Maintenance
0.050			By: <u>SLH</u> Date: <u>12/12/25</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other: Fuse exchange did not resolve
0.080			
0.200			
0.080 DGS	N/A		

Optical Bench Adjustment By: _____	Department Inspection By: _____			
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: _____			
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____
0.040				Acetone Stock Solution Lot #: _____ Exp: _____
0.100				Simulator _____ Serial Number _____
0.200				0.000
0.300				Interferent
0.080 DGS	N/A			0.050
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080
0.050				0.200
Attachments				
<input type="checkbox"/> Form 41 <input type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other: Repair			
Barometric Pressure Gauge: _____ ID#: _____				

Notes/Suggested Service: Sending to repair due power issue. SLH 12/9/2025 12/12/2025 Tech Review - corrected date for Note. SLH 12/17/2025	<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Digitally signed by Taylor Gutschow Date: 2025.12.19 11:22:00 -05'00'	Digitally signed by Wen-Chi Pierson Date: 2025.12.19 11:56:36 -05'00'
Taylor Gutschow	Wen-Chi Pierson
Tech Review	Admin Review

Return Material Authorization

Ship to:

CMI, Inc.

Enforcement Electronics

Shipment to repair facility authorized by: Keith Kelley on 12/12/2025

Items Returned: Instrument Supplies Other Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001205

Bill To Address: NAS PENSACOLA PD 541 John H Tower Rd Pensacola, FL	Ship to Address: FDLE Offsite Mail Facility c/o FDLE HQ 813 B Lake Bradford Rd Tallahassee, FL 32304
--	--

Reason for Return:
Agency discussed power issues in need of repair: red power button does light up, but will not initialize and clicks off.

I require an estimate **BEFORE** any repairs will be authorized and/or conducted

Please contact: Keith Kelley Phone #: 850-497-2935

Email: keith.h.kelley.civ@us.navy.mil

ATP Contact Name: LeAndra Higginbotham ATP Email: leandrahigginbotham@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD
Time of Inspection: 08:38

Date of Inspection: 08/27/2025

Serial Number: 80-001205
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
BYPASSED AI TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


KATIE T SPEARIN

 Signature and Printed Name

08/27/2025
Date

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
08/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.079	08:40
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:41
Control Test	0.079	08:42
Air Blank	0.000	08:42
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature

DGS. KTS 9/9/25

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
08/27/2025
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	08:44
Control Test	0.049	08:45
Air Blank	0.000	08:45
Control Test	0.050	08:46
Air Blank	0.000	08:46
Control Test	0.050	08:47
Air Blank	0.000	08:47
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	



Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
08/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.080	08:52
Air Blank	0.000	08:52
Control Test	0.080	08:53
Air Blank	0.000	08:53
Control Test	0.080	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

NAS PENSACOLA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001205
08/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:56
Control Test	0.198	08:56
Air Blank	0.000	08:57
Control Test	0.198	08:57
Air Blank	0.000	08:58
Control Test	0.198	08:59
Air Blank	0.000	08:59
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD
Time of Inspection: 13:30

Date of Inspection: 09/04/2025

Serial Number: 80-001205
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.080	0.198	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.050	0.081	0.198	0.079
0.000	0.050	0.081	0.198	0.079

Standard Deviations	0.0004	0.0004	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

09/04/2025
Date



Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001205, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001205</u>	UNCERTAINTY* ±	
Owning Agency:	<u>NAS PENSACOLA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/04/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:30</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3)
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/04/2025 Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality