

Alcohol Testing Program - Instrument Processing Sheet

Agency: Hialeah Police Department

Instrument Serial Number: 80-002463

Date In: 5/15/2026

DI Completion Date: 6/30/2026

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: TDG Date: 6/23/2026		Quality Checks By: TDG Date: 6/24/2026		Flow Adjustment By:																																																																																																																																																													
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Left keyboard pin and portion of casing has broken off. Agency Inspector reports the printer attachment port at the back of the instrument is loose. A screw was missing from the printer port. Called the Agency Inspector to notify him about the broken pin.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 110 ☒ Flow Verification (L/s) Flow Column #: ATP106 32 mm 0.156 (.139-.169) 36 mm 0.171 (.156-.190) 53 mm 0.246 (.228-.278) 103 mm 0.503 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 33363 Gauge: 1019 Instrument: 1007 ☒ Stability Checks		Flow Column #: ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: ☐ Post Adjustment Verification (L/S) Flow Column #: 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																																																																																																																																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr></thead><tbody><tr><td rowspan="3">0.050</td><td rowspan="3">MP6286</td><td>202603A</td></tr><tr><td>3/3/2028</td></tr><tr><td>202603B</td></tr><tr><td rowspan="2">0.080</td><td rowspan="2">MP6287</td><td>3/3/2028</td></tr><tr><td>202603D</td></tr><tr><td rowspan="2">0.200</td><td rowspan="2">MP6288</td><td>3/4/2028</td></tr><tr><td>AG526603</td></tr><tr><td rowspan="2">0.080 DGS</td><td rowspan="2">N/A</td><td>9/23/2027</td></tr></tbody></table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202603A	3/3/2028	202603B	0.080	MP6287	3/3/2028	202603D	0.200	MP6288	3/4/2028	AG526603	0.080 DGS	N/A	9/23/2027	Maintenance By: TDG Date: 6/23/2026 ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: Added screw to printer port.																																																																																																																																										
Simulator	Serial #	Lot#/Exp																																																																																																																																																															
0.050	MP6286	202603A																																																																																																																																																															
		3/3/2028																																																																																																																																																															
		202603B																																																																																																																																																															
0.080	MP6287	3/3/2028																																																																																																																																																															
		202603D																																																																																																																																																															
0.200	MP6288	3/4/2028																																																																																																																																																															
		AG526603																																																																																																																																																															
0.080 DGS	N/A	9/23/2027																																																																																																																																																															
		<table border="1"><thead><tr><th colspan="2">Optical Bench Adjustment</th><th colspan="2">By: TDG</th><th colspan="2">Department Inspection</th><th colspan="2">By: TDG</th></tr></thead><tbody><tr><td colspan="4">Barometric Pressure Gauge: 1019 ID#: 33364</td><td colspan="4">Barometric Pressure ID#: 33363</td></tr><tr><td colspan="2">Simulator</td><td colspan="2">Serial #</td><td colspan="2">Gauge: 1017 Instrument: 1018</td><td colspan="2"></td></tr><tr><td colspan="2">0.000</td><td colspan="2">MP5097 N/A</td><td colspan="2">Mouth Alcohol Solution Lot #: 2025-D</td><td colspan="2">Exp: 9/25/2027</td></tr><tr><td colspan="2">0.040</td><td colspan="2">MP5098 25090</td><td colspan="2">Acetone Stock Solution Lot #: 2025-B</td><td colspan="2">Exp: 9/22/2027</td></tr><tr><td colspan="2">0.100</td><td colspan="2">MP5099 24390</td><td colspan="2">Simulator</td><td colspan="2">Serial Number</td></tr><tr><td colspan="2">0.200</td><td colspan="2">MP5100 25020</td><td colspan="2">0.000</td><td colspan="2">MP6284</td></tr><tr><td colspan="2">0.300</td><td colspan="2">MP5101 24430</td><td colspan="2">Interferent</td><td colspan="2">MP6285</td></tr><tr><td colspan="2">0.080 DGS</td><td colspan="2">N/A 28424080A3</td><td colspan="2">0.050</td><td colspan="2">MP6286</td></tr><tr><td colspan="4">☒ Post Optical Bench Adjustment Stability Checks</td><td colspan="2">0.080</td><td colspan="2">MP6287</td></tr><tr><td colspan="2">Simulator</td><td colspan="2">Serial #</td><td colspan="2">0.200</td><td colspan="2">MP6288</td></tr><tr><td colspan="2">0.050</td><td colspan="2">MP6286 202603A</td><td colspan="4">Attachments</td></tr><tr><td colspan="2">0.080</td><td colspan="2">MP6287 202603B</td><td colspan="2">☒ Form 41</td><td colspan="2">☒ Post-Stability Checks</td></tr><tr><td colspan="2">0.200</td><td colspan="2">MP6288 202603D</td><td colspan="2">☒ Stability Checks</td><td colspan="2">☐ Flow Adjustment</td></tr><tr><td colspan="2">0.080 DGS</td><td colspan="2">N/A AG526603</td><td colspan="2">☒ Calibration Certificate</td><td colspan="2">☐ Form 40</td></tr><tr><td colspan="4">Gauge ID #: 33363</td><td colspan="2">☒ Optical Bench Adjustment</td><td colspan="2">☐ Other:</td></tr><tr><td colspan="4">Gauge: 1017 Instrument: 1018</td><td colspan="4"></td></tr><tr><td colspan="4">Notes/Suggested Service: Tech Review: Added checkmark to Post-Stability Checks in Attachments. (TDG 7/1/26)</td><td colspan="4">☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use</td></tr><tr><td colspan="4">Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.01 16:00:20 -04'00'</td><td colspan="4">Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.09 09:58:00 -04'00'</td></tr><tr><td colspan="4">Tech Review</td><td colspan="4">Admin Review</td></tr></tbody></table>		Optical Bench Adjustment		By: TDG		Department Inspection		By: TDG		Barometric Pressure Gauge: 1019 ID#: 33364				Barometric Pressure ID#: 33363				Simulator		Serial #		Gauge: 1017 Instrument: 1018				0.000		MP5097 N/A		Mouth Alcohol Solution Lot #: 2025-D		Exp: 9/25/2027		0.040		MP5098 25090		Acetone Stock Solution Lot #: 2025-B		Exp: 9/22/2027		0.100		MP5099 24390		Simulator		Serial Number		0.200		MP5100 25020		0.000		MP6284		0.300		MP5101 24430		Interferent		MP6285		0.080 DGS		N/A 28424080A3		0.050		MP6286		☒ Post Optical Bench Adjustment Stability Checks				0.080		MP6287		Simulator		Serial #		0.200		MP6288		0.050		MP6286 202603A		Attachments				0.080		MP6287 202603B		☒ Form 41		☒ Post-Stability Checks		0.200		MP6288 202603D		☒ Stability Checks		☐ Flow Adjustment		0.080 DGS		N/A AG526603		☒ Calibration Certificate		☐ Form 40		Gauge ID #: 33363				☒ Optical Bench Adjustment		☐ Other:		Gauge: 1017 Instrument: 1018								Notes/Suggested Service: Tech Review: Added checkmark to Post-Stability Checks in Attachments. (TDG 7/1/26)				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.01 16:00:20 -04'00'				Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.09 09:58:00 -04'00'				Tech Review				Admin Review	
Optical Bench Adjustment		By: TDG		Department Inspection		By: TDG																																																																																																																																																											
Barometric Pressure Gauge: 1019 ID#: 33364				Barometric Pressure ID#: 33363																																																																																																																																																													
Simulator		Serial #		Gauge: 1017 Instrument: 1018																																																																																																																																																													
0.000		MP5097 N/A		Mouth Alcohol Solution Lot #: 2025-D		Exp: 9/25/2027																																																																																																																																																											
0.040		MP5098 25090		Acetone Stock Solution Lot #: 2025-B		Exp: 9/22/2027																																																																																																																																																											
0.100		MP5099 24390		Simulator		Serial Number																																																																																																																																																											
0.200		MP5100 25020		0.000		MP6284																																																																																																																																																											
0.300		MP5101 24430		Interferent		MP6285																																																																																																																																																											
0.080 DGS		N/A 28424080A3		0.050		MP6286																																																																																																																																																											
☒ Post Optical Bench Adjustment Stability Checks				0.080		MP6287																																																																																																																																																											
Simulator		Serial #		0.200		MP6288																																																																																																																																																											
0.050		MP6286 202603A		Attachments																																																																																																																																																													
0.080		MP6287 202603B		☒ Form 41		☒ Post-Stability Checks																																																																																																																																																											
0.200		MP6288 202603D		☒ Stability Checks		☐ Flow Adjustment																																																																																																																																																											
0.080 DGS		N/A AG526603		☒ Calibration Certificate		☐ Form 40																																																																																																																																																											
Gauge ID #: 33363				☒ Optical Bench Adjustment		☐ Other:																																																																																																																																																											
Gauge: 1017 Instrument: 1018																																																																																																																																																																	
Notes/Suggested Service: Tech Review: Added checkmark to Post-Stability Checks in Attachments. (TDG 7/1/26)				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																																																																																																													
Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.07.01 16:00:20 -04'00'				Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.09 09:58:00 -04'00'																																																																																																																																																													
Tech Review				Admin Review																																																																																																																																																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✗																																																																																																																																																
0.003 of Wet ✓																																																																																																																																																			
0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✗ 0.003 of Wet ✓																																																																																																																																																
HIALERH PD Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:12	Control Test	0.049	12:13	Air Blank	0.000	12:13	Control Test	0.049	12:14	Air Blank	0.000	12:14	Control Test	0.048	12:15	Air Blank	0.000	12:16	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		HIALERH PD Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:56	Control Test	0.078	11:56	Air Blank	0.000	11:57	Control Test	0.077	11:58	Air Blank	0.000	11:58	Control Test	0.077	11:59	Air Blank	0.000	11:59	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466		HIALERH PD Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	12:04	Control Test	0.198	12:04	Air Blank	0.000	12:05	Control Test	0.197	12:06	Air Blank	0.000	12:06	Control Test	0.197	12:07	Air Blank	0.000	12:07	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		HIALERH PD Intoxilyzer - Alcohol Analyzer Model 8000 06/24/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:48</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0763</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7564</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:46	Control Test	0.077	11:46	Air Blank	0.000	11:47	Control Test	0.076	11:47	Air Blank	0.000	11:48	Control Test	0.076	11:48	Air Blank	0.000	11:49	Control Test Stats			Average	0.0763		Std Dev	0.0006		Rel Std Dev(%)	0.7564	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:12																																																																																																																																																	
Control Test	0.049	12:13																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.049	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.048	12:15																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.078	11:56																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test	0.077	11:58																																																																																																																																																	
Air Blank	0.000	11:58																																																																																																																																																	
Control Test	0.077	11:59																																																																																																																																																	
Air Blank	0.000	11:59																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.198	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.197	12:06																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.197	12:07																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1973																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:46																																																																																																																																																	
Control Test	0.077	11:46																																																																																																																																																	
Air Blank	0.000	11:47																																																																																																																																																	
Control Test	0.076	11:47																																																																																																																																																	
Air Blank	0.000	11:48																																																																																																																																																	
Control Test	0.076	11:48																																																																																																																																																	
Air Blank	0.000	11:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0763																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7564																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.7920 (-0.0010)
Sample #2 = 6.7670 (-0.0190)
Sample #3 = 6.7770 (-0.0240)
Sample #4 = 6.7540 (-0.0290)
Avg % Abs = 6.7660 (-0.0240)
STD DEV = 0.0115 (-0.0050)
REL STD DEV = 0.170 (20.833)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5110 (-0.0110)
Sample #2 = 1.5110 (-0.0020)
Sample #3 = 1.4950 (-0.0070)
Sample #4 = 1.5190 (-0.0110)
Avg % Abs = 1.5083 (-0.0033)
STD DEV = 0.0122 (-0.0032)
REL STD DEV = 0.810 (96.437)

H14LEAH PD
Intoxilyzer - Alconol Analyser
Model 8000
06/29/2026
SN 80-002463
13:38:56

Auto Calibration
Max Power Res Value = 50
Auto Range Res Value = 32

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12674, Sum Io = 13554
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1040 (-0.0210)
Sample #2 = 0.0990 (-0.0190)
Sample #3 = 0.1160 (-0.0200)
Sample #4 = 0.1060 (-0.0280)
Avg % Abs = 0.1070 (-0.0037)
STD DEV = 0.0085 (-0.0235)
REL STD DEV = 7.985 (642.117)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1220 (-0.0190)
Sample #2 = 0.1380 (-0.0280)
Sample #3 = 0.1120 (-0.0020)
Sample #4 = 0.1190 (-0.0030)
Avg % Abs = 0.1230 (-0.0090)
STD DEV = 0.0135 (-0.0166)
REL STD DEV = 10.938 (184.926)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12657, Sum Io = 13547
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8320 (-0.0190)
Sample #2 = 0.8430 (-0.0020)
Sample #3 = 0.8520 (-0.0140)
Sample #4 = 0.8430 (-0.0100)
Avg % Abs = 0.8460 (-0.0087)
STD DEV = 0.0052 (-0.0061)
REL STD DEV = 0.614 (70.501)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12649, Sum Io = 13546
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9570 (-0.0160)
Sample #2 = 1.9270 (-0.0250)
Sample #3 = 1.9380 (-0.0190)
Sample #4 = 1.9310 (-0.0420)
Avg % Abs = 1.9320 (-0.0287)
STD DEV = 0.0056 (-0.0119)
REL STD DEV = 0.288 (41.618)

Channel 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5550 (-0.0000)
Sample #2 = 3.5300 (-0.0370)
Sample #3 = 3.5450 (-0.0280)
Sample #4 = 3.5380 (-0.0350)
Avg % Abs = 3.5377 (-0.0333)
STD DEV = 0.0075 (-0.0047)
REL STD DEV = 0.212 (14.177)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12642, Sum Io = 13541
Channel 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7320 (-0.0180)
Sample #2 = 3.6980 (-0.0110)
Sample #3 = 3.7030 (-0.0380)
Sample #4 = 3.6980 (-0.0350)
Avg % Abs = 3.6997 (-0.0280)
STD DEV = 0.0029 (-0.0148)
REL STD DEV = 0.078 (52.852)

Optical Calibration Adjustment

By: TDG

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
1.000 0.000 -0.0000
0.040 0.040 0.0001
0.100 0.100 -0.0001
0.200 0.200 0.0000
0.300 0.300 -0.0000
Sol Value = 0.090 g/210L ***
Fit Value = 0.3610 mg/l %%%
Samples Taken = 4, Discarded = 1
Channel 1 >>>>
Sample #1 = 3123.00
Sample #2 = 3101.00
Sample #3 = 3081.00
Sample #4 = 3097.00
Average Result = 3093.0000
STD DEV = 10.5830
REL STD DEV = 0.342
Channel 2 >>>>
Sample #1 = 3361.00
Sample #2 = 3368.00
Sample #3 = 3376.00
Sample #4 = 3376.00
Average Result = 3373.3333
STD DEV = 4.6188
REL STD DEV = 0.137
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 716
9 um H2O Adjust (mg/l*10,000) = 436
**** AUTO CAL PRESS

Channel 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.107
Std Dev = 0.01 Rel Std Dev = 7.99
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.846
Std Dev = 0.01 Rel Std Dev = 0.61
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.932
Std Dev = 0.01 Rel Std Dev = 0.29
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.700
Std Dev = 0.00 Rel Std Dev = 0.08
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.428
Std Dev = 0.01 Rel Std Dev = 0.23
Zero Order Coef = -280.15
First Order Coef = 2571.43
Second Order Coef = 20.75
Standard Deviation = 5.260761

Channel 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.123
Std Dev = 0.01 Rel Std Dev = 10.94
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.508
Std Dev = 0.01 Rel Std Dev = 0.61
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.538
Std Dev = 0.01 Rel Std Dev = 0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.766
Std Dev = 0.01 Rel Std Dev = 0.17
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.842
Std Dev = 0.02 Rel Std Dev = 0.17
Zero Order Coef = -164.68
First Order Coef = 1351.47
Second Order Coef = 11.87
Standard Deviation = 2.690411

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0001
0.040 0.040 -0.0001
0.100 0.100 -0.0001
0.200 0.200 0.0001
0.300 0.300 -0.0000

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓ HIALEAH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002463 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:01</td></tr> <tr><td>Control Test</td><td>0.051</td><td>16:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:12</td></tr> <tr><td>Control Test</td><td>0.050</td><td>16:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:04</td></tr> <tr><td>Control Test</td><td>0.050</td><td>16:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	16:01	Control Test	0.051	16:02	Air Blank	0.000	16:12	Control Test	0.050	16:03	Air Blank	0.000	16:04	Control Test	0.050	16:14	Air Blank	0.000	16:05	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		✓ HIALEAH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002463 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>16:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	16:12	Control Test	0.080	16:13	Air Blank	0.000	16:14	Control Test	0.079	16:14	Air Blank	0.000	16:15	Control Test	0.079	16:16	Air Blank	0.000	16:16	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		✓ HIALEAH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002463 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:17</td></tr> <tr><td>Control Test</td><td>0.199</td><td>16:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:18</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:19</td></tr> <tr><td>Control Test</td><td>0.198</td><td>16:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	16:17	Control Test	0.199	16:18	Air Blank	0.000	16:18	Control Test	0.198	16:19	Air Blank	0.000	16:19	Control Test	0.198	16:20	Air Blank	0.000	16:21	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		✓ HIALEAH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002463 06/29/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>16:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	16:08	Control Test	0.079	16:08	Air Blank	0.000	16:09	Control Test	0.079	16:09	Air Blank	0.000	16:10	Control Test	0.079	16:10	Air Blank	0.000	16:11	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:01																																																																																																																																																	
Control Test	0.051	16:02																																																																																																																																																	
Air Blank	0.000	16:12																																																																																																																																																	
Control Test	0.050	16:03																																																																																																																																																	
Air Blank	0.000	16:04																																																																																																																																																	
Control Test	0.050	16:14																																																																																																																																																	
Air Blank	0.000	16:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:12																																																																																																																																																	
Control Test	0.080	16:13																																																																																																																																																	
Air Blank	0.000	16:14																																																																																																																																																	
Control Test	0.079	16:14																																																																																																																																																	
Air Blank	0.000	16:15																																																																																																																																																	
Control Test	0.079	16:16																																																																																																																																																	
Air Blank	0.000	16:16																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:17																																																																																																																																																	
Control Test	0.199	16:18																																																																																																																																																	
Air Blank	0.000	16:18																																																																																																																																																	
Control Test	0.198	16:19																																																																																																																																																	
Air Blank	0.000	16:19																																																																																																																																																	
Control Test	0.198	16:20																																																																																																																																																	
Air Blank	0.000	16:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:08																																																																																																																																																	
Control Test	0.079	16:08																																																																																																																																																	
Air Blank	0.000	16:09																																																																																																																																																	
Control Test	0.079	16:09																																																																																																																																																	
Air Blank	0.000	16:10																																																																																																																																																	
Control Test	0.079	16:10																																																																																																																																																	
Air Blank	0.000	16:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD
Time of Inspection: 14:08

Date of Inspection: 06/30/2026

Serial Number: 80-002463
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.077	0.196	0.079
0.000	0.049	0.077	0.197	0.079
0.000	0.049	0.077	0.197	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.049	0.077	0.197	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.049	0.077	0.197	0.079
0.000	0.049	0.077	0.197	0.078
0.000	0.050	0.077	0.197	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002463, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002463</u>	Serial Number:	<u>80-002463</u>
Owning Agency:	<u>HIALEAH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:08</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.06.30 14:35:48
-04'00'

06/30/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality