

Stability Checks

0.05g/210L 0.047 to 0.053			0.08g/210L 0.077 to 0.083			0.20g/210L 0.194 to 0.206			DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet		
✓			✓			✓			✓		
Performed Root Case Analysis			Performed Root Case Analysis			Performed Root Case Analysis			Performed Root Case Analysis		
PALM 84V PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 60-001721 07/10/2025 Software: 8100.27			PALM 84V PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 60-001721 07/10/2025 Software: 8100.27			PALM 84V PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 60-001721 07/10/2025 Software: 8100.27			PALM 84V PD Intoxilyzer - Alcohol Analyzer Model: 8100 SN: 60-001721 07/10/2025 Software: 8100.27		
Test	g/210L	Time									
Air Blank	0.000	11:03	Air Blank	0.000	11:11	Air Blank	0.000	11:18	Air Blank	0.000	10:49
Control Test	0.050	11:04	Control Test	0.079	11:11	Control Test	0.198	11:18	Control Test	0.069	10:50
Air Blank	0.000	11:04	Air Blank	0.000	11:12	Air Blank	0.000	11:19	Air Blank	0.000	10:50
Control Test	0.050	11:05	Control Test	0.079	11:12	Control Test	0.198	11:20	Control Test	0.071	10:51
Air Blank	0.000	11:06	Air Blank	0.000	11:13	Air Blank	0.000	11:20	Air Blank	0.000	10:51
Control Test	0.050	11:06	Control Test	0.080	11:14	Control Test	0.198	11:21	Control Test	0.073	10:51
Air Blank	0.000	11:07	Air Blank	0.000	11:14	Air Blank	0.000	11:21	Air Blank	0.000	10:52
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0500		Average	0.0793		Average	0.1980		Average	0.0710	
Std Dev	0.0000		Std Dev	0.0006		Std Dev	0.0000		Std Dev	0.0020	
Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.7277		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	2.8169	
 Operator's Signature			 Operator's Signature			 Operator's Signature			 Operator's Signature		

0.08g/210L

DGIS STABILITY

CHECK #1

o Performed RCA
repeated test

PEM BAY PD
Intoxilyzer - Alcohol Analyzer
Model: 8010 SN 60-001721
07/10/2025
Software: 8100.07

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.169	10:39
Air Blank	0.000	10:39
Control Test	0.168	10:40
Air Blank	0.000	10:40
Control Test	0.070	10:40
Air Blank	0.000	10:41
Control Test Stats		
Average	0.1690	
Std Dev	0.0010	
Rel Std Dev(%)	1.4493	


Operator's Signature

Inadvertently
prompted Dry
Control Test

PALM BEY RD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 60-001721
07/10/2025
Software: B100.27

DRY CONTROL TEST

Test	g/210L	Time
Air Blank	ABT*	11:07
Air Blank	0.000	11:08

*Sequence Aborted



Operator's Signature

80-001721
Sub 7/28/25

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001721
07/28/2025 10:40:15

Auto Calibration
Max Power Res Value = 29
Auto Range Res Value = 17

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12829, Num Io = 12204

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1490 (0.0080)
Sample #2 = 0.1230 (0.0560)
Sample #3 = 0.1240 (0.0720)
Sample #4 = 0.1320 (0.1090)
Avg % Abs = 0.1263 (0.0790)
STD DEV = 0.0049 (0.0272)
REL STD DEV = 3.905 (34.411)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.1520 (0.0110)
Sample #2 = 0.1480 (0.0340)
Sample #3 = 0.1400 (0.0400)
Sample #4 = 0.1240 (0.0720)
Avg % Abs = 0.1373 (0.0487)
STD DEV = 0.0122 (0.0204)
REL STD DEV = 8.898 (41.977)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12805, Num Io = 12189

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.8460 (-0.0110)
Sample #2 = 0.8520 (0.0220)
Sample #3 = 0.8540 (0.0360)
Sample #4 = 0.8610 (0.0600)
Avg % Abs = 0.8557 (0.0393)
STD DEV = 0.0047 (0.0192)
REL STD DEV = 0.552 (48.859)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.5200 (-0.0270)
Sample #2 = 1.5690 (-0.0270)
Sample #3 = 1.5120 (0.0950)
Sample #4 = 1.5350 (0.0070)
Avg % Abs = 1.5387 (-0.0050)
STD DEV = 0.0287 (0.0191)
REL STD DEV = 1.864 (381.576)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12778, Num Io = 12174

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 1.9350 (-0.0100)
Sample #2 = 1.8990 (0.0260)
Sample #3 = 1.9260 (0.0140)
Sample #4 = 1.8790 (0.0610)
Avg % Abs = 1.9013 (0.0337)
STD DEV = 0.0236 (0.0244)
REL STD DEV = 1.241 (72.534)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.5620 (-0.0090)
Sample #2 = 3.5310 (-0.0050)
Sample #3 = 3.5500 (-0.0160)
Sample #4 = 3.5290 (0.0030)
Avg % Abs = 3.5367 (-0.0060)
STD DEV = 0.0116 (0.0095)
REL STD DEV = 0.328 (158.990)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12771, Num Io = 12173

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 3.6510 (0.0070)
Sample #2 = 3.6540 (0.0180)
Sample #3 = 3.6210 (0.0350)
Sample #4 = 3.6370 (0.0600)
Avg % Abs = 3.6373 (0.0377)
STD DEV = 0.0165 (0.0211)
REL STD DEV = 0.454 (56.088)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 6.7460 (0.0080)
Sample #2 = 6.7950 (-0.0080)
Sample #3 = 6.7540 (0.0000)
Sample #4 = 6.7630 (0.0160)
Avg % Abs = 6.7707 (0.0027)
STD DEV = 0.0215 (0.0122)
REL STD DEV = 0.318 (458.258)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12762, Num Io = 12171

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 5.3070 (-0.0140)
Sample #2 = 5.3210 (0.0220)
Sample #3 = 5.3080 (0.0430)
Sample #4 = 5.3110 (0.0250)
Avg % Abs = 5.3133 (0.0300)
STD DEV = 0.0068 (0.0114)
REL STD DEV = 0.128 (37.859)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 9.7920 (0.0010)
Sample #2 = 9.8010 (0.0220)
Sample #3 = 9.8150 (0.0330)
Sample #4 = 9.8010 (0.0250)
Avg % Abs = 9.8057 (0.0267)
STD DEV = 0.0081 (0.0057)
REL STD DEV = 0.082 (21.323)

Auto Cal Data Channel 1:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.126
Std Dev = 0.00 Rel Std Dev = 3.90
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.856
Std Dev = 0.00 Rel Std Dev = 0.55
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.901
Std Dev = 0.02 Rel Std Dev = 1.24
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.637
Std Dev = 0.02 Rel Std Dev = 0.45
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.313
Std Dev = 0.01 Rel Std Dev = 0.13
Zero Order Coef = -341.41
First Order Coef = 2630.70
Second Order Coef = 22.94
Standard Deviation = 15.324927

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.137
Std Dev = 0.01 Rel Std Dev = 8.90
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.539
Std Dev = 0.03 Rel Std Dev = 1.86
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.537
Std Dev = 0.01 Rel Std Dev = 0.33
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.771
Std Dev = 0.02 Rel Std Dev = 0.32
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.806
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -188.00
First Order Coef = 1348.36
Second Order Coef = 12.98
Standard Deviation = 13.184261

Solution Stats Quadratic Fit Chan 1:
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0002
0.040 0.040 -0.0005
0.100 0.100 0.0004
0.200 0.200 -0.0001
0.300 0.300 0.0000

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0003
0.100	0.100	0.0004
0.200	0.200	-0.0003
0.300	0.300	0.0001

70^{Set}
80-001721
Set
7/28/25

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 2530.00
Sample #2 = 2636.00
Sample #3 = 2721.00
Sample #4 = 2883.00
Average Result = 2746.6667
STD DEV = 125.4844
REL STD DEV = 4.569

***** CHANNEL 2
Sample #1 = 2875.00
Sample #2 = 3103.00
Sample #3 = 3057.00
Sample #4 = 3165.00
Average Result = 3108.3333
STD DEV = 54.1972
REL STD DEV = 1.744

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1
Sample #1 = 2792.00
Sample #2 = 2896.00
Sample #3 = 2969.00
Sample #4 = 2850.00
Average Result = 2905.0000
STD DEV = 60.0083
REL STD DEV = 2.066

***** CHANNEL 2
Sample #1 = 3151.00
Sample #2 = 3195.00
Sample #3 = 3245.00
Sample #4 = 3266.00
Average Result = 3235.3333
STD DEV = 36.4737
REL STD DEV = 1.127

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 904
9 um H2O Adjust (mg/l*10,000) = 574
**** AUTO CAL PASS

80-001721

Post stability

DGS

PALM BAY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001721
07/28/2025
Software: 8100.27

RCA -

gas cylinder was empty -
Reran with new cylinder

slut

7/28/25

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.000	11:44
Air Blank	0.000	11:44
Control Test	0.000	11:45
Air Blank	0.000	11:45
Control Test	0.000	11:46
Air Blank	0.000	11:46
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

Post-Calibration Adjustment Stability Checks

SWA 1721 sub
80-007366

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L SD.003 g/210L of Wet																																																																																																
<p>Performed Root Case Analysis</p> <p>PALM BAY PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001721 07/29/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>PALM BAY PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001721 07/28/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>PALM BAY PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001721 07/29/2025 Software: 8100.27</p>	<p>Performed Root Case Analysis</p> <p>PALM BAY PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001721 07/28/2025 Software: 8100.27</p>																																																																																																
<p>Test</p> <table border="1"> <thead> <tr> <th>3/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>09:37</td></tr> <tr><td>Control Test</td><td>09:38</td></tr> <tr><td>Air Blank</td><td>09:39</td></tr> <tr><td>Control Test</td><td>09:39</td></tr> <tr><td>Air Blank</td><td>09:40</td></tr> <tr><td>Control Test</td><td>09:40</td></tr> <tr><td>Air Blank</td><td>09:41</td></tr> <tr><td>Control Test Stats</td><td></td></tr> <tr><td>Average</td><td>0.0493</td></tr> <tr><td>Std Dev</td><td>0.0016</td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1713</td></tr> </tbody> </table>	3/210L	Time	Air Blank	09:37	Control Test	09:38	Air Blank	09:39	Control Test	09:39	Air Blank	09:40	Control Test	09:40	Air Blank	09:41	Control Test Stats		Average	0.0493	Std Dev	0.0016	Rel. Std Dev(%)	1.1713	<p>Test</p> <table border="1"> <thead> <tr> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>11:58</td></tr> <tr><td>Control Test</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>11:59</td></tr> <tr><td>Control Test</td><td>11:59</td></tr> <tr><td>Air Blank</td><td>12:00</td></tr> <tr><td>Control Test</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>12:01</td></tr> <tr><td>Control Test Stats</td><td></td></tr> <tr><td>Average</td><td>0.0803</td></tr> <tr><td>Std Dev</td><td>0.0006</td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7187</td></tr> </tbody> </table>	g/210L	Time	Air Blank	11:58	Control Test	11:58	Air Blank	11:59	Control Test	11:59	Air Blank	12:00	Control Test	12:01	Air Blank	12:01	Control Test Stats		Average	0.0803	Std Dev	0.0006	Rel. Std Dev(%)	0.7187	<p>Test</p> <table border="1"> <thead> <tr> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>09:32</td></tr> <tr><td>Control Test</td><td>09:33</td></tr> <tr><td>Air Blank</td><td>09:33</td></tr> <tr><td>Control Test</td><td>09:34</td></tr> <tr><td>Air Blank</td><td>09:34</td></tr> <tr><td>Control Test</td><td>09:35</td></tr> <tr><td>Air Blank</td><td>09:36</td></tr> <tr><td>Control Test Stats</td><td></td></tr> <tr><td>Average</td><td>0.1977</td></tr> <tr><td>Std Dev</td><td>0.0016</td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2921</td></tr> </tbody> </table>	g/210L	Time	Air Blank	09:32	Control Test	09:33	Air Blank	09:33	Control Test	09:34	Air Blank	09:34	Control Test	09:35	Air Blank	09:36	Control Test Stats		Average	0.1977	Std Dev	0.0016	Rel. Std Dev(%)	0.2921	<p>Test</p> <table border="1"> <thead> <tr> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>09:00</td></tr> <tr><td>Control Test</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>09:00</td></tr> <tr><td>Control Test</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>09:00</td></tr> <tr><td>Control Test</td><td>09:07</td></tr> <tr><td>Air Blank</td><td>09:00</td></tr> <tr><td>Control Test Stats</td><td></td></tr> <tr><td>Average</td><td>0.0787</td></tr> <tr><td>Std Dev</td><td>0.0006</td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7339</td></tr> </tbody> </table>	g/210L	Time	Air Blank	09:00	Control Test	09:07	Air Blank	09:00	Control Test	09:07	Air Blank	09:00	Control Test	09:07	Air Blank	09:00	Control Test Stats		Average	0.0787	Std Dev	0.0006	Rel. Std Dev(%)	0.7339
3/210L	Time																																																																																																		
Air Blank	09:37																																																																																																		
Control Test	09:38																																																																																																		
Air Blank	09:39																																																																																																		
Control Test	09:39																																																																																																		
Air Blank	09:40																																																																																																		
Control Test	09:40																																																																																																		
Air Blank	09:41																																																																																																		
Control Test Stats																																																																																																			
Average	0.0493																																																																																																		
Std Dev	0.0016																																																																																																		
Rel. Std Dev(%)	1.1713																																																																																																		
g/210L	Time																																																																																																		
Air Blank	11:58																																																																																																		
Control Test	11:58																																																																																																		
Air Blank	11:59																																																																																																		
Control Test	11:59																																																																																																		
Air Blank	12:00																																																																																																		
Control Test	12:01																																																																																																		
Air Blank	12:01																																																																																																		
Control Test Stats																																																																																																			
Average	0.0803																																																																																																		
Std Dev	0.0006																																																																																																		
Rel. Std Dev(%)	0.7187																																																																																																		
g/210L	Time																																																																																																		
Air Blank	09:32																																																																																																		
Control Test	09:33																																																																																																		
Air Blank	09:33																																																																																																		
Control Test	09:34																																																																																																		
Air Blank	09:34																																																																																																		
Control Test	09:35																																																																																																		
Air Blank	09:36																																																																																																		
Control Test Stats																																																																																																			
Average	0.1977																																																																																																		
Std Dev	0.0016																																																																																																		
Rel. Std Dev(%)	0.2921																																																																																																		
g/210L	Time																																																																																																		
Air Blank	09:00																																																																																																		
Control Test	09:07																																																																																																		
Air Blank	09:00																																																																																																		
Control Test	09:07																																																																																																		
Air Blank	09:00																																																																																																		
Control Test	09:07																																																																																																		
Air Blank	09:00																																																																																																		
Control Test Stats																																																																																																			
Average	0.0787																																																																																																		
Std Dev	0.0006																																																																																																		
Rel. Std Dev(%)	0.7339																																																																																																		
<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 																																																																																																

DGS #2

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BAY PD

Time of Inspection: 17:47

Date of Inspection: 07/29/2025

Serial Number: 80-001721

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.195	0.073 / 0.080
0.000	0.048	0.078	0.196	0.077 / 0.081
0.000	0.048	0.078	0.196	0.079 / 0.081
0.000	0.048	0.078	0.196	0.079 / 0.081
0.000	0.048	0.078	0.197	0.080 / 0.081
0.000	0.048	0.079	0.196	0.080 / 0.081
0.000	0.049	0.079	0.196	0.081 / 0.081
0.000	0.049	0.079	0.196	0.080 / 0.081
0.000	0.049	0.079	0.197	0.081 / 0.081
0.000	0.048	0.079	0.196	0.081 / 0.081

Standard Deviations	0.0004	0.0005	0.0005	0.0024 / 0.0003
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. TIGHTENED CYLINDER, RERAN

SUA


The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Leandra Higginbotham

LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/29/2025
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001721, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001721</u>	UNCERTAINTY* ±	
Owning Agency:	<u>PALM BAY PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>07/29/2025</u>	0.080 g/210 L	0.004
Calibration Time:	<u>17:47</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/29/2025 Date
Leandra Higginbotham
 LEANDRA HIGGINBOTHAM,
 Department Inspector