

Alcohol Testing Program - Instrument Processing Sheet

Agency: Charlotte County Sheriff's Office **Instrument Serial Number:** 80-001739

Date In: 4/2/2026 **DI Completion Date:** N/A ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>4/23/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>4/23/2026</u>		Flow Adjustment By: _____																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☐ Instrument Set Up Verified ☐ R-Value: _____ ☐ Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) ☐ Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ ☐ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: _____	
Simulator	Serial #	Lot#/Exp																		
0.050																				
0.080																				
0.200																				
0.080 DGS	N/A																			
Optical Bench Adjustment By: _____		Department Inspection By: _____																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: _____																		
Simulator	Serial #	Lot #	Expiration	Gauge: _____ Instrument: _____																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: _____ Exp: _____																
0.040				Acetone Stock Solution Lot #: _____ Exp: _____																
0.100				Simulator	Serial Number															
0.200				0.000																
0.300				Interferent																
0.080 DGS	N/A			0.050																
☐ Post Optical Bench Adjustment Stability Checks				0.080																
Simulator	Serial #	Lot #	Expiration	0.200																
0.050				Attachments																
0.080				☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Return Material Authorization																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____ Gauge: _____ Instrument: _____																				
Notes/Suggested Service: Instrument shuts off midway through powering up. Could not complete Quality Checks. Sending to CMI for evaluation. (TDG 5/15/26)				☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																
				<table style="width: 100%;"> <tr> <td style="width: 50%;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.05.18 17:08:22 -04'00' </td> <td style="width: 50%;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.05.19 06:44:20 -04'00' </td> </tr> </table>		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.05.18 17:08:22 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.05.19 06:44:20 -04'00'													
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.05.18 17:08:22 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.05.19 06:44:20 -04'00'																			
Tech Review				Admin Review																

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Vanessa Chandler on 5/12/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-001739

Bill To Address:

Charlotte County Sheriff's Office
Attn: Vanessa Chandler

Ship to Address:

Florida Department of Law Enforcement
Fort Myers Regional Operations Center
Attn: Taylor Gutschow
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

Reason for Return:

Instrument shuts off midway through powering up. Tried multiple power cords and plugs.

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Vanessa Chandler Phone #: 941-833-6446Email: vchandler@ccsofl.netATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us