



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FFWCC

Instrument Serial Number: 80-007165

AGENCY INSPECTION DISCREPENCY

Agency Inspector: Richard Marroquin

Inspection Date: 11/01/2023

A test was repeated without reason for repeat or the corrective action given

Alcohol Free/Mouth Alcohol Alcohol Free Test Interferent Detect

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Agency Inspection not conducted for calendar month

Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Lot Number and/or Expiration Date for Standards entered incorrectly

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Other Electronic Data Review/Comments: Please include a remark about the corrective action that was taken prior to repeating the 0.05 and 0.08 ARS tests.

REQUESTED CORRECTIVE ACTION

Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

Provide written explanation regarding referenced item(s) to Department Inspector

Remove instrument from evidentiary use until otherwise directed by the Department

Other:

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2024.01.03 13:17:10 -05'00'

01/03/2024

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC
Time of Inspection: 14:16

Date of Inspection: 11/01/2023

Serial Number: 80-007165
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2022	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.050	0.078	0.200	0.081
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080

Number of Simulators Used: 5

Remarks:

2ND INSPECTION - DRY GAS WAS NOT GOING THROUGH ^Q Dry Gas accidentally connected during .08 g/210L test.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

RICHARD J MARROQUIN

11/01/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 13:30

Date of Inspection: 11/01/2023

Serial Number: 80-007165

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: AG222203 Exp: 08/10/2024	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.045 / 0.049	0.000 / 0.000		
0.000	0.044 / 0.049	0.000 / 0.000		
0.000	0.044 / 0.050	0.000 / 0.000		

Number of Simulators Used: 5

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance.

Dry Gas accidently connected for .08 g/210L test.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 N926

Signature and Printed Name

RICHARD J MARROQUIN

11/01/2023
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FFWCC

Instrument Serial Number: 80-007165

AGENCY INSPECTION DISCREPENCY

Agency Inspector: Richard Marroquin

Inspection Date: 05/16/2023

A test was repeated without reason for repeat or the corrective action given

Alcohol Free/Mouth Alcohol Alcohol Free Test Interferent Detect

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Agency Inspection not conducted for calendar month

Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Lot Number and/or Expiration Date for Standards entered incorrectly

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Other Electronic Data Review/Comments: The lot number was typed with an "O" instead of a "0".

REQUESTED CORRECTIVE ACTION

Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

Provide written explanation regarding referenced item(s) to Department Inspector

Remove instrument from evidentiary use until otherwise directed by the Department

Other:

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2023.06.15 12:18:58 -04'00'

06/15/2023

Signature of Alcohol Testing Program Staff Member

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC
Time of Inspection: 12:25

Date of Inspection: 05/16/2023

Serial Number: 80-007165
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) 202108D ^{202108D} Lot#: 202108D ^{202108D} Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#: 202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 1380119 Exp: 09/22/2023
0.000	0.050	0.078	0.199	0.081
0.000	0.050	0.078	0.199	0.080
0.000	0.050	0.078	0.199	0.080


Number of Simulators Used: 5

Remarks:
AI OK

Lot number "Amended" - (P)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



RICHARD J MARROQUIN

Signature and Printed Name

05/16/2023
Date