

Alcohol Testing Program - Instrument Processing Sheet

Agency: FL HIGHWAY PATROL Instrument Serial Number: 80-006635  
 Date In: 1/29/2026 DI Completion Date: 2/12/2026  Ship  P/U  H/D  CMI  EE

Intake By: <u>KTS</u> Date: <u>1/29/2026</u>		Quality Checks By: <u>SLH</u> Date: <u>2/11/2026</u>		Flow Adjustment By: _____																													
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>241</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 102</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.179</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.496</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28421</u> Gauge: <u>1019</u> Instrument: <u>1021</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																													
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K 06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202406L 06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N 06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 04/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202406L 06/19/2026	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG510701 04/17/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:														
Simulator	Serial #	Lot#/Exp																															
0.050	MP6291	202406K 06/19/2026																															
0.080	MP6292	202406L 06/19/2026																															
0.200	MP6293	202406N 06/20/2026																															
0.080 DGS	N/A	AG510701 04/17/2027																															
Optical Bench Adjustment By: _____		Department Inspection By: <u>SLH</u>																															
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34418</u>																															
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1015</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>0.300</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	0.300	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP6293																		
Simulator	Serial Number																																
0.000	MP5086																																
0.300	MP6290																																
0.050	MP5088																																
0.080	MP5089																																
0.200	MP6293																																
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																															
Gauge ID #: _____ Gauge: _____ Instrument: _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
Notes/Suggested Service: Stability checks were performed on 2/12/2026. SLH 2/12/26		Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2026.02.23 14:53:29 -05'00'</small>		Wen-Chi Pierson <small>Digitally signed by Wen-Chi Pierson Date: 2026.03.09 12:46:05 -04'00'</small>																													
		Tech Review		Admin Review																													

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 11:01

Date of Inspection: 02/11/2026

Serial Number: 80-006635  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 LEANDRA HIGGINBOTHAM  
 Signature and Printed Name

02/11/2026  
Date

# Stability Checks

80-006635  
SUA 2/12/26

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006635 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;"><i>M. Sigel</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:37	Control Test	0.050	10:38	Air Blank	0.000	10:38	Control Test	0.050	10:39	Air Blank	0.000	10:39	Control Test	0.050	10:40	Air Blank	0.000	10:41	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006635 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </tbody> </table> <p style="text-align: right;"><i>M. Sigel</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:08	Control Test	0.080	10:09	Air Blank	0.000	10:09	Control Test	0.080	10:10	Air Blank	0.000	10:10	Control Test	0.080	10:11	Air Blank	0.000	10:12	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006635 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2020</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;"><i>M. Sigel</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	10:28	Control Test	0.200	10:29	Air Blank	0.000	10:29	Control Test	0.200	10:30	Air Blank	0.000	10:30	Control Test	0.200	10:31	Air Blank	0.000	10:32	Control Test Stats			Average	0.2020		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p style="text-align: center;"><i>DGP</i></p> <p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006635 02/12/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:39</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:40</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </tbody> </table> <p style="text-align: right;"><i>M. Sigel</i> Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	09:39	Control Test	0.080	09:39	Air Blank	0.000	09:40	Control Test	0.080	09:40	Air Blank	0.000	09:41	Control Test	0.080	09:41	Air Blank	0.000	09:41	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:37																																																																																																																																																	
Control Test	0.050	10:38																																																																																																																																																	
Air Blank	0.000	10:38																																																																																																																																																	
Control Test	0.050	10:39																																																																																																																																																	
Air Blank	0.000	10:39																																																																																																																																																	
Control Test	0.050	10:40																																																																																																																																																	
Air Blank	0.000	10:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.080	10:09																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.080	10:10																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.080	10:11																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0813																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7099																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.200	10:29																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.200	10:30																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test	0.200	10:31																																																																																																																																																	
Air Blank	0.000	10:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2020																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:39																																																																																																																																																	
Control Test	0.080	09:39																																																																																																																																																	
Air Blank	0.000	09:40																																																																																																																																																	
Control Test	0.080	09:40																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test	0.080	09:41																																																																																																																																																	
Air Blank	0.000	09:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0813																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7099																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL  
Time of Inspection: 17:23

Date of Inspection: 02/12/2026

Serial Number: 80-006635  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.051	0.080	0.201	0.081
0.000	0.051	0.080	0.201	0.081
0.000	0.051	0.080	0.201	0.081
0.000	0.051	0.081	0.202	0.080
0.000	0.051	0.081	0.201	0.081
0.000	0.052	0.081	0.202	0.080
0.000	0.053	0.081	0.201	0.080
0.000	0.053	0.081	0.201	0.080
0.000	0.052	0.081	0.201	0.080
0.000	0.053	0.081	0.201	0.080


Standard Deviations	0.0009	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:



The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM  
Signature and Printed Name

02/12/2026  
Date



Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2331 Phillips Road  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006635, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006635</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/12/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>17:23</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005


All results are reported in g/ 210 L.  
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
 \*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3)  
 The instrument results before and after any adjustment are found in the associated pre and post stability checks

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

  
 LEANDRA HIGGINBOTHAM,  
 Department Inspector

02/12/2026

Date

FDLE/ATP Form 69 January 2026  
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality