

Instrument Processing Sheet

Agency: Charlotte County Sheriff's Office Instrument Serial Number: 80-001363
 Date In: 10/28/2025 DI Completion Date: 12/10/2025 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>12/9/25</u>	Quality Checks By: <u>TDG</u> Date: <u>12/9/25</u>	Flow Adjustment By: <u>TDG</u> Date: <u>12/9/25</u>																														
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>208</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm <u>0.128*</u> (.139-.169) 36 mm <u>0.144*</u> (.156-.190) 53 mm <u>0.210*</u> (.228-.278) 103 mm <u>0.488</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1016</u> Instrument: <u>1015</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Flow Column #: <u>ATP106</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>208</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP101</u> 32 mm <u>0.140</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.488</u> (.447-.547)															
Simulator	Serial #	Lot#/Exp																														
0.050	MP6286	202406K 6/19/2026																														
0.080	MP6287	202406L 6/19/2026																														
0.200	MP6288	202406N 6/20/2026																														
0.080 DGS	N/A	AG526603 9/23/2027																														
	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> <th>Maintenance</th> <th>By:</th> <th>Date:</th> </tr> </thead> <tbody> <tr> <td></td> <td></td> <td></td> <td><input type="checkbox"/> Battery Replacement</td> <td></td> <td></td> </tr> <tr> <td></td> <td></td> <td></td> <td><input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare</td> <td></td> <td></td> </tr> <tr> <td></td> <td></td> <td></td> <td><input type="checkbox"/> Breath Tube Replacement</td> <td></td> <td></td> </tr> <tr> <td></td> <td></td> <td></td> <td><input type="checkbox"/> Other:</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:				<input type="checkbox"/> Battery Replacement						<input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare						<input type="checkbox"/> Breath Tube Replacement						<input type="checkbox"/> Other:			
Simulator	Serial #	Lot#/Exp	Maintenance	By:	Date:																											
			<input type="checkbox"/> Battery Replacement																													
			<input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare																													
			<input type="checkbox"/> Breath Tube Replacement																													
			<input type="checkbox"/> Other:																													

Optical Bench Adjustment By: _____	Department Inspection By: <u>TDG</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1017</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Barometric Pressure Gauge: _____ ID#: _____	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								

Notes/Suggested Service: *Outside nominal. Root cause analysis did not identify any user or equipment error. (TDG 12/9/25)	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
Digitally signed by Shayla Platt Date: 2025.12.14 13:47:50 -05'00'	Digitally signed by LeAndra Higginbotham Date: 2025.12.15 12:01:18 -05'00'
Tech Review	Admin Review

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 13:06

Date of Inspection: 12/09/2025

Serial Number: 80-001363
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/09/2025
Date

Flow Calibration Adjustment(s)

Performed by TDG

CHARLOTTE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001363
12/09/2025
Software: 8100.27 ✓

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

√ SQRT(Diff)) = 8.059

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.660

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.141

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 720

Rounded Intercept = -845000

Correlation = 0.99037

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001363 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:35	Control Test	0.048	13:35	Air Blank	0.000	13:36	Control Test	0.049	13:36	Air Blank	0.000	13:37	Control Test	0.048	13:38	Air Blank	0.000	13:38	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001363 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:17	Control Test	0.079	13:18	Air Blank	0.000	13:18	Control Test	0.079	13:19	Air Blank	0.000	13:19	Control Test	0.079	13:20	Air Blank	0.000	13:20	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001363 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:24	Control Test	0.200	13:24	Air Blank	0.000	13:25	Control Test	0.199	13:25	Air Blank	0.000	13:26	Control Test	0.199	13:27	Air Blank	0.000	13:27	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		<p>CHARLOTTE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001363 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:40	Control Test	0.078	13:41	Air Blank	0.000	13:41	Control Test	0.078	13:42	Air Blank	0.000	13:42	Control Test	0.079	13:42	Air Blank	0.000	13:43	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:35																																																																																																																																																	
Control Test	0.048	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.049	13:36																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test	0.048	13:38																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.079	13:18																																																																																																																																																	
Air Blank	0.000	13:18																																																																																																																																																	
Control Test	0.079	13:19																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.079	13:20																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.200	13:24																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test	0.199	13:25																																																																																																																																																	
Air Blank	0.000	13:26																																																																																																																																																	
Control Test	0.199	13:27																																																																																																																																																	
Air Blank	0.000	13:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2896																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test	0.078	13:41																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.078	13:42																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.079	13:42																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CHARLOTTE COUNTY SO
Time of Inspection: 13:04

Date of Inspection: 12/10/2025

Serial Number: 80-001363
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.079	0.198	0.078
0.000	0.048	0.078	0.198	0.078


Standard Deviations	0.0000	0.0003	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/10/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001363, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001363</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>CHARLOTTE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/10/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Date: 2025.12.10 15:05:24
-05'00'

12/10/2025
Date
TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality