

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 16:08

Date of Inspection: 03/31/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.202	0.080
0.000	0.049	0.081	0.203	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


JESSE L BURKETT  
 \_\_\_\_\_  
 Signature and Printed Name

03/31/2022  
 Date

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
03/31/2022  
Software: 8100.27

-----  
DIAGNOSTICS

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
03/31/2022  
Software: 8100.27

-----  
WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	13:48
Control Test	0.049	13:49
Air Blank	0.000	13:49



Operator's Signature

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
03/31/2022  
Software: 8100.27

-----  
WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	13:46
Control Test	0.000	13:47
Air Blank	0.000	13:47



Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 13:42

Date of Inspection: 03/31/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.049	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		


Number of Simulators Used: 5

**Remarks:**

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



\_\_\_\_\_  
Signature and Printed Name


JESSE L BURKETT

03/31/2022  
Date

PORT ST JOE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-000908  
03/31/2022  
Software: 8100.27

-----  
WET CONTROL TEST

Test	g/210L	Time
Air Blank	0.000	13:51
Control Test	0.200	13:51
Air Blank	0.000	13:52

  
-----  
Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 18:20

Date of Inspection: 05/12/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202007A Exp: 07/21/2022	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.043 / 0.049	0.079	0.200	0.079
0.000	0.044 / 0.049	0.079	0.202	0.080
0.000	0.045 / 0.049	0.080	0.203	0.079

Number of Simulators Used: 5

**Remarks:**

05: Control Outside Tolerance.

*.05 simulator top accidentally turned off*

*Lot# 202101C  
Exp: 01-12-23*

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Dalton J Brammer*

DALTON J BRAMMER

Signature and Printed Name

05/12/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 14:39

Date of Inspection: 08/12/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:

Unintended verification of compliance / Issue corrected instrument is in compliance error made by This Agency Inspector.

The above instrument complies ( SD ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JESSE L BURKETT

\_\_\_\_\_  
Signature and Printed Name

08/12/2022  
Date



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 15:46

Date of Inspection: 08/12/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.050	0.078	0.202	0.079

Number of Simulators Used: 5

**Remarks:**

A F / M A: Range Exceeded.

Over saturation of M/A during initial test / issue resolved

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JESSE L BURKETT

Signature and Printed Name

08/12/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 11:43

Date of Inspection: 09/11/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____
0.000				
0.000				
0.000				


Number of Simulators Used: 5

**Remarks:**

Int Det:CONTAMINATED SAMPLE, . Non-compliance: UNABLE TO OBTAIN INTERFER DETECT.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JESSE L BURKETT

Signature and Printed Name

09/11/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 14:12

Date of Inspection: 09/12/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202101C Exp: 01/12/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.078	0.202	0.079
0.000	0.049	0.078	0.201	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JESSE L BURKETT

Signature and Printed Name

09/12/2022  
Date

**From:** [Jesse Burkett](#)  
**To:** [Soto, Israel](#)  
**Subject:** SEPT 2022 INTOX AFF PSJPD  
**Date:** Wednesday, September 21, 2022 5:18:14 AM  
**Attachments:** [INTOX AFF 091122.pdf](#)  
[INTOX AFF 091222.pdf](#)

---

**CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.**

First test 09/11/22 was unable to obtain interf det. Second test 09/12/22 was successful after adjustments to simulator.

Sent from [Mail](#) for Windows

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 17:55

Date of Inspection: 12/18/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.051	0.087 / 0.083	0.218 / 0.201	0.080
0.000	0.051	0.086 / 0.082	0.215 / 0.201	0.080
0.000	0.051	0.085 / 0.081	0.202 / 0.201	0.080

Number of Simulators Used: 5

**Remarks:**

Int Det: INLET TUBE NOT SECURED FIRST TEST. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

*17:55 Test did not comply due to M/A Error*

The above instrument complies ( ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



JESSE L BURKETT

Signature and Printed Name

12/18/2022  
Date

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD  
Time of Inspection: 22:49

Date of Inspection: 12/18/2022

Serial Number: 80-000908  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202106B Exp: 06/22/2023	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1327124 Exp: 05/14/2023
0.000	0.049	0.080	0.202	0.080
0.000	0.050	0.084	0.204	0.080
0.000	0.050	0.081	0.204	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

JESSE L BURKETT

Signature and Printed Name

12/18/2022  
Date