

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000000

Agency: TALLAHASSEE TRAINING

Date/Time: 10/13/2025 17:23:05

Data Download Date/Time: 12/09/2025 10:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000200

Agency: FDLE TRAINER

Date/Time: 02/27/2025 14:34:23

Data Download Date/Time: 05/01/2025 14:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINER

Date/Time: 11/03/2025 12:02:10

Data Download Date/Time: 12/09/2025 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000200

Agency: FDLE TRAINER

Date/Time: 11/03/2025 12:02:55

Data Download Date/Time: 12/09/2025 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE

Date/Time: 04/11/2025 14:13:26

Data Download Date/Time: 06/13/2025 17:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE

Date/Time: 07/01/2025 09:32:59

Data Download Date/Time: 07/01/2025 10:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000219

Agency: FDLE

Date/Time: 08/04/2025 09:44:47

Data Download Date/Time: 08/04/2025 10:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000220

Agency: FDLE TRAINER

Date/Time: 02/24/2025 11:53:59

Data Download Date/Time: 10/07/2025 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FDLE TRAINER

Date/Time: 04/02/2025 09:06:49

Data Download Date/Time: 10/07/2025 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000223

Agency: FDLE

Date/Time: 04/05/2025 08:51:12

Data Download Date/Time: 04/30/2025 22:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000224

Agency: FDLE

Date/Time: 09/25/2025 23:37:05

Data Download Date/Time: 10/16/2025 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

Agency: FDLE

Date/Time: 03/31/2025 17:11:31

Data Download Date/Time: 04/02/2025 05:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000228

Agency: FDLE

Date/Time: 11/24/2025 08:46:59

Data Download Date/Time: 12/05/2025 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

Agency: ALACHUA COUNTY SO

Date/Time: 11/04/2025 14:19:12

Data Download Date/Time: 11/04/2025 14:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 09/03/2025 13:20:21

Data Download Date/Time: 09/04/2025 13:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: VALPARAISO PD

Date/Time: 09/21/2025 02:50:10

Data Download Date/Time: 10/10/2025 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 09/22/2025 01:29:55

Data Download Date/Time: 10/10/2025 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: VALPARAISO PD

Date/Time: 09/22/2025 01:31:39

Data Download Date/Time: 10/10/2025 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

Agency: VALPARAISO PD

Date/Time: 09/22/2025 01:38:29

Data Download Date/Time: 10/10/2025 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

Agency: COLUMBIA COUNTY SO

Date/Time: 02/17/2025 13:52:14

Data Download Date/Time: 02/25/2025 14:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: COLUMBIA COUNTY SO

Date/Time: 12/12/2025 10:02:06

Data Download Date/Time: 12/12/2025 17:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000777

Agency: COLUMBIA COUNTY SO

Date/Time: 10/24/2025 08:54:57

Data Download Date/Time: 10/24/2025 09:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

Agency: SUWANNEE COUNTY SO

Date/Time: 06/21/2025 22:51:36

Data Download Date/Time: 06/23/2025 13:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000781

Agency: FL HIGHWAY PATROL

Date/Time: 04/04/2025 12:57:22

Data Download Date/Time: 05/12/2025 14:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 04/04/2025 13:06:08

Data Download Date/Time: 05/12/2025 14:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

Agency: ST LUCIE COUNTY SO

Date/Time: 05/25/2025 19:06:46

Data Download Date/Time: 07/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST LUCIE COUNTY SO

Date/Time: 06/12/2025 09:41:36

Data Download Date/Time: 07/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000790

Agency: VERO BEACH P.D.

Date/Time: 02/17/2025 09:36:19

Data Download Date/Time: 02/17/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: VERO BEACH P.D.

Date/Time: 02/17/2025 09:43:46

Data Download Date/Time: 02/17/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

Agency: LARGO PD

Date/Time: 11/20/2025 09:46:15

Data Download Date/Time: 11/20/2025 12:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 11/20/2025 09:48:57

Data Download Date/Time: 11/20/2025 12:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000791

Agency: LARGO PD

Date/Time: 11/20/2025 09:49:53

Data Download Date/Time: 11/20/2025 12:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 11/20/2025 11:38:42

Data Download Date/Time: 11/20/2025 12:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000794

Agency: ST LUCIE COUNTY SO

Date/Time: 05/29/2025 15:17:31

Data Download Date/Time: 05/29/2025 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 03/21/2025 23:14:47

Data Download Date/Time: 03/22/2025 00:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GULF COUNTY SO

Date/Time: 09/15/2025 13:49:28

Data Download Date/Time: 11/03/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

Agency: GULF COUNTY SO

Date/Time: 11/01/2025 19:57:14

Data Download Date/Time: 11/03/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000796

Agency: HAINES CITY PD

Date/Time: 02/09/2025 20:05:11

Data Download Date/Time: 02/09/2025 20:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAINES CITY PD

Date/Time: 06/12/2025 15:39:48

Data Download Date/Time: 06/13/2025 17:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

Agency: HERNANDO COUNTY SO

Date/Time: 04/16/2025 21:57:13

Data Download Date/Time: 04/29/2025 05:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 04/30/2025 09:47:58

Data Download Date/Time: 04/30/2025 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 07/01/2025 09:33:26

Data Download Date/Time: 07/01/2025 10:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

Agency: SUMTER COUNTY SO

Date/Time: 08/04/2025 09:43:52

Data Download Date/Time: 08/04/2025 10:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

Agency: CITRUS COUNTY SO

Date/Time: 12/23/2025 02:29:26

Data Download Date/Time: 12/23/2025 03:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CITRUS COUNTY SO

Date/Time: 12/23/2025 03:58:25

Data Download Date/Time: 12/23/2025 03:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000823

Agency: LAKE COUNTY SO

Date/Time: 08/15/2025 16:58:07

Data Download Date/Time: 08/15/2025 16:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

Agency: HILLSBOROUGH CO SO

Date/Time: 07/30/2025 12:22:51

Data Download Date/Time: 07/30/2025 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HILLSBOROUGH CO SO

Date/Time: 11/04/2025 14:49:40

Data Download Date/Time: 11/05/2025 13:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 05/30/2025 14:49:28

Data Download Date/Time: 06/12/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PANAMA CITY BEACH PD

Date/Time: 07/04/2025 14:16:51

Data Download Date/Time: 07/04/2025 16:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 08/20/2025 15:10:19

Data Download Date/Time: 08/29/2025 21:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PANAMA CITY BEACH PD

Date/Time: 10/17/2025 04:38:13

Data Download Date/Time: 10/17/2025 15:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000838

Agency: PANAMA CITY BEACH PD

Date/Time: 10/17/2025 05:35:13

Data Download Date/Time: 10/17/2025 15:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:34:46

Data Download Date/Time: 03/25/2025 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 09/27/2025 07:51:45

Data Download Date/Time: 09/27/2025 09:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000860

Agency: POLK COUNTY SO

Date/Time: 12/23/2025 08:14:30

Data Download Date/Time: 12/23/2025 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency: MIAMI PD

Date/Time: 09/30/2025 06:26:39

Data Download Date/Time: 10/30/2025 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MIAMI PD

Date/Time: 09/30/2025 06:33:13

Data Download Date/Time: 10/30/2025 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

Agency: MIAMI PD

Date/Time: 09/30/2025 06:40:05

Data Download Date/Time: 10/30/2025 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

Agency: MIAMI DADE SO

Date/Time: 09/16/2025 17:32:53

Data Download Date/Time: 10/23/2025 01:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

Agency: MIAMI DADE PD

Date/Time: 06/25/2025 13:29:06

Data Download Date/Time: 07/29/2025 13:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

Agency: FFWCC

Date/Time: 10/07/2025 07:29:30

Data Download Date/Time: 10/14/2025 14:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 02/13/2025 20:48:04

Data Download Date/Time: 02/13/2025 21:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 03/21/2025 22:19:32

Data Download Date/Time: 04/21/2025 11:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:51:59

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:52:36

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 17:55:28

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 19:11:37

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 05/13/2025 19:12:22

Data Download Date/Time: 06/23/2025 23:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST JOE PD

Date/Time: 12/08/2025 23:12:39

Data Download Date/Time: 12/30/2025 21:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

Agency: PORT ST JOE PD

Date/Time: 12/08/2025 23:13:18

Data Download Date/Time: 12/30/2025 21:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ST JOE PD

Date/Time: 12/08/2025 23:13:50

Data Download Date/Time: 12/30/2025 21:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

Agency: CHARLOTTE COUNTY SO

Date/Time: 07/07/2025 12:16:43

Data Download Date/Time: 10/06/2025 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CHARLOTTE COUNTY SO

Date/Time: 07/07/2025 12:17:24

Data Download Date/Time: 10/06/2025 11:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

Agency: LEON COUNTY SO

Date/Time: 02/17/2025 17:06:13

Data Download Date/Time: 02/25/2025 10:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000959

Agency: GADSDEN COUNTY SO

Date/Time: 02/26/2025 10:56:06

Data Download Date/Time: 02/26/2025 12:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

Agency: WAKULLA COUNTY SO

Date/Time: 12/30/2025 16:09:47

Data Download Date/Time: 01/20/2026 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 09:58:09

Data Download Date/Time: 01/13/2026 11:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 10:08:55

Data Download Date/Time: 01/13/2026 11:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 22:48:33

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 22:49:05

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 01/23/2025 23:19:36

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 01/24/2025 00:17:01

Data Download Date/Time: 01/24/2025 01:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 02/28/2025 20:58:38

Data Download Date/Time: 02/28/2025 23:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 03/31/2025 02:55:10

Data Download Date/Time: 03/31/2025 04:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 00:15:33

Data Download Date/Time: 04/30/2025 20:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 00:17:03

Data Download Date/Time: 04/30/2025 20:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/27/2025 01:29:23

Data Download Date/Time: 04/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 19:33:49

Data Download Date/Time: 04/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:36:36

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:43:05

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 04/30/2025 23:43:49

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/01/2025 05:55:11

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/10/2025 20:47:36

Data Download Date/Time: 05/28/2025 20:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/11/2025 00:18:48

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/24/2025 03:00:58

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/24/2025 03:44:47

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/25/2025 05:07:37

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: TITUSVILLE PD

Date/Time: 05/25/2025 06:52:07

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 05/26/2025 00:55:34

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 05/28/2025 19:56:29

Data Download Date/Time: 05/28/2025 20:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/04/2025 04:51:00

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 06/04/2025 04:52:53

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/07/2025 23:57:38

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 06/22/2025 21:46:31

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

Agency: TITUSVILLE PD

Date/Time: 06/30/2025 19:16:31

Data Download Date/Time: 06/30/2025 20:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TITUSVILLE PD

Date/Time: 07/29/2025 18:29:34

Data Download Date/Time: 07/29/2025 20:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY SO

Date/Time: 10/03/2025 04:22:00

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SEMINOLE COUNTY SO

Date/Time: 10/03/2025 04:22:59

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY SO

Date/Time: 10/03/2025 04:36:44

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SEMINOLE COUNTY SO

Date/Time: 10/03/2025 04:38:04

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY SO

Date/Time: 10/10/2025 21:09:16

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SEMINOLE COUNTY SO

Date/Time: 10/18/2025 04:07:29

Data Download Date/Time: 10/26/2025 02:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000998

Agency: SEMINOLE COUNTY SO

Date/Time: 10/19/2025 03:41:39

Data Download Date/Time: 10/26/2025 02:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 01/26/2025 20:06:55

Data Download Date/Time: 01/31/2025 23:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:23:28

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:28:43

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HARDEE COUNTY SO

Date/Time: 03/06/2025 00:46:14

Data Download Date/Time: 03/18/2025 03:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:13:55

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:26:07

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

Agency: HAINES CITY PD

Date/Time: 05/21/2025 18:41:51

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HAINES CITY PD

Date/Time: 06/14/2025 03:53:04

Data Download Date/Time: 06/14/2025 05:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE PD

Date/Time: 04/19/2025 12:08:10

Data Download Date/Time: 04/19/2025 15:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WEST MELBOURNE PD

Date/Time: 06/10/2025 12:00:25

Data Download Date/Time: 06/10/2025 13:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

Agency: WEST MELBOURNE PD

Date/Time: 08/22/2025 08:57:52

Data Download Date/Time: 08/22/2025 10:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WEST MELBOURNE PD

Date/Time: 10/09/2025 12:41:02

Data Download Date/Time: 10/14/2025 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 03/27/2025 11:32:38

Data Download Date/Time: 03/27/2025 13:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 04/30/2025 07:57:42

Data Download Date/Time: 04/30/2025 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

Agency: PALM BAY P.D.

Date/Time: 06/26/2025 13:55:01

Data Download Date/Time: 06/26/2025 15:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 06:59:22

Data Download Date/Time: 01/27/2025 07:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 02/27/2025 06:58:12

Data Download Date/Time: 02/28/2025 06:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 14:46:52

Data Download Date/Time: 03/26/2025 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 14:48:13

Data Download Date/Time: 03/26/2025 14:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 04/24/2025 07:38:04

Data Download Date/Time: 04/24/2025 09:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 10:56:57

Data Download Date/Time: 05/25/2025 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 11:03:31

Data Download Date/Time: 05/25/2025 11:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 06/28/2025 10:06:45

Data Download Date/Time: 06/28/2025 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 07/29/2025 17:16:17

Data Download Date/Time: 07/29/2025 18:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 08/18/2025 00:47:39

Data Download Date/Time: 09/22/2025 10:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 10/22/2025 00:00:04

Data Download Date/Time: 10/22/2025 00:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 10/22/2025 00:03:32

Data Download Date/Time: 10/22/2025 00:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001046

Agency: CLAY COUNTY SO

Date/Time: 12/30/2025 23:28:31

Data Download Date/Time: 12/30/2025 23:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 09:36:12

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 09:36:50

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001048

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 10:02:18

Data Download Date/Time: 04/11/2025 16:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE ALFRED PD

Date/Time: 04/11/2025 10:51:18

Data Download Date/Time: 04/11/2025 16:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 07/25/2025 11:57:02

Data Download Date/Time: 07/25/2025 13:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 08/29/2025 11:55:55

Data Download Date/Time: 08/29/2025 13:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 09/26/2025 11:51:49

Data Download Date/Time: 09/26/2025 13:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PASCO COUNTY SO

Date/Time: 09/26/2025 12:04:21

Data Download Date/Time: 09/26/2025 13:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 10/17/2025 11:24:23

Data Download Date/Time: 10/17/2025 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PASCO COUNTY SO

Date/Time: 10/17/2025 11:26:14

Data Download Date/Time: 10/17/2025 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 10/17/2025 11:37:58

Data Download Date/Time: 10/17/2025 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 10/17/2025 11:54:17

Data Download Date/Time: 10/17/2025 14:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 11/21/2025 13:24:28

Data Download Date/Time: 11/21/2025 14:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/21/2025 13:25:40

Data Download Date/Time: 11/21/2025 14:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001049

Agency: PASCO COUNTY SO

Date/Time: 12/31/2025 13:42:58

Data Download Date/Time: 12/31/2025 15:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

Agency: CORAL SPRINGS PD

Date/Time: 02/11/2025 14:06:51

Data Download Date/Time: 02/28/2025 18:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

Agency: POLK COUNTY SO

Date/Time: 11/22/2025 06:15:40

Data Download Date/Time: 11/22/2025 07:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 12/23/2025 06:38:30

Data Download Date/Time: 12/23/2025 07:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:39:31

Data Download Date/Time: 02/05/2025 13:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 05/20/2025 08:19:51

Data Download Date/Time: 05/20/2025 09:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK CSO

Date/Time: 05/22/2025 08:44:20

Data Download Date/Time: 06/26/2025 06:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 07/29/2025 08:08:22

Data Download Date/Time: 07/29/2025 08:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK CSO

Date/Time: 08/26/2025 08:03:13

Data Download Date/Time: 08/26/2025 10:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 09/27/2025 07:52:43

Data Download Date/Time: 09/27/2025 09:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK CSO

Date/Time: 11/22/2025 06:14:03

Data Download Date/Time: 11/22/2025 07:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001058

Agency: POLK CSO

Date/Time: 11/22/2025 06:15:33

Data Download Date/Time: 11/22/2025 07:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK CSO

Date/Time: 12/23/2025 08:15:35

Data Download Date/Time: 12/23/2025 09:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:42:17

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:44:27

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 03/12/2025 17:49:22

Data Download Date/Time: 03/12/2025 18:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLEARWATER PD

Date/Time: 04/29/2025 19:33:56

Data Download Date/Time: 05/03/2025 00:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:11:50

Data Download Date/Time: 05/31/2025 03:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 06/30/2025 05:41:26

Data Download Date/Time: 07/04/2025 00:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 07/17/2025 18:06:24

Data Download Date/Time: 07/17/2025 19:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 08/29/2025 15:06:10

Data Download Date/Time: 09/26/2025 21:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

Agency: CLEARWATER PD

Date/Time: 11/22/2025 21:14:26

Data Download Date/Time: 11/22/2025 22:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 12/17/2025 13:05:52

Data Download Date/Time: 12/17/2025 14:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

Agency: CLAY COUNTY SO

Date/Time: 11/30/2025 11:12:16

Data Download Date/Time: 11/30/2025 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CLAY COUNTY SO

Date/Time: 12/30/2025 20:58:36

Data Download Date/Time: 12/30/2025 23:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 01/17/2025 12:19:58

Data Download Date/Time: 01/17/2025 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 02/21/2025 10:34:27

Data Download Date/Time: 02/21/2025 11:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 03/26/2025 13:09:32

Data Download Date/Time: 03/28/2025 11:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:37:54

Data Download Date/Time: 04/30/2025 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:39:10

Data Download Date/Time: 04/30/2025 10:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 08/29/2025 11:53:46

Data Download Date/Time: 08/29/2025 13:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 09/26/2025 11:45:51

Data Download Date/Time: 09/26/2025 12:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/11/2025 10:27:00

Data Download Date/Time: 11/21/2025 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 11/11/2025 10:28:47

Data Download Date/Time: 11/21/2025 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/12/2025 10:52:27

Data Download Date/Time: 11/21/2025 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 11/21/2025 13:11:14

Data Download Date/Time: 11/21/2025 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/21/2025 13:28:28

Data Download Date/Time: 11/21/2025 14:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

Agency: PASCO COUNTY SO

Date/Time: 11/28/2025 11:36:36

Data Download Date/Time: 12/31/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 12/31/2025 13:45:42

Data Download Date/Time: 12/31/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

Agency: FHP TROOP E MIAMI

Date/Time: 06/13/2025 02:22:50

Data Download Date/Time: 12/18/2025 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP TROOP E MIAMI

Date/Time: 06/13/2025 02:24:55

Data Download Date/Time: 12/18/2025 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

Agency: FHP TROOP E MIAMI

Date/Time: 06/20/2025 08:41:44

Data Download Date/Time: 12/18/2025 14:27

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FHP TROOP E MIAMI

Date/Time: 12/05/2025 10:17:59

Data Download Date/Time: 12/18/2025 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001077

Agency: HARDEE COUNTY SO

Date/Time: 02/17/2025 15:52:03

Data Download Date/Time: 02/24/2025 11:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

Agency: LARGO PD

Date/Time: 11/20/2025 09:18:51

Data Download Date/Time: 11/20/2025 11:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LARGO PD

Date/Time: 11/20/2025 09:29:30

Data Download Date/Time: 11/20/2025 11:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 00:51:53

Data Download Date/Time: 01/15/2025 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 01:45:05

Data Download Date/Time: 01/15/2025 10:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 03/31/2025 07:13:00

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 09:21:52

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 09:40:46

Data Download Date/Time: 04/22/2025 09:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 08/26/2025 05:11:29

Data Download Date/Time: 08/26/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001080

Agency: POLK COUNTY SO

Date/Time: 12/11/2025 15:24:21

Data Download Date/Time: 12/12/2025 10:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 01/17/2025 12:39:01

Data Download Date/Time: 01/17/2025 13:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 02/21/2025 10:33:35

Data Download Date/Time: 02/21/2025 11:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 03/26/2025 13:09:11

Data Download Date/Time: 03/28/2025 10:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 04/30/2025 09:37:51

Data Download Date/Time: 04/30/2025 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 08/29/2025 12:09:44

Data Download Date/Time: 08/29/2025 13:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 08/29/2025 12:38:47

Data Download Date/Time: 08/29/2025 13:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 09/26/2025 11:48:36

Data Download Date/Time: 09/26/2025 13:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/11/2025 10:12:13

Data Download Date/Time: 11/28/2025 12:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 11/11/2025 10:41:03

Data Download Date/Time: 11/28/2025 12:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PASCO COUNTY SO

Date/Time: 11/12/2025 10:42:39

Data Download Date/Time: 11/28/2025 12:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

Agency: PASCO COUNTY SO

Date/Time: 11/28/2025 11:08:58

Data Download Date/Time: 11/28/2025 12:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PASCO COUNTY SO

Date/Time: 11/28/2025 11:33:26

Data Download Date/Time: 11/28/2025 12:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001116

Agency: EGLIN AFB

Date/Time: 05/21/2025 10:00:19

Data Download Date/Time: 06/22/2025 15:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: EGLIN AFB

Date/Time: 12/25/2025 05:52:34

Data Download Date/Time: 12/30/2025 08:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001118

Agency: FL HIGHWAY PATROL

Date/Time: 04/18/2025 07:16:01

Data Download Date/Time: 05/12/2025 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

Agency: FL HIGHWAY PATROL

Date/Time: 10/22/2025 08:31:16

Data Download Date/Time: 12/02/2025 10:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 10/22/2025 09:06:01

Data Download Date/Time: 12/02/2025 10:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

Agency: FL HIGHWAY PATROL

Date/Time: 10/30/2025 09:01:46

Data Download Date/Time: 12/02/2025 10:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001123

Agency: FHP TROOP D ORLANDO

Date/Time: 04/10/2025 23:10:54

Data Download Date/Time: 06/09/2025 14:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FHP TROOP D ORLANDO

Date/Time: 07/18/2025 14:03:27

Data Download Date/Time: 07/18/2025 14:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001127

Agency: POLK COUNTY SO

Date/Time: 01/14/2025 09:53:02

Data Download Date/Time: 01/14/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

Agency: PUTNAM COUNTY SO

Date/Time: 08/30/2025 19:29:23

Data Download Date/Time: 08/30/2025 20:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

Agency: VOLUSIA COUNTY SO

Date/Time: 07/03/2025 12:51:21

Data Download Date/Time: 07/30/2025 16:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

Agency: NEW SMYRNA BEACH PD

Date/Time: 03/06/2025 08:37:54

Data Download Date/Time: 04/16/2025 10:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: NEW SMYRNA BEACH PD

Date/Time: 06/25/2025 13:16:21

Data Download Date/Time: 07/11/2025 12:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

Agency: EGLIN AFB

Date/Time: 03/12/2025 23:34:06

Data Download Date/Time: 03/13/2025 03:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: EGLIN AFB

Date/Time: 10/11/2025 12:19:08

Data Download Date/Time: 10/11/2025 15:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001157

Agency: POLK COUNTY SO

Date/Time: 11/22/2025 06:06:01

Data Download Date/Time: 11/22/2025 07:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 12/23/2025 08:14:53

Data Download Date/Time: 12/23/2025 09:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

Agency: ORANGE CITY P.D.

Date/Time: 02/17/2025 18:03:36

Data Download Date/Time: 02/24/2025 11:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

Agency: GROVELAND PD

Date/Time: 02/09/2025 13:24:47

Data Download Date/Time: 03/03/2025 16:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001178

Agency: MIAMI PD

Date/Time: 08/07/2025 11:45:37

Data Download Date/Time: 08/22/2025 16:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency: MIAMI PD

Date/Time: 01/31/2025 21:00:36

Data Download Date/Time: 02/22/2025 03:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MIAMI PD

Date/Time: 03/16/2025 19:00:56

Data Download Date/Time: 03/26/2025 22:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

Agency: MIAMI PD

Date/Time: 11/21/2025 11:28:29

Data Download Date/Time: 12/18/2025 13:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

Agency: PANAMA CITY BEACH PD

Date/Time: 04/04/2025 16:20:40

Data Download Date/Time: 04/14/2025 06:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PANAMA CITY BEACH PD

Date/Time: 10/17/2025 03:42:14

Data Download Date/Time: 10/17/2025 15:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

Agency: PANAMA CITY BEACH PD

Date/Time: 10/17/2025 03:51:04

Data Download Date/Time: 10/17/2025 15:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PANAMA CITY BEACH PD

Date/Time: 10/17/2025 03:56:21

Data Download Date/Time: 10/17/2025 15:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

Agency: COLLIER COUNTY SO

Date/Time: 07/19/2025 13:26:38

Data Download Date/Time: 07/19/2025 15:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

Agency: POLK COUNTY SO

Date/Time: 02/26/2025 13:57:20

Data Download Date/Time: 02/27/2025 08:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: LAKE COUNTY SO

Date/Time: 10/27/2025 14:19:30

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKE COUNTY SO

Date/Time: 10/27/2025 14:21:00

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: LAKE COUNTY SO

Date/Time: 10/27/2025 14:25:43

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKE COUNTY SO

Date/Time: 10/27/2025 14:26:33

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: LAKE COUNTY SO

Date/Time: 11/28/2025 16:05:59

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKE COUNTY SO

Date/Time: 11/28/2025 16:07:24

Data Download Date/Time: 12/06/2025 11:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: LAKE COUNTY SO

Date/Time: 12/06/2025 11:58:37

Data Download Date/Time: 01/01/2026 12:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE COUNTY SO

Date/Time: 12/06/2025 12:00:25

Data Download Date/Time: 01/01/2026 12:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

Agency: LAKE COUNTY SO

Date/Time: 12/06/2025 12:01:06

Data Download Date/Time: 01/01/2026 12:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LAKE COUNTY SO

Date/Time: 12/06/2025 12:04:46

Data Download Date/Time: 01/01/2026 12:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER PD

Date/Time: 08/11/2025 17:23:16

Data Download Date/Time: 08/25/2025 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: EDGEWATER PD

Date/Time: 08/11/2025 17:29:53

Data Download Date/Time: 08/25/2025 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

Agency: EDGEWATER PD

Date/Time: 10/17/2025 12:47:53

Data Download Date/Time: 10/31/2025 09:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: EDGEWATER PD

Date/Time: 11/04/2025 16:56:27

Data Download Date/Time: 12/29/2025 12:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

Agency: PUTNAM COUNTY SO

Date/Time: 12/23/2025 18:57:04

Data Download Date/Time: 12/23/2025 20:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

Agency: SOUTH DAYTONA PD

Date/Time: 08/05/2025 07:32:31

Data Download Date/Time: 08/05/2025 08:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:28:21

Data Download Date/Time: 04/05/2025 19:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:36:21

Data Download Date/Time: 04/05/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

Agency: FL HIGHWAY PATROL

Date/Time: 03/04/2025 16:37:33

Data Download Date/Time: 04/05/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency: MELBOURNE PD

Date/Time: 10/23/2025 07:31:16

Data Download Date/Time: 10/23/2025 08:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MELBOURNE PD

Date/Time: 11/25/2025 07:41:19

Data Download Date/Time: 11/25/2025 08:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

Agency: MELBOURNE PD

Date/Time: 12/30/2025 07:45:55

Data Download Date/Time: 01/14/2026 08:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

Agency: INDIAN HARBOUR PD

Date/Time: 09/30/2025 15:53:47

Data Download Date/Time: 11/14/2025 07:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

Agency: BREVARD COUNTY S.O.

Date/Time: 02/17/2025 10:52:01

Data Download Date/Time: 02/17/2025 18:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COUNTY S.O.

Date/Time: 12/26/2025 07:45:32

Data Download Date/Time: 01/20/2026 10:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY PD

Date/Time: 04/30/2025 07:57:03

Data Download Date/Time: 04/30/2025 11:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:53:17

Data Download Date/Time: 06/26/2025 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:54:34

Data Download Date/Time: 06/26/2025 15:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY PD

Date/Time: 10/14/2025 17:37:53

Data Download Date/Time: 10/23/2025 19:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

Agency: PALM BAY P.D.

Date/Time: 02/26/2025 11:06:29

Data Download Date/Time: 02/26/2025 12:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY P.D.

Date/Time: 06/26/2025 13:53:48

Data Download Date/Time: 06/26/2025 15:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

Agency: SEMINOLE COUNTY SO

Date/Time: 02/06/2025 00:15:03

Data Download Date/Time: 02/26/2025 23:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

Agency: NASSAU COUNTY SO

Date/Time: 10/20/2025 10:45:48

Data Download Date/Time: 10/30/2025 09:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001284

Agency: GILCHRIST COUNTY SO

Date/Time: 09/01/2025 10:15:38

Data Download Date/Time: 12/16/2025 13:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

Agency: BAKER COUNTY SO

Date/Time: 06/23/2025 16:20:58

Data Download Date/Time: 07/17/2025 13:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency: LEVY COUNTY SO

Date/Time: 08/24/2025 15:27:05

Data Download Date/Time: 01/08/2026 09:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEVY COUNTY SO

Date/Time: 08/24/2025 15:32:21

Data Download Date/Time: 01/08/2026 09:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

Agency: LEVY COUNTY SO

Date/Time: 08/24/2025 15:40:37

Data Download Date/Time: 01/08/2026 09:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/09/2025 11:24:53

Data Download Date/Time: 11/19/2025 20:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SANTA ROSA COUNTY SO

Date/Time: 11/09/2025 11:28:54

Data Download Date/Time: 11/19/2025 20:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

Agency: NICEVILLE PD

Date/Time: 10/09/2025 14:44:58

Data Download Date/Time: 10/09/2025 15:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

Agency: WALTON COUNTY SO

Date/Time: 09/30/2025 20:45:28

Data Download Date/Time: 10/17/2025 09:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WALTON COUNTY SO

Date/Time: 09/30/2025 21:09:01

Data Download Date/Time: 10/17/2025 09:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

Agency: WALTON COUNTY SO

Date/Time: 10/15/2025 06:37:50

Data Download Date/Time: 10/17/2025 09:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: WALTON COUNTY SO

Date/Time: 10/17/2025 12:47:29

Data Download Date/Time: 11/21/2025 17:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001314

Agency: BREVARD COMM COLLEG

Date/Time: 10/16/2025 15:20:56

Data Download Date/Time: 11/19/2025 13:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	Fail
Modem Test	OK
Temperature Regulation Test	OK

Agency: BREVARD COMM COLLEG

Date/Time: 10/16/2025 15:22:45

Data Download Date/Time: 11/19/2025 13:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001317

Agency: BREVARD COMM COLLEG

Date/Time: 02/26/2025 10:26:46

Data Download Date/Time: 11/19/2025 13:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

Agency: OKEECHOBEE COUNTY SO

Date/Time: 03/13/2025 18:16:42

Data Download Date/Time: 03/13/2025 19:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE COUNTY SO

Date/Time: 05/07/2025 04:41:51

Data Download Date/Time: 05/07/2025 05:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

Agency: OKEECHOBEE COUNTY SO

Date/Time: 08/07/2025 12:56:42

Data Download Date/Time: 08/07/2025 12:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	Fail
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE COUNTY SO

Date/Time: 01/22/2025 10:01:41

Data Download Date/Time: 01/22/2025 11:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OKEECHOBEE COUNTY SO

Date/Time: 03/13/2025 18:14:50

Data Download Date/Time: 03/13/2025 19:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

Agency: OKEECHOBEE COUNTY SO

Date/Time: 05/20/2025 18:19:39

Data Download Date/Time: 05/20/2025 19:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OKEECHOBEE COUNTY SO

Date/Time: 06/26/2025 17:10:08

Data Download Date/Time: 06/26/2025 17:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

Agency: PORT ST LUCIE PD

Date/Time: 02/18/2025 10:20:13

Data Download Date/Time: 02/18/2025 11:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST LUCIE PD

Date/Time: 07/10/2025 10:42:19

Data Download Date/Time: 07/10/2025 11:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

Agency: PORT ST LUCIE PD

Date/Time: 10/28/2025 14:26:40

Data Download Date/Time: 10/28/2025 16:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

Agency: INDIAN RIVER CSO

Date/Time: 08/01/2025 11:49:48

Data Download Date/Time: 09/08/2025 14:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:45:33

Data Download Date/Time: 02/03/2025 14:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKELAND PD

Date/Time: 04/14/2025 08:10:39

Data Download Date/Time: 04/28/2025 10:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:46:01

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:51:30

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

Agency: LAKELAND PD

Date/Time: 04/28/2025 13:57:13

Data Download Date/Time: 05/27/2025 10:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency: BARTOW PD

Date/Time: 01/16/2025 18:59:51

Data Download Date/Time: 02/09/2025 08:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency: GAINESVILLE PD

Date/Time: 03/25/2025 19:58:16

Data Download Date/Time: 03/25/2025 20:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: GAINESVILLE PD

Date/Time: 03/25/2025 20:07:07

Data Download Date/Time: 03/25/2025 20:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency: GAINESVILLE PD

Date/Time: 08/20/2025 10:24:34

Data Download Date/Time: 08/26/2025 09:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency: LEVY COUNTY SO

Date/Time: 09/08/2025 10:35:57

Data Download Date/Time: 09/08/2025 13:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: LEVY COUNTY SO

Date/Time: 09/08/2025 10:48:31

Data Download Date/Time: 09/08/2025 13:43

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency: ORANGE COUNTY SO

Date/Time: 10/25/2025 16:30:12

Data Download Date/Time: 10/25/2025 16:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 04/30/2025 09:47:49

Data Download Date/Time: 04/30/2025 11:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 07/01/2025 09:26:17

Data Download Date/Time: 07/01/2025 10:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 08/04/2025 09:48:22

Data Download Date/Time: 08/04/2025 11:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: SUMTER COUNTY SO

Date/Time: 11/03/2025 10:59:04

Data Download Date/Time: 11/03/2025 12:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 07:58:54

Data Download Date/Time: 11/29/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 07:59:36

Data Download Date/Time: 11/29/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 08:00:34

Data Download Date/Time: 11/29/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 08:02:32

Data Download Date/Time: 11/29/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 08:05:00

Data Download Date/Time: 11/29/2025 08:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SUMTER COUNTY SO

Date/Time: 11/29/2025 08:37:48

Data Download Date/Time: 01/30/2026 08:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency: SUMTER COUNTY SO

Date/Time: 12/03/2025 08:43:01

Data Download Date/Time: 01/30/2026 08:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency: CAPE CORAL PD

Date/Time: 08/18/2025 17:54:46

Data Download Date/Time: 10/07/2025 16:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CAPE CORAL PD

Date/Time: 08/26/2025 13:46:55

Data Download Date/Time: 10/07/2025 16:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency: PUTNAM COUNTY SO

Date/Time: 08/30/2025 19:29:10

Data Download Date/Time: 08/30/2025 20:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 08/13/2025 10:38:53

Data Download Date/Time: 08/13/2025 12:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST PETERSBURG PD

Date/Time: 10/25/2025 21:57:04

Data Download Date/Time: 01/21/2026 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency: ST PETERSBURG PD

Date/Time: 10/29/2025 07:58:45

Data Download Date/Time: 01/21/2026 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 10:11:29

Data Download Date/Time: 11/21/2025 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 10:12:25

Data Download Date/Time: 11/21/2025 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 10:14:18

Data Download Date/Time: 11/21/2025 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: OSCEOLA COUNTY SO

Date/Time: 11/18/2025 10:41:29

Data Download Date/Time: 11/21/2025 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

Agency: OSCEOLA COUNTY SO

Date/Time: 12/24/2025 09:20:28

Data Download Date/Time: 12/24/2025 09:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency: PALM BAY PD

Date/Time: 04/30/2025 07:55:57

Data Download Date/Time: 04/30/2025 10:54

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PALM BAY PD

Date/Time: 06/26/2025 13:54:29

Data Download Date/Time: 07/10/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency: SOUTH MIAMI PD

Date/Time: 01/13/2025 11:23:40

Data Download Date/Time: 01/13/2025 16:44

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 01/29/2025 16:15:45

Data Download Date/Time: 02/03/2025 23:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 01/29/2025 21:32:36

Data Download Date/Time: 02/03/2025 23:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 04/29/2025 19:36:10

Data Download Date/Time: 05/03/2025 00:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:14:07

Data Download Date/Time: 05/31/2025 03:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 05/27/2025 21:37:31

Data Download Date/Time: 05/31/2025 03:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 06/30/2025 05:44:36

Data Download Date/Time: 07/04/2025 00:02

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 07/17/2025 18:07:29

Data Download Date/Time: 07/17/2025 19:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 09/26/2025 17:48:40

Data Download Date/Time: 09/26/2025 21:53

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency: CLEARWATER PD

Date/Time: 11/22/2025 21:14:53

Data Download Date/Time: 11/22/2025 22:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLEARWATER PD

Date/Time: 12/17/2025 13:12:18

Data Download Date/Time: 12/17/2025 14:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001865

Agency: LAKELAND PD

Date/Time: 02/14/2025 14:53:18

Data Download Date/Time: 02/17/2025 12:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency: PORT ST LUCIE PD

Date/Time: 01/24/2025 19:34:36

Data Download Date/Time: 02/18/2025 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST LUCIE PD

Date/Time: 01/26/2025 08:01:57

Data Download Date/Time: 02/18/2025 11:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency: PORT ST LUCIE PD

Date/Time: 04/04/2025 08:10:25

Data Download Date/Time: 04/27/2025 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: PORT ST LUCIE PD

Date/Time: 11/24/2025 09:41:04

Data Download Date/Time: 11/24/2025 10:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency: PORT ST LUCIE PD

Date/Time: 11/24/2025 10:58:12

Data Download Date/Time: 01/30/2026 08:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002988

Agency: MIAMI GARDENS PD

Date/Time: 10/28/2025 11:31:25

Data Download Date/Time: 01/16/2026 14:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: MIAMI GARDENS PD

Date/Time: 10/28/2025 11:53:17

Data Download Date/Time: 01/16/2026 14:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency: FL HIGHWAY PATROL

Date/Time: 06/11/2025 19:24:46

Data Download Date/Time: 06/11/2025 21:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 07/16/2025 20:11:39

Data Download Date/Time: 09/10/2025 09:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003409

Agency: FL HIGHWAY PATROL

Date/Time: 07/16/2025 20:15:41

Data Download Date/Time: 09/10/2025 09:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

Agency: OSCEOLA COUNTY SO

Date/Time: 05/14/2025 11:35:09

Data Download Date/Time: 05/14/2025 13:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA COUNTY SO

Date/Time: 10/02/2025 20:46:40

Data Download Date/Time: 10/29/2025 13:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

Agency: OSCEOLA COUNTY SO

Date/Time: 05/14/2025 13:59:06

Data Download Date/Time: 06/23/2025 14:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003945

Agency: LAKELAND PD

Date/Time: 01/15/2025 14:31:14

Data Download Date/Time: 01/29/2025 14:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency: OSCEOLA COUNTY SO

Date/Time: 10/01/2025 22:36:29

Data Download Date/Time: 10/29/2025 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: OSCEOLA COUNTY SO

Date/Time: 10/01/2025 22:38:16

Data Download Date/Time: 10/29/2025 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

Agency: OSCEOLA COUNTY SO

Date/Time: 10/01/2025 22:39:25

Data Download Date/Time: 10/29/2025 13:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004770

Agency: FORT WALTON BEACH PD

Date/Time: 08/14/2025 17:22:08

Data Download Date/Time: 08/31/2025 16:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FORT WALTON BEACH PD

Date/Time: 08/31/2025 15:13:49

Data Download Date/Time: 10/01/2025 00:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency:

Date/Time: 06/20/2025 11:59:20

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency:

Date/Time: 06/20/2025 12:08:54

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

Agency:

Date/Time: 06/20/2025 12:09:41

Data Download Date/Time: 07/09/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 01/15/2025 13:11:25

Data Download Date/Time: 01/21/2025 09:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 02/20/2025 09:30:46

Data Download Date/Time: 02/20/2025 11:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 03/26/2025 10:54:54

Data Download Date/Time: 04/23/2025 11:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 04/23/2025 09:21:07

Data Download Date/Time: 04/23/2025 11:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 05/16/2025 07:50:24

Data Download Date/Time: 05/16/2025 10:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 06/21/2025 16:33:14

Data Download Date/Time: 06/21/2025 18:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 15:50:23

Data Download Date/Time: 07/02/2025 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 16:04:47

Data Download Date/Time: 07/02/2025 17:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/02/2025 16:36:32

Data Download Date/Time: 07/02/2025 18:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 07/10/2025 01:51:31

Data Download Date/Time: 07/23/2025 16:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 07/23/2025 14:09:42

Data Download Date/Time: 07/23/2025 16:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 08/25/2025 18:24:40

Data Download Date/Time: 08/26/2025 19:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 09/30/2025 19:55:11

Data Download Date/Time: 10/27/2025 19:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 10/27/2025 19:04:18

Data Download Date/Time: 11/30/2025 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

Agency: CRESTVIEW PD

Date/Time: 11/30/2025 12:28:18

Data Download Date/Time: 11/30/2025 14:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CRESTVIEW PD

Date/Time: 12/25/2025 08:36:35

Data Download Date/Time: 12/25/2025 10:15

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

Agency: PORT ORANGE PD

Date/Time: 11/07/2025 16:49:29

Data Download Date/Time: 11/11/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: PORT ORANGE PD

Date/Time: 11/07/2025 16:50:01

Data Download Date/Time: 11/11/2025 16:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005404

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:22:06

Data Download Date/Time: 04/22/2025 09:47

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005538

Agency: FERNANDINA BEACH PD

Date/Time: 05/29/2025 17:25:27

Data Download Date/Time: 06/25/2025 17:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FERNANDINA BEACH PD

Date/Time: 05/29/2025 17:31:07

Data Download Date/Time: 06/25/2025 17:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006218

Agency: FDLE TRAINER

Date/Time: 06/17/2025 21:14:01

Data Download Date/Time: 12/09/2025 11:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Agency: BREVARD COUNTY SO

Date/Time: 12/12/2025 16:11:51

Data Download Date/Time: 12/14/2025 15:36

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

Agency: HENDRY COUNTY SO

Date/Time: 04/29/2025 11:49:22

Data Download Date/Time: 06/02/2025 13:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:32:02

Data Download Date/Time: 01/27/2025 08:18

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:33:56

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:35:33

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:36:26

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 01/27/2025 07:37:12

Data Download Date/Time: 01/27/2025 08:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 15:27:51

Data Download Date/Time: 03/27/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 03/25/2025 15:28:59

Data Download Date/Time: 03/27/2025 10:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 04/24/2025 09:06:20

Data Download Date/Time: 06/28/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 05/25/2025 10:38:48

Data Download Date/Time: 06/28/2025 12:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 07/29/2025 16:26:36

Data Download Date/Time: 07/29/2025 17:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 08/24/2025 10:35:29

Data Download Date/Time: 08/24/2025 12:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 09/29/2025 00:00:35

Data Download Date/Time: 09/29/2025 00:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 10/22/2025 00:59:59

Data Download Date/Time: 10/23/2025 09:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: CLAY COUNTY SO

Date/Time: 11/30/2025 10:36:52

Data Download Date/Time: 11/30/2025 11:23

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

Agency: CLAY COUNTY SO

Date/Time: 12/30/2025 20:55:01

Data Download Date/Time: 12/30/2025 22:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

Agency: ESCAMBIA COUNTY SO

Date/Time: 08/13/2025 09:54:43

Data Download Date/Time: 08/13/2025 11:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

Agency: MIAMI PD

Date/Time: 01/15/2025 08:27:23

Data Download Date/Time: 01/30/2025 20:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

Agency: MONROE COUNTY SO

Date/Time: 10/07/2025 15:29:07

Data Download Date/Time: 10/08/2025 10:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

Agency: KENNEDY SPACE CENTER

Date/Time: 12/22/2025 02:05:42

Data Download Date/Time: 12/23/2025 08:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 12/22/2025 02:19:51

Data Download Date/Time: 12/23/2025 08:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 01/22/2025 15:29:29

Data Download Date/Time: 01/22/2025 16:04

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 02/20/2025 09:57:30

Data Download Date/Time: 02/20/2025 10:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 03/20/2025 08:50:30

Data Download Date/Time: 03/20/2025 09:26

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 05/09/2025 17:09:18

Data Download Date/Time: 05/09/2025 18:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 06/03/2025 09:54:00

Data Download Date/Time: 06/03/2025 10:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 07/22/2025 08:23:06

Data Download Date/Time: 07/22/2025 09:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 08/25/2025 08:01:30

Data Download Date/Time: 08/25/2025 08:33

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 08/28/2025 14:18:55

Data Download Date/Time: 08/28/2025 15:01

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 10/28/2025 14:43:25

Data Download Date/Time: 10/28/2025 15:20

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: BOCA RATON PD

Date/Time: 11/24/2025 15:09:52

Data Download Date/Time: 11/25/2025 09:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

Agency: BOCA RATON PD

Date/Time: 12/09/2025 14:33:03

Data Download Date/Time: 12/09/2025 15:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

Agency: FHP

Date/Time: 11/19/2025 23:35:29

Data Download Date/Time: 11/19/2025 23:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 02/04/2025 01:39:23

Data Download Date/Time: 02/17/2025 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/04/2025 01:55:20

Data Download Date/Time: 02/17/2025 20:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

Agency: FL HIGHWAY PATROL

Date/Time: 07/06/2025 21:00:11

Data Download Date/Time: 07/30/2025 01:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

Agency: FL HIGHWAY PATROL

Date/Time: 07/26/2025 21:24:39

Data Download Date/Time: 07/30/2025 01:48

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 04/10/2025 17:48:17

Data Download Date/Time: 04/24/2025 16:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 08/18/2025 15:30:58

Data Download Date/Time: 08/27/2025 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 08/25/2025 15:30:45

Data Download Date/Time: 08/27/2025 08:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 12/12/2025 08:50:43

Data Download Date/Time: 12/12/2025 14:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 12/12/2025 08:55:12

Data Download Date/Time: 12/12/2025 14:45

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 12/12/2025 08:55:40

Data Download Date/Time: 12/12/2025 14:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Agency: FL HIGHWAY PATROL

Date/Time: 12/14/2025 13:19:45

Data Download Date/Time: 12/14/2025 19:56

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006637

Agency: FL HIGHWAY PATROL

Date/Time: 06/17/2025 20:31:07

Data Download Date/Time: 07/30/2025 01:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 10:51:26

Data Download Date/Time: 06/08/2025 21:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 10:54:01

Data Download Date/Time: 06/08/2025 21:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:51:25

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:54:44

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 01/13/2025 08:56:09

Data Download Date/Time: 02/07/2025 11:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HOLMES BEACH PD

Date/Time: 03/06/2025 11:55:01

Data Download Date/Time: 04/22/2025 17:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

Agency: HOLMES BEACH PD

Date/Time: 07/15/2025 08:10:34

Data Download Date/Time: 07/15/2025 09:16

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HOLMES BEACH PD

Date/Time: 12/14/2025 16:05:27

Data Download Date/Time: 12/15/2025 15:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

Agency: MANATEE COUNTY SO

Date/Time: 01/31/2025 13:33:41

Data Download Date/Time: 01/31/2025 13:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2025 18:19:27

Data Download Date/Time: 02/07/2025 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 02/01/2025 18:20:53

Data Download Date/Time: 02/07/2025 09:05

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 05/22/2025 11:22:34

Data Download Date/Time: 06/25/2025 08:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 06/14/2025 23:03:41

Data Download Date/Time: 06/25/2025 08:52

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 07/22/2025 06:45:22

Data Download Date/Time: 07/22/2025 07:34

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 09/16/2025 22:24:42

Data Download Date/Time: 09/28/2025 03:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 09/21/2025 06:58:48

Data Download Date/Time: 09/28/2025 03:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: KENNEDY SPACE CENTER

Date/Time: 09/28/2025 02:31:35

Data Download Date/Time: 09/28/2025 03:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006759

Agency: KENNEDY SPACE CENTER

Date/Time: 12/22/2025 02:19:09

Data Download Date/Time: 12/22/2025 03:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 11:16:12

Data Download Date/Time: 06/11/2025 17:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

Agency: TOXL

Date/Time: 09/03/2025 14:36:03

Data Download Date/Time: 10/07/2025 09:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	Fail
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: TOXL

Date/Time: 09/03/2025 14:44:28

Data Download Date/Time: 10/07/2025 09:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	Fail
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006766

Agency: TOXL

Date/Time: 09/03/2025 14:50:31

Data Download Date/Time: 10/07/2025 09:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	Fail
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

Agency: FL HIGHWAY PATROL

Date/Time: 11/09/2025 00:52:50

Data Download Date/Time: 11/12/2025 15:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006771

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:49:04

Data Download Date/Time: 04/15/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:50:58

Data Download Date/Time: 04/15/2025 11:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007008

Agency: BROWARD COUNTY SO

Date/Time: 09/03/2025 10:42:40

Data Download Date/Time: 09/03/2025 11:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007158

Agency: HILLSBOROUGH CO SO

Date/Time: 12/04/2025 14:10:10

Data Download Date/Time: 12/04/2025 15:45

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007159

Agency: PENSACOLA P.D.

Date/Time: 04/30/2025 11:18:47

Data Download Date/Time: 04/30/2025 13:35

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007169

Agency: FFWCC

Date/Time: 11/11/2025 17:31:43

Data Download Date/Time: 11/11/2025 17:39

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007170

Agency: MONROE COUNTY SO

Date/Time: 10/09/2025 14:38:57

Data Download Date/Time: 11/25/2025 12:24

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: MONROE COUNTY SO

Date/Time: 11/21/2025 10:01:35

Data Download Date/Time: 11/25/2025 12:25

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 01/31/2025 13:19:06

Data Download Date/Time: 06/12/2025 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 01/31/2025 13:20:34

Data Download Date/Time: 06/12/2025 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007173

Agency: WALTON COUNTY SO

Date/Time: 12/11/2025 13:02:24

Data Download Date/Time: 12/11/2025 13:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007252

Agency: GROVELAND PD

Date/Time: 11/25/2025 12:21:00

Data Download Date/Time: 11/25/2025 13:08

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 09/28/2025 07:46:56

Data Download Date/Time: 10/16/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: WALTON COUNTY SO

Date/Time: 09/30/2025 20:47:27

Data Download Date/Time: 10/16/2025 10:27

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007364

Agency: WALTON COUNTY SO

Date/Time: 10/17/2025 09:21:19

Data Download Date/Time: 11/04/2025 14:09

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007366

Agency: ST JOHNS SO

Date/Time: 07/31/2025 10:45:18

Data Download Date/Time: 08/06/2025 09:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: ST JOHNS SO

Date/Time: 08/28/2025 20:55:10

Data Download Date/Time: 09/30/2025 22:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007373

Agency: BROWARD COUNTY SO

Date/Time: 10/01/2025 11:12:19

Data Download Date/Time: 10/01/2025 12:07

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007382

Agency: SUWANNEE COUNTY SO

Date/Time: 04/20/2025 15:41:15

Data Download Date/Time: 04/20/2025 15:40

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: SUWANNEE COUNTY SO

Date/Time: 09/19/2025 11:25:01

Data Download Date/Time: 09/29/2025 16:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007440

Agency: FL HIGHWAY PATROL

Date/Time: 09/20/2025 02:18:32

Data Download Date/Time: 10/29/2025 21:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007442

Agency: FHP TRP B LAKE CITY

Date/Time: 08/14/2025 16:01:50

Data Download Date/Time: 08/18/2025 14:55

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007443

Agency: FL HIGHWAY PATROL

Date/Time: 07/07/2025 15:12:10

Data Download Date/Time: 07/11/2025 10:38

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:22:55

Data Download Date/Time: 02/14/2025 13:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 02/12/2025 11:49:36

Data Download Date/Time: 02/14/2025 13:49

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 05/30/2025 11:24:01

Data Download Date/Time: 06/08/2025 21:14

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 07/02/2025 19:51:02

Data Download Date/Time: 08/20/2025 15:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 07/02/2025 19:53:07

Data Download Date/Time: 08/20/2025 15:28

Voltage/Current Test	Fail
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FL HIGHWAY PATROL

Date/Time: 07/02/2025 19:55:39

Data Download Date/Time: 08/20/2025 15:28

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007455

Agency: FL HIGHWAY PATROL

Date/Time: 11/01/2025 01:03:53

Data Download Date/Time: 11/20/2025 19:12

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: FL HIGHWAY PATROL

Date/Time: 11/20/2025 19:26:32

Data Download Date/Time: 12/27/2025 18:21

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007459

Agency: FL HIGHWAY PATROL

Date/Time: 01/20/2025 03:46:54

Data Download Date/Time: 02/03/2025 12:31

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007478

Agency: BAY COUNTY SO

Date/Time: 08/04/2025 10:30:41

Data Download Date/Time: 08/04/2025 13:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

Agency: FFWCC

Date/Time: 07/24/2025 18:47:26

Data Download Date/Time: 08/07/2025 19:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FFWCC

Date/Time: 07/24/2025 18:48:23

Data Download Date/Time: 08/07/2025 19:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007494

Agency: FFWCC

Date/Time: 07/24/2025 19:16:11

Data Download Date/Time: 08/07/2025 19:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: FFWCC

Date/Time: 11/11/2025 17:45:16

Data Download Date/Time: 12/11/2025 18:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007495

Agency: FFWCC

Date/Time: 10/02/2025 10:18:01

Data Download Date/Time: 10/31/2025 11:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007496

Agency: FFWCC

Date/Time: 06/27/2025 21:37:54

Data Download Date/Time: 06/28/2025 16:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007503

Agency: POLK COUNTY SO

Date/Time: 11/22/2025 06:12:02

Data Download Date/Time: 11/22/2025 07:00

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 12/23/2025 08:16:01

Data Download Date/Time: 12/23/2025 09:19

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:32:59

Data Download Date/Time: 03/25/2025 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:33:57

Data Download Date/Time: 03/25/2025 09:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:21:50

Data Download Date/Time: 04/22/2025 09:10

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 07/02/2025 15:35:15

Data Download Date/Time: 07/29/2025 08:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 07/29/2025 08:10:10

Data Download Date/Time: 07/29/2025 08:58

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 09/27/2025 08:05:45

Data Download Date/Time: 09/27/2025 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 09/27/2025 08:06:56

Data Download Date/Time: 09/27/2025 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 11/18/2025 13:31:29

Data Download Date/Time: 11/22/2025 05:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007504

Agency: POLK COUNTY SO

Date/Time: 11/18/2025 13:50:40

Data Download Date/Time: 11/22/2025 05:29

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/19/2025 21:01:55

Data Download Date/Time: 01/19/2025 21:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 02/16/2025 21:25:36

Data Download Date/Time: 02/16/2025 22:11

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 03/20/2025 21:55:01

Data Download Date/Time: 03/20/2025 22:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 05/10/2025 19:54:48

Data Download Date/Time: 05/10/2025 20:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/16/2025 04:09:07

Data Download Date/Time: 10/08/2025 15:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	Fail
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 11/28/2025 02:17:41

Data Download Date/Time: 02/06/2026 09:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007505

Agency: HIGHLANDS COUNTY SO

Date/Time: 12/07/2025 22:15:15

Data Download Date/Time: 02/06/2026 09:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HIGHLANDS COUNTY SO

Date/Time: 12/07/2025 22:15:46

Data Download Date/Time: 02/06/2026 09:37

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 01/24/2025 09:28:29

Data Download Date/Time: 01/24/2025 09:46

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	Fail
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 03/25/2025 08:35:31

Data Download Date/Time: 03/25/2025 09:30

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement

Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 04/22/2025 08:21:39

Data Download Date/Time: 04/22/2025 09:13

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 05/20/2025 08:13:15

Data Download Date/Time: 05/20/2025 09:03

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 07/29/2025 08:08:59

Data Download Date/Time: 07/29/2025 08:51

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: POLK COUNTY SO

Date/Time: 08/26/2025 08:02:59

Data Download Date/Time: 08/26/2025 09:59

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007506

Agency: POLK COUNTY SO

Date/Time: 09/27/2025 07:49:52

Data Download Date/Time: 09/27/2025 09:17

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 01/19/2025 21:03:40

Data Download Date/Time: 01/19/2025 22:22

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/16/2025 04:37:22

Data Download Date/Time: 08/16/2025 05:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Florida Department of Law Enforcement
Alcohol Testing Program

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/16/2025 04:44:25

Data Download Date/Time: 08/16/2025 05:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	Fail

Agency: HIGHLANDS COUNTY SO

Date/Time: 08/16/2025 05:05:38

Data Download Date/Time: 08/16/2025 05:50

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007507

Agency: HIGHLANDS COUNTY SO

Date/Time: 09/13/2025 20:56:52

Data Download Date/Time: 09/13/2025 21:57

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007510

Agency:

Date/Time: 07/07/2025 15:03:02

Data Download Date/Time: 02/04/2026 16:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

Agency:

Date/Time: 07/09/2025 08:54:52

Data Download Date/Time: 02/04/2026 16:06

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007516

Agency: MIAMI PD

Date/Time: 09/10/2025 06:16:35

Data Download Date/Time: 09/12/2025 11:42

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK

**Florida Department of Law Enforcement
Alcohol Testing Program**

Diagnostic Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-007517

Agency: MIAMI PD

Date/Time: 01/31/2025 17:14:55

Data Download Date/Time: 02/22/2025 00:32

Voltage/Current Test	OK
RAM Test	OK
EEPROM Checksum Test	OK
Real Time Clock Test	OK
DSP Test	OK
Analytical Stability Test	OK
Internal Printer Test	OK
Modem Test	OK
Temperature Regulation Test	OK
