

Alcohol Testing Program - Instrument Processing Sheet

Agency: Miami Dade Sheriff's Office Instrument Serial Number: 80-000883

Date In: 3/18/2026 DI Completion Date: N/A ☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>3/19/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>3/19/2026</u>		Flow Adjustment By: _____																																							
☒ Annual ☐ Dropped Off ☒ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Update registration from Miami Dade PD to Miami Dade SO.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☐ R-Value: _____ ☐ Flow Verification (L/s) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547) ☐ Barometric Pressure Check Gauge ID #: _____ Gauge: _____ Instrument: _____ ☐ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																							
						<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </table>		Simulator	Serial #	Lot#/Exp	0.050			0.080			0.200			0.080 DGS	N/A		Maintenance By: <u>TDG</u> Date: <u>3/19/2026</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☒ Other: CMI was unable to remotely resolve the DSP Fail.																				
						Simulator	Serial #	Lot#/Exp																																			
						0.050																																					
						0.080																																					
0.200																																											
0.080 DGS	N/A																																										
Optical Bench Adjustment By: _____ Department Inspection By: _____ Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: _____																																											
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 20%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: _____ Instrument: _____ Mouth Alcohol Solution Lot #: _____ Exp: _____ Acetone Stock Solution Lot #: _____ Exp: _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </table>		Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																								
0.000		N/A	N/A																																								
0.040																																											
0.100																																											
0.200																																											
0.300																																											
0.080 DGS	N/A																																										
Simulator	Serial Number																																										
0.000																																											
Interferent																																											
0.050																																											
0.080																																											
0.200																																											
☐ Post Optical Bench Adjustment Stability Checks		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 20%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Gauge ID #: _____ Gauge: _____ Instrument: _____																			
Simulator	Serial #	Lot #	Expiration																																								
0.050																																											
0.080																																											
0.200																																											
0.080 DGS	N/A																																										
Notes/Suggested Service: Instrument gave DSP Fail during warmup diagnostics. Message persisted even after restarting. CMI was unable to resolve. Sending to repair. (TDG 3/19/26)		Attachments ☐ Form 41 ☐ Stability Checks ☐ Calibration Certificate ☐ Optical Bench Adjustment		☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☒ Other: Return Material Authorization																																							
		☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use		Digitally signed by LeAndra Higginbotham Date: 2026.03.30 11:04:02 -04'00'																																							
		Digitally signed by Shayla Platt Date: 2026.04.28 10:29:29 -04'00'																																									
		Tech Review		Admin Review																																							

Return Material Authorization

Ship to:☒ CMI, Inc.☐ Enforcement ElectronicsShipment to repair facility authorized by: Richard Closius on 3/20/2026Items Returned: ☒ Instrument ☐ Supplies ☐ Other Describe: _____Instrument Model: Intoxilyzer 8000 Serial Number: 80-000883

Bill To Address:

Miami Dade Sheriff's Office
Attn: Richard Closius

Ship to Address:

Florida Department of Law Enforcement
Fort Myers Regional Operations Center
Attn: Taylor Gutschow
4700 Terminal Drive, Suite 1
Fort Myers, FL 33907

Reason for Return:

Instrument displays DSP Fail and will not go into Ready Mode. CMI could not resolve remotely.

I require an estimate **BEFORE** any repairs will be authorized and/or conductedPlease contact: Richard Closius Phone #: 1-305-458-3991Email: rlclosius@mdso.comATP Contact Name: Taylor GutschowATP Email: TaylorGutschow@fdle.state.fl.us