

Alcohol Testing Program - Instrument Processing Sheet

Agency: Monroe County Sheriff's Office Instrument Serial Number: 80-006471
 Date In: 10/14/2025 DI Completion Date: 1/23/2026 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>1/14/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>1/22/2026</u>		Flow Adjustment By: _____																																	
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>230</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.507</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1018</u> Instrument: <u>1018</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																	
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:																		
Simulator	Serial #	Lot#/Exp																																			
0.050	MP6286	202406K 6/19/2026																																			
0.080	MP6287	202406L 6/19/2026																																			
0.200	MP6288	202406N 6/20/2026																																			
0.080 DGS	N/A	AG526603 9/23/2027																																			
Optical Bench Adjustment By: _____		Department Inspection By: <u>TDG</u>																																			
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>33364</u>																																			
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1021</u> Instrument: <u>1021</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>							
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Simulator	Serial Number																																				
0.000	MP6284																																				
Interferent	MP6285																																				
0.050	MP6286																																				
0.080	MP6287																																				
0.200	MP6288																																				
		Attachments																																			
		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment		<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other:																																	
Gauge ID #: _____ Gauge: _____ Instrument: _____																																					
Notes/Suggested Service:		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																			
		Digitally signed by: Shayla Platt Date: 2026.01.30 15:36:21 -05'00'		Digitally signed by: Wen-Chi Pierson Date: 2026.02.02 08:45:12 -05'00'																																	
		Tech Review		Admin Review																																	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 11:05

Date of Inspection: 01/22/2026

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/22/2026
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 SN 80-006471 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>----- Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	14:21	Control Test	0.049	14:22	Air Blank	0.000	14:22	Control Test	0.049	14:23	Air Blank	0.000	14:23	Control Test	0.049	14:24	Air Blank	0.000	14:25	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 SN 80-006471 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>----- Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	14:15	Control Test	0.080	14:16	Air Blank	0.000	14:16	Control Test	0.079	14:17	Air Blank	0.000	14:17	Control Test	0.079	14:18	Air Blank	0.000	14:19	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 SN 80-006471 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> <p>----- Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	14:08	Control Test	0.197	14:09	Air Blank	0.000	14:09	Control Test	0.196	14:10	Air Blank	0.000	14:11	Control Test	0.196	14:11	Air Blank	0.000	14:12	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		<p>MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model: 8000 SN 80-006471 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>----- Operator's Signature <i>MW</i></p>	Test	g/210L	Time	Air Blank	0.000	13:51	Control Test	0.077	13:52	Air Blank	0.000	13:52	Control Test	0.078	13:53	Air Blank	0.000	13:53	Control Test	0.078	13:53	Air Blank	0.000	13:54	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.049	14:22																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.049	14:23																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test	0.049	14:24																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.080	14:16																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test	0.079	14:17																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.079	14:18																																																																																																																																																	
Air Blank	0.000	14:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test	0.197	14:09																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.196	14:10																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.196	14:11																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test	0.077	13:52																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.078	13:53																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test	0.078	13:53																																																																																																																																																	
Air Blank	0.000	13:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 13:02

Date of Inspection: 01/23/2026

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.050	0.078	0.197	0.077
0.000	0.050	0.078	0.196	0.077
0.000	0.050	0.078	0.197	0.077
0.000	0.050	0.079	0.197	0.077
0.000	0.050	0.079	0.197	0.077
0.000	0.050	0.078	0.197	0.077
0.000	0.051	0.079	0.197	0.077


Standard Deviations	0.0006	0.0005	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/23/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006471</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MONROE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/23/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:02</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Date: 2026.01.25 20:04:37
-05'00'

01/23/2026
Date
TAYLOR D GUTSCHOW,

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality