

Alcohol Testing Program - Instrument Processing Sheet

Agency: Marion CO SO

Instrument Serial Number: 80-007172

Date In: 7/17/2026

DI Completion Date: 7/23/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>IAS</u> Date: <u>7/17/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>7/22/2026</u>		Flow Adjustment By: <u>WKP</u>																
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>117</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.117</u> (.139-.169) 36 mm <u>0.132</u> (.156-.190) 53 mm <u>0.218</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1012</u> Instrument: <u>1010</u> ☒ Stability Checks		Flow Column #: <u>ATP 102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value: <u>118</u> ☒ Post Adjustment Verification (L/S) Flow Column #: <u>ATP 105</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.523</u> (.447-.547)																
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP6291	202603A 3/3/2028	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	
Simulator	Serial #	Lot#/Exp																		
0.050	MP6291	202603A 3/3/2028																		
0.080	MP6292	202603B 3/3/2028																		
0.200	MP6293	202603D 3/4/2028																		
0.080 DGS	N/A	23825080A1 10/5/2027																		
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																		
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																		
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1014</u> Instrument: <u>1014</u>																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																
0.100				Simulator	Serial Number															
0.200				0.000	MP6289															
0.300				Interferent	MP6290															
0.080 DGS	N/A			0.050	MP6291															
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6292															
Simulator	Serial #	Lot #	Expiration	0.200	MP6293															
0.050				Attachments																
0.080				☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Adjustment ☒ Calibration Certificate ☐ Form 40 ☐ Optical Bench Adjustment ☐ Other:																
0.200																				
0.080 DGS	N/A																			
Gauge ID #: _____																				
Gauge: _____ Instrument: _____																				
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																
				Digitally signed by LeAndra Higginbotham Date: 2026.07.27 10:47:56 -04'00'																
				Digitally signed by Shayla Platt Date: 2026.07.29 12:57:51 -04'00'																
				Tech Review Admin Review																

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007172 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	12:54	Control Test	0.049	12:55	Air Blank	0.000	12:56	Control Test	0.049	12:56	Air Blank	0.000	12:57	Control Test	0.049	12:57	Air Blank	0.000	12:58	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007172 07/22/2026 Software: 8110.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:04	Control Test	0.080	13:05	Air Blank	0.000	13:05	Control Test	0.079	13:06	Air Blank	0.000	13:07	Control Test	0.079	13:07	Air Blank	0.000	13:08	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 3000 SN 80-007172 17/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:09	Control Test	0.197	13:10	Air Blank	0.000	13:10	Control Test	0.196	13:11	Air Blank	0.000	13:12	Control Test	0.196	13:12	Air Blank	0.000	13:13	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 3000 SN 80-007172 07/22/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:50	Control Test	0.079	13:51	Air Blank	0.000	13:51	Control Test	0.078	13:51	Air Blank	0.000	13:52	Control Test	0.078	13:52	Air Blank	0.000	13:53	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.049	12:55																																																																																																																																																	
Air Blank	0.000	12:56																																																																																																																																																	
Control Test	0.049	12:56																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.049	12:57																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.080	13:05																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.079	13:06																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test	0.079	13:07																																																																																																																																																	
Air Blank	0.000	13:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.197	13:10																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.196	13:11																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.196	13:12																																																																																																																																																	
Air Blank	0.000	13:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1963																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test	0.079	13:51																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test	0.078	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.078	13:52																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Flow Calibration Adjustment

WKP
7/22/2026

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007172
07/22/2026
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 5.656

2: Rate (Liters/min) = 15

SQRT(Diff) = 10.629

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.047

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 671

Rounded Intercept = -414385

Correlation = 0.99811

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 12:33

Date of Inspection: 07/23/2026

Serial Number: 80-007172
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.077
0.000	0.050	0.080	0.199	0.078
Standard Deviations	0.0000	0.0003	0.0000	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi Pierson

WEN-CHI K PIERSON

Signature and Printed Name

07/23/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007172, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007172</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/23/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:33</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/23/2026

Date

Wen-Chi
Pierson

Digitally signed by Wen-Chi Pierson
Date: 2026.07.23
15:01:09 -04'00'

WEN-CHI K PIERSON,
Department Inspector