

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,SKY,O
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WICKHAM,DAVID
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: WALKLING

DATE/TIME OF TEST: 04/02/2026 03:38:57

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.165	03:34
Breath Volume	4.511	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.169	03:37
Breath Volume	5.007	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/29/2026 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BABA,TANNER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: TATO DIAZ

DATE/TIME OF TEST: 04/12/2026 01:52:15

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.080	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.221	01:46
Breath Volume	1.378	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.219	01:50
Breath Volume	1.437	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 04/29/2026 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATTISTA,BRITTNEY,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: KING,BRANDON
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403342588
DRY GAS EXPIRY DATE: 05/09/2028
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: TRAYLOR

DATE/TIME OF TEST: 04/22/2026 02:34:23

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.115	02:28
Breath Volume	2.171	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.117	02:32
Breath Volume	1.421	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 04/29/2026 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 04/02/2026 03:01:59

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.082	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.085	02:57
Breath Volume	1.210	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.076	03:00
Breath Volume	1.320	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: 2C11,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/03/2026 05:07:01

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.104	05:02
Breath Volume	1.894	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.096	05:05
Breath Volume	1.214	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:07

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HACKER

DATE/TIME OF TEST: 04/04/2026 22:38:28

Diagnostics Check	OK	22:37	
Air Blank	0.000	22:38	
Control Test	0.000	22:38	RFI Detect
Air Blank		22:38	RFI Detect

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: HACKER

DATE/TIME OF TEST: 04/04/2026 22:56:25

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:41	
Control Test	0.080	22:41	
Air Blank	0.000	22:42	
Subject Sample	# 1 0.126	22:45	Volume Not Met
Breath Volume	0.238	22:45	
Air Blank	0.000	22:45	
Air Blank	0.000	22:47	
Subject Sample	# 2 0.151	22:50	
Breath Volume	1.597	22:50	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 3 0.139	22:54	
Breath Volume	1.250	22:54	
Air Blank	0.000	22:55	
Control Test	0.078	22:55	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WALLER

DATE/TIME OF TEST: 04/08/2026 23:49:32

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.082	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.111	23:45
Breath Volume	2.246	23:45
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.102	23:48
Breath Volume	2.156	23:48
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLAHERTY,Z
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403374313
DRY GAS EXPIRY DATE: 06/25/2028
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 04/13/2026 19:34:17

Diagnostics Check	OK	19:28
Air Blank	0.000	19:28
Control Test	0.081	19:28
Air Blank	0.000	19:29
Subject Sample # 1	0.171	19:29
Breath Volume	2.769	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:32
Subject Sample # 2	0.168	19:32
Breath Volume	4.007	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Diagnostics Check	OK	19:34

Data Download Date/Time: 04/29/2026 20:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SAMHOURI

DATE/TIME OF TEST: 04/04/2026 22:59:24

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.030	22:55
Breath Volume	1.335	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.028	22:57
Breath Volume	2.191	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Diagnostics Check	OK	22:59

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: QUICKENTON

DATE/TIME OF TEST: 04/11/2026 23:00:27

Diagnostics Check	OK	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:54
Air Blank	0.000	22:55
Subject Sample # 1	0.178	22:55
Breath Volume	2.523	22:55
Air Blank	0.000	22:56
Air Blank	0.000	22:58
Subject Sample # 2	0.180	22:58
Breath Volume	2.453	22:58
Air Blank	0.000	22:59
Control Test	0.080	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/16/2026 00:12:29

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.034	00:08
Breath Volume	2.382	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.031	00:11
Breath Volume	1.742	00:11
Air Blank	0.000	00:11
Control Test	0.079	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SISOFO

DATE/TIME OF TEST: 04/18/2026 00:13:42

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.045	00:09
Breath Volume	2.367	00:09
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.043	00:12
Breath Volume	2.964	00:12
Air Blank	0.000	00:12
Control Test	0.079	00:13
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RAYMUNDO

DATE/TIME OF TEST: 04/18/2026 02:56:43

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.234	02:52
Breath Volume	2.074	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.229	02:55
Breath Volume	1.738	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMDEN,DENNIS,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CAMDEN,DENNIS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: FAUERBACH

DATE/TIME OF TEST: 04/18/2026 20:01:21

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.182	19:55
Breath Volume	1.539	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.182	19:59
Breath Volume	1.882	19:59
Air Blank	0.000	20:00
Control Test	0.078	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: VEREEN,STEVEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: VALENZUELA

DATE/TIME OF TEST: 04/23/2026 19:10:45

Diagnostics Check	OK	19:04
Air Blank	0.000	19:04
Control Test	0.080	19:05
Air Blank	0.000	19:05
Subject Sample # 1	0.250	19:06
Breath Volume	1.863	19:06
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.255	19:09
Breath Volume	3.765	19:09
Air Blank	0.000	19:09
Control Test	0.080	19:10
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: LOESCHEN

DATE/TIME OF TEST: 04/25/2026 01:44:29

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.124	01:39
Breath Volume	2.992	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.127	01:42
Breath Volume	3.679	01:42
Air Blank	0.000	01:43
Control Test	0.080	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/29/2026 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,OSCAR,E
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BARNES,SAMUEL
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 04/02/2026 20:15:09

Diagnostics Check	OK	20:07
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.159	20:10
Breath Volume	1.179	20:10
Air Blank	0.000	20:11
Air Blank	0.000	20:12
Subject Sample # 2	0.160	20:13
Breath Volume	1.367	20:13
Air Blank	0.000	20:14
Control Test	0.079	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 04/27/2026 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,KEELAN,D
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WILLIAMS,M
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 04/03/2026 20:51:04

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.079	20:45
Air Blank	0.000	20:46
Subject Sample # 1	0.000	20:46
Breath Volume	2.492	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.000	20:49
Breath Volume	2.179	20:49
Air Blank	0.000	20:50
Control Test	0.079	20:50
Air Blank	0.000	20:50
Diagnostics Check	OK	20:51

Data Download Date/Time: 04/27/2026 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,AARON,J
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: LEVINE,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 04/19/2026 01:10:12

Diagnostics Check	OK	01:02
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.196	01:05
Breath Volume	2.070	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.192	01:08
Breath Volume	1.730	01:08
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Diagnostics Check	OK	01:10

Data Download Date/Time: 04/27/2026 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISBEE PAULL,SHEALYNN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DEPORTER,A
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403207508
DRY GAS EXPIRY DATE: 12/02/2027
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 04/24/2026 03:38:33

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.182	03:33
Breath Volume	4.101	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.179	03:37
Breath Volume	1.699	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 04/27/2026 18:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSHWORTH,JOSEPH,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: MSWIA,KANYIKA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 04/04/2026 02:59:53

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.076	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.112	02:55
Breath Volume	1.574	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.106	02:58
Breath Volume	1.894	02:58
Air Blank	0.000	02:58
Control Test	0.076	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ALFONSO,RAFAEL
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 04/08/2026 04:41:09

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.076	04:28	
Air Blank	0.000	04:28	
Subject Sample	# 1 0.157	04:31	Volume Not Met
Breath Volume	0.671	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2 0.173	04:35	
Breath Volume	1.746	04:35	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 3 0.170	04:40	
Breath Volume	1.457	04:40	
Air Blank		04:40	RFI Detect
Air Blank	0.000	04:41	

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ALFONSO,RAFAEL
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: CAIN

DATE/TIME OF TEST: 04/08/2026 05:01:07

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.075	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.144	04:56
Breath Volume	1.484	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.160	04:59
Breath Volume	1.519	04:59
Air Blank	0.000	05:00
Control Test	0.075	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSHWORTH,JOSEPH,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LOPEZ,REGINA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: DENBY

DATE/TIME OF TEST: 04/24/2026 00:43:16

Diagnostics Check	OK	00:36
Air Blank	0.000	00:37
Control Test	0.075	00:37
Air Blank	0.000	00:37
Subject Sample # 1	0.139	00:38
Breath Volume	2.003	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.148	00:41
Breath Volume	2.628	00:41
Air Blank	0.000	00:42
Control Test	0.075	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WHIGHAM,ZACH
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RAINEY

DATE/TIME OF TEST: 04/29/2026 22:01:12

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.076	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1 0.097	21:57	
Breath Volume	2.636	21:57	
Air Blank	0.000	21:57	
Air Blank	0.000	21:59	
Subject Sample	# 2	21:59	RFI Detect
Breath Volume	0.000	21:59	
Air Blank	0.000	22:00	
Control Test	0.075	22:00	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIAS,MARK,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WHIGHAM,ZACH
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 03/31/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: RAINEY

DATE/TIME OF TEST: 04/29/2026 22:09:50

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.076	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.096	22:05
Breath Volume	2.648	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:07
Subject Sample # 2	0.092	22:08
Breath Volume	2.722	22:08
Air Blank	0.000	22:08
Control Test	0.075	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 04/30/2026 17:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHERR, EVAN, M
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: TORRES, DEVIN
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 03/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: HUGO

DATE/TIME OF TEST: 04/03/2026 21:49:47

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.080	21:49	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 04/17/2026 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LULU,ABD,U
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LULU,ABD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: CASH

DATE/TIME OF TEST: 04/08/2026 00:58:26

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.079	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.079	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 04/17/2026 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,KEATON,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: RAY,KEATON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 03/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: COOPMAN

DATE/TIME OF TEST: 04/09/2026 00:28:15

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:25	
Control Test	0.080	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:26	Subject Test Refused
Breath Volume	0.000	00:26	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 04/17/2026 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREVATT,JENNIFER,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: VANANTWERP

DATE/TIME OF TEST: 04/11/2026 03:33:58

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 04/17/2026 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKINSON,RYAN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WILKINSON,RYAN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/16/2026 22:23:28

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.080	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.170	22:19
Breath Volume	2.554	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.169	22:21
Breath Volume	1.187	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 04/17/2026 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIRSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MEDRANO PALMER

DATE/TIME OF TEST: 04/04/2026 05:03:47

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:52	
Control Test	0.081	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1 0.183	04:56	Volume Not Met
Breath Volume	0.082	04:56	
Air Blank	0.000	04:57	
Air Blank	0.000	04:59	
Subject Sample	# 2 0.168	05:02	Volume Not Met
Breath Volume	0.355	05:02	
Air Blank	0.000	05:02	
Control Test	0.079	05:03	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WRIGHT,KIRSTEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: MEDRANO PALMER

DATE/TIME OF TEST: 04/04/2026 05:12:41

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.080	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.179	05:08
Breath Volume	2.050	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.176	05:11
Breath Volume	2.171	05:11
Air Blank	0.000	05:11
Control Test	0.080	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:04

SUBJECT'S LAST NAME: EDOUARD

DATE/TIME OF TEST: 04/04/2026 08:36:26

Diagnostics Check	OK	08:30
Air Blank	0.000	08:30
Control Test	0.081	08:31
Air Blank	0.000	08:31
Subject Sample # 1	0.127	08:32
Breath Volume	1.160	08:32
Air Blank	0.000	08:32
Air Blank	0.000	08:34
Subject Sample # 2	0.136	08:34
Breath Volume	2.488	08:34
Air Blank	0.000	08:35
Control Test	0.080	08:35
Air Blank	0.000	08:36
Diagnostics Check	OK	08:36

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KENNY,MICHAEL
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 04/06/2026 23:23:17

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.081	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.265	23:18
Breath Volume	3.597	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.266	23:21
Breath Volume	3.535	23:21
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Diagnostics Check	OK	23:23

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 04/07/2026 01:09:03

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:03
Subject Sample # 1	0.103	01:04
Breath Volume	1.554	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.108	01:07
Breath Volume	1.683	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNINGSEN,KENNETH
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 04/07/2026 01:42:23

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:41	Sequence Aborted
Breath Volume	0.027	01:41	
Air Blank	0.000	01:42	

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNINGSSEN,KENNETH
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 04/07/2026 02:11:45

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.186	02:07
Breath Volume	1.792	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.183	02:10
Breath Volume	1.742	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 05/15/2026 14:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: COLLOVA

DATE/TIME OF TEST: 04/07/2026 02:33:14

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.080	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.175	02:28
Breath Volume	1.429	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.182	02:31
Breath Volume	1.777	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: EMILE,EDOUARD
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: PLATT

DATE/TIME OF TEST: 04/13/2026 00:21:40

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.081	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.168	00:17
Breath Volume	1.335	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.165	00:20
Breath Volume	2.285	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: WERNER

DATE/TIME OF TEST: 04/13/2026 05:07:16

Diagnostics Check	OK	05:00
Air Blank	0.000	05:01
Control Test	0.080	05:01
Air Blank	0.000	05:02
Subject Sample # 1	0.194	05:02
Breath Volume	2.414	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.185	05:05
Breath Volume	1.406	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: STEWART JR.

DATE/TIME OF TEST: 04/20/2026 02:25:47

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:21
Breath Volume	2.910	02:21
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	1.343	02:24
Air Blank	0.000	02:24
Control Test	0.082	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PEREZ,CHELSEA
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 04/21/2026 22:07:43

Diagnostics Check	OK	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Subject Sample # 1	0.301	22:03
Breath Volume	2.117	22:03
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.314	22:06
Breath Volume	2.343	22:06
Air Blank	0.000	22:06
Control Test	0.081	22:07
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GLEZIL

DATE/TIME OF TEST: 04/22/2026 03:14:45

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.199	03:10
Breath Volume	1.785	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.197	03:13
Breath Volume	1.824	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: DIEGUEZ RODRIGUEZ

DATE/TIME OF TEST: 04/22/2026 05:38:21

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.156	05:33
Breath Volume	1.472	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.142	05:36
Breath Volume	1.648	05:36
Air Blank	0.000	05:37
Control Test	0.080	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GRAZIOSE,MADISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: PACELLA

DATE/TIME OF TEST: 04/23/2026 00:27:36

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:17	
Control Test	0.080	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1 0.188	00:21	Volume Not Met
Breath Volume	1.015	00:21	
Air Blank	0.000	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 2	00:26	Subject Test Refused
Breath Volume	0.386	00:26	
Air Blank	0.000	00:26	
Control Test	0.080	00:27	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,CARLOS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/23/2026 01:41:16

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.185	01:36
Breath Volume	3.367	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.187	01:39
Breath Volume	2.894	01:39
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SAKKA

DATE/TIME OF TEST: 04/23/2026 02:09:57

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.133	02:04
Breath Volume	1.664	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.145	02:08
Breath Volume	2.324	02:08
Air Blank	0.000	02:09
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DUQUE,BRYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ROMEU

DATE/TIME OF TEST: 04/24/2026 02:53:22

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.153	02:48
Breath Volume	1.722	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.150	02:51
Breath Volume	2.121	02:51
Air Blank	0.000	02:52
Control Test	0.081	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: AMICO

DATE/TIME OF TEST: 04/24/2026 22:22:30

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.081	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.000	22:18
Breath Volume	2.472	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.000	22:21
Breath Volume	3.058	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PARKER,BRYCE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 04/25/2026 00:40:30

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:35
Air Blank	0.000	00:35
Subject Sample # 1	0.280	00:36
Breath Volume	1.785	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.280	00:39
Breath Volume	1.722	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 04/25/2026 04:28:20

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.114	04:23
Breath Volume	2.898	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.113	04:26
Breath Volume	2.812	04:26
Air Blank	0.000	04:27
Control Test	0.081	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403345775
DRY GAS EXPIRY DATE: 05/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: LESMES QUINTERO

DATE/TIME OF TEST: 04/26/2026 23:14:07

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.199	23:08
Breath Volume	1.812	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.198	23:12
Breath Volume	2.675	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/15/2026 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MARTINEAU,ARIC
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 09:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:06

SUBJECT'S LAST NAME: BOGACZ

DATE/TIME OF TEST: 04/01/2026 10:32:54

Diagnostics Check	OK	10:26
Air Blank	0.000	10:27
Control Test	0.081	10:27
Air Blank	0.000	10:28
Subject Sample # 1	0.000	10:28
Breath Volume	1.621	10:28
Air Blank	0.000	10:29
Air Blank	0.000	10:31
Subject Sample # 2	0.000	10:31
Breath Volume	1.496	10:31
Air Blank	0.000	10:31
Control Test	0.081	10:32
Air Blank	0.000	10:32
Diagnostics Check	OK	10:32

Data Download Date/Time: 04/27/2026 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: FEARS

DATE/TIME OF TEST: 04/03/2026 20:48:09

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.082	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:47	
Control Test	0.081	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 04/27/2026 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ADEN

DATE/TIME OF TEST: 04/04/2026 17:56:22

Diagnostics Check	OK	17:50
Air Blank	0.000	17:50
Control Test	0.081	17:51
Air Blank	0.000	17:51
Subject Sample # 1	0.062	17:51
Breath Volume	2.414	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.061	17:54
Breath Volume	2.531	17:54
Air Blank	0.000	17:55
Control Test	0.080	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 04/27/2026 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MIELE,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: KANDAH

DATE/TIME OF TEST: 04/05/2026 04:03:43

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Subject Sample # 1	0.256	03:58
Breath Volume	1.707	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.259	04:02
Breath Volume	2.164	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/27/2026 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: COVINGTON

DATE/TIME OF TEST: 04/05/2026 21:07:26

Diagnostics Check	OK	20:58
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.214	20:59
Breath Volume	3.937	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.181	21:02
Breath Volume	1.511	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 3	0.208	21:05
Breath Volume	2.914	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:07
Diagnostics Check	OK	21:07

Data Download Date/Time: 04/27/2026 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,AUSTIN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:57

SUBJECT'S LAST NAME: RUSS

DATE/TIME OF TEST: 04/08/2026 14:26:42

Diagnostics Check	OK	14:21	
Air Blank	0.000	14:22	
Control Test	0.081	14:22	
Air Blank	0.000	14:23	
Subject Sample	# 1	14:25	Subject Test Refused
Breath Volume	0.000	14:25	
Air Blank	0.000	14:25	
Control Test	0.080	14:26	
Air Blank	0.000	14:26	
Diagnostics Check	OK	14:26	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 04/09/2026 05:01:37

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:00	
Control Test	0.079	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: BLEVIN

DATE/TIME OF TEST: 04/10/2026 06:09:37

Diagnostics Check	OK	06:05	
Air Blank	0.000	06:05	
Control Test	0.081	06:06	
Air Blank	0.000	06:06	
Subject Sample	# 1	06:08	Subject Test Refused
Breath Volume	0.000	06:08	
Air Blank	0.000	06:08	
Control Test	0.081	06:09	
Air Blank	0.000	06:09	
Diagnostics Check	OK	06:09	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PONS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/11/2026 01:42:58

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:30	
Control Test	0.081	01:30	
Air Blank	0.000	01:31	
Subject Sample	# 1	01:34	No Sample Provided
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2	0.000	01:38
Breath Volume	1.621	01:38	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample	# 3	0.000	01:41
Breath Volume	1.574	01:41	
Air Blank	0.000	01:42	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BOWERS,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: MUTSCHLER

DATE/TIME OF TEST: 04/11/2026 07:34:12

Diagnostics Check	OK	07:30	
Air Blank	0.000	07:30	
Control Test	0.080	07:31	
Air Blank	0.000	07:31	
Subject Sample	# 1	07:32	Subject Test Refused
Breath Volume	0.000	07:32	
Air Blank	0.000	07:33	
Control Test	0.079	07:33	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:34	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKINNON,CHRISTIAN,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCKINNON,CHRISTIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: LOWRY

DATE/TIME OF TEST: 04/11/2026 12:13:04

Diagnostics Check	OK	12:07
Air Blank	0.000	12:07
Control Test	0.081	12:07
Air Blank	0.000	12:08
Subject Sample # 1	0.000	12:08
Breath Volume	2.257	12:08
Air Blank	0.000	12:09
Air Blank	0.000	12:11
Subject Sample # 2	0.000	12:11
Breath Volume	2.351	12:11
Air Blank	0.000	12:12
Control Test	0.081	12:12
Air Blank	0.000	12:12
Diagnostics Check	OK	12:13

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,KRISTEN,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 04/12/2026 02:44:43

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:37	No Sample Provided
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:43	No Sample Provided
Breath Volume	0.000	02:43	
Air Blank	0.000	02:43	
Control Test	0.081	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:44	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWERS,OFFICER,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LINGG,OFFICER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: DUKE

DATE/TIME OF TEST: 04/12/2026 06:54:46

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:50	
Control Test	0.081	06:51	
Air Blank	0.000	06:51	
Subject Sample	# 1	06:53	Subject Test Refused
Breath Volume	0.000	06:53	
Air Blank	0.000	06:53	
Control Test	0.080	06:54	
Air Blank	0.000	06:54	
Diagnostics Check	OK	06:54	

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,AUSTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:06

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 04/13/2026 08:34:45

Diagnostics Check	OK	08:28
Air Blank	0.000	08:29
Control Test	0.080	08:29
Air Blank	0.000	08:30
Subject Sample # 1	0.000	08:30
Breath Volume	3.726	08:30
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.000	08:33
Breath Volume	2.945	08:33
Air Blank	0.000	08:33
Control Test	0.080	08:34
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOUGH,SEAN,J
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MCCULLOUGH,SEAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 03/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:58

SUBJECT'S LAST NAME: BAXTER

DATE/TIME OF TEST: 04/13/2026 15:44:34

Diagnostics Check	OK	15:38
Air Blank	0.000	15:38
Control Test	0.080	15:39
Air Blank	0.000	15:39
Subject Sample # 1	0.221	15:40
Breath Volume	3.671	15:40
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.215	15:42
Breath Volume	3.601	15:42
Air Blank	0.000	15:43
Control Test	0.079	15:43
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 04/27/2026 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISBEE PAULL,SHEALYNN,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HICKS,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: JAEGER

DATE/TIME OF TEST: 04/05/2026 00:14:27

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.113	00:09
Breath Volume	1.714	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.109	00:12
Breath Volume	1.566	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/29/2026 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITKE,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: MARTIN,CODY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: VASQUEZPASAYE

DATE/TIME OF TEST: 04/20/2026 00:55:42

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.079	00:50
Air Blank	0.000	00:51
Subject Sample # 1	0.000	00:51
Breath Volume	2.328	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.000	00:54
Breath Volume	2.535	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 04/29/2026 18:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITKE,WILLIAM,P
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: MARTIN,CODY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/25/2026 03:11:37

Diagnostics Check	OK	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.132	03:07
Breath Volume	3.484	03:07
Air Blank	0.000	03:07
Air Blank	0.000	03:09
Subject Sample # 2	0.134	03:09
Breath Volume	2.671	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/29/2026 18:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,RICHARD,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: STACY,NIKOLAS
ARRESTING OFFICER AGENCY: Holly Hills PD
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: DAVIDSON-BRATCHER

DATE/TIME OF TEST: 04/02/2026 23:14:10

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.081	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.164	23:09
Breath Volume	1.910	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.164	23:12
Breath Volume	1.460	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/18/2026 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,RICHARD,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SKINNER,RYAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: RAYBURN

DATE/TIME OF TEST: 04/03/2026 04:48:16

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.139	04:43
Breath Volume	2.085	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.146	04:46
Breath Volume	2.167	04:46
Air Blank	0.000	04:47
Control Test	0.081	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 05/18/2026 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,RICHARD,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SALZL,GUNTHER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 04/10/2026 20:25:50

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.081	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:20
Breath Volume	4.132	20:20
Air Blank	0.000	20:21
Air Blank	0.000	20:23
Subject Sample # 2	0.000	20:24
Breath Volume	3.820	20:24
Air Blank	0.000	20:24
Control Test	0.081	20:25
Air Blank	0.000	20:25
Diagnostics Check	OK	20:25

Data Download Date/Time: 05/18/2026 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,RICHARD,T
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 04/22/2026 01:39:30

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.116	01:34
Breath Volume	3.453	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.120	01:38
Breath Volume	3.558	01:38
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 05/18/2026 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENTE,JORGE,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: OGLE,SHAWN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 04/29/2026 22:58:07

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.081	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.215	22:53
Breath Volume	4.007	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.220	22:56
Breath Volume	3.371	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 05/18/2026 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 04/07/2026 20:45:48

Diagnostics Check	OK	20:39
Air Blank	0.000	20:39
Control Test	0.081	20:40
Air Blank	0.000	20:40
Subject Sample # 1	0.128	20:41
Breath Volume	1.863	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.132	20:44
Breath Volume	3.152	20:44
Air Blank	0.000	20:44
Control Test	0.081	20:45
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 04/17/2026 09:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BEENER,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MORALES_CARDENAS

DATE/TIME OF TEST: 04/18/2026 04:58:02

Diagnostics Check	OK	04:57	
Air Blank		04:57	RFI Detect
Air Blank		04:58	RFI Detect

Data Download Date/Time: 06/02/2026 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BEENER,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MORALES_CARDENAS

DATE/TIME OF TEST: 04/18/2026 05:00:12

Diagnostics Check	OK	04:59	
Air Blank		04:59	RFI Detect
Air Blank		05:00	RFI Detect

Data Download Date/Time: 06/02/2026 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BEENER,K
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MORALES_CARDENAS

DATE/TIME OF TEST: 04/18/2026 05:09:49

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.167	05:04
Breath Volume	2.667	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.167	05:08
Breath Volume	2.398	05:08
Air Blank	0.000	05:08
Control Test	0.080	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 06/02/2026 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACAVOY,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: TREVINO,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PICARD

DATE/TIME OF TEST: 04/19/2026 01:43:14

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.222	01:38
Breath Volume	3.148	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.226	01:41
Breath Volume	2.929	01:41
Air Blank	0.000	01:42
Control Test	0.081	01:42
Air Blank	0.000	01:43
Diagnostics Check	OK	01:43

Data Download Date/Time: 06/02/2026 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BUSHONG,NICOLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: OELHOFFEN

DATE/TIME OF TEST: 04/21/2026 22:19:18

Diagnostics Check	OK	22:17	
Air Blank	0.000	22:18	
Control Test	0.081	22:18	RFI Detect
Air Blank	0.000	22:19	

Data Download Date/Time: 06/02/2026 10:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BUSHONG,NICOLE,R

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BUSHONG,NICOLE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No

DRY GAS STANDARD LOT #: 403008479

DRY GAS EXPIRY DATE: 03/28/2027

VIOLATION CODE: DUI

ARREST TIME: 21:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: OELHOFFEN

DATE/TIME OF TEST: 04/21/2026 22:27:14

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.081	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.214	22:22
Breath Volume	2.824	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.206	22:25
Breath Volume	1.464	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 06/02/2026 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APONTE,J,
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: APONTE,J
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 04/17/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403008479
DRY GAS EXPIRY DATE: 03/28/2027
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: M233417990000

DATE/TIME OF TEST: 04/29/2026 19:08:24

Diagnostics Check	OK	19:04	
Air Blank	0.000	19:05	
Control Test	0.081	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.082	19:07	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 06/02/2026 10:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWENDEMAN, ELIZABETH, P
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: HECTOR, BRADY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 192543
DRY GAS EXPIRY DATE: 12/13/2022
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SCHWENDEMAN

DATE/TIME OF TEST: 04/19/2026 04:11:11

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.182	04:06
Breath Volume	2.777	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.180	04:09
Breath Volume	2.671	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 05/29/2026 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEXIS,VLADIMIR,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: ALEXIS,VLADIMIR
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 250514
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/23/2026 02:52:19

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.177	02:47
Breath Volume	2.726	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.173	02:50
Breath Volume	2.070	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 05/29/2026 17:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLE,MAX,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BARTLE,MAX
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: LEATHERS

DATE/TIME OF TEST: 04/02/2026 01:41:16

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:38	
Control Test	0.079	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.079	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 04/04/2026 23:32:20

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.017	23:27
Breath Volume	1.156	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.017	23:30
Breath Volume	2.011	23:30
Air Blank	0.000	23:31
Control Test	0.081	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCKENZIE

DATE/TIME OF TEST: 04/05/2026 04:01:34

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.161	03:53
Breath Volume	1.492	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.190	03:57
Breath Volume	1.519	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 3	0.181	04:00
Breath Volume	1.996	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/05/2026 23:19:05

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:14
Breath Volume	2.417	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	2.304	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:18
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: SIEG

DATE/TIME OF TEST: 04/06/2026 00:13:00

Diagnostics Check	OK	00:09	
Air Blank	0.000	00:10	
Control Test	0.081	00:10	
Air Blank	0.000	00:11	
Subject Sample	# 1	00:11	Subject Test Refused
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Control Test	0.081	00:12	
Air Blank	0.000	00:12	
Diagnostics Check	OK	00:13	

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: OTWELL

DATE/TIME OF TEST: 04/06/2026 01:32:40

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.081	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.150	01:28
Breath Volume	1.941	01:28
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.161	01:31
Breath Volume	2.656	01:31
Air Blank	0.000	01:31
Control Test	0.080	01:31
Air Blank	0.000	01:32
Diagnostics Check	OK	01:32

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARLE,MAX,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BARTLE,MAX
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 04/07/2026 04:40:37

Diagnostics Check	OK	04:36	
Air Blank	0.000	04:37	
Control Test	0.080	04:37	
Air Blank	0.000	04:38	
Subject Sample	# 1	04:39	Subject Test Refused
Breath Volume	0.000	04:39	
Air Blank	0.000	04:39	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Diagnostics Check	OK	04:40	

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: CORNELIOUS

DATE/TIME OF TEST: 04/09/2026 20:52:22

Diagnostics Check	OK	20:46
Air Blank	0.000	20:46
Control Test	0.081	20:47
Air Blank	0.000	20:47
Subject Sample # 1	0.000	20:48
Breath Volume	2.382	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.000	20:50
Breath Volume	2.808	20:50
Air Blank	0.000	20:51
Control Test	0.080	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 04/14/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CLINTON

DATE/TIME OF TEST: 04/09/2026 22:07:47

Diagnostics Check	OK	22:06	
Air Blank		22:07	RFI Detect
Air Blank	0.000	22:07	

Data Download Date/Time: 04/14/2026 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CLINTON

DATE/TIME OF TEST: 04/09/2026 22:15:12

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.100	22:10
Breath Volume	2.687	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.099	22:13
Breath Volume	1.558	22:13
Air Blank	0.000	22:14
Control Test	0.081	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 04/14/2026 17:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: HARRISON,CHRISTIAN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 04/10/2026 23:43:45

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.065	23:39
Breath Volume	2.035	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.064	23:42
Breath Volume	1.996	23:42
Air Blank	0.000	23:42
Control Test	0.081	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/14/2026 17:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 03/03/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: RODNEY

DATE/TIME OF TEST: 04/11/2026 02:02:47

Diagnostics Check	OK	01:50	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1 0.122	01:54	Volume Not Met
Breath Volume	0.019	01:54	
Air Blank	0.000	01:55	
Air Blank	0.000	01:56	
Subject Sample	# 2 0.140	01:57	
Breath Volume	1.285	01:57	
Air Blank	0.000	01:58	
Air Blank	0.000	02:00	
Subject Sample	# 3 0.138	02:01	
Breath Volume	1.468	02:01	
Air Blank	0.000	02:01	
Control Test	0.081	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 04/14/2026 17:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 04/18/2026 18:59:42

Diagnostics Check	OK	18:53
Air Blank	0.000	18:53
Control Test	0.081	18:54
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:55
Breath Volume	1.574	18:55
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:58
Breath Volume	1.585	18:58
Air Blank	0.000	18:58
Control Test	0.080	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 04/19/2026 23:02:11

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.082	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.171	22:57
Breath Volume	2.105	22:57
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.166	23:00
Breath Volume	1.886	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HELLINGS

DATE/TIME OF TEST: 04/20/2026 03:30:06

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.068	03:25
Breath Volume	2.722	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.068	03:28
Breath Volume	2.386	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:30

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FERRER KENNEDY

DATE/TIME OF TEST: 04/22/2026 20:55:56

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.080	20:50
Air Blank	0.000	20:51
Subject Sample # 1	0.000	20:51
Breath Volume	1.507	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.000	20:54
Breath Volume	1.769	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLE,MAX,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SWAIN,CECIL
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LEIGHTON

DATE/TIME OF TEST: 04/25/2026 02:10:10

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.076	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:05
Breath Volume	2.101	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	1.929	02:08
Air Blank	0.000	02:09
Control Test	0.078	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOME,PATRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROOME,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: POITEVINT

DATE/TIME OF TEST: 04/25/2026 03:42:00

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.058	03:37
Breath Volume	2.128	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:39
Subject Sample # 2	0.057	03:40
Breath Volume	3.023	03:40
Air Blank	0.000	03:41
Control Test	0.079	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:42

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSEPH,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: SMITH,JOSEPH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/25/2026 21:29:31

Diagnostics Check	OK	21:22
Air Blank	0.000	21:23
Control Test	0.079	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.213	21:24
Breath Volume	1.550	21:24
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.194	21:27
Breath Volume	1.226	21:27
Air Blank	0.000	21:28
Control Test	0.079	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLE,MAX,S
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: BARTLE,MAX
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 04/29/2026 21:11:41

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.080	21:06
Air Blank	0.000	21:06
Subject Sample # 1	0.200	21:07
Breath Volume	1.132	21:07
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.196	21:10
Breath Volume	1.109	21:10
Air Blank	0.000	21:10
Control Test	0.079	21:11
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COZAD,AUNDREA
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 04/30/2026 17:15:05

Diagnostics Check	OK	17:07
Air Blank	0.000	17:07
Control Test	0.080	17:07
Air Blank	0.000	17:08
Subject Sample # 1	0.000	17:09
Breath Volume	1.281	17:09
Air Blank	0.000	17:09
Air Blank	0.000	17:11
Subject Sample # 2	0.000	17:13
Breath Volume	1.523	17:13
Air Blank	0.000	17:14
Control Test	0.080	17:14
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOSS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: MOSS,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: BURKS

DATE/TIME OF TEST: 04/30/2026 20:29:24

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.078	20:23
Air Blank	0.000	20:24
Subject Sample # 1	0.000	20:24
Breath Volume	2.382	20:24
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.000	20:27
Breath Volume	1.472	20:27
Air Blank	0.000	20:28
Control Test	0.078	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLWELL,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: COLWELL,NICHOLAS
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/13/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403516665
DRY GAS EXPIRY DATE: 12/16/2028
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: LIEBERT

DATE/TIME OF TEST: 04/30/2026 21:22:55

Diagnostics Check	OK	21:14
Air Blank	0.000	21:14
Control Test	0.081	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.257	21:17
Breath Volume	1.546	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.277	21:21
Breath Volume	2.480	21:21
Air Blank	0.000	21:21
Control Test	0.078	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 05/07/2026 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSSO,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARMONA,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 04/02/2026 01:39:23

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.082	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.068	01:34
Breath Volume	1.988	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.064	01:37
Breath Volume	1.156	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,ANTONIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VERREAULT,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MARANGI

DATE/TIME OF TEST: 04/02/2026 02:02:29

Diagnostics Check	OK	02:01	
Air Blank		02:02	RFI Detect
Air Blank	0.000	02:02	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VERREAULT,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MARANGI

DATE/TIME OF TEST: 04/02/2026 02:13:49

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.081	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.061	02:08
Breath Volume	1.402	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.063	02:12
Breath Volume	1.625	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CAMPBELL,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/03/2026 04:26:17

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.000	04:21
Breath Volume	2.097	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.000	04:24
Breath Volume	2.257	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: FURILLO

DATE/TIME OF TEST: 04/04/2026 00:12:50

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:06
Air Blank	0.000	00:07
Subject Sample # 1	0.045	00:08
Breath Volume	1.976	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.048	00:11
Breath Volume	2.816	00:11
Air Blank	0.000	00:11
Control Test	0.081	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SYMIRNIOS,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 04/05/2026 21:02:42

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:47	
Control Test	0.081	20:47	
Air Blank	0.000	20:48	
Subject Sample	# 1 0.148	20:51	Volume Not Met
Breath Volume	0.152	20:51	
Air Blank	0.000	20:52	
Air Blank	0.000	20:53	
Subject Sample	# 2 0.158	20:55	
Breath Volume	1.113	20:55	
Air Blank	0.000	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 3 0.138	21:00	Volume Not Met
Breath Volume	0.000	21:00	
Air Blank	0.000	21:01	
Control Test	0.079	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SYMIRNIOS,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 04/05/2026 21:17:43

Diagnostics Check	OK	21:05	
Air Blank	0.000	21:06	
Control Test	0.078	21:06	
Air Blank	0.000	21:07	
Subject Sample	# 1 0.153	21:10	Volume Not Met
Breath Volume	0.175	21:10	
Air Blank	0.000	21:11	
Air Blank	0.000	21:12	
Subject Sample	# 2 0.139	21:15	Volume Not Met
Breath Volume	0.066	21:15	
Air Blank	0.000	21:16	
Control Test	0.078	21:17	
Air Blank	0.000	21:17	
Diagnostics Check	OK	21:17	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORIETY,G,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SYMIRNIOS,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 04/05/2026 21:30:35

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.078	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.126	21:25	Volume Not Met
Breath Volume	0.117	21:25	
Air Blank	0.000	21:25	
Air Blank	0.000	21:27	
Subject Sample	# 2	21:28	Subject Test Refused
Breath Volume	0.144	21:28	
Air Blank	0.000	21:29	
Control Test	0.077	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNOS,AVERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNOS,AVERY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: RENAN

DATE/TIME OF TEST: 04/09/2026 01:36:52

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.000	01:32
Breath Volume	2.027	01:32
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.000	01:35
Breath Volume	1.968	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACUNA,V,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HARRISON,Z
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: LOPEZ DE LEON

DATE/TIME OF TEST: 04/09/2026 23:13:30

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.080	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.177	23:08
Breath Volume	2.875	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.177	23:11
Breath Volume	2.890	23:11
Air Blank	0.000	23:12
Control Test	0.080	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSTABLE,AVRY,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WALKER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LIAS

DATE/TIME OF TEST: 04/11/2026 03:48:46

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.218	03:43
Breath Volume	1.683	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.214	03:47
Breath Volume	1.675	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDICINO,KAYLA,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DAVIS,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 04/11/2026 19:25:12

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.081	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.336	19:20
Breath Volume	1.292	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:22
Subject Sample # 2	0.323	19:23
Breath Volume	1.503	19:23
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: OBRIEN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/11/2026 19:33:31

Diagnostics Check	OK	19:30	
Air Blank	0.000	19:30	
Control Test	0.080	19:31	
Air Blank	0.000	19:31	
Subject Sample	# 1	19:32	Subject Test Refused
Breath Volume	0.000	19:32	
Air Blank	0.000	19:32	
Control Test	0.081	19:32	
Air Blank	0.000	19:33	
Diagnostics Check	OK	19:33	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOUSSAINT,S,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STANLEY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ANDREOLI

DATE/TIME OF TEST: 04/12/2026 03:50:57

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.074	03:46
Breath Volume	2.058	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.068	03:49
Breath Volume	1.437	03:49
Air Blank	0.000	03:50
Control Test	0.081	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROCA,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 205240B0A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HERNANDEZ-MORA

DATE/TIME OF TEST: 04/12/2026 04:02:18

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.247	03:57
Breath Volume	1.519	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.230	04:00
Breath Volume	1.468	04:00
Air Blank	0.000	04:01
Control Test	0.081	04:01
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARON,ZACHARY,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FAREZESPIN,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 04/16/2026 02:30:09

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.081	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.041	02:25
Breath Volume	1.304	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:27
Subject Sample # 2	0.042	02:28
Breath Volume	1.468	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUGG,MEGAN,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HILL,G
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BECERRA CARMONA

DATE/TIME OF TEST: 04/18/2026 03:23:23

Diagnostics Check	OK	03:15
Air Blank	0.000	03:16
Control Test	0.082	03:16
Air Blank	0.000	03:16
Subject Sample # 1	0.164	03:18
Breath Volume	1.425	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.148	03:21
Breath Volume	1.890	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVER,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PLMISANO,DEPUTY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: CARREJO

DATE/TIME OF TEST: 04/18/2026 12:34:48

Diagnostics Check	OK	12:27
Air Blank	0.000	12:28
Control Test	0.080	12:28
Air Blank	0.000	12:29
Subject Sample # 1	0.244	12:29
Breath Volume	3.683	12:29
Air Blank	0.000	12:30
Air Blank	0.000	12:32
Subject Sample # 2	0.232	12:33
Breath Volume	3.632	12:33
Air Blank	0.000	12:33
Control Test	0.080	12:34
Air Blank	0.000	12:34
Diagnostics Check	OK	12:34

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FRAGA,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 205240B0A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: ROKHVARGER

DATE/TIME OF TEST: 04/19/2026 00:30:10

Diagnostics Check	OK	00:19	
Air Blank	0.000	00:19	
Control Test	0.081	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1 0.000	00:23	Volume Not Met
Breath Volume	0.148	00:23	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2	00:28	Subject Test Refused
Breath Volume	0.167	00:28	
Air Blank	0.000	00:29	
Control Test	0.080	00:29	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAHAGIAN,TARA,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PELLEGRINO,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: CALIFANO

DATE/TIME OF TEST: 04/19/2026 00:46:38

Diagnostics Check	OK	00:37
Air Blank	0.000	00:38
Control Test	0.080	00:38
Air Blank	0.000	00:39
Subject Sample # 1	0.184	00:40
Breath Volume	2.535	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.176	00:44
Breath Volume	1.472	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:46
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERALEZ,R,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TROY,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MACHICAO

DATE/TIME OF TEST: 04/19/2026 01:53:43

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:47
Air Blank	0.000	01:48
Subject Sample # 1	0.164	01:49
Breath Volume	1.839	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.152	01:52
Breath Volume	1.285	01:52
Air Blank	0.000	01:52
Control Test	0.080	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,M,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BLANK,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: PEREZ DE LEON

DATE/TIME OF TEST: 04/19/2026 21:31:22

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.248	21:25
Breath Volume	1.980	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.243	21:29
Breath Volume	1.285	21:29
Air Blank	0.000	21:30
Control Test	0.079	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,ARTHUR,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROELL,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BOZANO

DATE/TIME OF TEST: 04/21/2026 02:59:30

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.082	02:48	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.143	02:49	
Breath Volume	1.152	02:49	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2 0.124	02:54	Volume Not Met
Breath Volume	1.011	02:54	
Air Blank	0.000	02:55	
Air Blank	0.000	02:57	
Subject Sample	# 3 0.124	02:57	
Breath Volume	1.480	02:57	
Air Blank	0.000	02:58	
Control Test	0.079	02:58	
Air Blank	0.000	02:59	
Diagnostics Check	OK	02:59	

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODI,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FARQUHARSON

DATE/TIME OF TEST: 04/21/2026 04:01:05

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.216	03:56
Breath Volume	3.093	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.218	03:59
Breath Volume	3.179	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/26/2026 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPISE,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CLARK,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BUSBY

DATE/TIME OF TEST: 04/23/2026 00:29:00

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.081	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.125	00:22
Breath Volume	2.082	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.124	00:27
Breath Volume	1.300	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/26/2026 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREGOSO,DANIEL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MERCADO,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: BARNHILL

DATE/TIME OF TEST: 04/23/2026 11:43:41

Diagnostics Check	OK	11:37
Air Blank	0.000	11:37
Control Test	0.080	11:38
Air Blank	0.000	11:38
Subject Sample # 1	0.164	11:39
Breath Volume	2.347	11:39
Air Blank	0.000	11:39
Air Blank	0.000	11:41
Subject Sample # 2	0.157	11:41
Breath Volume	2.839	11:41
Air Blank	0.000	11:42
Control Test	0.080	11:43
Air Blank	0.000	11:43
Diagnostics Check	OK	11:43

Data Download Date/Time: 04/26/2026 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDERE,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HOWELL,H
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: LAUARIEGA

DATE/TIME OF TEST: 04/24/2026 21:33:23

Diagnostics Check	OK	21:26
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.201	21:28
Breath Volume	2.046	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.192	21:31
Breath Volume	2.175	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 04/26/2026 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MADDEN,M,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GUTIEREZ,I
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SCHELL

DATE/TIME OF TEST: 04/25/2026 01:20:56

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.081	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.176	01:16
Breath Volume	1.550	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.175	01:19
Breath Volume	1.503	01:19
Air Blank	0.000	01:19
Control Test	0.081	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 04/26/2026 08:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001202

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAGBY,A,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARBAJAL,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20524080A2
DRY GAS EXPIRY DATE: 09/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/25/2026 01:35:00

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.179	01:30
Breath Volume	1.425	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.176	01:33
Breath Volume	1.242	01:33
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:35

Data Download Date/Time: 04/26/2026 08:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 09:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: ETIENNE WHITEHALL

DATE/TIME OF TEST: 04/04/2026 10:03:34

Diagnostics Check	OK	09:57
Air Blank	0.000	09:57
Control Test	0.080	09:58
Air Blank	0.000	09:58
Subject Sample # 1	0.000	09:59
Breath Volume	1.839	09:59
Air Blank	0.000	09:59
Air Blank	0.000	10:01
Subject Sample # 2	0.000	10:02
Breath Volume	2.093	10:02
Air Blank	0.000	10:02
Control Test	0.078	10:02
Air Blank	0.000	10:03
Diagnostics Check	OK	10:03

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOSA,RONALD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CEDILLO ZAVALA

DATE/TIME OF TEST: 04/17/2026 22:24:42

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.079	22:19	
Air Blank	0.000	22:20	
Subject Sample # 1	0.134	22:21	
Breath Volume	1.878	22:21	
Air Blank	0.000	22:21	
Air Blank	0.000	22:23	
Subject Sample # 2	0.127	22:24	
Breath Volume	1.160	22:24	
Air Blank		22:24	RFI Detect
Air Blank	0.000	22:24	

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DREW,LIAM,S
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GOSA,RONALD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CEDILLO ZAVALA

DATE/TIME OF TEST: 04/17/2026 22:34:11

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.136	22:29
Breath Volume	2.367	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.123	22:32
Breath Volume	1.132	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:34
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: JOQCEUS

DATE/TIME OF TEST: 04/18/2026 16:03:52

Diagnostics Check	OK	15:58	
Air Blank	0.000	15:58	
Control Test	0.079	15:58	
Air Blank	0.000	15:59	
Subject Sample	# 1 0.336	16:02	Volume Not Met
Breath Volume	0.574	16:02	
Air Blank		16:03	Purge Fail
Air Blank	0.000	16:03	

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ GARCIA,HERNAN,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CONANT,ROBERT
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: JOQCEUS

DATE/TIME OF TEST: 04/18/2026 16:16:08

Diagnostics Check	OK	16:06
Air Blank	0.000	16:07
Control Test	0.078	16:07
Air Blank	0.000	16:07
Subject Sample # 1	0.348	16:09
Breath Volume	2.492	16:09
Air Blank	0.000	16:10
Air Blank	0.000	16:11
Subject Sample # 2	0.357	16:14
Breath Volume	2.101	16:14
Air Blank	0.000	16:15
Control Test	0.078	16:15
Air Blank	0.000	16:16
Diagnostics Check	OK	16:16

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, LAURA,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: FARRAJ, SANAD
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 03/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ESCOBAR

DATE/TIME OF TEST: 04/19/2026 02:03:59

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.155	01:59
Breath Volume	3.699	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:01
Subject Sample # 2	0.150	02:02
Breath Volume	3.875	02:02
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 04/24/2026 09:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOW,DAIMON,D
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: DOW,DAIMON
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 04/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403268835
DRY GAS EXPIRY DATE: 02/13/2028
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 04/30/2026 23:44:38

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.173	23:40
Breath Volume	1.996	23:40
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.171	23:43
Breath Volume	2.218	23:43
Air Blank	0.000	23:43
Control Test	0.078	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 05/20/2026 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUERNFEIND,BRANDON,M
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BAUERNFEIND,BRANDON
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 04/05/2026 03:04:23

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.143	02:59
Breath Volume	1.910	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.136	03:02
Breath Volume	2.238	03:02
Air Blank	0.000	03:03
Control Test	0.078	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 05/02/2026 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMPIERCHIO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: COMPIERCHIO,ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: VALERIO

DATE/TIME OF TEST: 04/11/2026 03:44:27

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.173	03:39
Breath Volume	1.550	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:41
Subject Sample # 2	0.173	03:42
Breath Volume	1.425	03:42
Air Blank	0.000	03:43
Control Test	0.079	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/02/2026 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMPIERCHIO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: COMPIERCHIO, ANTHONY
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: MALEC

DATE/TIME OF TEST: 04/12/2026 01:26:48

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.192	01:22
Breath Volume	2.128	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.183	01:25
Breath Volume	1.554	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 05/02/2026 04:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMPIERCHIO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: IGNOFFO,ZACH
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 04/12/2026 04:09:20

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.166	04:04
Breath Volume	1.464	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.171	04:07
Breath Volume	1.636	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/02/2026 04:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEURTSEN,ANTHONY,P
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: PIGNATARO,THOMAS
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: RAWLINGS

DATE/TIME OF TEST: 04/18/2026 00:15:02

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.180	00:09
Breath Volume	1.433	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.170	00:13
Breath Volume	2.007	00:13
Air Blank	0.000	00:14
Control Test	0.078	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:15

Data Download Date/Time: 05/02/2026 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEINEKE,TIMOTHY,N
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: PIERCE,NATHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LAMPING

DATE/TIME OF TEST: 04/18/2026 03:02:34

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.165	02:57
Breath Volume	1.156	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.166	03:00
Breath Volume	2.257	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 05/02/2026 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMIN, PEYTON, C
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CARMIN, PEYTON
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: YARBROUGH

DATE/TIME OF TEST: 04/29/2026 23:42:10

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.198	23:37
Breath Volume	1.453	23:37
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.198	23:40
Breath Volume	1.515	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 05/02/2026 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COMPIERCHIO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: BYERS, JARED
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32324080A2
DRY GAS EXPIRY DATE: 12/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DAUBAR

DATE/TIME OF TEST: 04/30/2026 01:11:16

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.080	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.203	01:06
Breath Volume	1.582	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.196	01:09
Breath Volume	1.609	01:09
Air Blank	0.000	01:10
Control Test	0.079	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/02/2026 04:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,TREVOR,W
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GABOURY,SCOTT
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 04/20/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403419188
DRY GAS EXPIRY DATE: 08/13/2028
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: ARMATO

DATE/TIME OF TEST: 04/25/2026 02:44:21

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Subject Sample # 1	0.195	02:39
Breath Volume	2.175	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.186	02:42
Breath Volume	1.750	02:42
Air Blank	0.000	02:43
Control Test	0.077	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/24/2026 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLEIMAN,ERIC,M
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: HARMAN,MARIA
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 03/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: SOLIS GARCIA

DATE/TIME OF TEST: 04/25/2026 02:06:01

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.198	02:01
Breath Volume	2.882	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.201	02:04
Breath Volume	1.929	02:04
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:06

Data Download Date/Time: 04/28/2026 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATINELLA,GIANFRANCO,
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WATTERS,JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: MCCLOUD

DATE/TIME OF TEST: 04/11/2026 07:37:36

Diagnostics Check	OK	07:36	
Air Blank		07:37	RFI Detect
Air Blank	0.000	07:37	

Data Download Date/Time: 04/27/2026 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATINELLA,GIANFRANCO,
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: WATTERS,JONATHAN
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 03/30/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: MCCLLOUD

DATE/TIME OF TEST: 04/11/2026 07:49:31

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.082	07:42
Air Blank	0.000	07:42
Subject Sample # 1	0.140	07:43
Breath Volume	1.339	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:45
Subject Sample # 2	0.139	07:47
Breath Volume	2.828	07:47
Air Blank	0.000	07:48
Control Test	0.082	07:48
Air Blank	0.000	07:49
Diagnostics Check	OK	07:49

Data Download Date/Time: 04/27/2026 17:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUERNFEIND,BRANDON,M
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: BAUERNFEIND,BRANDON
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/27/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403411042
DRY GAS EXPIRY DATE: 08/02/2028
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/27/2026 21:53:52

Diagnostics Check	OK	21:51	
Air Blank	0.000	21:51	
Control Test	0.081	21:51	
Air Blank	0.000	21:52	
Subject Sample	# 1	21:52	Subject Test Refused
Breath Volume	0.000	21:52	
Air Blank	0.000	21:52	
Control Test	0.082	21:53	
Air Blank	0.000	21:53	
Diagnostics Check	OK	21:53	

Data Download Date/Time: 05/29/2026 05:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS, LOURDES, P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MAYNARD, DARREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 04/10/2026 21:28:38

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:20	
Control Test	0.082	21:20	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.000	21:24	Volume Not Met
Breath Volume	0.160	21:24	
Air Blank	0.000	21:25	
Air Blank	0.000	21:26	
Subject Sample	# 2	21:27	Subject Test Refused
Breath Volume	0.000	21:27	
Air Blank	0.000	21:27	
Control Test	0.081	21:28	
Air Blank	0.000	21:28	
Diagnostics Check	OK	21:28	

Data Download Date/Time: 04/27/2026 10:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ISOLA, CHRISTOPHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DIAGNAULT

DATE/TIME OF TEST: 04/14/2026 02:42:28

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.216	02:37
Breath Volume	2.027	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.220	02:40
Breath Volume	1.867	02:40
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/27/2026 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS, LOURDES, P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GILMORE, CAGER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BRITTAIN

DATE/TIME OF TEST: 04/21/2026 03:26:56

Diagnostics Check	OK	03:26	
Air Blank		03:26	RFI Detect
Air Blank	0.000	03:26	

Data Download Date/Time: 04/27/2026 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS, LOURDES, P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GILMORE, CAGER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BRITTAIN

DATE/TIME OF TEST: 04/21/2026 03:35:35

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.082	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.161	03:31
Breath Volume	1.320	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.152	03:34
Breath Volume	1.152	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 04/27/2026 10:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001253

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS, LOURDES, P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS, LOURDES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PASSIO

DATE/TIME OF TEST: 04/26/2026 01:31:55

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.083	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.195	01:27
Breath Volume	1.148	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.186	01:30
Breath Volume	1.980	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/27/2026 10:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA, LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 04/04/2026 01:39:23

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.120	01:34
Breath Volume	1.253	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.123	01:37
Breath Volume	1.375	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 04/08/2026 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTRERAS, JOHNATHAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CASTANEDA ALVARDO

DATE/TIME OF TEST: 04/04/2026 04:05:35

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.147	04:00
Breath Volume	1.464	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.157	04:04
Breath Volume	1.878	04:04
Air Blank	0.000	04:04
Control Test	0.077	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 04/08/2026 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MUNIZ MORALES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: CARDENAS MENDOZA

DATE/TIME OF TEST: 04/05/2026 03:18:34

Diagnostics Check	OK	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:07	
Air Blank	0.000	03:07	
Subject Sample	# 1 0.093	03:10	Volume Not Met
Breath Volume	0.367	03:10	
Air Blank	0.000	03:11	
Air Blank	0.000	03:13	
Subject Sample	# 2 0.131	03:13	
Breath Volume	1.292	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:16	
Subject Sample	# 3 0.131	03:17	
Breath Volume	2.187	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 04/08/2026 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 04/07/2026 03:14:34

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.173	03:10
Breath Volume	2.937	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.169	03:13
Breath Volume	3.312	03:13
Air Blank	0.000	03:13
Control Test	0.078	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 04/08/2026 13:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCALES,BRANDON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 04/01/2026 07:16:25

Diagnostics Check	OK	07:10
Air Blank	0.000	07:10
Control Test	0.083	07:10
Air Blank	0.000	07:11
Subject Sample # 1	0.236	07:11
Breath Volume	1.265	07:11
Air Blank	0.000	07:12
Air Blank	0.000	07:14
Subject Sample # 2	0.230	07:14
Breath Volume	1.554	07:14
Air Blank	0.000	07:15
Control Test	0.082	07:15
Air Blank	0.000	07:16
Diagnostics Check	OK	07:16

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 04/01/2026 18:17:55

Diagnostics Check	OK	18:11
Air Blank	0.000	18:12
Control Test	0.083	18:12
Air Blank	0.000	18:13
Subject Sample # 1	0.000	18:13
Breath Volume	3.023	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 2	0.000	18:16
Breath Volume	2.683	18:16
Air Blank	0.000	18:17
Control Test	0.082	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/01/2026 22:28:07

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.083	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.023	22:23
Breath Volume	1.511	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.020	22:26
Breath Volume	1.105	22:26
Air Blank	0.000	22:27
Control Test	0.083	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEAGALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MENDEZ YOC

DATE/TIME OF TEST: 04/02/2026 03:00:45

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.083	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.189	02:52
Breath Volume	1.339	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:54
Subject Sample # 2	0.212	02:56
Breath Volume	1.902	02:56
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 3	0.218	02:59
Breath Volume	2.664	02:59
Air Blank	0.000	02:59
Control Test	0.079	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:00

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DORVILUS

DATE/TIME OF TEST: 04/02/2026 04:00:36

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.126	03:54
Breath Volume	1.382	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.124	03:59
Breath Volume	1.527	03:59
Air Blank	0.000	03:59
Control Test	0.082	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROMAN, MARIANO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: SAUCIER

DATE/TIME OF TEST: 04/02/2026 16:09:36

Diagnostics Check	OK	16:03
Air Blank	0.000	16:04
Control Test	0.082	16:04
Air Blank	0.000	16:04
Subject Sample # 1	0.220	16:05
Breath Volume	1.257	16:05
Air Blank	0.000	16:05
Air Blank	0.000	16:07
Subject Sample # 2	0.231	16:08
Breath Volume	1.207	16:08
Air Blank	0.000	16:08
Control Test	0.083	16:08
Air Blank	0.000	16:09
Diagnostics Check	OK	16:09

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PADILLA
SANTILLAN,ALBERTO
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 04/03/2026 03:21:30

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.244	03:16
Breath Volume	1.632	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.239	03:19
Breath Volume	2.355	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KALIKAPERSAUD,NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 04/03/2026 04:07:25

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.082	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.188	04:02
Breath Volume	1.257	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.191	04:05
Breath Volume	1.660	04:05
Air Blank	0.000	04:06
Control Test	0.082	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CONTRERAS,JONATHAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: RAMIREZMACARENO

DATE/TIME OF TEST: 04/03/2026 21:40:39

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.082	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.118	21:36
Breath Volume	1.121	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.114	21:39
Breath Volume	1.574	21:39
Air Blank	0.000	21:39
Control Test	0.082	21:40
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: GOMEZCAX

DATE/TIME OF TEST: 04/04/2026 01:16:35

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.082	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.243	01:11
Breath Volume	2.054	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.236	01:15
Breath Volume	1.234	01:15
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,PHILLIP
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 04/04/2026 03:01:14

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.083	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.150	02:56
Breath Volume	1.515	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.146	02:59
Breath Volume	2.003	02:59
Air Blank	0.000	03:00
Control Test	0.083	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCMAHAN,JARED
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: NAJDUCH

DATE/TIME OF TEST: 04/04/2026 03:30:27

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.082	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.180	03:26
Breath Volume	2.386	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.177	03:28
Breath Volume	2.152	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 04/08/2026 13:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 04/04/2026 07:36:22

Diagnostics Check	OK	07:30
Air Blank	0.000	07:30
Control Test	0.082	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.000	07:31
Breath Volume	1.863	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.000	07:34
Breath Volume	1.312	07:34
Air Blank	0.000	07:35
Control Test	0.083	07:35
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARRERO,GABRIEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 04/04/2026 23:40:59

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.083	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.169	23:36
Breath Volume	4.132	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.169	23:39
Breath Volume	4.125	23:39
Air Blank	0.000	23:40
Control Test	0.082	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELLA VOIR,JOHN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 04/05/2026 17:46:36

Diagnostics Check	OK	17:36
Air Blank	0.000	17:36
Control Test	0.082	17:36
Air Blank	0.000	17:37
Subject Sample # 1	0.211	17:38
Breath Volume	2.625	17:38
Air Blank	0.000	17:38
Air Blank	0.000	17:40
Subject Sample # 2	0.190	17:41
Breath Volume	1.335	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:44
Subject Sample # 3	0.197	17:45
Breath Volume	1.652	17:45
Air Blank	0.000	17:45
Control Test	0.081	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 04/06/2026 00:12:44

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.081	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.186	00:08
Breath Volume	2.593	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.193	00:11
Breath Volume	3.175	00:11
Air Blank	0.000	00:11
Control Test	0.082	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ, BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/06/2026 03:57:47

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.142	03:53
Breath Volume	1.328	03:53
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.139	03:56
Breath Volume	1.601	03:56
Air Blank	0.000	03:56
Control Test	0.082	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ENGLISH,LEE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: BRIONES

DATE/TIME OF TEST: 04/06/2026 21:28:05

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.082	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.315	21:23
Breath Volume	2.527	21:23
Air Blank	0.000	21:24
Air Blank	0.000	21:25
Subject Sample # 2	0.296	21:26
Breath Volume	1.312	21:26
Air Blank	0.000	21:27
Control Test	0.080	21:27
Air Blank	0.000	21:28
Diagnostics Check	OK	21:28

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN LOPEZ,BIENVENIDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNOWLTON,BIRANNA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 04/07/2026 02:05:31

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.082	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.000	02:01
Breath Volume	3.484	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.000	02:04
Breath Volume	3.214	02:04
Air Blank	0.000	02:04
Control Test	0.082	02:04
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANTERMAN,TAYLOR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/08/2026 02:12:53

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.082	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.098	02:08
Breath Volume	3.707	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.088	02:11
Breath Volume	1.476	02:11
Air Blank	0.000	02:11
Control Test	0.081	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 04/08/2026 02:29:22

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.082	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.000	02:25
Breath Volume	1.656	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	2.773	02:28
Air Blank	0.000	02:28
Control Test	0.082	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 04/08/2026 13:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HARTNETT

DATE/TIME OF TEST: 04/09/2026 23:40:22

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.084	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.000	23:36
Breath Volume	2.468	23:36
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.000	23:39
Breath Volume	3.117	23:39
Air Blank	0.000	23:39
Control Test	0.083	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATKINS,ANTHONY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: ERAZO

DATE/TIME OF TEST: 04/10/2026 00:55:47

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.083	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.083	00:55	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: YULFO VARGAS, DAIRA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: FIGUEROA FLORES

DATE/TIME OF TEST: 04/10/2026 03:21:00

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.083	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.060	03:16
Breath Volume	2.113	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:18
Subject Sample # 2	0.061	03:19
Breath Volume	1.472	03:19
Air Blank	0.000	03:20
Control Test	0.082	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:21

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHEWS,ZACHARY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/11/2026 02:25:51

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.083	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.000	02:21
Breath Volume	3.671	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:24
Breath Volume	3.773	02:24
Air Blank	0.000	02:24
Control Test	0.083	02:25
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/12/2026 03:43:18

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.084	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.143	03:38
Breath Volume	1.902	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.138	03:41
Breath Volume	1.218	03:41
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SEJOUR,NICO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/18/2026 04:41:47

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.082	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1 0.222	04:34	Volume Not Met
Breath Volume	0.031	04:34	
Air Blank	0.000	04:35	
Air Blank	0.000	04:37	
Subject Sample	# 2 0.172	04:40	Volume Not Met
Breath Volume	0.429	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:41	
Air Blank	0.000	04:41	
Diagnostics Check	OK	04:41	

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEON-VAZQUEZ,BIANCA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: FOWLER

DATE/TIME OF TEST: 04/19/2026 23:51:34

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.082	23:38	
Air Blank	0.000	23:38	
Subject Sample	# 1 0.360	23:39	
Breath Volume	2.156	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 2	23:44	No Sample Provided
Breath Volume	0.000	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 3 0.341	23:50	
Breath Volume	1.167	23:50	
Air Blank	0.000	23:50	
Control Test	0.083	23:50	
Air Blank	0.000	23:51	
Diagnostics Check	OK	23:51	

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: PRESCOTT

DATE/TIME OF TEST: 04/20/2026 03:58:10

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.242	03:53
Breath Volume	2.187	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.241	03:56
Breath Volume	1.527	03:56
Air Blank	0.000	03:57
Control Test	0.082	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SENA,HECTOR
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SELF

DATE/TIME OF TEST: 04/25/2026 03:07:47

Diagnostics Check	OK	03:01
Air Blank	0.000	03:02
Control Test	0.083	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.208	03:03
Breath Volume	1.761	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.190	03:06
Breath Volume	1.167	03:06
Air Blank	0.000	03:06
Control Test	0.083	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/14/2026 13:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: COOKS

DATE/TIME OF TEST: 04/27/2026 03:08:40

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.084	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1 0.169	03:01	Volume Not Met
Breath Volume	0.296	03:01	
Air Blank	0.000	03:02	
Air Blank	0.000	03:04	
Subject Sample	# 2 0.161	03:07	Volume Not Met
Breath Volume	0.125	03:07	
Air Blank	0.000	03:07	
Control Test	0.083	03:08	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 05/14/2026 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,LAWRENCE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: BOUNCE

DATE/TIME OF TEST: 04/27/2026 04:37:19

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.084	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.000	04:33
Breath Volume	3.917	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.000	04:35
Breath Volume	3.195	04:35
Air Blank	0.000	04:36
Control Test	0.083	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 05/14/2026 13:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AVILA,ARNOLD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WASHBURN

DATE/TIME OF TEST: 04/29/2026 03:55:40

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.083	03:40	
Air Blank	0.000	03:40	
Subject Sample # 1	0.223	03:43	
Breath Volume	1.300	03:43	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample # 2	0.172	03:48	Volume Not Met
Breath Volume	0.351	03:48	
Air Blank	0.000	03:49	
Air Blank	0.000	03:51	
Subject Sample # 3		03:54	No Sample Provided
Breath Volume	0.000	03:54	
Air Blank	0.000	03:54	
Control Test	0.081	03:55	
Air Blank	0.000	03:55	
Diagnostics Check	OK	03:55	

Data Download Date/Time: 05/14/2026 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AVILA,ARNOLD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/08/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: WASHBURN

DATE/TIME OF TEST: 04/29/2026 04:25:11

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.082	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.224	04:20
Breath Volume	2.453	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.226	04:23
Breath Volume	2.187	04:23
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 05/14/2026 13:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYNY,ERIC,C
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HERNANDEZ,D
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/03/2026 06:29:20

Diagnostics Check	OK	06:21
Air Blank	0.000	06:22
Control Test	0.079	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.000	06:24
Breath Volume	2.550	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.000	06:27
Breath Volume	1.558	06:27
Air Blank	0.000	06:28
Control Test	0.078	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 04/14/2026 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: LAMB

DATE/TIME OF TEST: 04/03/2026 23:46:28

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:44	
Control Test	0.079	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:45	Subject Test Refused
Breath Volume	0.000	23:45	
Air Blank	0.000	23:45	
Control Test	0.079	23:45	
Air Blank	0.000	23:46	
Diagnostics Check	OK	23:46	

Data Download Date/Time: 04/14/2026 21:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNAMARA,K,
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MCNAMARA,K
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 03/18/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A2
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: CODY

DATE/TIME OF TEST: 04/05/2026 21:25:37

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:21	No Sample Provided
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Air Blank	0.000	21:23	
Subject Sample	# 2	21:24	Subject Test Refused
Breath Volume	0.000	21:24	
Air Blank	0.000	21:24	
Control Test	0.079	21:25	
Air Blank	0.000	21:25	
Diagnostics Check	OK	21:25	

Data Download Date/Time: 04/14/2026 21:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCNAMARA,K,

BREATH TEST OPERATOR AGENCY: Cocoa PD

ARRESTING OFFICER: MCNAMARA,K

ARRESTING OFFICER AGENCY: Cocoa PD

LAST AGENCY INSPECTION: 04/14/2026

Date or Time Changed: No

DRY GAS STANDARD LOT #: 15025080A2

DRY GAS EXPIRY DATE: 07/05/2027

VIOLATION CODE: DUI

ARREST TIME: 04:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: NAVEDO JR

DATE/TIME OF TEST: 04/24/2026 04:44:38

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:42	
Control Test	0.080	04:42	
Air Blank	0.000	04:42	
Subject Sample	# 1	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.079	04:44	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 05/26/2026 20:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,JOCELYN,L
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BELK,R
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 13:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/20/2026 14:37:41

Diagnostics Check	OK	14:31
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:32
Subject Sample # 1	0.341	14:33
Breath Volume	2.480	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:35
Subject Sample # 2	0.338	14:36
Breath Volume	2.339	14:36
Air Blank	0.000	14:36
Control Test	0.078	14:37
Air Blank	0.000	14:37
Diagnostics Check	OK	14:37

Data Download Date/Time: 05/28/2026 08:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: SUMNER,IAN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 04/01/2026 00:11:16

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.220	00:06
Breath Volume	2.324	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.224	00:09
Breath Volume	2.550	00:09
Air Blank	0.000	00:10
Control Test	0.080	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WOCK,MICHAEL
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 04/08/2026 00:08:45

Diagnostics Check	OK	00:02
Air Blank	0.000	00:02
Control Test	0.080	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.190	00:04
Breath Volume	3.566	00:04
Air Blank	0.000	00:05
Air Blank	0.000	00:06
Subject Sample # 2	0.194	00:07
Breath Volume	2.746	00:07
Air Blank	0.000	00:07
Control Test	0.080	00:08
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,DERICK,
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: SOMMER,RUSSELL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: SCIASCIA

DATE/TIME OF TEST: 04/08/2026 22:54:05

Diagnostics Check	OK	22:47
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Subject Sample # 1	0.205	22:49
Breath Volume	1.765	22:49
Air Blank	0.000	22:50
Air Blank	0.000	22:51
Subject Sample # 2	0.208	22:52
Breath Volume	1.500	22:52
Air Blank	0.000	22:53
Control Test	0.080	22:53
Air Blank	0.000	22:54
Diagnostics Check	OK	22:54

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: RAMOS,JHONATHAN
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/11/2026 03:43:18

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.207	03:36
Breath Volume	1.468	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.217	03:41
Breath Volume	1.312	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: JOSEPH,DYLAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/12/2026 22:00:07

Diagnostics Check	OK	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:53
Air Blank	0.000	21:53
Subject Sample # 1	0.140	21:55
Breath Volume	1.984	21:55
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.139	21:58
Breath Volume	1.582	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: ROCK,JAY-QUAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 03/12/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOREL

DATE/TIME OF TEST: 04/13/2026 01:28:09

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.081	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.009	01:23
Breath Volume	1.320	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.010	01:26
Breath Volume	1.246	01:26
Air Blank	0.000	01:27
Control Test	0.080	01:27
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 04/15/2026 21:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: RAMOS,JHONATHAN
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: OUELLETTE

DATE/TIME OF TEST: 04/15/2026 22:31:05

Diagnostics Check	OK	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:29	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:29	Subject Test Refused
Breath Volume	0.000	22:29	
Air Blank	0.000	22:30	
Control Test	0.081	22:30	
Air Blank	0.000	22:31	
Diagnostics Check	OK	22:31	

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZERVAS

DATE/TIME OF TEST: 04/18/2026 02:08:54

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1 0.096	02:04	
Breath Volume	2.703	02:04	
Air Blank	0.000	02:05	
Air Blank	0.000	02:07	
Subject Sample	# 2	02:07	RFI Detect
Breath Volume	0.000	02:07	
Air Blank	0.000	02:08	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ZERVAS

DATE/TIME OF TEST: 04/18/2026 02:16:56

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.099	02:12
Breath Volume	3.058	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.095	02:15
Breath Volume	2.886	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,D
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: IMEL,JOSHUA
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JAGIELSKI

DATE/TIME OF TEST: 04/19/2026 04:09:13

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.294	04:04
Breath Volume	2.667	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.281	04:07
Breath Volume	2.097	04:07
Air Blank	0.000	04:08
Control Test	0.077	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,D
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: JENKINSON,T
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MC KOWEN

DATE/TIME OF TEST: 04/19/2026 23:11:25

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.273	23:06
Breath Volume	2.957	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.272	23:09
Breath Volume	2.617	23:09
Air Blank	0.000	23:10
Control Test	0.080	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: RAMOS,JHONATHAN
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: KARIUKI

DATE/TIME OF TEST: 04/26/2026 01:23:43

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.081	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Subject Test Refused
Breath Volume	0.000	01:22	
Air Blank	0.000	01:22	
Control Test	0.081	01:23	
Air Blank	0.000	01:23	
Diagnostics Check	OK	01:23	

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: MAHLER,CARLITA
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/26/2026 20:32:06

Diagnostics Check	OK	20:23
Air Blank	0.000	20:24
Control Test	0.081	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.063	20:26
Breath Volume	1.156	20:26
Air Blank	0.000	20:26
Air Blank	0.000	20:28
Subject Sample # 2	0.063	20:30
Breath Volume	1.542	20:30
Air Blank	0.000	20:31
Control Test	0.081	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: ROCK,JAYQUAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 04/27/2026 02:25:22

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.081	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:24	
Control Test	0.081	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,COREY,D
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: KOUTROS,PATRICK
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 04/15/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 2375080A2
DRY GAS EXPIRY DATE: 10/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 04/29/2026 04:00:19

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.034	03:55
Breath Volume	1.503	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.031	03:58
Breath Volume	1.410	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/29/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORIARTY,RORY,P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BUSTILLO,ZANE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: HUFFAKER

DATE/TIME OF TEST: 04/04/2026 20:33:49

Diagnostics Check	OK	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Subject Sample # 1	0.163	20:29
Breath Volume	2.785	20:29
Air Blank	0.000	20:30
Air Blank	0.000	20:31
Subject Sample # 2	0.170	20:32
Breath Volume	3.046	20:32
Air Blank	0.000	20:32
Control Test	0.080	20:33
Air Blank	0.000	20:33
Diagnostics Check	OK	20:33

Data Download Date/Time: 04/27/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRAEUTER,DARIAN,R
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: POTTRATZ,KEITH
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: WITHROW

DATE/TIME OF TEST: 04/05/2026 08:09:25

Diagnostics Check	OK	08:00
Air Blank	0.000	08:01
Control Test	0.079	08:01
Air Blank	0.000	08:02
Subject Sample # 1	0.143	08:03
Breath Volume	3.304	08:03
Air Blank	0.000	08:03
Air Blank	0.000	08:05
Subject Sample # 2	0.155	08:07
Breath Volume	3.171	08:07
Air Blank	0.000	08:08
Control Test	0.078	08:08
Air Blank	0.000	08:09
Diagnostics Check	OK	08:09

Data Download Date/Time: 04/27/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOWICKI,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: BIRCHFIELD,JERMEY
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ESPARZA

DATE/TIME OF TEST: 04/05/2026 22:12:02

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:07	No Sample Provided
Breath Volume	0.000	22:07	
Air Blank	0.000	22:07	
Air Blank	0.000	22:09	
Subject Sample	# 2	22:11	Sequence Aborted
Breath Volume	0.000	22:11	
Air Blank	0.000	22:12	

Data Download Date/Time: 04/27/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOWICKI,JUSTIN,T
BREATH TEST OPERATOR AGENCY: Kennedy Space Center
ARRESTING OFFICER: BIRCHFIELD,JEREMY
ARRESTING OFFICER AGENCY: Kennedy Space Center
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ESPARZA VELASCO

DATE/TIME OF TEST: 04/05/2026 22:18:29

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.080	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:17	Sequence Aborted
Breath Volume	0.000	22:17	
Air Blank	0.000	22:18	

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MAYER,JEREMIAH
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: TORRES JUSTINIANO

DATE/TIME OF TEST: 04/06/2026 17:14:56

Diagnostics Check	OK	17:06
Air Blank	0.000	17:06
Control Test	0.080	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.229	17:08
Breath Volume	1.589	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:10
Subject Sample # 2	0.231	17:13
Breath Volume	1.835	17:13
Air Blank	0.000	17:14
Control Test	0.078	17:14
Air Blank	0.000	17:14
Diagnostics Check	OK	17:14

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUNTAIN,DALE,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: SANTANA,LUCIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/09/2026 04:03:08

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.259	03:58
Breath Volume	1.179	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.275	04:01
Breath Volume	1.253	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINICIS, LOURDES, P
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOMINICIS, LOURDES
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: AGOSTO

DATE/TIME OF TEST: 04/16/2026 06:34:52

Diagnostics Check	OK	06:28
Air Blank	0.000	06:29
Control Test	0.078	06:29
Air Blank	0.000	06:30
Subject Sample # 1	0.177	06:30
Breath Volume	3.347	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:32
Subject Sample # 2	0.179	06:33
Breath Volume	3.269	06:33
Air Blank	0.000	06:33
Control Test	0.078	06:34
Air Blank	0.000	06:34
Diagnostics Check	OK	06:34

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HEDERMAN,CASEY,L

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: LEWIS,SHAQUILLE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29525080A1

DRY GAS EXPIRY DATE: 11/05/2027

VIOLATION CODE: DUI

ARREST TIME: 22:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 04/17/2026 23:30:27

Diagnostics Check	OK	23:23
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:24
Subject Sample # 1	0.263	23:25
Breath Volume	1.949	23:25
Air Blank	0.000	23:25
Air Blank	0.000	23:27
Subject Sample # 2	0.252	23:28
Breath Volume	2.093	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAROCHE, CHARLES, W
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIRCHFIELD, JEREMY
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: ZUROMS

DATE/TIME OF TEST: 04/19/2026 19:00:14

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.080	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.048	18:55
Breath Volume	1.714	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.048	18:58
Breath Volume	1.574	18:58
Air Blank	0.000	18:59
Control Test	0.080	18:59
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LOGAN, TYRONE, P

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: LOGAN, TYRONE

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No

DRY GAS STANDARD LOT #: 29525080A1

DRY GAS EXPIRY DATE: 11/05/2027

VIOLATION CODE: DUI

ARREST TIME: 20:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: MCCAIN

DATE/TIME OF TEST: 04/19/2026 21:16:53

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:11
Subject Sample # 1	0.185	21:12
Breath Volume	3.734	21:12
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.165	21:15
Breath Volume	1.351	21:15
Air Blank	0.000	21:15
Control Test	0.079	21:16
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDRO,DARIEN,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CARRERA,CALEO
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: SERRET

DATE/TIME OF TEST: 04/20/2026 11:42:35

Diagnostics Check	OK	11:36
Air Blank	0.000	11:37
Control Test	0.079	11:37
Air Blank	0.000	11:38
Subject Sample # 1	0.000	11:38
Breath Volume	4.214	11:38
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.000	11:41
Breath Volume	3.476	11:41
Air Blank	0.000	11:41
Control Test	0.079	11:42
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 04/27/2026 10:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLARD,F,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: POLLARD,F
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 04/26/2026 01:17:53

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.162	01:13
Breath Volume	1.562	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.165	01:16
Breath Volume	1.261	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 04/27/2026 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTIZ,AXEL,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: ORTIZ,AXEL
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 03/25/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29525080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RENTZ

DATE/TIME OF TEST: 04/26/2026 03:49:29

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.211	03:44
Breath Volume	2.453	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.210	03:48
Breath Volume	2.437	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 04/27/2026 10:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BASTOS,KLAYTON
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/03/2026 03:48:16

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.080	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1 0.193	03:42	
Breath Volume	1.808	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample	# 2 0.186	03:46	
Breath Volume	2.050	03:46	
Air Blank	0.000	03:47	
Control Test	0.078	03:47	Interferent Detect
Air Blank	0.000	03:48	

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: APRILE,MICHAEL
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: SLAUGHTER

DATE/TIME OF TEST: 04/03/2026 04:07:53

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.120	04:02
Breath Volume	1.171	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.119	04:06
Breath Volume	1.730	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BYRD,RYAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: VELASCO ROJAS

DATE/TIME OF TEST: 04/04/2026 02:14:15

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.080	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1 0.000	02:05	Volume Not Met
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2 0.280	02:09	No .020 agreement
Breath Volume	2.937	02:09	
Air Blank	0.000	02:10	
Air Blank	0.000	02:12	
Subject Sample	# 3 0.247	02:12	No .020 agreement
Breath Volume	1.855	02:12	
Air Blank	0.000	02:13	
Control Test	0.076	02:13	
Air Blank	0.000	02:14	
Diagnostics Check	OK	02:14	

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BYRD,RYAN
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: VELASCO ROJAS

DATE/TIME OF TEST: 04/04/2026 02:24:29

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:17	
Control Test	0.078	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1 0.260	02:20	
Breath Volume	1.843	02:20	
Air Blank	0.000	02:20	
Air Blank	0.000	02:22	
Subject Sample	# 2	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:23	
Control Test	0.076	02:23	
Air Blank	0.000	02:24	
Diagnostics Check	OK	02:24	

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BURDGE,DALTON
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/04/2026 04:09:36

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.166	04:05
Breath Volume	2.460	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.174	04:08
Breath Volume	2.734	04:08
Air Blank	0.000	04:08
Control Test	0.075	04:09
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ENNS,JOHN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 04/04/2026 05:34:06

Diagnostics Check	OK	05:26
Air Blank	0.000	05:27
Control Test	0.080	05:27
Air Blank	0.000	05:28
Subject Sample # 1	0.165	05:29
Breath Volume	1.250	05:29
Air Blank	0.000	05:29
Air Blank	0.000	05:31
Subject Sample # 2	0.166	05:32
Breath Volume	1.195	05:32
Air Blank	0.000	05:33
Control Test	0.080	05:33
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WATT,KEIONDRE
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 04/05/2026 00:18:50

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.240	00:14
Breath Volume	3.628	00:14
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.235	00:17
Breath Volume	3.660	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:18
Air Blank	0.000	00:18
Diagnostics Check	OK	00:18

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VANROMPA,VITO
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GOBER

DATE/TIME OF TEST: 04/06/2026 03:51:18

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.194	03:45
Breath Volume	2.519	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.193	03:49
Breath Volume	1.816	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHMIEDER,N
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: VAZQUEZ VAZQUEZ

DATE/TIME OF TEST: 04/08/2026 02:13:55

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.159	02:05
Breath Volume	1.828	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.138	02:09
Breath Volume	1.410	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:11
Subject Sample # 3	0.145	02:12
Breath Volume	1.453	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARMSTRONG,ISAAC,C
BREATH TEST OPERATOR AGENCY: Winter Springs PD
ARRESTING OFFICER: BLASKOWSKI,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 04/08/2026 19:52:32

Diagnostics Check	OK	19:42
Air Blank	0.000	19:43
Control Test	0.080	19:43
Air Blank	0.000	19:44
Subject Sample # 1	0.145	19:46
Breath Volume	1.261	19:46
Air Blank	0.000	19:47
Air Blank	0.000	19:49
Subject Sample # 2	0.139	19:50
Breath Volume	1.507	19:50
Air Blank	0.000	19:51
Control Test	0.076	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ROSE,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 04/09/2026 01:07:39

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.177	01:03
Breath Volume	1.566	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.190	01:06
Breath Volume	1.953	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MALDONADO,K
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 04/09/2026 01:52:17

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.153	01:47
Breath Volume	2.335	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.148	01:50
Breath Volume	1.832	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COOK,A
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DOMENECH MALDONADO

DATE/TIME OF TEST: 04/10/2026 23:16:11

Diagnostics Check	OK	23:09
Air Blank	0.000	23:10
Control Test	0.081	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.096	23:11
Breath Volume	3.449	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:14
Subject Sample # 2	0.098	23:14
Breath Volume	2.660	23:14
Air Blank	0.000	23:15
Control Test	0.080	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 04/29/2026 23:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOORHEAD,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: WIDDALL

DATE/TIME OF TEST: 04/11/2026 01:31:31

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.081	01:25
Air Blank	0.000	01:26
Subject Sample # 1	0.232	01:27
Breath Volume	1.699	01:27
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.226	01:29
Breath Volume	2.691	01:29
Air Blank	0.000	01:30
Control Test	0.079	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MCMULLEN,M
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: CARBO

DATE/TIME OF TEST: 04/11/2026 02:31:40

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.117	02:27
Breath Volume	2.765	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.115	02:30
Breath Volume	2.867	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOXLEY,M
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ABURTO

DATE/TIME OF TEST: 04/11/2026 04:20:47

Diagnostics Check	OK	04:12
Air Blank	0.000	04:13
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.029	04:15
Breath Volume	1.421	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:18
Subject Sample # 2	0.029	04:19
Breath Volume	1.292	04:19
Air Blank	0.000	04:19
Control Test	0.079	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAYNE,N
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 04/11/2026 05:36:22

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:20	
Control Test	0.080	05:21	
Air Blank	0.000	05:21	
Subject Sample	# 1 0.144	05:24	Volume Not Met
Breath Volume	0.093	05:24	
Air Blank	0.000	05:25	
Air Blank	0.000	05:27	
Subject Sample	# 2 0.134	05:29	
Breath Volume	1.324	05:29	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 3 0.123	05:34	
Breath Volume	1.183	05:34	
Air Blank	0.000	05:35	
Control Test	0.077	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 04/12/2026 01:11:49

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:05
Subject Sample # 1	0.153	01:06
Breath Volume	2.011	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.157	01:10
Breath Volume	2.085	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CULHANE,C
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 04/12/2026 03:12:07

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:06
Subject Sample # 1	0.166	03:07
Breath Volume	1.273	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.163	03:10
Breath Volume	1.246	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/12/2026 04:00:52

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.081	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.070	04:00	Control Outside Tolerance
Air Blank	0.000	04:00	

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: FRITSCH,A
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/12/2026 04:48:26

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.173	04:41
Breath Volume	1.316	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.165	04:46
Breath Volume	1.187	04:46
Air Blank	0.000	04:47
Control Test	0.077	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MATEO,Y
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/14/2026 23:41:13

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.163	23:36
Breath Volume	2.421	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.163	23:39
Breath Volume	1.800	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAYNE,NOAH
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: HAYS

DATE/TIME OF TEST: 04/15/2026 22:20:31

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.093	22:15
Breath Volume	1.257	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.089	22:19
Breath Volume	1.574	22:19
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RITTGER,N
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MASSEY

DATE/TIME OF TEST: 04/16/2026 04:14:27

Diagnostics Check	OK	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.229	04:09
Breath Volume	1.949	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.227	04:12
Breath Volume	1.285	04:12
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SHIMPENO,JASON
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: SHUKLA

DATE/TIME OF TEST: 04/16/2026 23:53:05

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.244	23:48
Breath Volume	3.281	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.244	23:51
Breath Volume	3.207	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:53

Data Download Date/Time: 04/29/2026 23:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VELAZQUEZ,RODERICK
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROOMFIELD

DATE/TIME OF TEST: 04/17/2026 04:17:39

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:13	
Control Test	0.079	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:16	Interferent Detect
Breath Volume	1.378	04:16	
Air Blank	0.000	04:16	
Control Test	0.080	04:17	Interferent Detect
Air Blank	0.000	04:17	

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VELAZQUEZ,RODERICK
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROOMFIELD

DATE/TIME OF TEST: 04/17/2026 04:21:56

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.078	04:21	Interferent Detect
Air Blank	0.000	04:21	

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: VELAZQUEZ,RODERICK
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BROOMFIELD

DATE/TIME OF TEST: 04/17/2026 04:32:47

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:27
Subject Sample # 1	0.165	04:28
Breath Volume	1.132	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.155	04:31
Breath Volume	1.382	04:31
Air Blank	0.000	04:31
Control Test	0.076	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,N
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NEUSCH,H
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/17/2026 04:57:02

Diagnostics Check	OK	04:50	
Air Blank	0.000	04:51	
Control Test	0.079	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:55	Interferent Detect
Breath Volume	1.519	04:55	
Air Blank	0.000	04:56	
Control Test	0.080	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: EWING,HALEY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ECHEVERRIA

DATE/TIME OF TEST: 04/18/2026 03:53:21

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.179	03:48
Breath Volume	1.562	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.191	03:51
Breath Volume	1.644	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: YOUNG,N
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: SURINA

DATE/TIME OF TEST: 04/19/2026 23:55:00

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.187	23:50
Breath Volume	1.746	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.176	23:53
Breath Volume	1.554	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Diagnostics Check	OK	23:55

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NORTHCUTT,E
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: MOKHOV

DATE/TIME OF TEST: 04/20/2026 01:46:17

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.166	01:41
Breath Volume	2.066	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.167	01:44
Breath Volume	1.156	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SUAREZ,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/21/2026 04:39:57

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.135	04:33
Breath Volume	1.320	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:35
Subject Sample # 2	0.126	04:38
Breath Volume	1.468	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES,V
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: SOULE

DATE/TIME OF TEST: 04/21/2026 22:18:14

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.296	22:13
Breath Volume	1.699	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.301	22:16
Breath Volume	1.820	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ACKERMANN,S
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 04/22/2026 00:11:18

Diagnostics Check	OK	00:03
Air Blank	0.000	00:04
Control Test	0.080	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.000	00:05
Breath Volume	2.437	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.000	00:09
Breath Volume	2.054	00:09
Air Blank	0.000	00:10
Control Test	0.081	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 04/29/2026 23:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOURI,TAMY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: JOHNSON,M
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 03/02/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403314418
DRY GAS EXPIRY DATE: 04/14/2028
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TORRES AGUILAR

DATE/TIME OF TEST: 04/22/2026 03:44:32

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.227	03:40
Breath Volume	2.171	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.219	03:43
Breath Volume	1.976	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 04/29/2026 23:50