



# Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000229 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:54</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:55</td></tr> <tr><td>Control Test</td><td>0.047</td><td>13:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:57</td></tr> <tr><td>Control Test</td><td>0.047</td><td>13:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0473</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2198</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:54	Control Test	0.048	13:55	Air Blank	0.000	13:55	Control Test	0.047	13:56	Air Blank	0.000	13:57	Control Test	0.047	13:57	Air Blank	0.000	13:58	Control Test Stats			Average	0.0473		Std Dev	0.0016		Rel. Std Dev(%)	1.2198		<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000229 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.100</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.100	14:01	Control Test	0.078	14:02	Air Blank	0.000	14:03	Control Test	0.078	14:03	Air Blank	0.000	14:04	Control Test	0.077	14:05	Air Blank	0.000	14:05	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel. Std Dev(%)	0.7434		<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000229 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.199</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	12:03	Control Test	0.199	12:04	Air Blank	0.000	12:05	Control Test	0.198	12:05	Air Blank	0.000	12:06	Control Test	0.198	12:07	Air Blank	0.000	12:07	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel. Std Dev(%)	0.2911		<p>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000229 01/22/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:50	Control Test	0.079	13:50	Air Blank	0.000	13:51	Control Test	0.079	13:51	Air Blank	0.000	13:52	Control Test	0.079	13:52	Air Blank	0.000	13:53	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel. Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:54																																																																																																																																																	
Control Test	0.048	13:55																																																																																																																																																	
Air Blank	0.000	13:55																																																																																																																																																	
Control Test	0.047	13:56																																																																																																																																																	
Air Blank	0.000	13:57																																																																																																																																																	
Control Test	0.047	13:57																																																																																																																																																	
Air Blank	0.000	13:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0473																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel. Std Dev(%)	1.2198																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.100	14:01																																																																																																																																																	
Control Test	0.078	14:02																																																																																																																																																	
Air Blank	0.000	14:03																																																																																																																																																	
Control Test	0.078	14:03																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test	0.077	14:05																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.199	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test	0.198	12:05																																																																																																																																																	
Air Blank	0.000	12:06																																																																																																																																																	
Control Test	0.198	12:07																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test	0.079	13:50																																																																																																																																																	
Air Blank	0.000	13:51																																																																																																																																																	
Control Test	0.079	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.079	13:52																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 11:56

Date of Inspection: 01/23/2026

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.047	0.078	0.200	0.078
0.000	0.047	0.078	0.200	0.078
0.000	0.047	0.078	0.199	0.079
0.000	0.047	0.078	0.200	0.078
0.000	0.047	0.078	0.199	0.078
0.000	0.047	0.078	0.200	0.079
0.000	0.047	0.078	0.200	0.078
0.000	0.047	0.078	0.200	0.079
0.000	0.047	0.078	0.200	0.079
0.000	0.047	0.078	0.200	0.079
0.000	0.048	0.079	0.199	0.079

Standard Deviations	0.0003	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/23/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/23/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:56</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor  
Gutschow  
Digitally signed by Taylor  
Gutschow  
Date: 2026.01.25 19:44:55  
-05'00'

01/23/2026

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality