

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,ISRAEL,I
BREATH TEST OPERATOR AGENCY: FHP TROOP E
ARRESTING OFFICER: CRUZ,ISRAEL
ARRESTING OFFICER AGENCY: FHP TROOP E
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: ARBELO

DATE/TIME OF TEST: 01/01/2026 20:36:55

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.151	20:32
Breath Volume	3.542	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:34
Subject Sample # 2	0.155	20:35
Breath Volume	3.843	20:35
Air Blank	0.000	20:36
Control Test	0.079	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLAVERDE,ORLANDO,R
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: VILLAVERDE,ORLANDO
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SOBALVARRO

DATE/TIME OF TEST: 01/03/2026 04:03:17

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.178	03:58
Breath Volume	3.589	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.171	04:01
Breath Volume	3.507	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ORTEGA,WILFREDO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: BURON DIAZ

DATE/TIME OF TEST: 01/03/2026 08:19:09

Diagnostics Check	OK	08:16	
Air Blank	0.000	08:16	
Control Test	0.078	08:17	
Air Blank	0.000	08:17	
Subject Sample	# 1	08:17	Subject Test Refused
Breath Volume	0.000	08:17	
Air Blank	0.000	08:18	
Control Test	0.079	08:18	
Air Blank	0.000	08:19	
Diagnostics Check	OK	08:19	

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXIS,J
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GARCIA,ALEXIS
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 01/04/2026 02:53:14

Diagnostics Check		OK	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.201	02:48
Breath Volume		1.464	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:50
Subject Sample	# 2	0.198	02:51
Breath Volume		1.679	02:51
Air Blank		0.000	02:52
Control Test		0.078	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLAVERDE,ORLANDO,R
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: GONZALEZ,ANACELIA
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KHULLAR

DATE/TIME OF TEST: 01/07/2026 00:22:26

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.270	00:17
Breath Volume		2.925	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.260	00:20
Breath Volume		1.746	00:20
Air Blank		0.000	00:21
Control Test		0.077	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/08/2026 21:58:04

Diagnostics Check	OK	21:56	
Air Blank		21:57	Ambient Fail
Air Blank		21:58	Purge Fail

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,RANDY,R
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: VAZQUEZ,RANDY
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/08/2026 22:01:47

Diagnostics Check	OK	22:00	
Air Blank	0.000	22:00	
Control Test	0.063	22:01	Control Outside Tolerance
Air Blank	0.000	22:01	

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CALVO MULGADO

DATE/TIME OF TEST: 01/09/2026 02:21:41

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.078	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.157	02:17
Breath Volume	2.253	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.150	02:20
Breath Volume	1.910	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MARQUEZ CATANO

DATE/TIME OF TEST: 01/10/2026 02:55:01

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:50
Breath Volume		2.308	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:53
Breath Volume		2.730	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:55

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLAVERDE,ORLANDO,R
BREATH TEST OPERATOR AGENCY: SWEETWATER PD
ARRESTING OFFICER: GONZALEZ,ANACELIA
ARRESTING OFFICER AGENCY: SWEETWATER PD
LAST AGENCY INSPECTION: 12/03/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: DUARTE VELAZQUEZ

DATE/TIME OF TEST: 01/11/2026 05:53:16

Diagnostics Check	OK	05:50	
Air Blank	0.000	05:50	
Control Test	0.079	05:51	
Air Blank	0.000	05:51	
Subject Sample	# 1	05:51	Subject Test Refused
Breath Volume	0.000	05:51	
Air Blank	0.000	05:52	
Control Test	0.079	05:52	
Air Blank	0.000	05:53	
Diagnostics Check	OK	05:53	

Data Download Date/Time: 01/16/2026 02:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,SILVIO,A
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: QUINTANA,SILVIO
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/16/2026 04:09:55

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.223	04:05
Breath Volume		2.003	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.236	04:08
Breath Volume		2.437	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/05/2026 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORTEGA,WILFREDO,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 01/17/2026 08:39:51

Diagnostics Check		OK	08:33
Air Blank		0.000	08:33
Control Test		0.079	08:33
Air Blank		0.000	08:34
Subject Sample	# 1	0.156	08:35
Breath Volume		1.777	08:35
Air Blank		0.000	08:35
Air Blank		0.000	08:37
Subject Sample	# 2	0.145	08:38
Breath Volume		1.937	08:38
Air Blank		0.000	08:38
Control Test		0.078	08:39
Air Blank		0.000	08:39
Diagnostics Check		OK	08:39

Data Download Date/Time: 02/05/2026 04:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RUBIDO GONZALEZ

DATE/TIME OF TEST: 01/21/2026 22:18:07

Diagnostics Check	OK	22:11
Air Blank	0.000	22:11
Control Test	0.078	22:12
Air Blank	0.000	22:12
Subject Sample # 1	0.222	22:13
Breath Volume	2.113	22:13
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 2	0.222	22:16
Breath Volume	2.414	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:18
Diagnostics Check	OK	22:18

Data Download Date/Time: 02/05/2026 04:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAPE,RYAN,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: LAPE,RYAN
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 01/23/2026 20:59:04

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.078	20:53
Air Blank	0.000	20:54
Subject Sample # 1	0.000	20:54
Breath Volume	2.878	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:57
Breath Volume	4.003	20:57
Air Blank	0.000	20:58
Control Test	0.079	20:58
Air Blank	0.000	20:58
Diagnostics Check	OK	20:59

Data Download Date/Time: 02/05/2026 04:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: SAENZ

DATE/TIME OF TEST: 01/24/2026 05:52:02

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.078	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.227	05:47
Breath Volume		2.882	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:49
Subject Sample	# 2	0.238	05:50
Breath Volume		3.015	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:52

Data Download Date/Time: 02/05/2026 04:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANOS,LUIS,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: DIAZ,DANIEL
ARRESTING OFFICER AGENCY: HIALEAH GARDENS PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: BRAVO ORTIZ

DATE/TIME OF TEST: 01/27/2026 21:17:09

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.078	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.265	21:12
Breath Volume		3.417	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:14
Subject Sample	# 2	0.262	21:15
Breath Volume		3.605	21:15
Air Blank		0.000	21:16
Control Test		0.078	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 02/05/2026 04:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GATTORNO,ALEXANDER,F
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: GATTORNO,ALEXANDER
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: OLIVERA

DATE/TIME OF TEST: 01/31/2026 07:55:26

Diagnostics Check	OK	07:47
Air Blank	0.000	07:48
Control Test	0.078	07:48
Air Blank	0.000	07:49
Subject Sample # 1	0.128	07:50
Breath Volume	3.316	07:50
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.125	07:53
Breath Volume	3.070	07:53
Air Blank	0.000	07:54
Control Test	0.078	07:54
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 02/05/2026 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IBANEZ,HAREL,
BREATH TEST OPERATOR AGENCY: DORAL PD
ARRESTING OFFICER: GONZALEZ,FRANK
ARRESTING OFFICER AGENCY: DORAL PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: RUEDA

DATE/TIME OF TEST: 01/31/2026 21:24:31

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.078	21:21	
Air Blank	0.000	21:22	
Subject Sample	# 1	21:23	Subject Test Refused
Breath Volume	0.000	21:23	
Air Blank	0.000	21:23	
Control Test	0.079	21:23	
Air Blank	0.000	21:24	
Diagnostics Check	OK	21:24	

Data Download Date/Time: 02/05/2026 04:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: VILLAGOMEZ

DATE/TIME OF TEST: 01/01/2026 03:58:47

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.141	03:54
Breath Volume	1.496	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.142	03:57
Breath Volume	1.585	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 01/16/2026 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: MESA RODRIGUEZ

DATE/TIME OF TEST: 01/01/2026 05:28:09

Diagnostics Check	OK	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:23
Subject Sample # 1	0.103	05:23
Breath Volume	2.378	05:23
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.111	05:26
Breath Volume	2.285	05:26
Air Blank	0.000	05:27
Control Test	0.081	05:27
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 01/16/2026 00:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: MORRONE

DATE/TIME OF TEST: 01/01/2026 07:15:38

Diagnostics Check	OK	07:09
Air Blank	0.000	07:09
Control Test	0.080	07:09
Air Blank	0.000	07:10
Subject Sample # 1	0.166	07:11
Breath Volume	1.593	07:11
Air Blank	0.000	07:11
Air Blank	0.000	07:13
Subject Sample # 2	0.170	07:14
Breath Volume	2.667	07:14
Air Blank	0.000	07:14
Control Test	0.080	07:15
Air Blank	0.000	07:15
Diagnostics Check	OK	07:15

Data Download Date/Time: 01/16/2026 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ESTOL

DATE/TIME OF TEST: 01/03/2026 01:15:45

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.081	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.216	01:11
Breath Volume		2.300	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.225	01:14
Breath Volume		2.144	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 01/16/2026 00:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI DADE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTUNAY,MUHAMMET,
BREATH TEST OPERATOR AGENCY: MIAMI-DADE PD
ARRESTING OFFICER: ALTUNAY,MUHAMMET
ARRESTING OFFICER AGENCY: MIAMI-DADE PD
LAST AGENCY INSPECTION: 12/04/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A4
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/03/2026 07:20:37

Diagnostics Check		OK	07:13
Air Blank		0.000	07:14
Control Test		0.080	07:14
Air Blank		0.000	07:14
Subject Sample	# 1	0.201	07:16
Breath Volume		2.089	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.201	07:18
Breath Volume		1.984	07:18
Air Blank		0.000	07:19
Control Test		0.080	07:19
Air Blank		0.000	07:20
Diagnostics Check		OK	07:20

Data Download Date/Time: 01/16/2026 00:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ZELAYA AVILA

DATE/TIME OF TEST: 01/01/2026 02:03:18

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.081	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.078	01:55	Volume Not Met
Breath Volume		0.222	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.074	01:58	
Breath Volume		1.578	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 3	0.072	02:01	
Breath Volume		1.601	02:01	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 01/09/2026 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCARBOROUGH,W
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/01/2026 05:08:01

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:07	
Control Test	0.081	05:07	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 01/09/2026 10:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 01/03/2026 23:56:27

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.079	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.083	23:51
Breath Volume	1.464	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.087	23:54
Breath Volume	1.636	23:54
Air Blank	0.000	23:55
Control Test	0.078	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 01/09/2026 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: BOOKER

DATE/TIME OF TEST: 01/04/2026 02:51:49

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.080	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 01/09/2026 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STOINIOFF,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 01/04/2026 04:50:20

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Subject Sample # 1	0.122	04:45
Breath Volume	1.457	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.126	04:48
Breath Volume	1.753	04:48
Air Blank	0.000	04:49
Control Test	0.079	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 01/09/2026 10:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHAVEZ,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/04/2026 19:46:25

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:44	
Control Test	0.080	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:45	
Control Test	0.080	19:45	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 01/09/2026 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: PEDRO ANTONIO

DATE/TIME OF TEST: 01/05/2026 01:01:04

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.166	00:56
Breath Volume	1.812	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.171	00:59
Breath Volume	1.691	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:01

Data Download Date/Time: 01/09/2026 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KIRKENDOLL

DATE/TIME OF TEST: 01/06/2026 02:52:18

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.148	02:47
Breath Volume	1.136	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.163	02:50
Breath Volume	1.781	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 01/09/2026 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KOHLER,E
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: SCHENDER

DATE/TIME OF TEST: 01/07/2026 03:28:00

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 01/09/2026 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,D,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/10/2026 01:50:07

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.078	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.043	01:45
Breath Volume	1.660	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.043	01:48
Breath Volume	1.113	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 02/20/2026 13:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BOBANGO

DATE/TIME OF TEST: 01/11/2026 01:34:08

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.079	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:33	
Control Test	0.081	01:33	
Air Blank	0.000	01:34	
Diagnostics Check	OK	01:34	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/11/2026 03:05:38

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:04	
Control Test	0.080	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,M,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SUTKUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: RICHMAN

DATE/TIME OF TEST: 01/11/2026 22:25:03

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.211	22:20
Breath Volume		2.300	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.218	22:23
Breath Volume		1.562	22:23
Air Blank		0.000	22:24
Control Test		0.079	22:24
Air Blank		0.000	22:24
Diagnostics Check		OK	22:25

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: HYDLEBERG

DATE/TIME OF TEST: 01/11/2026 23:11:34

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/12/2026 02:03:08

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.000	01:58
Breath Volume	3.281	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.000	02:01
Breath Volume	3.441	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 01/18/2026 01:41:13

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.081	01:37	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 01/18/2026 05:01:57

Diagnostics Check	OK	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.080	05:01	
Air Blank	0.000	05:01	
Diagnostics Check	OK	05:01	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DEJESUS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: TOUSSAINT-BOOTHE

DATE/TIME OF TEST: 01/20/2026 02:36:36

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.189	02:31
Breath Volume	1.246	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.190	02:35
Breath Volume	1.597	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: NORTHRUP

DATE/TIME OF TEST: 01/23/2026 21:40:27

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.099	21:35
Breath Volume	3.015	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.093	21:38
Breath Volume	3.105	21:38
Air Blank	0.000	21:39
Control Test	0.078	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DUGGAN

DATE/TIME OF TEST: 01/24/2026 02:38:17

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.079	02:37	RFI Detect
Air Blank	0.000	02:38	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HARRELL,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DUGGAN

DATE/TIME OF TEST: 01/24/2026 02:43:46

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUPTILL,HANNAH,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HART,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: KANICKI

DATE/TIME OF TEST: 01/25/2026 06:05:17

Diagnostics Check	OK	05:58
Air Blank	0.000	05:58
Control Test	0.079	05:59
Air Blank	0.000	05:59
Subject Sample # 1	0.298	06:00
Breath Volume	3.019	06:00
Air Blank	0.000	06:01
Air Blank	0.000	06:03
Subject Sample # 2	0.299	06:03
Breath Volume	4.023	06:03
Air Blank	0.000	06:04
Control Test	0.079	06:04
Air Blank	0.000	06:05
Diagnostics Check	OK	06:05

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HILLIER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 01/25/2026 23:20:52

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.094	23:15
Breath Volume	1.171	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.094	23:19
Breath Volume	1.347	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,J,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAGYAR,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 01/28/2026 03:12:30

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.081	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.163	03:07
Breath Volume	1.839	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.170	03:11
Breath Volume	1.738	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON,K,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/30/2026 01:40:33

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.264	01:35
Breath Volume		2.648	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.258	01:39
Breath Volume		2.929	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRIKHAM,CHRISTIAN,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DYER

DATE/TIME OF TEST: 01/30/2026 03:38:37

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.080	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 01/30/2026 19:26:31

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:21	
Control Test	0.080	19:21	
Air Blank	0.000	19:22	
Subject Sample	# 1 0.208	19:22	
Breath Volume	2.750	19:22	
Air Blank	0.000	19:23	
Air Blank	0.000	19:25	
Subject Sample	# 2 0.216	19:25	
Breath Volume	2.984	19:25	
Air Blank		19:26	RFI Detect
Air Blank	0.000	19:26	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VALERIO,E
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 01/30/2026 19:34:56

Diagnostics Check		OK	19:28
Air Blank		0.000	19:29
Control Test		0.080	19:29
Air Blank		0.000	19:30
Subject Sample	# 1	0.209	19:30
Breath Volume		2.937	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:32
Subject Sample	# 2	0.205	19:33
Breath Volume		3.015	19:33
Air Blank		0.000	19:34
Control Test		0.079	19:34
Air Blank		0.000	19:34
Diagnostics Check		OK	19:34

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKHAM,CHRISTIAN,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARTIN,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28424080A3
DRY GAS EXPIRY DATE: 11/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FILS AIME

DATE/TIME OF TEST: 01/31/2026 02:49:54

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.081	02:49	
Air Blank	0.000	02:49	
Diagnostics Check	OK	02:49	

Data Download Date/Time: 02/20/2026 13:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,CODY,T
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: GIVENS,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/01/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/09/2026 03:29:53

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.097	03:25
Breath Volume	3.503	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.095	03:28
Breath Volume	2.957	03:28
Air Blank	0.000	03:28
Control Test	0.078	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 02/06/2026 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,CODY,T
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MURRAY,B
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/01/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 01/17/2026 02:00:35

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.078	01:55
Air Blank	0.000	01:55
Subject Sample # 1	0.157	01:55
Breath Volume	1.757	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.167	01:58
Breath Volume	2.164	01:58
Air Blank	0.000	01:59
Control Test	0.078	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 02/06/2026 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,GEORGE,W
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WARD,GLENN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: CAVANAGH

DATE/TIME OF TEST: 01/19/2026 07:24:58

Diagnostics Check	OK	07:21	
Air Blank	0.000	07:22	
Control Test	0.081	07:22	
Air Blank	0.000	07:23	
Subject Sample	# 1	07:23	Subject Test Refused
Breath Volume	0.000	07:23	
Air Blank	0.000	07:24	
Control Test	0.080	07:24	
Air Blank	0.000	07:24	
Diagnostics Check	OK	07:24	

Data Download Date/Time: 02/06/2026 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONKEL,SCOTT,M
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: JONES,ZACHARY
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GERDING

DATE/TIME OF TEST: 01/20/2026 01:42:35

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.191	01:38
Breath Volume	1.812	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.188	01:40
Breath Volume	1.183	01:40
Air Blank	0.000	01:41
Control Test	0.081	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 02/06/2026 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001283

INSTRUMENT REGISTERED TO: HIGH SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELLS,CODY,T
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MURRAY,BILL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/01/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: DENMARK

DATE/TIME OF TEST: 01/30/2026 23:02:00

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.078	22:56
Air Blank	0.000	22:57
Subject Sample # 1	0.203	22:57
Breath Volume	3.742	22:57
Air Blank	0.000	22:58
Air Blank	0.000	22:59
Subject Sample # 2	0.194	23:00
Breath Volume	3.480	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:02

Data Download Date/Time: 02/06/2026 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JESSICA,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: NELSON,JESSICA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: OTHER
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: KEYS

DATE/TIME OF TEST: 01/10/2026 12:42:36

Diagnostics Check	OK	12:35
Air Blank	0.000	12:36
Control Test	0.078	12:36
Air Blank	0.000	12:37
Subject Sample # 1	0.000	12:38
Breath Volume	3.558	12:38
Air Blank	0.000	12:38
Air Blank	0.000	12:40
Subject Sample # 2	0.000	12:41
Breath Volume	3.210	12:41
Air Blank	0.000	12:41
Control Test	0.078	12:42
Air Blank	0.000	12:42
Diagnostics Check	OK	12:42

Data Download Date/Time: 02/09/2026 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,JAMES,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KELLY,JAMES
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 01/13/2026 20:20:06

Diagnostics Check	OK	20:16	
Air Blank	0.000	20:17	
Control Test	0.078	20:17	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:18	Subject Test Refused
Breath Volume	0.000	20:18	
Air Blank	0.000	20:19	
Control Test	0.078	20:19	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 02/09/2026 14:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEANS-REINECKE,KATRINA,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MEANS-REINECKE,KATRINA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: G'SELL

DATE/TIME OF TEST: 01/14/2026 00:14:28

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.078	00:00	
Air Blank	0.000	00:00	
Subject Sample # 1	0.168	00:01	
Breath Volume	1.976	00:01	
Air Blank	0.000	00:02	
Air Blank	0.000	00:03	
Subject Sample # 2		00:07	No Sample Provided
Breath Volume	0.000	00:07	
Air Blank	0.000	00:07	
Air Blank	0.000	00:09	
Subject Sample # 3	0.160	00:12	Slope Not Level
Breath Volume	1.304	00:12	
Air Blank	0.000	00:13	
Control Test	0.078	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 02/09/2026 14:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNSFORD,LEIGHA,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BERRY,TYLER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 12:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: DUNSFORD

DATE/TIME OF TEST: 01/17/2026 01:41:22

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.078	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.194	01:37
Breath Volume	2.027	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.194	01:39
Breath Volume	1.613	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINS II,JERRY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FORD,JEREMIAH
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: HA

DATE/TIME OF TEST: 01/18/2026 07:46:45

Diagnostics Check	OK	07:39
Air Blank	0.000	07:40
Control Test	0.079	07:40
Air Blank	0.000	07:41
Subject Sample # 1	0.109	07:42
Breath Volume	1.738	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:44
Subject Sample # 2	0.105	07:45
Breath Volume	1.828	07:45
Air Blank	0.000	07:45
Control Test	0.077	07:46
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSLEY,KYLE,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOSLEY,KYLE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 01/18/2026 23:15:17

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.077	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:14	
Control Test	0.078	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,PATRICK,W
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: STEPHENSON,PATRICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: BEEKS

DATE/TIME OF TEST: 01/19/2026 04:27:53

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.187	04:22
Breath Volume	1.703	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.184	04:26
Breath Volume	1.457	04:26
Air Blank	0.000	04:26
Control Test	0.078	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CONKELL,LINDSEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 01/22/2026 17:51:10

Diagnostics Check	OK	17:47	
Air Blank	0.000	17:47	
Control Test	0.078	17:48	
Air Blank	0.000	17:48	
Subject Sample	# 1 0.278	17:49	
Breath Volume	2.304	17:49	
Air Blank		17:49	RFI Detect
Air Blank	0.000	17:50	
Control Test	0.077	17:50	
Air Blank	0.000	17:51	
Diagnostics Check	OK	17:51	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CONKELL,LINDSEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 01/22/2026 18:01:26

Diagnostics Check	OK	17:52	
Air Blank	0.000	17:52	
Control Test	0.077	17:53	
Air Blank	0.000	17:53	
Subject Sample	# 1 0.270	17:54	
Breath Volume	2.464	17:54	
Air Blank	0.000	17:55	
Air Blank	0.000	17:56	
Subject Sample	# 2 0.245	17:57	
Breath Volume	2.179	17:57	
Air Blank	0.000	17:58	
Air Blank	0.000	18:00	
Subject Sample	# 3 0.273	18:00	
Breath Volume	2.828	18:00	
Air Blank		18:00	RFI Detect
Air Blank	0.000	18:01	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAREY,ANDREW,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CONKELL,LINDSEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 16:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:25

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 01/22/2026 18:20:25

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.076	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.281	18:15
Breath Volume	2.644	18:15
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.277	18:18
Breath Volume	2.503	18:18
Air Blank	0.000	18:19
Control Test	0.076	18:19
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUISE,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MUISE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RIDLEY

DATE/TIME OF TEST: 01/23/2026 02:43:01

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:37
Air Blank	0.000	02:37
Subject Sample # 1	0.172	02:38
Breath Volume	1.625	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.169	02:41
Breath Volume	2.023	02:41
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:42
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMILLION,LANCE,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: GOMILLION,LANCE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 01/25/2026 01:52:02

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.078	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:47
Breath Volume	2.519	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:50
Breath Volume	2.113	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:52

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLIFFT,CORY
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 01/27/2026 01:49:01

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.071	01:48	RFI Detect
Air Blank	0.000	01:49	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLIFFT,CORY
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 01/27/2026 01:52:37

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
		01:51	Improper Sample
Air Blank	0.000	01:52	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLIFFT,CORY
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 01/27/2026 01:55:58

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:55	
Air Blank		01:55	RFI Detect
Air Blank	0.000	01:55	

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,SHELBY,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CLIFFT,CORY
ARRESTING OFFICER AGENCY: Milton PD
LAST AGENCY INSPECTION: 01/09/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403369407
DRY GAS EXPIRY DATE: 06/23/2028
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 01/27/2026 02:07:50

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.153	02:03
Breath Volume		4.105	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:05
Subject Sample	# 2	0.158	02:06
Breath Volume		4.621	02:06
Air Blank		0.000	02:06
Control Test		0.077	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/09/2026 14:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,MADISON,L
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: SANSOM,ADAM
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/19/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403476659
DRY GAS EXPIRY DATE: 10/28/2028
VIOLATION CODE: DUI
ARREST TIME: 15:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: BOOTH

DATE/TIME OF TEST: 01/05/2026 15:23:44

Diagnostics Check	OK	15:20	
Air Blank	0.000	15:21	
Control Test	0.080	15:21	
Air Blank	0.000	15:22	
Subject Sample	# 1	15:22	Subject Test Refused
Breath Volume	0.000	15:22	
Air Blank	0.000	15:22	
Control Test	0.080	15:23	
Air Blank	0.000	15:23	
Diagnostics Check	OK	15:23	

Data Download Date/Time: 01/29/2026 14:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,ROBERT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LABARRE,BEARY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: CASTLEBERRY JR

DATE/TIME OF TEST: 01/01/2026 00:36:17

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:32	No Sample Provided
Breath Volume	0.000	00:32	
Air Blank	0.000	00:32	
Air Blank	0.000	00:34	
Subject Sample	# 2	00:34	Subject Test Refused
Breath Volume	0.000	00:34	
Air Blank	0.000	00:35	
Control Test	0.080	00:35	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STALNAKER, TEDDY, L
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: IRONS, MASON
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SQUIER

DATE/TIME OF TEST: 01/01/2026 03:29:51

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.130	03:24
Breath Volume	1.191	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.122	03:28
Breath Volume	1.468	03:28
Air Blank	0.000	03:28
Control Test	0.082	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIPF,RICHARD,K
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ZIPF,RICHARD
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 01/02/2026 23:28:51

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:14	
Control Test		0.081	23:15	
Air Blank		0.000	23:15	
Subject Sample	# 1	0.153	23:16	
Breath Volume		1.437	23:16	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.128	23:21	Volume Not Met
Breath Volume		1.042	23:21	
Air Blank		0.000	23:22	
Air Blank		0.000	23:24	
Subject Sample	# 3	0.120	23:27	Volume Not Met
Breath Volume		0.589	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE,SEVEN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SIPE,SEVEN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CHESHIRE

DATE/TIME OF TEST: 01/07/2026 01:31:36

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.082	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.000	01:27
Breath Volume		2.410	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.000	01:30
Breath Volume		2.347	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,DYLAN,J
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCOTT,DYLAN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: VINEYARD

DATE/TIME OF TEST: 01/10/2026 05:30:51

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.080	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.143	05:26
Breath Volume	3.210	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.147	05:29
Breath Volume	4.023	05:29
Air Blank	0.000	05:29
Control Test	0.080	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LOCKMAN,HEATHER
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: NANCE

DATE/TIME OF TEST: 01/10/2026 13:48:52

Diagnostics Check	OK	13:42
Air Blank	0.000	13:43
Control Test	0.081	13:43
Air Blank	0.000	13:44
Subject Sample # 1	0.000	13:44
Breath Volume	2.480	13:44
Air Blank	0.000	13:45
Air Blank	0.000	13:47
Subject Sample # 2	0.000	13:47
Breath Volume	2.214	13:47
Air Blank	0.000	13:47
Control Test	0.081	13:48
Air Blank	0.000	13:48
Diagnostics Check	OK	13:48

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,EMILY,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: FORD,GARRETT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: CHARBONNEAU IV

DATE/TIME OF TEST: 01/11/2026 06:10:39

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.079	06:05
Air Blank	0.000	06:05
Subject Sample # 1	0.230	06:06
Breath Volume	3.769	06:06
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.244	06:09
Breath Volume	3.167	06:09
Air Blank	0.000	06:09
Control Test	0.080	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FURMAN,FREDDIE,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: FORD,GARRETT
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: PAIZ-PEREZ

DATE/TIME OF TEST: 01/11/2026 16:28:02

Diagnostics Check		OK	16:16	
Air Blank		0.000	16:17	
Control Test		0.082	16:17	
Air Blank		0.000	16:17	
Subject Sample	# 1	0.000	16:21	Volume Not Met
Breath Volume		0.343	16:21	
Air Blank		0.000	16:21	
Air Blank		0.000	16:23	
Subject Sample	# 2	0.000	16:26	Volume Not Met
Breath Volume		0.707	16:26	
Air Blank		0.000	16:27	
Control Test		0.080	16:27	
Air Blank		0.000	16:27	
Diagnostics Check		OK	16:28	

Data Download Date/Time: 01/21/2026 11:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOSTER,ALEX,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: FOSTER,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 01/23/2026 00:55:31

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:50	
Control Test	0.081	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:54	Subject Test Refused
Breath Volume	0.000	00:54	
Air Blank	0.000	00:54	
Control Test	0.080	00:54	
Air Blank	0.000	00:55	
Diagnostics Check	OK	00:55	

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,ROBERT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LABARRE,BEARY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 01/25/2026 21:55:00

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.081	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	1.007	21:53	
Air Blank	0.000	21:54	
Control Test	0.081	21:54	
Air Blank	0.000	21:54	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,ROBERT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: LABARRE,BEARY
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 01/25/2026 22:05:57

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.085	21:59
Breath Volume	1.847	21:59
Air Blank	0.000	22:00
Air Blank	0.000	22:02
Subject Sample # 2	0.084	22:04
Breath Volume	2.156	22:04
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERMAN,DEVON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SMITH,KANE
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 10:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/27/2026 12:09:14

Diagnostics Check	OK	12:03
Air Blank	0.000	12:03
Control Test	0.083	12:04
Air Blank	0.000	12:04
Subject Sample # 1	0.000	12:04
Breath Volume	1.703	12:04
Air Blank	0.000	12:05
Air Blank	0.000	12:07
Subject Sample # 2	0.000	12:07
Breath Volume	2.453	12:07
Air Blank	0.000	12:08
Control Test	0.081	12:08
Air Blank	0.000	12:09
Diagnostics Check	OK	12:09

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,ROBERT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CHESTNUT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KRETZENBACHER

DATE/TIME OF TEST: 01/29/2026 01:29:01

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.081	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:28	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,ROBERT,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CHESTNUT,CHRISTOPHER
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: KRETZENBACHER

DATE/TIME OF TEST: 01/29/2026 01:46:19

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.081	01:33	
Air Blank	0.000	01:34	
Subject Sample # 1	0.171	01:35	
Breath Volume	1.304	01:35	
Air Blank	0.000	01:36	
Air Blank	0.000	01:38	
Subject Sample # 2	0.168	01:41	Volume Not Met
Breath Volume	1.085	01:41	
Air Blank	0.000	01:41	
Air Blank	0.000	01:43	
Subject Sample # 3	0.162	01:44	
Breath Volume	1.562	01:44	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 02/12/2026 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EAKINS,SIERRA,M
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MACBETH,MATTHEW
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/01/2026 02:00:18

Diagnostics Check		OK	01:45	
Air Blank		0.000	01:46	
Control Test		0.077	01:46	
Air Blank		0.000	01:46	
Subject Sample	# 1	0.157	01:50	Volume Not Met
Breath Volume		0.273	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 2	0.155	01:52	
Breath Volume		1.812	01:52	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 3	0.157	01:58	Volume Not Met
Breath Volume		0.890	01:58	
Air Blank		0.000	01:59	
Control Test		0.073	01:59	Control Outside Tolerance
Air Blank		0.000	02:00	

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,MARC,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: TEETS,SHAWN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: NORDMAN

DATE/TIME OF TEST: 01/02/2026 05:38:58

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.077	05:27	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:30	No Sample Provided
Breath Volume	0.000	05:30	
Air Blank	0.000	05:31	
Air Blank	0.000	05:33	
Subject Sample	# 2	0.199	05:34
Breath Volume	1.496	05:34	
Air Blank	0.000	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 3	0.206	05:37
Breath Volume	2.429	05:37	
Air Blank	0.000	05:38	
Control Test	0.078	05:38	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULFORD,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: FULFORD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HENDERSHOT

DATE/TIME OF TEST: 01/06/2026 22:32:06

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.076	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.172	22:24
Breath Volume		1.863	22:24
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.195	22:27
Breath Volume		2.519	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 3	0.187	22:30
Breath Volume		1.261	22:30
Air Blank		0.000	22:31
Control Test		0.076	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MOYER,JUSTON
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GUNTER

DATE/TIME OF TEST: 01/11/2026 00:53:53

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.077	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:52	No Sample Provided
Breath Volume	0.000	00:52	
Air Blank	0.000	00:52	
Control Test	0.077	00:53	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: MOYER,JUSTON
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: CHAMPION LOWE

DATE/TIME OF TEST: 01/16/2026 04:26:32

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.077	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.237	04:17
Breath Volume		1.710	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.216	04:21
Breath Volume		1.148	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 3	0.220	04:24
Breath Volume		1.546	04:24
Air Blank		0.000	04:25
Control Test		0.077	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 01/24/2026 04:18:28

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.077	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 01/27/2026 08:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON-GARCIA,TRISTAN,A
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: JACKSON-GARCIA,TRISTAN
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 01/24/2026 04:30:22

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.077	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.069	04:25
Breath Volume	2.562	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.068	04:28
Breath Volume	2.632	04:28
Air Blank	0.000	04:29
Control Test	0.077	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 01/27/2026 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULFORD,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: FULFORD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: PEEL

DATE/TIME OF TEST: 01/30/2026 05:03:58

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.077	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.170	04:59
Breath Volume		4.019	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.174	05:02
Breath Volume		4.226	05:02
Air Blank		0.000	05:03
Control Test		0.076	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 02/05/2026 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUTHERFORD, SYDNEY, N
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SANTIAGO DIAZ, EDGARDO
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 01/26/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 222583
DRY GAS EXPIRY DATE: 06/27/2026
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: GILLAND

DATE/TIME OF TEST: 01/31/2026 04:43:22

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.077	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:42	No Sample Provided
Breath Volume	0.000	04:42	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Diagnostics Check	OK	04:43	

Data Download Date/Time: 02/05/2026 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/01/2026 23:43:52

Diagnostics Check	OK	23:41	
Air Blank	0.000	23:41	
Control Test	0.081	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1	23:42	Subject Test Refused
Breath Volume	0.000	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:43	
Air Blank	0.000	23:43	
Diagnostics Check	OK	23:43	

Data Download Date/Time: 01/07/2026 16:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRABTREE,JEREMIAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CRABTREE,JEREMIAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: GINN

DATE/TIME OF TEST: 01/03/2026 04:15:38

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:13	
Control Test	0.081	04:13	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:14	Subject Test Refused
Breath Volume	0.000	04:14	
Air Blank	0.000	04:14	
Control Test	0.081	04:15	
Air Blank	0.000	04:15	
Diagnostics Check	OK	04:15	

Data Download Date/Time: 01/07/2026 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRABTREE,JEREMIAH,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HELMS,SAMUEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PLEDGER ADAMS

DATE/TIME OF TEST: 01/03/2026 22:57:58

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.091	22:53
Breath Volume	4.328	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.084	22:56
Breath Volume	4.144	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 01/07/2026 16:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DAVEMPORT,DANIEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: SCHUMACHER

DATE/TIME OF TEST: 01/04/2026 20:18:20

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.081	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.000	20:13
Breath Volume	2.796	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.000	20:16
Breath Volume	6.171	20:16
Air Blank	0.000	20:17
Control Test	0.080	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 01/07/2026 16:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: FRASER

DATE/TIME OF TEST: 01/09/2026 20:38:49

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.081	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.000	20:34
Breath Volume	2.402	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.000	20:37
Breath Volume	2.753	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/10/2026 02:51:09

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.081	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.150	02:46
Breath Volume	2.378	02:46
Air Blank	0.000	02:46
Air Blank	0.000	02:48
Subject Sample # 2	0.148	02:49
Breath Volume	2.261	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 01/15/2026 00:26:25

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.082	00:15	
Air Blank		0.000	00:16	
Subject Sample	# 1	0.000	00:19	Volume Not Met
Breath Volume		0.714	00:19	
Air Blank		0.000	00:19	
Air Blank		0.000	00:21	
Subject Sample	# 2		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,KARSON,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACOBSON,KARSON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: PERUSSE

DATE/TIME OF TEST: 01/16/2026 23:48:37

Diagnostics Check	OK	23:42
Air Blank	0.000	23:42
Control Test	0.082	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.184	23:43
Breath Volume	1.843	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.185	23:47
Breath Volume	1.308	23:47
Air Blank	0.000	23:47
Control Test	0.080	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SWANSON,HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ARNOL

DATE/TIME OF TEST: 01/19/2026 03:11:02

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:08	
Control Test	0.081	03:08	
Air Blank	0.000	03:09	
Subject Sample	# 1	03:09	Subject Test Refused
Breath Volume	0.000	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:10	
Diagnostics Check	OK	03:11	

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/20/2026 19:25:01

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.082	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:20
Breath Volume	2.507	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.000	19:23
Breath Volume	2.031	19:23
Air Blank	0.000	19:24
Control Test	0.081	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:25

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BLEVINS,DEREK
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MAPLES

DATE/TIME OF TEST: 01/25/2026 02:54:46

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:52	
Control Test	0.080	02:52	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:53	
Control Test	0.081	02:54	
Air Blank	0.000	02:54	
Diagnostics Check	OK	02:54	

Data Download Date/Time: 02/05/2026 16:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMATEER,TRACY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PALMATEER,TRACY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:31

SUBJECT'S LAST NAME: FALCK

DATE/TIME OF TEST: 01/28/2026 16:57:49

Diagnostics Check	OK	16:51
Air Blank	0.000	16:51
Control Test	0.083	16:52
Air Blank	0.000	16:52
Subject Sample # 1	0.000	16:53
Breath Volume	1.402	16:53
Air Blank	0.000	16:54
Air Blank	0.000	16:56
Subject Sample # 2	0.000	16:56
Breath Volume	2.019	16:56
Air Blank	0.000	16:56
Control Test	0.081	16:57
Air Blank	0.000	16:57
Diagnostics Check	OK	16:57

Data Download Date/Time: 02/05/2026 16:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORRESTER, TRAVIS, D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: CARROLL, STEPHEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 01/29/2026 20:00:46

Diagnostics Check	OK	19:53
Air Blank	0.000	19:54
Control Test	0.081	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.290	19:55
Breath Volume	1.281	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.276	19:59
Breath Volume	2.285	19:59
Air Blank	0.000	19:59
Control Test	0.081	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 02/05/2026 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: WARNICK,KEATON
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/30/2026 00:32:21

Diagnostics Check	OK	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:28	
Air Blank	0.000	00:28	
Subject Sample	# 1 0.146	00:28	
Breath Volume	1.785	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:31	RFI Detect
Breath Volume	1.417	00:31	
Air Blank		00:32	RFI Detect
Air Blank		00:32	RFI Detect

Data Download Date/Time: 02/05/2026 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: WARNICK,KEATON
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403135737
DRY GAS EXPIRY DATE: 08/30/2027
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/30/2026 00:40:34

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.079	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.155	00:36
Breath Volume	2.328	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.140	00:39
Breath Volume	1.382	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 02/05/2026 16:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARNICK,KEATON,L
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: ELLRICK,JOSHUA
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/11/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MEZ PAAU

DATE/TIME OF TEST: 01/05/2026 23:45:49

Diagnostics Check	OK	23:39
Air Blank	0.000	23:40
Control Test	0.079	23:40
Air Blank	0.000	23:41
Subject Sample # 1	0.130	23:41
Breath Volume	1.726	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.127	23:44
Breath Volume	2.101	23:44
Air Blank	0.000	23:44
Control Test	0.077	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/07/2026 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SWANSON,HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PERUSSE

DATE/TIME OF TEST: 01/16/2026 23:42:35

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:37
Air Blank	0.000	23:37
Subject Sample # 1	0.154	23:38
Breath Volume	2.765	23:38
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.153	23:41
Breath Volume	3.187	23:41
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Diagnostics Check	OK	23:42

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DRAKE,KELBY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GAMBLE

DATE/TIME OF TEST: 01/17/2026 00:00:46

Diagnostics Check	OK	23:54
Air Blank	0.000	23:54
Control Test	0.079	23:55
Air Blank	0.000	23:55
Subject Sample # 1	0.230	23:56
Breath Volume	3.347	23:56
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.226	23:59
Breath Volume	3.726	23:59
Air Blank	0.000	23:59
Control Test	0.079	00:00
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 01/19/2026 22:06:48

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.080	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1	22:05	Subject Test Refused
Breath Volume	0.000	22:05	
Air Blank	0.000	22:05	
Control Test	0.080	22:06	
Air Blank	0.000	22:06	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TULLEY,DUSTIN,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: MATAACCHIERO

DATE/TIME OF TEST: 01/21/2026 12:49:16

Diagnostics Check	OK	12:42
Air Blank	0.000	12:43
Control Test	0.080	12:43
Air Blank	0.000	12:44
Subject Sample # 1	0.000	12:44
Breath Volume	3.820	12:44
Air Blank	0.000	12:45
Air Blank	0.000	12:47
Subject Sample # 2	0.000	12:47
Breath Volume	3.562	12:47
Air Blank	0.000	12:48
Control Test	0.080	12:48
Air Blank	0.000	12:49
Diagnostics Check	OK	12:49

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DRAKE,KELBY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JIMENEZ ALVAREZ

DATE/TIME OF TEST: 01/22/2026 04:08:16

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARDIF,HAMPTON,G
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TARDIF,HAMPTON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HITE

DATE/TIME OF TEST: 01/25/2026 23:07:35

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.080	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:07	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA CRUZ,JUAN,C
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DRAKE,KELBY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/31/2026 03:07:59

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.081	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.000	03:03
Breath Volume	3.085	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.000	03:06
Breath Volume	3.058	03:06
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVER,S,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/06/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: COURT ORD
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:53

SUBJECT'S LAST NAME: MATAACCHIERO

DATE/TIME OF TEST: 01/31/2026 11:19:38

Diagnostics Check		OK	11:13
Air Blank		0.000	11:14
Control Test		0.082	11:14
Air Blank		0.000	11:14
Subject Sample	# 1	0.000	11:15
Breath Volume		2.433	11:15
Air Blank		0.000	11:15
Air Blank		0.000	11:17
Subject Sample	# 2	0.000	11:18
Breath Volume		1.621	11:18
Air Blank		0.000	11:18
Control Test		0.082	11:19
Air Blank		0.000	11:19
Diagnostics Check		OK	11:19

Data Download Date/Time: 02/05/2026 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: RHODES,A
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: COSTELLO

DATE/TIME OF TEST: 01/01/2026 02:10:29

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.081	02:02	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:06	No Sample Provided
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Air Blank	0.000	02:08	
Subject Sample	# 2	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 01/30/2026 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRTRACE,JENNIFER,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: FAIRTRACE,JENNIFER
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 01/03/2026 19:51:18

Diagnostics Check	OK	19:48	
Air Blank	0.000	19:48	
Control Test	0.079	19:49	
Air Blank	0.000	19:49	
Subject Sample	# 1	19:49	Subject Test Refused
Breath Volume	0.000	19:49	
Air Blank	0.000	19:50	
Control Test	0.077	19:50	
Air Blank	0.000	19:51	
Diagnostics Check	OK	19:51	

Data Download Date/Time: 01/30/2026 18:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SWEAT

DATE/TIME OF TEST: 01/10/2026 03:08:33

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.081	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:07	Subject Test Refused
Breath Volume	0.000	03:07	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MERLINO,L
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 01/10/2026 03:36:42

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.080	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:35	Subject Test Refused
Breath Volume	0.000	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:36	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTTARI,LOGAN,H
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: MERLINO,LAWERENCE
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: C412810990810

DATE/TIME OF TEST: 01/14/2026 17:33:04

Diagnostics Check	OK	17:27
Air Blank	0.000	17:27
Control Test	0.082	17:27
Air Blank	0.000	17:28
Subject Sample # 1	0.268	17:28
Breath Volume	1.437	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.260	17:31
Breath Volume	2.914	17:31
Air Blank	0.000	17:32
Control Test	0.080	17:32
Air Blank	0.000	17:32
Diagnostics Check	OK	17:33

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNADORE,R
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CLAY

DATE/TIME OF TEST: 01/15/2026 22:05:03

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.081	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Control Test	0.080	22:04	
Air Blank	0.000	22:04	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,MISEI
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: DOTSON

DATE/TIME OF TEST: 01/16/2026 23:47:12

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.081	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.169	23:42
Breath Volume		2.242	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.165	23:45
Breath Volume		2.460	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,ALEJANDRA,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: ESQUIVEL,ALEJANDRA
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: KNOX

DATE/TIME OF TEST: 01/17/2026 00:17:30

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESQUIVEL,MISEI,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CONROY,N
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: LANGFORD

DATE/TIME OF TEST: 01/23/2026 22:56:04

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.112	22:49
Breath Volume	1.699	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.114	22:54
Breath Volume	2.304	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:56

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULL,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: HERNANDEZ,RICARDO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/20/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402886996
DRY GAS EXPIRY DATE: 11/01/2026
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: LUCAS CASTRO

DATE/TIME OF TEST: 01/24/2026 19:54:48

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.081	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.153	19:50
Breath Volume	1.824	19:50
Air Blank	0.000	19:50
Air Blank	0.000	19:52
Subject Sample # 2	0.155	19:53
Breath Volume	1.308	19:53
Air Blank	0.000	19:53
Control Test	0.081	19:54
Air Blank	0.000	19:54
Diagnostics Check	OK	19:54

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARK,A
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: WRIGHT,MARK
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:44

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/30/2026 15:07:11

Diagnostics Check	OK	15:04	
Air Blank	0.000	15:04	
Control Test	0.080	15:05	
Air Blank	0.000	15:05	
Subject Sample	# 1	15:05	Subject Test Refused
Breath Volume	0.000	15:05	
Air Blank	0.000	15:06	
Control Test	0.080	15:06	
Air Blank	0.000	15:07	
Diagnostics Check	OK	15:07	

Data Download Date/Time: 01/30/2026 18:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VARNADORE,RICHARD
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402872586
DRY GAS EXPIRY DATE: 10/12/2026
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: AIKEN

DATE/TIME OF TEST: 01/30/2026 19:00:54

Diagnostics Check	OK	18:54
Air Blank	0.000	18:54
Control Test	0.080	18:55
Air Blank	0.000	18:55
Subject Sample # 1	0.117	18:56
Breath Volume	1.644	18:56
Air Blank	0.000	18:57
Air Blank	0.000	18:58
Subject Sample # 2	0.113	18:59
Breath Volume	1.359	18:59
Air Blank	0.000	18:59
Control Test	0.079	19:00
Air Blank	0.000	19:00
Diagnostics Check	OK	19:00

Data Download Date/Time: 02/12/2026 10:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWTON,BRENNAN,P
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:42

SUBJECT'S LAST NAME: RAYMOND

DATE/TIME OF TEST: 01/02/2026 18:23:42

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.080	18:18
Air Blank		0.000	18:18
Subject Sample	# 1	0.271	18:19
Breath Volume		2.257	18:19
Air Blank		0.000	18:19
Air Blank		0.000	18:21
Subject Sample	# 2	0.262	18:22
Breath Volume		1.792	18:22
Air Blank		0.000	18:22
Control Test		0.080	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 01/30/2026 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,RODNESHIA,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHERISOL,BRUNO
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 01/08/2026 04:35:45

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.226	04:31
Breath Volume		2.332	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:33
Subject Sample	# 2	0.225	04:34
Breath Volume		1.445	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 01/30/2026 16:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUIS,RODNESHIA,R
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ISAAC

DATE/TIME OF TEST: 01/10/2026 02:11:57

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.217	02:07
Breath Volume	1.902	02:07
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.222	02:10
Breath Volume	2.191	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLENICH,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MACQUINN

DATE/TIME OF TEST: 01/11/2026 02:05:39

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.240	02:01
Breath Volume		4.445	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.242	02:03
Breath Volume		4.089	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUBITSCHEK,LEONARD,H
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: KUBITSCHEK,LEONARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BEYZER

DATE/TIME OF TEST: 01/13/2026 23:15:04

Diagnostics Check	OK	23:06
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.130	23:07
Breath Volume	1.710	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.152	23:10
Breath Volume	2.492	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 3	0.151	23:13
Breath Volume	1.804	23:13
Air Blank	0.000	23:14
Control Test	0.078	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:15

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUAGLIA,ERIC,S
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: QUAGLIA,ERIC
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: FOULK

DATE/TIME OF TEST: 01/14/2026 21:30:54

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.081	21:23
Air Blank	0.000	21:24
Subject Sample # 1	0.168	21:25
Breath Volume	3.144	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.171	21:29
Breath Volume	2.886	21:29
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHARTOFF,SCOTT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: FUKSMAN

DATE/TIME OF TEST: 01/15/2026 02:07:04

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:05	Subject Test Refused
Breath Volume	0.000	02:05	
Air Blank	0.000	02:06	
Control Test	0.081	02:06	
Air Blank	0.000	02:06	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RENDEROS,WENDY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: IANIA

DATE/TIME OF TEST: 01/23/2026 02:31:40

Diagnostics Check	OK	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:25
Subject Sample # 1	0.166	02:26
Breath Volume	2.808	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.165	02:30
Breath Volume	2.843	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Diagnostics Check	OK	02:31

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VAUGHN,RHEANNA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 01/24/2026 00:08:12

Diagnostics Check	Fail	00:07	Temperature Regulation Test
Air Blank	0.000	00:08	

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: VAUGHN,RHEANNA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 01/24/2026 00:17:25

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.080	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.186	00:12
Breath Volume		1.335	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.173	00:15
Breath Volume		1.785	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLENICH,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 01/26/2026 01:24:48

Diagnostics Check	OK	01:21	
Air Blank	0.000	01:21	
Control Test	0.080	01:22	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:24	Sequence Aborted
Breath Volume	0.000	01:24	
Air Blank	0.000	01:24	

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLENICH,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: BLACK,EDWARD
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/24/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 01/26/2026 01:51:38

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.080	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:47
Breath Volume	1.968	01:47
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.000	01:50
Breath Volume	3.699	01:50
Air Blank	0.000	01:50
Control Test	0.078	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 01/30/2026 16:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPELBAUM,KEITH,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: DAIGNEAULT,CODY
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 01/30/2026 17:04:33

Diagnostics Check	OK	16:58
Air Blank	0.000	16:59
Control Test	0.080	16:59
Air Blank	0.000	16:59
Subject Sample # 1	0.000	17:00
Breath Volume	2.246	17:00
Air Blank	0.000	17:00
Air Blank	0.000	17:02
Subject Sample # 2	0.000	17:03
Breath Volume	2.195	17:03
Air Blank	0.000	17:03
Control Test	0.080	17:03
Air Blank	0.000	17:04
Diagnostics Check	OK	17:04

Data Download Date/Time: 02/20/2026 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,FABIAN,A
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RIOS,JESSICA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: FRAZIER

DATE/TIME OF TEST: 01/31/2026 03:55:32

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.081	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.083	03:51
Breath Volume		3.636	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.080	03:54
Breath Volume		3.093	03:54
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 02/20/2026 13:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHARTOFF,SCOTT,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: CHARTOFF,SCOTT
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15924080A2
DRY GAS EXPIRY DATE: 07/05/2026
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: HARTNETT

DATE/TIME OF TEST: 01/31/2026 22:23:14

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.203	22:18
Breath Volume		3.613	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.192	22:21
Breath Volume		3.417	22:21
Air Blank		0.000	22:22
Control Test		0.076	22:22
Air Blank		0.000	22:23
Diagnostics Check		OK	22:23

Data Download Date/Time: 02/20/2026 13:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,R,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: REDDING

DATE/TIME OF TEST: 01/01/2026 03:32:24

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.084	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.000	03:27
Breath Volume	3.898	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.000	03:31
Breath Volume	4.765	03:31
Air Blank	0.000	03:31
Control Test	0.084	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANALES,ANGEL,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CIOFALO,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: IRVING

DATE/TIME OF TEST: 01/01/2026 21:01:11

Diagnostics Check	OK	20:51
Air Blank	0.000	20:51
Control Test	0.084	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.119	20:55
Breath Volume	1.156	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.119	20:59
Breath Volume	1.558	20:59
Air Blank	0.000	21:00
Control Test	0.083	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAZARES,MARGARITA,N
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAGILL,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 01/02/2026 00:10:16

Diagnostics Check	OK	00:02
Air Blank	0.000	00:03
Control Test	0.083	00:03
Air Blank	0.000	00:03
Subject Sample # 1	0.180	00:05
Breath Volume	1.867	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:07
Subject Sample # 2	0.180	00:08
Breath Volume	2.484	00:08
Air Blank	0.000	00:09
Control Test	0.082	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOXIE,B,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KEISER,T
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: BUSHEY

DATE/TIME OF TEST: 01/02/2026 14:14:06

Diagnostics Check	OK	14:09	
Air Blank	0.000	14:09	
Control Test	0.083	14:10	
Air Blank	0.000	14:10	
Subject Sample	# 1	14:12	Subject Test Refused
Breath Volume	0.523	14:12	
Air Blank	0.000	14:13	
Control Test	0.081	14:13	
Air Blank	0.000	14:14	
Diagnostics Check	OK	14:14	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 01/03/2026 02:13:05

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:07	
Control Test	0.082	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1 0.182	02:08	
Breath Volume	3.109	02:08	
Air Blank	0.000	02:09	
Air Blank	0.000	02:11	
Subject Sample	# 2	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.082	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PITTMAN,M
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/03/2026 18:36:56

Diagnostics Check		OK	18:24	
Air Blank		0.000	18:24	
Control Test		0.083	18:25	
Air Blank		0.000	18:25	
Subject Sample	# 1	0.237	18:28	Volume Not Met
Breath Volume		0.671	18:28	
Air Blank		0.000	18:29	
Air Blank		0.000	18:31	
Subject Sample	# 2	0.251	18:31	
Breath Volume		1.769	18:31	
Air Blank		0.000	18:32	
Air Blank		0.000	18:34	
Subject Sample	# 3	0.245	18:35	
Breath Volume		1.921	18:35	
Air Blank		0.000	18:36	
Control Test		0.081	18:36	
Air Blank		0.000	18:36	
Diagnostics Check		OK	18:36	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MICHAEL

DATE/TIME OF TEST: 01/04/2026 03:42:45

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:40	
Control Test	0.083	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:41	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHAMBERLAIN, J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MCLAURIN

DATE/TIME OF TEST: 01/07/2026 02:39:22

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.082	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.157	02:34
Breath Volume		1.492	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.169	02:37
Breath Volume		1.308	02:37
Air Blank		0.000	02:38
Control Test		0.082	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,OFC.
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/09/2026 02:45:16

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.083	02:34	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.128	02:38	Volume Not Met
Breath Volume		0.054	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.139	02:43	Volume Not Met
Breath Volume		0.335	02:43	
Air Blank		0.000	02:44	
Control Test		0.081	02:44	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CPL.,.
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,OFC.
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: D120012720500

DATE/TIME OF TEST: 01/09/2026 02:55:06

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:53	Subject Test Refused
Breath Volume	0.000	02:53	
Air Blank	0.000	02:54	
Control Test	0.081	02:54	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAZARES,MARGARITA,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GREEN,O
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: RIES

DATE/TIME OF TEST: 01/10/2026 22:08:33

Diagnostics Check	OK	22:01
Air Blank	0.000	22:02
Control Test	0.083	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.219	22:03
Breath Volume	2.921	22:03
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.221	22:06
Breath Volume	2.925	22:06
Air Blank	0.000	22:07
Control Test	0.082	22:07
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MAGILL,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: HUBERT

DATE/TIME OF TEST: 01/11/2026 20:03:53

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.083	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.127	19:58
Breath Volume	1.871	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.129	20:02
Breath Volume	1.828	20:02
Air Blank	0.000	20:02
Control Test	0.082	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, SOLEDAD,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: HERNANDEZ, SOLEDAD
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 01/12/2026 00:36:09

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.083	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.107	00:31
Breath Volume	1.109	00:31
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.104	00:34
Breath Volume	2.335	00:34
Air Blank	0.000	00:35
Control Test	0.083	00:35
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 01/24/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HISS,J,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: ARTMAN

DATE/TIME OF TEST: 01/13/2026 01:28:38

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.083	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.083	01:28	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WICKERT,P,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: WICKERT,P
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: TRASK

DATE/TIME OF TEST: 01/15/2026 04:36:10

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.083	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.070	04:31
Breath Volume	2.187	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.070	04:34
Breath Volume	1.851	04:34
Air Blank	0.000	04:35
Control Test	0.082	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DOMINGUEZ, ZACHARY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ZACEK

DATE/TIME OF TEST: 01/16/2026 21:33:23

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.083	21:21	
Air Blank	0.000	21:22	
Subject Sample # 1	0.197	21:23	
Breath Volume	1.234	21:23	
Air Blank	0.000	21:23	
Air Blank	0.000	21:25	
Subject Sample # 2	0.182	21:28	Volume Not Met
Breath Volume	0.148	21:28	
Air Blank	0.000	21:29	
Air Blank	0.000	21:31	
Subject Sample # 3	0.187	21:31	
Breath Volume	1.753	21:31	
Air Blank	0.000	21:32	
Control Test	0.080	21:32	
Air Blank	0.000	21:33	
Diagnostics Check	OK	21:33	

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EDGAR,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LANDES

DATE/TIME OF TEST: 01/17/2026 02:58:07

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.083	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.248	02:53
Breath Volume		3.667	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:55
Subject Sample	# 2	0.241	02:56
Breath Volume		3.562	02:56
Air Blank		0.000	02:57
Control Test		0.082	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,THOMAS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 01/17/2026 16:26:47

Diagnostics Check	OK	16:20
Air Blank	0.000	16:20
Control Test	0.083	16:21
Air Blank	0.000	16:21
Subject Sample # 1	0.105	16:22
Breath Volume	2.207	16:22
Air Blank	0.000	16:23
Air Blank	0.000	16:24
Subject Sample # 2	0.104	16:25
Breath Volume	1.667	16:25
Air Blank	0.000	16:25
Control Test	0.082	16:26
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HACKETT,B
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: JACQUIN

DATE/TIME OF TEST: 01/19/2026 01:32:42

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.083	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.119	01:27
Breath Volume		1.562	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.118	01:31
Breath Volume		1.402	01:31
Air Blank		0.000	01:31
Control Test		0.082	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDOLPH,B,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/19/2026 08:04:55

Diagnostics Check	OK	08:02	
Air Blank	0.000	08:02	
Control Test	0.084	08:02	
Air Blank	0.000	08:03	
Subject Sample	# 1	08:03	Subject Test Refused
Breath Volume	0.000	08:03	
Air Blank	0.000	08:04	
Control Test	0.083	08:04	
Air Blank	0.000	08:04	
Diagnostics Check	OK	08:04	

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHAMBERLAIN, J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VENNARO

DATE/TIME OF TEST: 01/21/2026 00:07:23

Diagnostics Check		OK	00:00
Air Blank		0.000	00:00
Control Test		0.083	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		1.714	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.000	00:06
Breath Volume		1.988	00:06
Air Blank		0.000	00:06
Control Test		0.083	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GOODRICH,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/27/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: EDMOND

DATE/TIME OF TEST: 01/23/2026 22:15:08

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:01	
Control Test		0.083	22:01	
Air Blank		0.000	22:02	
Subject Sample	# 1	0.135	22:05	Volume Not Met
Breath Volume		0.406	22:05	
Air Blank		0.000	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 2	0.142	22:09	
Breath Volume		1.718	22:09	
Air Blank		0.000	22:10	
Air Blank		0.000	22:11	
Subject Sample	# 3	0.137	22:13	
Breath Volume		1.496	22:13	
Air Blank		0.000	22:14	
Control Test		0.081	22:14	
Air Blank		0.000	22:15	
Diagnostics Check		OK	22:15	

Data Download Date/Time: 01/24/2026 15:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAZARES,MARGARITA,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BARTON,A
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 01/25/2026 02:30:34

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.082	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1 0.207	02:19	
Breath Volume	1.257	02:19	
Air Blank	0.000	02:20	
Air Blank	0.000	02:21	
Subject Sample	# 2 0.203	02:25	Volume Not Met
Breath Volume	0.089	02:25	
Air Blank	0.000	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 3	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.079	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MANGEL, DEPUTY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CASTELLANO

DATE/TIME OF TEST: 01/25/2026 23:06:19

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.082	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.083	23:01
Breath Volume	1.500	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.084	23:04
Breath Volume	1.398	23:04
Air Blank	0.000	23:05
Control Test	0.082	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THOMAS, R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: EXATUS

DATE/TIME OF TEST: 01/26/2026 01:24:56

Diagnostics Check	OK	01:24	
Air Blank		01:24	RFI Detect
Air Blank	0.000	01:24	

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: THOMAS, R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: EXATUS

DATE/TIME OF TEST: 01/26/2026 01:38:16

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:35	
Control Test	0.082	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.082	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANE,JAMES,G
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROLLANDO,T
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 14:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:20

SUBJECT'S LAST NAME: PALLATH

DATE/TIME OF TEST: 01/26/2026 14:59:16

Diagnostics Check		OK	14:52
Air Blank		0.000	14:52
Control Test		0.083	14:53
Air Blank		0.000	14:53
Subject Sample	# 1	0.211	14:54
Breath Volume		1.992	14:54
Air Blank		0.000	14:55
Air Blank		0.000	14:56
Subject Sample	# 2	0.201	14:57
Breath Volume		2.148	14:57
Air Blank		0.000	14:58
Control Test		0.082	14:58
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLESTER, ANDY
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: JUSTI

DATE/TIME OF TEST: 01/28/2026 23:54:30

Diagnostics Check	OK	23:44	
Air Blank	0.000	23:45	
Control Test	0.084	23:45	
Air Blank	0.000	23:45	
Subject Sample	# 1	23:48	No Sample Provided
Breath Volume	0.000	23:48	
Air Blank	0.000	23:49	
Air Blank	0.000	23:51	
Subject Sample	# 2	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.085	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,C
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: GRANLUND

DATE/TIME OF TEST: 01/30/2026 21:32:39

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.083	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.225	21:27
Breath Volume	1.835	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.224	21:30
Breath Volume	2.019	21:30
Air Blank	0.000	21:31
Control Test	0.082	21:32
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CSO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLOMON,F,L
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: KRAMER,K
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/24/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403141809
DRY GAS EXPIRY DATE: 09/10/2027
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: DAGOSTINO

DATE/TIME OF TEST: 01/30/2026 23:38:00

Diagnostics Check	OK	23:31
Air Blank	0.000	23:32
Control Test	0.082	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.247	23:33
Breath Volume	2.359	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.230	23:36
Breath Volume	2.000	23:36
Air Blank	0.000	23:37
Control Test	0.082	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:38

Data Download Date/Time: 02/22/2026 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEST,ALISHA,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: WEST,ALISHA

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04125080A5

DRY GAS EXPIRY DATE: 03/05/2027

VIOLATION CODE: DUI

ARREST TIME: 20:25

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 01/01/2026 21:15:30

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.082	21:08
Air Blank	0.000	21:09
Subject Sample # 1	0.213	21:09
Breath Volume	1.566	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.205	21:13
Breath Volume	1.621	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:15
Diagnostics Check	OK	21:15

Data Download Date/Time: 01/28/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/07/2026 03:31:58

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.060	03:26
Breath Volume	1.613	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.061	03:30
Breath Volume	2.632	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 01/28/2026 09:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAWGER,JACOB,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SHAWGER,JACOB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: LEVIN-EPSTEIN

DATE/TIME OF TEST: 01/14/2026 07:26:23

Diagnostics Check	OK	07:19
Air Blank	0.000	07:20
Control Test	0.081	07:20
Air Blank	0.000	07:21
Subject Sample # 1	0.135	07:21
Breath Volume	1.906	07:21
Air Blank	0.000	07:22
Air Blank	0.000	07:24
Subject Sample # 2	0.131	07:24
Breath Volume	1.667	07:24
Air Blank	0.000	07:25
Control Test	0.080	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 01/28/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SPOTO, MATTHEW, R

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SPOTO, MATTHEW

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No

DRY GAS STANDARD LOT #: 04125080A5

DRY GAS EXPIRY DATE: 03/05/2027

VIOLATION CODE: DUI

ARREST TIME: 06:20

GENDER: Male

OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: THAMES

DATE/TIME OF TEST: 01/16/2026 09:04:16

Diagnostics Check		OK	08:57
Air Blank		0.000	08:57
Control Test		0.082	08:58
Air Blank		0.000	08:58
Subject Sample	# 1	0.214	08:59
Breath Volume		3.183	08:59
Air Blank		0.000	09:00
Air Blank		0.000	09:02
Subject Sample	# 2	0.200	09:02
Breath Volume		1.664	09:02
Air Blank		0.000	09:03
Control Test		0.082	09:03
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 01/28/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAWGER,JACOB,L
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: WILSON,KALEB
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 12/17/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: HERENSTEIN

DATE/TIME OF TEST: 01/28/2026 03:33:13

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.020	03:28
Breath Volume	1.933	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.019	03:31
Breath Volume	1.828	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 01/28/2026 09:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAYNE,ABIGAIL,D
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: PAYNE,ABIGAIL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PLOSS

DATE/TIME OF TEST: 01/31/2026 03:28:35

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.082	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.035	03:23
Breath Volume	1.808	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.035	03:27
Breath Volume	1.984	03:27
Air Blank	0.000	03:27
Control Test	0.081	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 02/25/2026 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: COOK, BRENDAN
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MESSINA

DATE/TIME OF TEST: 01/01/2026 02:45:40

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.081	02:41	
Air Blank	0.000	02:42	
Subject Sample	# 1	02:45	Sequence Aborted
Breath Volume	1.628	02:45	
Air Blank	0.000	02:45	

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: COOK, BRENDAN
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: MESSINA

DATE/TIME OF TEST: 01/01/2026 02:55:12

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.161	02:50
Breath Volume		1.253	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.164	02:53
Breath Volume		1.800	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MADDEN

DATE/TIME OF TEST: 01/01/2026 05:29:17

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.080	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.250	05:24
Breath Volume	2.156	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.237	05:27
Breath Volume	1.773	05:27
Air Blank	0.000	05:28
Control Test	0.080	05:28
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RICHARDSON,RACHAEL
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: RUFFIN

DATE/TIME OF TEST: 01/01/2026 06:36:46

Diagnostics Check	OK	06:29
Air Blank	0.000	06:29
Control Test	0.080	06:29
Air Blank	0.000	06:30
Subject Sample # 1	0.184	06:32
Breath Volume	1.953	06:32
Air Blank	0.000	06:32
Air Blank	0.000	06:34
Subject Sample # 2	0.182	06:35
Breath Volume	2.347	06:35
Air Blank	0.000	06:35
Control Test	0.080	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:36

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ II,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ II,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GONZALEZ MUNOZ

DATE/TIME OF TEST: 01/04/2026 05:00:04

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.188	04:55
Breath Volume	1.277	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.185	04:58
Breath Volume	1.867	04:58
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: CRUZ ASTUDILLO

DATE/TIME OF TEST: 01/08/2026 01:17:05

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.149	01:12
Breath Volume	2.871	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.147	01:15
Breath Volume	1.898	01:15
Air Blank	0.000	01:16
Control Test	0.080	01:16
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 01/09/2026 00:53:43

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Subject Sample # 1	0.075	00:49
Breath Volume	1.269	00:49
Air Blank	0.000	00:49
Air Blank	0.000	00:51
Subject Sample # 2	0.079	00:52
Breath Volume	1.386	00:52
Air Blank	0.000	00:52
Control Test	0.079	00:53
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/09/2026 23:45:51

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.079	23:39
Air Blank	0.000	23:39
Subject Sample # 1	0.205	23:41
Breath Volume	1.953	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:43
Subject Sample # 2	0.214	23:44
Breath Volume	2.027	23:44
Air Blank	0.000	23:44
Control Test	0.081	23:45
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SPINELLI

DATE/TIME OF TEST: 01/10/2026 04:16:31

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.119	04:12
Breath Volume	1.832	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.108	04:15
Breath Volume	1.667	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LUCE,CARRIE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: PETERSEN

DATE/TIME OF TEST: 01/10/2026 15:07:59

Diagnostics Check	OK	15:01
Air Blank	0.000	15:02
Control Test	0.080	15:02
Air Blank	0.000	15:02
Subject Sample # 1	0.132	15:03
Breath Volume	1.972	15:03
Air Blank	0.000	15:04
Air Blank	0.000	15:05
Subject Sample # 2	0.132	15:06
Breath Volume	2.023	15:06
Air Blank	0.000	15:07
Control Test	0.080	15:07
Air Blank	0.000	15:07
Diagnostics Check	OK	15:07

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BATYCKI

DATE/TIME OF TEST: 01/11/2026 03:14:15

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.176	03:09
Breath Volume		1.996	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.187	03:12
Breath Volume		1.492	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNISTON,G
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HALE

DATE/TIME OF TEST: 01/11/2026 03:44:56

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.183	03:40
Breath Volume		3.437	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:42
Subject Sample	# 2	0.186	03:43
Breath Volume		3.796	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUTKO,MARIAM,L
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERGSTAD,BRANDON
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: LEITZ

DATE/TIME OF TEST: 01/11/2026 04:11:59

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.188	04:07
Breath Volume	1.800	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.190	04:10
Breath Volume	1.152	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETROWSKI,DAVID,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/12/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: PETTY

DATE/TIME OF TEST: 01/11/2026 19:04:30

Diagnostics Check	OK	18:57
Air Blank	0.000	18:57
Control Test	0.079	18:58
Air Blank	0.000	18:58
Subject Sample # 1	0.224	18:59
Breath Volume	1.535	18:59
Air Blank	0.000	18:59
Air Blank	0.000	19:01
Subject Sample # 2	0.220	19:03
Breath Volume	1.664	19:03
Air Blank	0.000	19:03
Control Test	0.076	19:03
Air Blank	0.000	19:04
Diagnostics Check	OK	19:04

Data Download Date/Time: 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 01/12/2026 01:24:43

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.280	01:16
Breath Volume		3.195	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.255	01:20
Breath Volume		2.164	01:20
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 3	0.255	01:23
Breath Volume		2.019	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ II,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MARTINEZ II,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: LOPEZ VASQUEZ

DATE/TIME OF TEST: 01/12/2026 20:14:10

Diagnostics Check	OK	20:08
Air Blank	0.000	20:08
Control Test	0.080	20:08
Air Blank	0.000	20:09
Subject Sample # 1	0.262	20:09
Breath Volume	2.972	20:09
Air Blank	0.000	20:10
Air Blank	0.000	20:12
Subject Sample # 2	0.249	20:12
Breath Volume	2.640	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:14
Diagnostics Check	OK	20:14

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,JARED
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: OMORODION

DATE/TIME OF TEST: 01/13/2026 01:06:10

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.081	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.000	01:01
Breath Volume	2.390	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.000	01:04
Breath Volume	2.667	01:04
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, BLAKE, C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS, BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BABYCH

DATE/TIME OF TEST: 01/13/2026 22:29:26

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.159	22:24
Breath Volume	1.585	22:24
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.156	22:27
Breath Volume	2.000	22:27
Air Blank	0.000	22:28
Control Test	0.081	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: BERGSTAD, BRANDON
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: WARENDA

DATE/TIME OF TEST: 01/14/2026 23:34:35

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.248	23:29
Breath Volume	2.292	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:31
Subject Sample # 2	0.247	23:33
Breath Volume	1.628	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,BLAKE,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: GARBER

DATE/TIME OF TEST: 01/16/2026 18:58:52

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.081	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.160	18:54
Breath Volume	1.339	18:54
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.163	18:57
Breath Volume	2.089	18:57
Air Blank	0.000	18:57
Control Test	0.080	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, BLAKE, C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS, BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WECKESSER

DATE/TIME OF TEST: 01/17/2026 02:08:01

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.081	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.180	02:03
Breath Volume	1.554	02:03
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.163	02:06
Breath Volume	1.882	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:08

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTINEZ II,OFFICER,
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: GOSVENER,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: SULT

DATE/TIME OF TEST: 01/17/2026 04:56:57

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.082	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.152	04:52
Breath Volume		1.859	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:54
Subject Sample	# 2	0.162	04:55
Breath Volume		1.902	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS,BLAKE,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS,BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 01/17/2026 21:48:36

Diagnostics Check	OK	21:41
Air Blank	0.000	21:42
Control Test	0.082	21:42
Air Blank	0.000	21:43
Subject Sample # 1	0.000	21:44
Breath Volume	2.691	21:44
Air Blank	0.000	21:44
Air Blank	0.000	21:46
Subject Sample # 2	0.000	21:47
Breath Volume	3.253	21:47
Air Blank	0.000	21:47
Control Test	0.082	21:48
Air Blank	0.000	21:48
Diagnostics Check	OK	21:48

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,JARED
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MORRISON JR

DATE/TIME OF TEST: 01/18/2026 03:08:14

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.081	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.164	03:03
Breath Volume		2.402	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.159	03:06
Breath Volume		2.554	03:06
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MORIARTY,MICHAEL
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LUNN

DATE/TIME OF TEST: 01/18/2026 03:27:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.230	03:23
Breath Volume	2.800	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:25
Subject Sample # 2	0.237	03:26
Breath Volume	2.753	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERSON,CORY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PIERSON,CORY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 01/18/2026 15:44:22

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.081	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.286	15:39
Breath Volume	2.527	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:42
Subject Sample # 2	0.290	15:42
Breath Volume	2.164	15:42
Air Blank	0.000	15:43
Control Test	0.080	15:43
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 02/11/2026 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLINS,JAMIE,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: RICHARDSON,RACHEL
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: CYR

DATE/TIME OF TEST: 01/18/2026 23:53:19

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.079	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 1	0.228	23:44	Volume Not Met
Breath Volume		0.703	23:44	
Air Blank		0.000	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 2	0.230	23:48	
Breath Volume		2.382	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 3	0.240	23:51	
Breath Volume		1.957	23:51	
Air Blank		0.000	23:52	
Control Test		0.080	23:52	
Air Blank		0.000	23:53	
Diagnostics Check		OK	23:53	

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HENKEL

DATE/TIME OF TEST: 01/19/2026 02:56:28

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.081	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.205	02:50
Breath Volume		1.226	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.207	02:54
Breath Volume		2.121	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERRIMAN,PAUL,I
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SOSA FERNANDEZ,NARCISO
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/19/2026 07:29:06

Diagnostics Check	OK	07:22
Air Blank	0.000	07:23
Control Test	0.082	07:23
Air Blank	0.000	07:23
Subject Sample # 1	0.214	07:24
Breath Volume	2.855	07:24
Air Blank	0.000	07:25
Air Blank	0.000	07:27
Subject Sample # 2	0.215	07:27
Breath Volume	2.382	07:27
Air Blank	0.000	07:28
Control Test	0.082	07:28
Air Blank	0.000	07:29
Diagnostics Check	OK	07:29

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETROWSKI,DAVID,S
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PETROWSKI,DAVID
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/19/2026 20:28:53

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:16	
Subject Sample	# 1	0.165	20:20	Volume Not Met
Breath Volume		0.402	20:20	
Air Blank		0.000	20:20	
Air Blank		0.000	20:22	
Subject Sample	# 2	0.199	20:23	
Breath Volume		1.187	20:23	
Air Blank		0.000	20:24	
Air Blank		0.000	20:26	
Subject Sample	# 3	0.195	20:27	
Breath Volume		1.890	20:27	
Air Blank		0.000	20:27	
Control Test		0.080	20:28	
Air Blank		0.000	20:28	
Diagnostics Check		OK	20:28	

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,DEVIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BERRY,DEVIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: POMERANTZ

DATE/TIME OF TEST: 01/19/2026 21:17:54

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.238	21:13
Breath Volume		2.273	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.235	21:16
Breath Volume		2.097	21:16
Air Blank		0.000	21:16
Control Test		0.080	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KEUSCH,KYLE
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SEE

DATE/TIME OF TEST: 01/20/2026 01:55:32

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.082	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.200	01:51
Breath Volume	1.851	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.209	01:54
Breath Volume	1.902	01:54
Air Blank	0.000	01:54
Control Test	0.081	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEATY,JARED,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BEATY,JARED
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: GIOCONDA

DATE/TIME OF TEST: 01/21/2026 23:34:56

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.081	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.147	23:29
Breath Volume	1.566	23:29
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.136	23:33
Breath Volume	2.328	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,DANIEL,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DAVENPORT,DANIEL
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 01/22/2026 02:33:11

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.000	02:28
Breath Volume	1.878	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.000	02:31
Breath Volume	2.757	02:31
Air Blank	0.000	02:32
Control Test	0.081	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: MAKI, CHRISTOPHER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: MESHERYAKOV

DATE/TIME OF TEST: 01/23/2026 23:04:55

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.081	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.151	23:00
Breath Volume		2.496	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.146	23:03
Breath Volume		1.394	23:03
Air Blank		0.000	23:04
Control Test		0.081	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIPE, RAYCHEL, E
BREATH TEST OPERATOR AGENCY: Sarasota PD
ARRESTING OFFICER: KENNISTON, GABRIELLE
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 01/23/2026 23:44:36

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.153	23:39
Breath Volume	2.285	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.147	23:43
Breath Volume	1.433	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:44
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CRAWFORD,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: EASTEP

DATE/TIME OF TEST: 01/24/2026 23:41:01

Diagnostics Check	OK	23:34
Air Blank	0.000	23:35
Control Test	0.081	23:35
Air Blank	0.000	23:35
Subject Sample # 1	0.191	23:36
Breath Volume	4.156	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:38
Subject Sample # 2	0.190	23:39
Breath Volume	3.824	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:41

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: FERRANTE

DATE/TIME OF TEST: 01/25/2026 02:25:17

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.081	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.214	02:20
Breath Volume	1.335	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.223	02:23
Breath Volume	1.152	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 02/11/2026 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LERNER

DATE/TIME OF TEST: 01/25/2026 04:29:55

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.081	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.060	04:20	Volume Not Met
Breath Volume		0.019	04:20	
Air Blank		0.000	04:20	
Air Blank		0.000	04:22	
Subject Sample	# 2	0.067	04:25	
Breath Volume		1.187	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 3		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LERNER

DATE/TIME OF TEST: 01/25/2026 04:43:20

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.059	04:38
Breath Volume	1.902	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.059	04:41
Breath Volume	1.761	04:41
Air Blank	0.000	04:42
Control Test	0.081	04:42
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MCMILLIN

DATE/TIME OF TEST: 01/25/2026 22:41:49

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.081	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.120	22:37
Breath Volume	1.906	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.130	22:40
Breath Volume	2.906	22:40
Air Blank	0.000	22:40
Control Test	0.082	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CRAWFORD,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ENOS

DATE/TIME OF TEST: 01/26/2026 01:04:29

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.076	00:59
Breath Volume		1.898	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.078	01:03
Breath Volume		1.871	01:03
Air Blank		0.000	01:03
Control Test		0.081	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTOS,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SANTOS,JOSHUA
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: BLAKENBILLER

DATE/TIME OF TEST: 01/26/2026 18:57:20

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:51
Subject Sample # 1	0.177	18:52
Breath Volume	1.851	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.175	18:55
Breath Volume	1.718	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, BLAKE, C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS, BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 01/26/2026 19:43:16

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.081	19:37
Air Blank	0.000	19:37
Subject Sample # 1	0.234	19:38
Breath Volume	2.046	19:38
Air Blank	0.000	19:39
Air Blank	0.000	19:40
Subject Sample # 2	0.222	19:41
Breath Volume	1.460	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Diagnostics Check	OK	19:43

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, BLAKE, C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS, BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15025080A1
DRY GAS EXPIRY DATE: 07/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MORLEY

DATE/TIME OF TEST: 01/27/2026 01:37:03

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.082	01:31
Air Blank	0.000	01:31
Subject Sample # 1	0.214	01:32
Breath Volume	1.281	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.211	01:35
Breath Volume	1.554	01:35
Air Blank	0.000	01:36
Control Test	0.081	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:37

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: TALLEY

DATE/TIME OF TEST: 01/28/2026 20:54:22

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.081	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.189	20:49
Breath Volume		2.417	20:49
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.194	20:52
Breath Volume		2.343	20:52
Air Blank		0.000	20:53
Control Test		0.082	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SALINAS BURRIOS

DATE/TIME OF TEST: 01/29/2026 01:17:12

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:03	
Control Test		0.082	01:04	
Air Blank		0.000	01:04	
Subject Sample	# 1	0.154	01:07	Volume Not Met
Breath Volume		1.089	01:07	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 2	0.146	01:12	
Breath Volume		1.238	01:12	
Air Blank		0.000	01:13	
Air Blank		0.000	01:14	
Subject Sample	# 3	0.153	01:15	
Breath Volume		1.472	01:15	
Air Blank		0.000	01:16	
Control Test		0.079	01:16	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:17	

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRENCKLE,TIMOTHY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GHALIEH

DATE/TIME OF TEST: 01/30/2026 01:16:30

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.000	01:12
Breath Volume	3.367	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.000	01:15
Breath Volume	3.468	01:15
Air Blank	0.000	01:15
Control Test	0.082	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DUSTIN,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,DUSTIN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BLANCH

DATE/TIME OF TEST: 01/30/2026 04:11:25

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.283	04:06	Volume Not Met
Breath Volume	0.882	04:06	
Air Blank	0.000	04:07	
Air Blank	0.000	04:09	
Subject Sample	# 2	04:10	Subject Test Refused
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Control Test	0.078	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENCKLE,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CODINA,ALDEN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: YODER

DATE/TIME OF TEST: 01/30/2026 04:29:18

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.117	04:24
Breath Volume	2.257	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.113	04:27
Breath Volume	3.085	04:27
Air Blank	0.000	04:28
Control Test	0.081	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENTS, BLAKE, C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CLEMENTS, BLAKE
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/11/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 10124080A2
DRY GAS EXPIRY DATE: 06/05/2026
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/31/2026 04:39:11

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.201	04:34
Breath Volume		2.628	04:34
Air Blank		0.000	04:35
Air Blank		0.000	04:37
Subject Sample	# 2	0.195	04:37
Breath Volume		2.777	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:39
Diagnostics Check		OK	04:39

Data Download Date/Time: 02/11/2026 08:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MORKAN

DATE/TIME OF TEST: 01/11/2026 21:51:04

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Subject Sample	# 1	21:50	Sequence Aborted
Breath Volume	0.000	21:50	
Air Blank	0.000	21:51	

Data Download Date/Time: 02/02/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CALDARELLI

DATE/TIME OF TEST: 01/18/2026 02:23:21

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.073	02:22	Control Outside Tolerance
Air Blank	0.000	02:23	

Data Download Date/Time: 02/02/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CALDARELLI

DATE/TIME OF TEST: 01/18/2026 02:33:28

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.075	02:26	
Air Blank	0.000	02:27	
Subject Sample # 1	0.169	02:29	
Breath Volume	1.375	02:29	
Air Blank	0.000	02:29	
Air Blank	0.000	02:31	
Subject Sample # 2	0.184	02:31	
Breath Volume	1.773	02:31	
Air Blank	0.000	02:32	
Control Test	0.074	02:32	Control Outside Tolerance
Air Blank	0.000	02:33	

Data Download Date/Time: 02/02/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WIGGINS, TYLER, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WIGGINS, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/31/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 403207507
DRY GAS EXPIRY DATE: 11/26/2027
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: CALDARELLI

DATE/TIME OF TEST: 01/18/2026 02:44:16

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.076	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.173	02:38
Breath Volume		2.207	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.169	02:42
Breath Volume		2.023	02:42
Air Blank		0.000	02:43
Control Test		0.075	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/02/2026 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JORDAN,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GONZALEZ,JORDAN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: MANSSUR-MURPHY

DATE/TIME OF TEST: 01/10/2026 04:25:36

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.078	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.267	04:20
Breath Volume		1.152	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.257	04:23
Breath Volume		2.613	04:23
Air Blank		0.000	04:24
Control Test		0.076	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 01/28/2026 20:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADER,JAMES,D
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MURRIELLO,MATTHEW
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 01/14/2026 02:04:57

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.208	02:00
Breath Volume		3.058	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.214	02:03
Breath Volume		3.558	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 01/28/2026 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,JASON,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: PEREZ,MIGUEL
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/22/2026 20:38:48

Diagnostics Check	OK	20:32
Air Blank	0.000	20:32
Control Test	0.079	20:33
Air Blank	0.000	20:33
Subject Sample # 1	0.211	20:34
Breath Volume	3.605	20:34
Air Blank	0.000	20:34
Air Blank	0.000	20:36
Subject Sample # 2	0.211	20:37
Breath Volume	3.308	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 01/28/2026 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTOX,STERLING,A
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KONSTANTIS,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/29/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 01/24/2026 22:58:24

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.196	22:53
Breath Volume	5.062	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.196	22:56
Breath Volume	5.625	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 01/28/2026 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JORDAN,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KONSTANTIS,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 01/29/2026 19:32:51

Diagnostics Check	OK	19:32	
Air Blank		19:32	RFI Detect
Air Blank	0.000	19:32	

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JORDAN,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KONSTANTIS,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 01/29/2026 19:38:35

Diagnostics Check	OK	19:37	
Air Blank		19:38	RFI Detect
Air Blank		19:38	RFI Detect

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JORDAN,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KONSTANTIS,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 01/29/2026 19:48:08

Diagnostics Check	OK	19:41
Air Blank	0.000	19:42
Control Test	0.080	19:42
Air Blank	0.000	19:42
Subject Sample # 1	0.173	19:43
Breath Volume	3.277	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:45
Subject Sample # 2	0.188	19:46
Breath Volume	3.445	19:46
Air Blank	0.000	19:47
Control Test	0.079	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JORDAN,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: KONSTANTIS,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/30/2026 02:33:17

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.080	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.130	02:28
Breath Volume	2.089	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.137	02:31
Breath Volume	1.539	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MCCAY,B
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: 663216000

DATE/TIME OF TEST: 01/31/2026 03:28:11

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.081	03:27	
Air Blank		03:28	RFI Detect
Air Blank		03:28	RFI Detect

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,D,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MCCAY,BRITTANY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/28/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19825080A1
DRY GAS EXPIRY DATE: 09/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HOLLINGSWORTH

DATE/TIME OF TEST: 01/31/2026 03:38:23

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:37	Subject Test Refused
Breath Volume	0.000	03:37	
Air Blank	0.000	03:37	
Control Test	0.080	03:37	
Air Blank	0.000	03:38	
Diagnostics Check	OK	03:38	

Data Download Date/Time: 02/25/2026 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LIEURANCE,COURTNEY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BONJOC

DATE/TIME OF TEST: 01/01/2026 02:58:37

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:56	RFI Detect
Breath Volume	0.847	02:56	
Air Blank	0.000	02:57	
Control Test	0.078	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 01/21/2026 13:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: LIEURANCE,COURTNEY
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BONJOC

DATE/TIME OF TEST: 01/01/2026 03:09:40

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.077	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.243	03:04
Breath Volume		1.757	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.237	03:08
Breath Volume		1.859	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/21/2026 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ABRAHAM,MARLO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PLANGE

DATE/TIME OF TEST: 01/02/2026 04:26:32

Diagnostics Check	OK	04:18	
Air Blank	0.000	04:18	
Control Test	0.078	04:19	
Air Blank	0.000	04:19	
Subject Sample # 1	0.181	04:21	
Breath Volume	2.671	04:21	
Air Blank	0.000	04:22	
Air Blank	0.000	04:24	
Subject Sample # 2	0.179	04:25	
Breath Volume	3.828	04:25	
Air Blank		04:26	RFI Detect
Air Blank	0.000	04:26	

Data Download Date/Time: 01/21/2026 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ABRAHAM,MARLO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RUACHO

DATE/TIME OF TEST: 01/10/2026 22:40:14

Diagnostics Check	OK	22:39	
Air Blank		22:40	RFI Detect
Air Blank		22:40	RFI Detect

Data Download Date/Time: 01/21/2026 13:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TYNDAL,ERICK,D
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: ABRAHAM,MARLO
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13225080A2
DRY GAS EXPIRY DATE: 06/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RUACHO

DATE/TIME OF TEST: 01/10/2026 22:49:32

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Subject Sample # 1	0.088	22:45
Breath Volume	3.503	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.072	22:48
Breath Volume	2.402	22:48
Air Blank	0.000	22:48
Control Test	0.078	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 01/21/2026 13:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEA GALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/03/2026 04:45:17

Diagnostics Check	OK	04:37
Air Blank	0.000	04:38
Control Test	0.080	04:38
Air Blank	0.000	04:39
Subject Sample # 1	0.092	04:40
Breath Volume	1.203	04:40
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.088	04:43
Breath Volume	1.132	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: KOHLYDARROCH

DATE/TIME OF TEST: 01/06/2026 18:38:10

Diagnostics Check	OK	18:32
Air Blank	0.000	18:32
Control Test	0.080	18:33
Air Blank	0.000	18:33
Subject Sample # 1	0.000	18:33
Breath Volume	2.585	18:33
Air Blank	0.000	18:34
Air Blank	0.000	18:36
Subject Sample # 2	0.000	18:36
Breath Volume	1.914	18:36
Air Blank	0.000	18:37
Control Test	0.080	18:37
Air Blank	0.000	18:38
Diagnostics Check	OK	18:38

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: JEANBAPTISTE

DATE/TIME OF TEST: 01/07/2026 21:29:58

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.079	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.000	21:25
Breath Volume	3.816	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.000	21:28
Breath Volume	4.097	21:28
Air Blank	0.000	21:29
Control Test	0.079	21:29
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NOVAK, TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: MENENDEZ COLLAZO

DATE/TIME OF TEST: 01/09/2026 02:01:24

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.067	01:57
Breath Volume	3.152	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.070	01:59
Breath Volume	3.332	01:59
Air Blank	0.000	02:00
Control Test	0.080	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUIZ,ANDRES
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LEGGS

DATE/TIME OF TEST: 01/09/2026 03:39:18

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.079	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.168	03:34
Breath Volume		2.117	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.170	03:37
Breath Volume		2.105	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 01/10/2026 03:12:32

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.133	03:08
Breath Volume	1.808	03:08
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.138	03:11
Breath Volume	1.652	03:11
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARC,VUKANSON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 01/10/2026 05:35:06

Diagnostics Check	OK	05:28
Air Blank	0.000	05:29
Control Test	0.079	05:29
Air Blank	0.000	05:30
Subject Sample # 1	0.116	05:30
Breath Volume	1.253	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.120	05:33
Breath Volume	1.515	05:33
Air Blank	0.000	05:34
Control Test	0.079	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FREY,ERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BLOOMER

DATE/TIME OF TEST: 01/13/2026 23:34:30

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.242	23:30
Breath Volume	2.070	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.248	23:32
Breath Volume	2.093	23:32
Air Blank	0.000	23:33
Control Test	0.079	23:33
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALL,KRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/15/2026 04:07:57

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.137	04:03
Breath Volume		1.207	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.139	04:06
Breath Volume		1.398	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ MARTINEZ,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ORTIZ RODRIGUEZ

DATE/TIME OF TEST: 01/16/2026 03:49:23

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.169	03:42	Volume Not Met
Breath Volume		1.003	03:42	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.141	03:47	Volume Not Met
Breath Volume		0.777	03:47	
Air Blank		0.000	03:48	
Control Test		0.079	03:48	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 01/19/2026 10:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ MARTINEZ,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ORTIZ RODRIGUEZ

DATE/TIME OF TEST: 01/16/2026 03:59:39

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.185	03:54
Breath Volume	1.636	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.171	03:58
Breath Volume	1.156	03:58
Air Blank	0.000	03:58
Control Test	0.077	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 01/19/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: ISAACSON

DATE/TIME OF TEST: 01/20/2026 23:56:06

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.195	23:51
Breath Volume	1.242	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.189	23:54
Breath Volume	1.636	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOWRONSKI,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,BENJAMIN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: CLYBURN

DATE/TIME OF TEST: 01/24/2026 04:10:19

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.079	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.117	04:03	Volume Not Met
Breath Volume		0.593	04:03	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.108	04:08	Volume Not Met
Breath Volume		0.976	04:08	
Air Blank		0.000	04:09	
Control Test		0.079	04:09	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZAMORA,LEONARDO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: PIERRELOUIS

DATE/TIME OF TEST: 01/25/2026 00:28:37

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.079	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.135	00:24
Breath Volume		2.550	00:24
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.123	00:27
Breath Volume		1.652	00:27
Air Blank		0.000	00:27
Control Test		0.079	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FENTON,NOAH
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: GUSTIN

DATE/TIME OF TEST: 01/25/2026 03:33:07

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:22	
Control Test	0.079	03:22	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:26	No Sample Provided
Breath Volume	0.000	03:26	
Air Blank	0.000	03:26	
Air Blank	0.000	03:28	
Subject Sample	# 2	03:31	No Sample Provided
Breath Volume	0.000	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Diagnostics Check	OK	03:33	

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIOS,JOSHUA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: ARANDA

DATE/TIME OF TEST: 01/25/2026 05:05:39

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.078	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.317	05:00
Breath Volume	1.257	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:02
Subject Sample # 2	0.312	05:04
Breath Volume	1.562	05:04
Air Blank	0.000	05:04
Control Test	0.077	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BERRIOS,MARK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: VIEIRA

DATE/TIME OF TEST: 01/26/2026 00:19:28

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.079	00:17	
Air Blank	0.000	00:17	
Subject Sample	# 1	00:18	Subject Test Refused
Breath Volume	0.000	00:18	
Air Blank	0.000	00:18	
Control Test	0.079	00:18	
Air Blank	0.000	00:19	
Diagnostics Check	OK	00:19	

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JULES II,DELINCE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 01/26/2026 03:34:44

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.079	03:24	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:27	No Sample Provided
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Air Blank	0.000	03:30	
Subject Sample	# 2	03:33	No Sample Provided
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TAVERAS,TYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: NUNES

DATE/TIME OF TEST: 01/26/2026 04:08:06

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.079	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Air Blank	0.000	04:03	
Subject Sample	# 2	04:06	No Sample Provided
Breath Volume	0.000	04:06	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 02/16/2026 15:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINAPASTORIZA,ARNALDO,H
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RYAN,LACEE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/19/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17725080A1
DRY GAS EXPIRY DATE: 08/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: FAUSTIN

DATE/TIME OF TEST: 01/31/2026 01:46:35

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.098	01:42
Breath Volume	1.902	01:42
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.102	01:45
Breath Volume	1.203	01:45
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 02/16/2026 15:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 01/01/2026 04:16:59

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.081	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.109	04:12
Breath Volume		2.214	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.098	04:15
Breath Volume		1.718	04:15
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALICEA GALLOZA,MARK
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: TURCIOS SANTOS

DATE/TIME OF TEST: 01/04/2026 02:24:40

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.122	02:20
Breath Volume	2.308	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.120	02:23
Breath Volume	3.218	02:23
Air Blank	0.000	02:23
Control Test	0.079	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOWERY,JOHN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: PRADO

DATE/TIME OF TEST: 01/10/2026 02:37:55

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.155	02:33
Breath Volume		1.269	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.161	02:36
Breath Volume		1.265	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PATTERSON,ZAKAI
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/10/2026 05:01:08

Diagnostics Check	OK	04:55
Air Blank	0.000	04:55
Control Test	0.080	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.000	04:56
Breath Volume	2.933	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.000	04:59
Breath Volume	2.828	04:59
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPARROW,CLARENCE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: DEPAZCAMPOS

DATE/TIME OF TEST: 01/11/2026 02:28:05

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:23
Subject Sample # 1	0.169	02:23
Breath Volume	2.699	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.165	02:26
Breath Volume	1.726	02:26
Air Blank	0.000	02:27
Control Test	0.078	02:27
Air Blank	0.000	02:27
Diagnostics Check	OK	02:28

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MIA,R
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SILVA,LUIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CALDERON JR

DATE/TIME OF TEST: 01/11/2026 04:38:29

Diagnostics Check	OK	04:27	
Air Blank	0.000	04:27	
Control Test	0.080	04:27	
Air Blank	0.000	04:28	
Subject Sample	# 1	04:31	No Sample Provided
Breath Volume	0.000	04:31	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:37	No Sample Provided
Breath Volume	0.000	04:37	
Air Blank	0.000	04:37	
Control Test	0.079	04:37	
Air Blank	0.000	04:38	
Diagnostics Check	OK	04:38	

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TRENDAFILOV,NICHOLAS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 01/11/2026 23:22:38

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.080	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.056	23:18
Breath Volume	3.105	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.059	23:21
Breath Volume	3.878	23:21
Air Blank	0.000	23:21
Control Test	0.080	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAZQUEZ,JOSEPH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: CAVANAUGH

DATE/TIME OF TEST: 01/17/2026 00:53:11

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.274	00:48
Breath Volume		2.125	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.279	00:51
Breath Volume		1.492	00:51
Air Blank		0.000	00:52
Control Test		0.079	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCKAHEN,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 01/17/2026 03:34:39

Diagnostics Check	OK	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:29
Air Blank	0.000	03:29
Subject Sample # 1	0.146	03:30
Breath Volume	2.085	03:30
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.148	03:33
Breath Volume	1.824	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:34

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,KATIE,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN,RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 01/17/2026 23:18:28

Diagnostics Check		OK	23:04	
Air Blank		0.000	23:05	
Control Test		0.080	23:05	
Air Blank		0.000	23:06	
Subject Sample	# 1	0.036	23:09	Volume Not Met
Breath Volume		0.339	23:09	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 2	0.034	23:14	
Breath Volume		1.175	23:14	
Air Blank		0.000	23:14	
Air Blank		0.000	23:16	
Subject Sample	# 3	0.036	23:16	
Breath Volume		1.285	23:16	
Air Blank		0.000	23:17	
Control Test		0.079	23:17	
Air Blank		0.000	23:18	
Diagnostics Check		OK	23:18	

Data Download Date/Time: 01/19/2026 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUIZ,ANDRES
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: EGOROV

DATE/TIME OF TEST: 01/18/2026 03:12:54

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.152	03:08
Breath Volume		1.742	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.157	03:11
Breath Volume		3.429	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/19/2026 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMPSEY,COURTNEY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANIELS,DAMON
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 12/10/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00925080A3
DRY GAS EXPIRY DATE: 02/05/2027
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GERVER

DATE/TIME OF TEST: 01/19/2026 00:18:04

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.081	00:12
Air Blank	0.000	00:12
Subject Sample # 1	0.096	00:13
Breath Volume	1.488	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.101	00:16
Breath Volume	1.187	00:16
Air Blank	0.000	00:17
Control Test	0.080	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:18

Data Download Date/Time: 01/19/2026 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISHOP,TREVOR,C
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ST VAL,JEFF
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: KERSEY

DATE/TIME OF TEST: 01/01/2026 04:52:25

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.081	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.105	04:47
Breath Volume	3.332	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.109	04:50
Breath Volume	3.003	04:50
Air Blank	0.000	04:51
Control Test	0.080	04:51
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 01/30/2026 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAYWARD,KYLE,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILLIAMS,SHELTON
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/07/2026 02:35:01

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.324	02:30
Breath Volume		2.292	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.312	02:33
Breath Volume		2.250	02:33
Air Blank		0.000	02:34
Control Test		0.075	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/30/2026 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARRIVEE-GRANT,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Palatka PD
ARRESTING OFFICER: ROBERTS,EAN
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/17/2026 03:52:22

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.189	03:45
Breath Volume	4.031	03:45
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.181	03:50
Breath Volume	3.675	03:50
Air Blank	0.000	03:51
Control Test	0.079	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 01/30/2026 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: ADAMS,PAUL
ARRESTING OFFICER AGENCY: Palatka PD
LAST AGENCY INSPECTION: 12/23/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 25925080A1
DRY GAS EXPIRY DATE: 11/05/2027
VIOLATION CODE: DUI
ARREST TIME: 18:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 01/20/2026 19:17:37

Diagnostics Check	OK	19:10
Air Blank	0.000	19:10
Control Test	0.081	19:10
Air Blank	0.000	19:11
Subject Sample # 1	0.130	19:13
Breath Volume	2.949	19:13
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.132	19:16
Breath Volume	3.257	19:16
Air Blank	0.000	19:16
Control Test	0.080	19:17
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 01/30/2026 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLMES,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 01/23/2026 23:24:40

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.080	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.106	23:19
Breath Volume	2.773	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.090	23:23
Breath Volume	1.582	23:23
Air Blank	0.000	23:23
Control Test	0.079	23:24
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 02/26/2026 08:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLMES,A,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CHOUINARD,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/21/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04125080A5
DRY GAS EXPIRY DATE: 03/05/2027
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MUNCHEL

DATE/TIME OF TEST: 01/24/2026 02:46:09

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.144	02:41
Breath Volume	3.175	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.145	02:44
Breath Volume	2.351	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 02/26/2026 08:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/04/2026 01:21:41

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.083	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.063	01:17
Breath Volume	2.660	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.062	01:20
Breath Volume	2.843	01:20
Air Blank	0.000	01:20
Control Test	0.083	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/04/2026 03:32:57

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.083	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.161	03:28
Breath Volume	3.035	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.163	03:31
Breath Volume	2.582	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: SOLAKAIN

DATE/TIME OF TEST: 01/07/2026 23:05:45

Diagnostics Check	OK	22:59
Air Blank	0.000	23:00
Control Test	0.082	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.084	23:01
Breath Volume	1.613	23:01
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.084	23:04
Breath Volume	1.695	23:04
Air Blank	0.000	23:04
Control Test	0.082	23:05
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: TROCHEZ

DATE/TIME OF TEST: 01/11/2026 02:25:17

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.083	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.151	02:20
Breath Volume	1.945	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.147	02:23
Breath Volume	1.957	02:23
Air Blank	0.000	02:24
Control Test	0.082	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BAILEY,DEREK
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/13/2026 23:30:19

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.083	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.220	23:25
Breath Volume	2.117	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.205	23:28
Breath Volume	2.171	23:28
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCCULLOUGH,JAMES
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: HECHAVARRIA CUBA

DATE/TIME OF TEST: 01/15/2026 01:59:44

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.083	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.187	01:55
Breath Volume	1.183	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.185	01:58
Breath Volume	3.343	01:58
Air Blank	0.000	01:58
Control Test	0.080	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: RAVAN

DATE/TIME OF TEST: 01/24/2026 03:26:26

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.083	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.139	03:22
Breath Volume		1.796	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.134	03:24
Breath Volume		3.554	03:24
Air Blank		0.000	03:25
Control Test		0.083	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BLACKMAN,ALBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 01/24/2026 04:35:50

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.083	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.171	04:31
Breath Volume	1.710	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.163	04:34
Breath Volume	1.167	04:34
Air Blank	0.000	04:34
Control Test	0.082	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: GRAYER

DATE/TIME OF TEST: 01/24/2026 19:10:24

Diagnostics Check	OK	19:03
Air Blank	0.000	19:04
Control Test	0.084	19:04
Air Blank	0.000	19:05
Subject Sample # 1	0.231	19:05
Breath Volume	2.675	19:05
Air Blank	0.000	19:06
Air Blank	0.000	19:08
Subject Sample # 2	0.234	19:08
Breath Volume	2.910	19:08
Air Blank	0.000	19:09
Control Test	0.083	19:09
Air Blank	0.000	19:10
Diagnostics Check	OK	19:10

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BLACKMAN,ALBERT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 01/25/2026 01:32:39

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.083	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.148	01:28
Breath Volume		2.402	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.148	01:31
Breath Volume		2.414	01:31
Air Blank		0.000	01:31
Control Test		0.083	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,SCOTT,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MILLS,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/16/2025

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/28/2026 22:35:16

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.085	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.134	22:30
Breath Volume	1.722	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:33
Subject Sample # 2	0.142	22:33
Breath Volume	1.410	22:33
Air Blank	0.000	22:34
Control Test	0.084	22:34
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,J,R
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: FELBINGER,C
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: REISER

DATE/TIME OF TEST: 01/31/2026 00:49:12

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.084	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.228	00:44
Breath Volume		2.238	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.217	00:47
Breath Volume		1.710	00:47
Air Blank		0.000	00:48
Control Test		0.083	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/27/2026 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLAHERTY,Z,R
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DAVIS-WELCH,M
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/29/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 402816921
DRY GAS EXPIRY DATE: 08/09/2026
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GAULT

DATE/TIME OF TEST: 01/31/2026 01:54:21

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.083	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.177	01:48
Breath Volume	1.550	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.172	01:52
Breath Volume	1.886	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 02/27/2026 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,JACOB,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: 2C14,JACOB
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ROURKS

DATE/TIME OF TEST: 01/17/2026 05:30:41

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Subject Sample	# 1	0.208	05:25
Breath Volume		2.140	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.204	05:28
Breath Volume		1.332	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 02/16/2026 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCWILLIAMS,JACOB,S
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MINER,M
ARRESTING OFFICER AGENCY: High Springs PD
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: HERNACKI

DATE/TIME OF TEST: 01/18/2026 19:19:32

Diagnostics Check	OK	19:12
Air Blank	0.000	19:13
Control Test	0.076	19:13
Air Blank	0.000	19:14
Subject Sample # 1	0.189	19:14
Breath Volume	3.375	19:14
Air Blank	0.000	19:15
Air Blank	0.000	19:17
Subject Sample # 2	0.192	19:17
Breath Volume	3.781	19:17
Air Blank	0.000	19:18
Control Test	0.075	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

Data Download Date/Time: 02/16/2026 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,MALCOLM,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: WILSON,MALCOLM
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/24/2026 09:26:24

Diagnostics Check	OK	09:19
Air Blank	0.000	09:20
Control Test	0.079	09:20
Air Blank	0.000	09:21
Subject Sample # 1	0.134	09:21
Breath Volume	1.316	09:21
Air Blank	0.000	09:22
Air Blank	0.000	09:24
Subject Sample # 2	0.145	09:24
Breath Volume	1.660	09:24
Air Blank	0.000	09:25
Control Test	0.079	09:25
Air Blank	0.000	09:26
Diagnostics Check	OK	09:26

Data Download Date/Time: 02/16/2026 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAHARAJ,KRISHNA,H
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: MAHARAJ,KRISHNA
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ACS

DATE/TIME OF TEST: 01/30/2026 01:39:35

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.080	01:34
Air Blank	0.000	01:34
Subject Sample # 1	0.000	01:35
Breath Volume	1.847	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.000	01:38
Breath Volume	1.687	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 02/16/2026 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

INSTRUMENT REGISTERED TO: ALACHUA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABDALE,JAMES',M
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: ABDALE,JAMES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/16/2026

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01524080A4
DRY GAS EXPIRY DATE: 03/05/2026
VIOLATION CODE: DUI
ARREST TIME: 08:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: GRACE

DATE/TIME OF TEST: 01/30/2026 09:26:59

Diagnostics Check	OK	09:20
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:22
Subject Sample # 1	0.168	09:22
Breath Volume	3.464	09:22
Air Blank	0.000	09:23
Air Blank	0.000	09:24
Subject Sample # 2	0.162	09:25
Breath Volume	3.269	09:25
Air Blank	0.000	09:26
Control Test	0.079	09:26
Air Blank	0.000	09:26
Diagnostics Check	OK	09:26

Data Download Date/Time: 02/16/2026 09:43