



INSTRUMENT PROCESSING SHEET

Agency Franklin County SO

S/N 80-000952

Florida Department of Law Enforcement

Date In 02/26/2025 DI Completion Date 3/4/2025

Ship P/U H/D CMI EE

Intake By ALL Date 02/26/2025 Quality Checks By PN Date 3/4/2025 Flow Calibration By Date
Annual Registration Return from CMI / EE
Visual Inspection: Case Handle Keyboard Dry Gas Shelf Feet Breath Tube Ports Screws Tight
Other Equipment/ Accessories: Power cord Printer Cable Static Bag 12V DC Cable
Notes: drop off, no box
Breath Tube Screen Replace External O-Rings Instrument Set Up Verified R-Value 211 Flow Verification (L/s)
Flow Column # ATP 102
32 mm .144 (.139 - .169)
36 mm .156 (.156 - .190)
53 mm .230 (.228 - .278)
103 mm .492 (.447 - .547)
Barometric Pressure Check Gauge ID # 30793
Stability Checks
Simulator Serial # Lot #/Exp
0.050 MP6291 202406K 06/19/2026
0.080 MP6292 202406L 06/19/2026
0.200 MP6293 202406N 06/20/2026
0.080 DGS N/A AG429602 10/22/2026
Flow Column # 5L/min - 17mm 15L/min - 53mm 30L/min - 103mm
R-Value Post Calibration Verification (L/s)
Flow Column # 32 mm 36 mm 53 mm 103 mm
Maintenance By Date
Battery Replacement Dry Gas Regulator Replacement Breath Tube Replacement Other


Calibration Adjustment By Department Inspection By PN
Barometric Pressure Gauge ID #
Simulator Serial # Lot # Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.300
0.080 DGS N/A
Post Calibration Adjustment Stability Checks
Simulator Serial # Lot # Expiration
0.050
0.080
0.200
0.080 DGS N/A
Barometric Pressure ID# 30793
Gauge 1013 Instrument 1013
Mouth Alcohol Solution Lot # 2024-A
Acetone Stock Solution Lot # 2024-B
Simulator Serial Number
0.000 MP6294
Interferent MP6290
0.050 MP6291
0.080 MP6292
0.200 MP6293
Attachments
Form 41 Post-Stability Checks
Stability Checks Flow Calibration
Calibration Certificate Form 40
Calibration Adjustment Other

Notes/Suggested Service:
Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use
Digitally signed by
Destinee Armstrong Date: 2025.03.05 14:40:38 -05'00'
Shayla Platt Date: 2025.03.06 10:58:36 -05'00'
Tech Review / Date Admin Review / Date

STABILITY CHECKS

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.100	10:52
Control Test	0.049	10:53
Air Blank	0.000	10:53
Control Test	0.049	10:54
Air Blank	0.000	10:54
Control Test	0.049	10:55
Air Blank	0.000	10:55
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0030	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:57
Control Test	0.078	10:58
Air Blank	0.000	10:58
Control Test	0.078	10:59
Air Blank	0.000	10:59
Control Test	0.078	11:00
Air Blank	0.000	11:01
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:03
Control Test	0.198	11:04
Air Blank	0.000	11:04
Control Test	0.197	11:05
Air Blank	0.000	11:06
Control Test	0.198	11:06
Air Blank	0.000	11:07
Control Test Stats		
Average	0.1977	
Std Dev	0.0006	
Rel Std Dev(%)	0.2921	


Operator's Signature

FRANKLIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000952
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:09
Control Test	0.079	11:09
Air Blank	0.000	11:10
Control Test	0.078	11:10
Air Blank	0.000	11:10
Control Test	0.079	11:11
Air Blank	0.000	11:11
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

D65


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FRANKLIN COUNTY SO
Time of Inspection: 14:25

Date of Inspection: 03/04/2025

Serial Number: 80-000952
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.047	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.077	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.047	0.078	0.197	0.079
0.000	0.047	0.078	0.198	0.079
0.000	0.047	0.078	0.197	0.078


Standard Deviations	0.0005	0.0004	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:



The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO
Signature and Printed Name

03/04/2025
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000952, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000952</u>	UNCERTAINTY* ±	
Owning Agency:	<u>FRANKLIN COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/04/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

PHIL NICODEMO,
 Department Inspector

03/04/2025
 Date

FDLE/ATP Form 69 October 2024
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality