

Alcohol Testing Program - Instrument Processing Sheet

Agency: FL HIGHWAY PATROL Instrument Serial Number: 80-000785
 Date In: 1/16/2025 DI Completion Date: 01/20/2026 Ship P/U H/D CMI EE

Intake By: <u>WKP</u> Date: <u>1/16/2026</u>	Quality Checks By: <u>WKP</u> Date: <u>01/16/2026</u>	Flow Adjustment By: _____
<input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>225</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.511</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28662</u> Gauge: <u>1017</u> Instrument: <u>1017</u>	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

Stability Checks			Maintenance By:	Date:
Simulator	Serial #	Lot#/Exp	<input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:	
0.050	MP5088	202406K 06/19/2026		
0.080	MP5089	202406L 06/19/2026		
0.200	MP5090	202406N 06/20/2026		
0.080 DGS	N/A	AG510701		
		04/17/2027		

Optical Bench Adjustment By: _____ **Department Inspection** By: WKP

Barometric Pressure Gauge: _____ ID#: _____				Barometric Pressure ID#: <u>28427</u>	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1026</u>	Instrument: <u>1028</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>09/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>09/22/2027</u>
0.100				Simulator	Serial Number
0.200				0.000	MP6289
0.300				Interferent	MP6290
0.080 DGS	N/A			0.050	MP5088

<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	MP5089
Simulator	Serial #	Lot #	Expiration	0.200	MP5090
0.050				Attachments	
0.080				<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
0.200				<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Adjustment
0.080 DGS	N/A			<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40
Gauge ID #: _____				<input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Other:
Gauge: _____ Instrument: _____					

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
	Digitally signed by Taylor Gutschow Date: 2026.01.27 12:37:06 -05'00'
	Digitally signed by Kaitlyn Spearin Date: 2026.01.28 12:52:10 -05'00'
Tech Review	Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	<input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <input checked="" type="checkbox"/> <small>≤0.003 of Wet</small>																																																																																																																																																		
<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000785 01/16/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:58	Control Test	0.048	13:58	Air Blank	0.000	13:59	Control Test	0.048	13:59	Air Blank	0.000	14:00	Control Test	0.048	14:00	Air Blank	0.000	14:01	Control Test	0.048	14:01	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000785 01/16/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.076</td><td>14:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:04	Control Test	0.079	14:05	Air Blank	0.000	14:05	Control Test	0.078	14:06	Air Blank	0.000	14:06	Control Test	0.076	14:07	Air Blank	0.000	14:07	Control Test Stats			Average	0.0783		Std Dev	0.0016		Rel. Std Dev(%)	0.7370		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000785 01/16/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.196</td><td>14:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.195</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.195</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2956</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:09	Control Test	0.196	14:10	Air Blank	0.000	14:10	Control Test	0.195	14:11	Air Blank	0.000	14:11	Control Test	0.195	14:12	Air Blank	0.000	14:13	Control Test Stats			Average	0.1953		Std Dev	0.0006		Rel. Std Dev(%)	0.2956		<p>FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-000785 01/16/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	14:15	Control Test	0.081	14:16	Air Blank	0.000	14:16	Control Test	0.081	14:17	Air Blank	0.000	14:17	Control Test	0.080	14:17	Air Blank	0.000	14:18	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel. Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	13:58																																																																																																																																																				
Control Test	0.048	13:58																																																																																																																																																				
Air Blank	0.000	13:59																																																																																																																																																				
Control Test	0.048	13:59																																																																																																																																																				
Air Blank	0.000	14:00																																																																																																																																																				
Control Test	0.048	14:00																																																																																																																																																				
Air Blank	0.000	14:01																																																																																																																																																				
Control Test	0.048	14:01																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0480																																																																																																																																																					
Std Dev	0.0000																																																																																																																																																					
Rel. Std Dev(%)	0.0000																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	14:04																																																																																																																																																				
Control Test	0.079	14:05																																																																																																																																																				
Air Blank	0.000	14:05																																																																																																																																																				
Control Test	0.078	14:06																																																																																																																																																				
Air Blank	0.000	14:06																																																																																																																																																				
Control Test	0.076	14:07																																																																																																																																																				
Air Blank	0.000	14:07																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0783																																																																																																																																																					
Std Dev	0.0016																																																																																																																																																					
Rel. Std Dev(%)	0.7370																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	14:09																																																																																																																																																				
Control Test	0.196	14:10																																																																																																																																																				
Air Blank	0.000	14:10																																																																																																																																																				
Control Test	0.195	14:11																																																																																																																																																				
Air Blank	0.000	14:11																																																																																																																																																				
Control Test	0.195	14:12																																																																																																																																																				
Air Blank	0.000	14:13																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.1953																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel. Std Dev(%)	0.2956																																																																																																																																																					
Test	g/210L	Time																																																																																																																																																				
Air Blank	0.000	14:15																																																																																																																																																				
Control Test	0.081	14:16																																																																																																																																																				
Air Blank	0.000	14:16																																																																																																																																																				
Control Test	0.081	14:17																																																																																																																																																				
Air Blank	0.000	14:17																																																																																																																																																				
Control Test	0.080	14:17																																																																																																																																																				
Air Blank	0.000	14:18																																																																																																																																																				
Control Test Stats																																																																																																																																																						
Average	0.0807																																																																																																																																																					
Std Dev	0.0006																																																																																																																																																					
Rel. Std Dev(%)	0.7157																																																																																																																																																					

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 60-010765
01/16/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:53
Control Test	0.044	13:54
Air Blank	0.000	13:54
Control Test	0.044	13:55
Air Blank	0.000	13:55
Control Test	0.044	13:56
Air Blank	0.000	13:56
Control Test Stats		
Average	0.0440	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

0.05g/210L
stability
outside acceptable
range, tightened
sim head and
secured sim
connection with
instrument. Repeated.
4/16/2026


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 13:53

Date of Inspection: 01/20/2026

Serial Number: 80-000785
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.046	0.078	0.197	0.080
0.000	0.047	0.079	0.197	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.078	0.197	0.080
0.000	0.047	0.078	0.197	0.081
0.000	0.047	0.078	0.197	0.080
0.000	0.047	0.079	0.197	0.081
0.000	0.048	0.079	0.197	0.080
0.000	0.046	0.079	0.198	0.080
0.000	0.047	0.079	0.197	0.080

Standard Deviations	0.0005	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson
Signature and Printed Name

01/20/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000785, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000785</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/20/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:53</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.01.26 10:49:55 -05'00'

01/20/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality