

Alcohol Testing Program - Instrument Processing Sheet

Agency: PENSACOLA PD Instrument Serial Number: 80-007159  
 Date In: 1/20/2026 DI Completion Date: 01/28/2026  Ship  P/U  H/D  CMI  EE

| Intake By: <u>WKP</u> Date: <u>1/20/2026</u><br><input checked="" type="checkbox"/> Annual<br><input checked="" type="checkbox"/> Dropped Off<br><input type="checkbox"/> Registration<br><input type="checkbox"/> Return from CMI / EE<br><input type="checkbox"/> Training Instrument<br>Visual Inspection<br><input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle<br><input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf<br><input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube<br><input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight<br>Other Equipment/Accessories<br><input checked="" type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable<br><input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable<br>Notes:<br>Instrument arrived with missing F7 key.<br>WKP   | Quality Checks By: <u>WKP</u> Date: <u>1/20/2026</u><br><input checked="" type="checkbox"/> Breath Tube Screen<br><input checked="" type="checkbox"/> Replace External O-Rings<br><input checked="" type="checkbox"/> Instrument Set Up Verified<br><input checked="" type="checkbox"/> R-Value: <u>199</u><br><input checked="" type="checkbox"/> Flow Verification (L/s)<br>Flow Column #: <u>ATP 105</u><br>32 mm <u>0.144</u> (.139-.169)<br>36 mm <u>0.160</u> (.156-.190)<br>53 mm <u>0.234</u> (.228-.278)<br>103 mm <u>0.511</u> (.447-.547)<br><input checked="" type="checkbox"/> Barometric Pressure Check<br>Gauge ID #: <u>28662</u><br>Gauge: <u>1028</u> Instrument: <u>1030</u><br><input checked="" type="checkbox"/> Stability Checks | Flow Adjustment By: _____<br>Flow Column #: _____<br><input type="checkbox"/> 5L/min – 17mm<br><input type="checkbox"/> 15L/min – 53mm<br><input type="checkbox"/> 30L/min – 103mm<br><input type="checkbox"/> R-Value: _____<br><input type="checkbox"/> Post Adjustment Verification (L/S)<br>Flow Column #: _____<br>32 mm _____ (.139-.169)<br>36 mm _____ (.156-.190)<br>53 mm _____ (.228-.278)<br>103 mm _____ (.447-.547) |  |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|---|---|---|--|----------|----------|-----------------|-------|-------|--------|---------|--|--|------------|-------|--------|---------|--|--|------------|-------|--------|---------|---|--|------------|-----------|-----|----------|--|--|------------|--|--|--|---------------------------------|--|
| <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> <th>Maintenance By:</th> <th>Date:</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202406K</td> <td rowspan="2"><input type="checkbox"/> Battery Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202406L</td> <td rowspan="2"><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>06/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202406N</td> <td rowspan="2"><input type="checkbox"/> Tank Sensor Tare</td> <td rowspan="2"></td> </tr> <tr> <td>06/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG510701</td> <td rowspan="2"><input type="checkbox"/> Breath Tube Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>04/17/2027</td> </tr> <tr> <td colspan="3"></td> <td><input type="checkbox"/> Other:</td> <td></td> </tr> </tbody> </table> |   |   | Simulator  | Serial # | Lot#/Exp | Maintenance By: | Date: | 0.050 | MP5088 | 202406K | <input type="checkbox"/> Battery Replacement |  | 06/19/2026 | 0.080 | MP5089 | 202406L | <input type="checkbox"/> Dry Gas Regulator Replacement |  | 06/19/2026 | 0.200 | MP5090 | 202406N | <input type="checkbox"/> Tank Sensor Tare |  | 06/20/2026 | 0.080 DGS | N/A | AG510701 | <input type="checkbox"/> Breath Tube Replacement |  | 04/17/2027 |  |  |  | <input type="checkbox"/> Other: |  |
| Simulator   | Serial #  | Lot#/Exp  | Maintenance By:  | Date:    |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
| 0.050   | MP5088  | 202406K   | <input type="checkbox"/> Battery Replacement           |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|   |   | 06/19/2026  |  |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
| 0.080   | MP5089  | 202406L   | <input type="checkbox"/> Dry Gas Regulator Replacement |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|   |   | 06/19/2026  |  |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
| 0.200   | MP5090  | 202406N   | <input type="checkbox"/> Tank Sensor Tare              |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|   |   | 06/20/2026  |  |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
| 0.080 DGS   | N/A   | AG510701  | <input type="checkbox"/> Breath Tube Replacement       |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|   |   | 04/17/2027  |  |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |
|   |   |   | <input type="checkbox"/> Other:                        |          |          |                 |       |       |        |         |  |  |            |       |        |         |  |  |            |       |        |         |   |  |            |           |     |          |  |  |            |  |  |  |                                 |  |

| <b>Optical Bench Adjustment</b> By: <u>WKP</u><br>Barometric Pressure Gauge: <u>1028</u> ID#: <u>34419</u>   | <b>Department Inspection</b> By: <u>WKP</u><br>Barometric Pressure ID#: <u>34418</u><br>Gauge: <u>1025</u> Instrument: <u>1024</u><br>Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u><br>Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u> |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
|--|--|------------|------------|------------|-------|--------|---------|------------|-------|--------|---------|------------|-------|--------|---------|------------|-----------|--------|----------|------------|--|--------|-------|----------|-----------|-----|------------|-----------|---|-----------|---------------|-------|--------|-------------|--------|-------|--------|-------|--------|-------|--------|
| <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>25282</td> <td>8/12/2027</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>5/6/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>28424080A3</td> <td>11/5/2026</td> </tr> </tbody> </table> | Simulator  | Serial #   | Lot #      | Expiration | 0.000 | MP6294 | N/A     | N/A        | 0.040 | MP6295 | 25090   | 3/11/2027  | 0.100 | MP6296 | 25282   | 8/12/2027  | 0.200     | MP6297 | 25020    | 1/14/2027  | 0.300  | MP6298 | 25150 | 5/6/2027 | 0.080 DGS | N/A | 28424080A3 | 11/5/2026 | <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table> | Simulator | Serial Number | 0.000 | MP6289 | Interferent | MP5087 | 0.050 | MP5088 | 0.080 | MP5089 | 0.200 | MP5090 |
| Simulator  | Serial #   | Lot #      | Expiration |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.000  | MP6294   | N/A        | N/A        |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.040  | MP6295   | 25090      | 3/11/2027  |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.100  | MP6296   | 25282      | 8/12/2027  |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.200  | MP6297   | 25020      | 1/14/2027  |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.300  | MP6298   | 25150      | 5/6/2027   |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.080 DGS  | N/A  | 28424080A3 | 11/5/2026  |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| Simulator  | Serial Number  |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.000  | MP6289   |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| Interferent  | MP5087   |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.050  | MP5088   |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.080  | MP5089   |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.200  | MP5090   |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| <input checked="" type="checkbox"/> Post Optical Bench Adjustment Stability Checks   |  |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202406K</td> <td>06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202406L</td> <td>06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202406N</td> <td>06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701</td> <td>04/17/2027</td> </tr> </tbody> </table>  | Simulator  | Serial #   | Lot #      | Expiration | 0.050 | MP6291 | 202406K | 06/19/2026 | 0.080 | MP6292 | 202406L | 06/19/2026 | 0.200 | MP6293 | 202406N | 06/20/2026 | 0.080 DGS | N/A    | AG510701 | 04/17/2027 | <b>Attachments</b><br><input checked="" type="checkbox"/> Form 41<br><input checked="" type="checkbox"/> Stability Checks<br><input checked="" type="checkbox"/> Calibration Certificate<br><input checked="" type="checkbox"/> Optical Bench Adjustment<br><input checked="" type="checkbox"/> Post-Stability Checks<br><input type="checkbox"/> Flow Adjustment<br><input type="checkbox"/> Form 40<br><input type="checkbox"/> Other: |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| Simulator  | Serial #   | Lot #      | Expiration |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.050  | MP6291   | 202406K    | 06/19/2026 |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.080  | MP6292   | 202406L    | 06/19/2026 |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.200  | MP6293   | 202406N    | 06/20/2026 |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| 0.080 DGS  | N/A  | AG510701   | 04/17/2027 |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |
| Gauge ID #: <u>34420</u><br>Gauge: <u>1026</u> Instrument: <u>1027</u>   |  |            |            |            |       |        |         |            |       |        |         |            |       |        |         |            |           |        |          |            |  |        |       |          |           |     |            |           |   |           |               |       |        |             |        |       |        |       |        |       |        |

|   |   |  |  |
|---|---|--|--|
| Notes/Suggested Service:  | <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC<br><input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC<br><input checked="" type="checkbox"/> Return to/Place into Evidentiary Use<br><input type="checkbox"/> Remain Out of Evidentiary Use<br><input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use |  |  |
| <table style="width:100%;"> <tr> <td style="width:50%; text-align: center;">                     Digitally signed by Taylor Gutschow<br/>                     Date: 2026.01.30 09:55:30 -05'00'                 </td> <td style="width:50%; text-align: center;">                     Digitally signed by Taylor Gutschow<br/>                     Date: 2026.01.30 10:18:52 -05'00'                 </td> </tr> </table> |   | Digitally signed by Taylor Gutschow<br>Date: 2026.01.30 09:55:30 -05'00' | Digitally signed by Taylor Gutschow<br>Date: 2026.01.30 10:18:52 -05'00' |
| Digitally signed by Taylor Gutschow<br>Date: 2026.01.30 09:55:30 -05'00'  | Digitally signed by Taylor Gutschow<br>Date: 2026.01.30 10:18:52 -05'00'  |  |  |
| Tech Review   | Admin Review  |  |  |

# Stability Checks

| <b>0.05g/210L</b><br>0.047 to 0.053 <input checked="" type="checkbox"/>  | <b>0.08g/210L</b><br>0.077 to 0.083 <input checked="" type="checkbox"/> | <b>0.20g/210L</b><br>0.194 to 0.206 <input checked="" type="checkbox"/> | <b>DGS 0.08g/210L</b><br>0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet <input checked="" type="checkbox"/> |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
|--|---|---|---|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|
| <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/20/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:55</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test  | g/210L  | Time  | Air Blank | 0.000 | 11:53 | Control Test | 0.049 | 11:53 | Air Blank | 0.000 | 11:54 | Control Test | 0.049 | 11:55 | Air Blank | 0.000 | 11:55 | Control Test | 0.048 | 11:55 | Air Blank | 0.000 | 11:56 | Control Test Stats |  |  | Average | 0.0487 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 1.1863 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/20/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 11:58 | Control Test | 0.079 | 11:58 | Air Blank | 0.000 | 11:59 | Control Test | 0.079 | 12:00 | Air Blank | 0.000 | 12:00 | Control Test | 0.079 | 12:01 | Air Blank | 0.000 | 12:01 | Control Test Stats |  |  | Average | 0.0790 |  | Std Dev | 0.0000 |  | Rel Std Dev(%) | 0.0000 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8100 SN 80-007159<br/>01/20/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 12:03 | Control Test | 0.197 | 12:03 | Air Blank | 0.000 | 12:04 | Control Test | 0.196 | 12:05 | Air Blank | 0.000 | 12:05 | Control Test | 0.196 | 12:06 | Air Blank | 0.000 | 12:06 | Control Test Stats |  |  | Average | 0.1963 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 0.2941 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/20/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.083</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.082</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.083</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0827</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.6984</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 12:14 | Control Test | 0.083 | 12:15 | Air Blank | 0.000 | 12:15 | Control Test | 0.082 | 12:16 | Air Blank | 0.000 | 12:16 | Control Test | 0.083 | 12:16 | Air Blank | 0.000 | 12:17 | Control Test Stats |  |  | Average | 0.0827 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 0.6984 |  |
| Test   | g/210L  | Time  |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:53   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.049   | 11:53   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:54   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.049   | 11:55   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:55   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.048   | 11:55   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:56   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |   |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0487  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 1.1863  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L  | Time  |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:58   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.079   | 11:58   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 11:59   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.079   | 12:00   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:00   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.079   | 12:01   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:01   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |   |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0790  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0000  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.0000  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L  | Time  |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:03   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.197   | 12:03   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:04   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.196   | 12:05   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:05   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.196   | 12:06   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:06   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |   |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.1963  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.2941  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L  | Time  |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:14   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.083   | 12:15   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:15   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.082   | 12:16   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:16   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.083   | 12:16   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000   | 12:17   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |   |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0827  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.6984  |   |   |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |

#1

PENSACOLA P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007159  
01/20/2026  
Software: 8100.27

| Test               | g/210L | Time  |
|--------------------|--------|-------|
| Air Blank          | 0.000  | 12:09 |
| Control Test       | 0.083  | 12:09 |
| Air Blank          | 0.000  | 12:10 |
| Control Test       | 0.082  | 12:10 |
| Air Blank          | 0.000  | 12:10 |
| Control Test       | 0.082  | 12:11 |
| Air Blank          | 0.000  | 12:11 |
| Control Test Stats |        |       |
| Average            | 0.0823 |       |
| Std Dev            | 0.0006 |       |
| Rel Std Dev(%)     | 0.7012 |       |

*WMP*

Operator's Signature

average of one 0.07g/210L  
met and the average of  
one 0.08g/210L DORS difference  
greater than 0.003. Repeated  
after reattaching one DORS.

WMP 1/20/2026

PENSACOLA P.D.  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-007159  
01/27/2026 12:01:56

Auto Calibration  
Max Power Res Value = 104  
Auto Range Res Value = 78

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12635, Sum lo = 13769

<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1740 (-0.0080)  
Sample #2 = 0.1640 (0.0100)  
Sample #3 = 0.1510 (0.0310)  
Sample #4 = 0.1500 (0.0440)  
Avg % Abs = 0.1550 (0.0283)  
STD DEV = 0.0078 (0.0172)  
REL STD DEV = 5.039 (60.551)

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1840 (-0.0060)  
Sample #2 = 0.1950 (-0.0070)  
Sample #3 = 0.1790 (0.0020)  
Sample #4 = 0.1810 (0.0000)  
Avg % Abs = 0.1850 (-0.0017)  
STD DEV = 0.0087 (0.0047)  
REL STD DEV = 4.712 (283.549)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12629, Sum lo = 13768

<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8960 (-0.0020)  
Sample #2 = 0.9010 (0.0010)  
Sample #3 = 0.9100 (0.0080)  
Sample #4 = 0.9040 (0.0140)  
Avg % Abs = 0.9050 (0.0077)  
STD DEV = 0.0046 (0.0065)  
REL STD DEV = 0.506 (84.866)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12623, Sum lo = 13764

<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.9970 (-0.0150)  
Sample #2 = 1.9660 (0.0220)  
Sample #3 = 1.9890 (0.0150)  
Sample #4 = 1.9780 (0.0260)  
Avg % Abs = 1.9777 (0.0210)  
STD DEV = 0.0115 (0.0056)  
REL STD DEV = 0.582 (26.513)

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.4980 (-0.0040)  
Sample #2 = 3.5090 (0.0040)  
Sample #3 = 3.5010 (0.0000)  
Sample #4 = 3.5120 (-0.0060)  
Avg % Abs = 3.5073 (-0.0007)  
STD DEV = 0.0057 (0.0050)  
REL STD DEV = 0.162 (754.984)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12620, Sum lo = 13763

<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.7790 (0.0000)  
Sample #2 = 3.7790 (0.0050)  
Sample #3 = 3.7830 (0.0120)  
Sample #4 = 3.7650 (0.0300)  
Avg % Abs = 3.7757 (0.0157)  
STD DEV = 0.0095 (0.0129)  
REL STD DEV = 0.250 (82.321)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum lo = 12626, Sum lo = 13768

<<<< CHANNEL 1 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.5100 (-0.0120)  
Sample #2 = 5.4890 (0.0070)  
Sample #3 = 5.5040 (0.0050)  
Sample #4 = 5.5020 (0.0100)  
Avg % Abs = 5.4983 (0.0073)  
STD DEV = 0.0081 (0.0025)  
REL STD DEV = 0.148 (34.317)

<<<< CHANNEL 2 >>>>  
Sample % Abs (% Abs Ref)  
Sample #1 = 9.6830 (-0.0020)  
Sample #2 = 9.6490 (0.0180)  
Sample #3 = 9.6550 (0.0140)  
Sample #4 = 9.6650 (0.0110)  
Avg % Abs = 9.6563 (0.0143)  
STD DEV = 0.0081 (0.0035)  
REL STD DEV = 0.084 (24.502)

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*  
<<<< CHANNEL 1 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.155  
Std Dev = 0.01 Rel Std Dev = 5.04  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.905  
Std Dev = 1.00 Rel Std Dev = 0.51  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.978  
Std Dev = 0.01 Rel Std Dev = 0.58  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.776  
Std Dev = 0.01 Rel Std Dev = 0.25  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.498  
Std Dev = 0.01 Rel Std Dev = 0.15  
Zero Order Coef = -401.60  
First Order Coef = 2553.80  
Second Order Coef = 21.17  
Standard Deviation = 21.262308

<<<< CHANNEL 2 >>>>  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.185  
Std Dev = 0.01 Rel Std Dev = 4.71  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.547  
Std Dev = 0.01 Rel Std Dev = 0.46  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.507  
Std Dev = 0.01 Rel Std Dev = 0.16  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.673  
Std Dev = 0.01 Rel Std Dev = 0.12  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.656  
Std Dev = 0.01 Rel Std Dev = 0.08  
Zero Order Coef = -255.56  
First Order Coef = 1381.49  
Second Order Coef = 12.84  
Standard Deviation = 11.001620

Solution Stats Quadratic Fit Chan 1

| Act    | Fit    | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L   |
| 0.000  | -0.000 | 0.0001   |
| 0.040  | 0.040  | -0.0005  |
| 0.100  | 0.099  | 0.0006   |
| 0.200  | 0.200  | -0.0004  |
| 0.300  | 0.300  | 0.0001   |

Solution Stats Quadratic Fit Chan 2

| Act    | Fit    | Residual |
|--------|--------|----------|
| g/210L | g/210L | g/210L   |
| 0.000  | 0.000  | -0.0000  |
| 0.040  | 0.040  | -0.0001  |
| 0.100  | 0.100  | 0.0003   |
| 0.200  | 0.200  | -0.0002  |
| 0.300  | 0.300  | 0.0001   |

Sol Value = 0.380 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1 \*\*\*\*\*  
Sample #1 = 3124.00  
Sample #2 = 3194.00  
Sample #3 = 3154.00  
Sample #4 = 3180.00  
Average Result = 3076.0000  
STD DEV = 20.2978  
REL STD DEV = 0.660  
\*\*\*\*\*  
\*\*\*\*\* CHANNEL 2 \*\*\*\*\*  
Sample #1 = 3383.00  
Sample #2 = 3326.00  
Sample #3 = 3332.00  
Sample #4 = 3307.00  
Average Result = 3321.6667  
STD DEV = 13.0512  
REL STD DEV = 0.393  
\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1326  
3 um H2O Adjust (ng/l x 10,000) = 733  
9 um H2O Adjust (ng/l x 10,000) = 488  
\*\*\*\* AUTO CAL PASS

Optical Bench  
calibration  
adjustment  
wep 1/27/2026

# Post-Cal Stability Checks

| 0.05g/210L<br>0.047 to 0.053   | 0.08g/210L<br>0.077 to 0.083 | 0.20g/210L<br>0.194 to 0.206 | DGS 0.08g/210L<br>0.077 to 0.083 ≤0.003 of Wet |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
|--|------------------------------|------------------------------|--|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|--|------|--------|------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------|-------|-------|-----------|-------|-------|--------------------|--|--|---------|--------|--|---------|--------|--|----------------|--------|--|
| <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/27/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test                         | g/210L                       | Time   | Air Blank | 0.000 | 12:53 | Control Test | 0.050 | 12:53 | Air Blank | 0.000 | 12:54 | Control Test | 0.049 | 12:55 | Air Blank | 0.000 | 12:55 | Control Test | 0.049 | 12:56 | Air Blank | 0.000 | 12:56 | Control Test Stats |  |  | Average | 0.0493 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 1.1703 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/27/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 12:57 | Control Test | 0.080 | 12:58 | Air Blank | 0.000 | 12:59 | Control Test | 0.080 | 12:59 | Air Blank | 0.000 | 13:00 | Control Test | 0.080 | 13:01 | Air Blank | 0.000 | 13:01 | Control Test Stats |  |  | Average | 0.0800 |  | Std Dev | 0.0000 |  | Rel Std Dev(%) | 0.0000 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/27/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 13:03 | Control Test | 0.200 | 13:04 | Air Blank | 0.000 | 13:05 | Control Test | 0.199 | 13:05 | Air Blank | 0.000 | 13:06 | Control Test | 0.199 | 13:06 | Air Blank | 0.000 | 13:07 | Control Test Stats |  |  | Average | 0.1993 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 0.2896 |  | <p>PENSACOLA P.D.<br/>Intoxilyzer - Alcohol Analyzer<br/>Model 8000 SN 80-007159<br/>01/27/2026<br/>Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0833</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p><i>menpi</i><br/>Operator's Signature</p> | Test | g/210L | Time | Air Blank | 0.000 | 13:11 | Control Test | 0.081 | 13:11 | Air Blank | 0.000 | 13:12 | Control Test | 0.080 | 13:12 | Air Blank | 0.000 | 13:13 | Control Test | 0.080 | 13:13 | Air Blank | 0.000 | 13:13 | Control Test Stats |  |  | Average | 0.0833 |  | Std Dev | 0.0006 |  | Rel Std Dev(%) | 0.7187 |  |
| Test   | g/210L                       | Time                         |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:53                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.050                        | 12:53                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:54                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.049                        | 12:55                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:55                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.049                        | 12:56                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:56                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |                              |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0493                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 1.1703                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L                       | Time                         |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:57                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.080                        | 12:58                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 12:59                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.080                        | 12:59                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:00                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.080                        | 13:01                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:01                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |                              |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0800                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0000                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.0000                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L                       | Time                         |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:03                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.200                        | 13:04                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:05                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.199                        | 13:05                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:06                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.199                        | 13:06                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:07                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |                              |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.1993                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.2896                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Test   | g/210L                       | Time                         |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:11                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.081                        | 13:11                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:12                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.080                        | 13:12                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:13                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test   | 0.080                        | 13:13                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Air Blank  | 0.000                        | 13:13                        |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Control Test Stats   |                              |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Average  | 0.0833                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Std Dev  | 0.0006                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |
| Rel Std Dev(%)   | 0.7187                       |                              |  |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |  |      |        |      |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |              |       |       |           |       |       |                    |  |  |         |        |  |         |        |  |                |        |  |

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PENSACOLA P.D.  
Time of Inspection: 11:54

Date of Inspection: 01/28/2026

Serial Number: 80-007159  
Software: 8100.27

| Check or Test                               | YES | NO | Check or Test                          | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK       | Yes |    | Date and/or Time Adjusted              |     | No |
| Minimum Sample Volume Check: OK             | Yes |    | Barometric Pressure Sensor Check: OK   | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    | Mouth Alcohol Test: Slope Not Met      | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    | Diagnostic Check (Post-Inspection): OK | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202406K<br>Exp: 06/19/2026 | 0.08g/210L Test (g/210L)<br>Lot#:202406L<br>Exp: 06/19/2026 | 0.20g/210L Test (g/210L)<br>Lot#:202406N<br>Exp: 06/20/2026 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:AG510701<br>Exp: 04/17/2027 |
|----------------------------|---|---|---|---|
| 0.000                      | 0.048   | 0.078   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.199   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.198   | 0.080   |
| 0.000                      | 0.048   | 0.079   | 0.199   | 0.080   |
| 0.000                      | 0.048   | 0.080   | 0.198   | 0.080   |
| 0.000                      | 0.049   | 0.080   | 0.199   | 0.080   |


|                     |        |        |        |        |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0003 | 0.0005 | 0.0004 | 0.0000 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

01/28/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007159, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

|                   |                       |                                |       |
|-------------------|-----------------------|--------------------------------|-------|
| Serial Number:    | <u>80-007159</u>      | UNCERTAINTY* ±                 |       |
| Owning Agency:    | <u>PENSACOLA P.D.</u> | 0.050 g/ 210 L                 | 0.004 |
| Calibration Date: | <u>01/28/2026</u>     | 0.080 g/ 210 L                 | 0.004 |
| Calibration Time: | <u>11:54</u>          | 0.200 g/ 210 L                 | 0.008 |
|                   |                       | 0.080 g/ 210 L Dry Gas Control | 0.005 |

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/28/2026

Date

Wen-Chi  
Pierson

Digitally signed by Wen-Chi  
Pierson  
Date: 2026.01.30 09:41:32  
-05'00'

WEN-CHI K PIERSON,  
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality