

Alcohol Testing Program - Instrument Processing Sheet

Agency: LAKE HELEN PD

Instrument Serial Number: 80-001149

Date In: 7/9/2026

DI Completion Date: 7/24/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: WKP Date: 7/9/2026		Quality Checks By: WKP Date: 7/13/2026		Flow Adjustment By: _____																													
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 231 ☒ Flow Verification (L/s) Flow Column #: ATP 105 32 mm 0.164 (.139-.169) 36 mm 0.179 (.156-.190) 53 mm 0.246 (.228-.278) 103 mm 0.503 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: 34419 Gauge: 1017 Instrument: 1014 ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																													
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202603A 3/3/2028</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202603D 3/4/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1 10/5/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 3/3/2028	0.080	MP5089	202603B 3/3/2028	0.200	MP5090	202603D 3/4/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Maintenance By: _____ Date: _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:														
Simulator	Serial #	Lot#/Exp																															
0.050	MP5088	202603A 3/3/2028																															
0.080	MP5089	202603B 3/3/2028																															
0.200	MP5090	202603D 3/4/2028																															
0.080 DGS	N/A	23825080A1 10/5/2027																															
Optical Bench Adjustment By: _____		Department Inspection By: SLH																															
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: 28427																															
<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: 1014 Instrument: 1013 Mouth Alcohol Solution Lot #: 2025-D Exp: 09/25/2027 Acetone Stock Solution Lot #: 2025-B Exp: 09/22/2027			
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Optical Bench Adjustment Stability Checks		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A										
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Gauge ID #: _____ Gauge: _____ Instrument: _____		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																
Simulator	Serial Number																																
0.000	MP5086																																
Interferent	MP5087																																
0.050	MP5088																																
0.080	MP5089																																
0.200	MP5090																																
Notes/Suggested Service:		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																															
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
		Digitally signed by Taylor Gutschow Date: 2026.07.27 14:45:49 -04'00'																															
		Digitally signed by Shayla Platt Date: 2026.07.29 12:45:05 -04'00'																															
		Tech Review Admin Review																															

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																												
LAKE HELEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001149 07/13/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:58</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:00</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	13:58	Control Test	0.049	13:59	Air Blank	0.000	14:00	Control Test	0.049	14:00	Air Blank	0.000	14:01	Control Test	0.049	14:01	Air Blank	0.000	14:02	Control Test	0.049	14:02	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LAKE HELEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001149 07/13/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2821</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:04	Control Test	0.079	14:05	Air Blank	0.000	14:06	Control Test	0.077	14:06	Air Blank	0.000	14:07	Control Test	0.078	14:08	Air Blank	0.000	14:08	Control Test	0.078	14:08	Control Test Stats			Average	0.0780		Std Dev	0.0010		Rel Std Dev(%)	1.2821		LAKE HELEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001149 07/13/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:14	Control Test	0.201	14:15	Air Blank	0.000	14:15	Control Test	0.199	14:16	Air Blank	0.000	14:17	Control Test	0.200	14:17	Air Blank	0.000	14:18	Control Test	0.200	14:18	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		LAKE HELEN PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001149 07/13/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:20	Control Test	0.080	14:20	Air Blank	0.000	14:20	Control Test	0.080	14:21	Air Blank	0.000	14:21	Control Test	0.079	14:22	Air Blank	0.000	14:22	Control Test	0.079	14:22	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	13:58																																																																																																																																																													
Control Test	0.049	13:59																																																																																																																																																													
Air Blank	0.000	14:00																																																																																																																																																													
Control Test	0.049	14:00																																																																																																																																																													
Air Blank	0.000	14:01																																																																																																																																																													
Control Test	0.049	14:01																																																																																																																																																													
Air Blank	0.000	14:02																																																																																																																																																													
Control Test	0.049	14:02																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0490																																																																																																																																																														
Std Dev	0.0000																																																																																																																																																														
Rel Std Dev(%)	0.0000																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	14:04																																																																																																																																																													
Control Test	0.079	14:05																																																																																																																																																													
Air Blank	0.000	14:06																																																																																																																																																													
Control Test	0.077	14:06																																																																																																																																																													
Air Blank	0.000	14:07																																																																																																																																																													
Control Test	0.078	14:08																																																																																																																																																													
Air Blank	0.000	14:08																																																																																																																																																													
Control Test	0.078	14:08																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0780																																																																																																																																																														
Std Dev	0.0010																																																																																																																																																														
Rel Std Dev(%)	1.2821																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	14:14																																																																																																																																																													
Control Test	0.201	14:15																																																																																																																																																													
Air Blank	0.000	14:15																																																																																																																																																													
Control Test	0.199	14:16																																																																																																																																																													
Air Blank	0.000	14:17																																																																																																																																																													
Control Test	0.200	14:17																																																																																																																																																													
Air Blank	0.000	14:18																																																																																																																																																													
Control Test	0.200	14:18																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.2000																																																																																																																																																														
Std Dev	0.0010																																																																																																																																																														
Rel Std Dev(%)	0.5000																																																																																																																																																														
Test	g/210L	Time																																																																																																																																																													
Air Blank	0.000	14:20																																																																																																																																																													
Control Test	0.080	14:20																																																																																																																																																													
Air Blank	0.000	14:20																																																																																																																																																													
Control Test	0.080	14:21																																																																																																																																																													
Air Blank	0.000	14:21																																																																																																																																																													
Control Test	0.079	14:22																																																																																																																																																													
Air Blank	0.000	14:22																																																																																																																																																													
Control Test	0.079	14:22																																																																																																																																																													
Control Test Stats																																																																																																																																																															
Average	0.0797																																																																																																																																																														
Std Dev	0.0006																																																																																																																																																														
Rel Std Dev(%)	0.7247																																																																																																																																																														

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE HELEN PD
Time of Inspection: 15:33

Date of Inspection: 07/24/2026

Serial Number: 80-001149
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.078	0.201	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.079	0.200	0.080


Standard Deviations	0.0003	0.0005	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/24/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001149, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001149</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LAKE HELEN PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/24/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:33</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.


Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Date

07/24/2026


LEANDRA HIGGINBOTHAM,

Department Inspector

Service • Integrity • Respect • Quality