

Alcohol Testing Program - Instrument Processing Sheet

Agency: WALTON COUNTY SO

Instrument Serial Number: 80-001312

Date In: 7/23/2026

DI Completion Date: 7/30/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>7/23/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>7/23/2026</u>		Flow Adjustment By: _____																											
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>225</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>34419</u> Gauge: <u>1016</u> Instrument: <u>1013</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																											
						<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A 3/3/2028</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D 3/4/2028</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> </tr> <tr> <td>10/5/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6291	202603A 3/3/2028	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP6292	202603B 3/3/2028	0.200	MP6293	202603D 3/4/2028	0.080 DGS	N/A	23825080A1	10/5/2027						
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																						
						0.050	MP6291	202603A 3/3/2028	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																						
						0.080	MP6292	202603B 3/3/2028																							
0.200	MP6293	202603D 3/4/2028																													
0.080 DGS	N/A	23825080A1																													
		10/5/2027																													
<table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> <tr> <td>0.000</td> <td>MP6294</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP6295</td> <td>25090</td> <td>3/11/2027</td> </tr> <tr> <td>0.100</td> <td>MP6296</td> <td>24390</td> <td>10/29/2026</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>25020</td> <td>1/14/2027</td> </tr> <tr> <td>0.300</td> <td>MP6298</td> <td>25150</td> <td>5/6/2027</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603</td> <td>9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	3/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	1/14/2027	0.300	MP6298	25150	5/6/2027	0.080 DGS	N/A	AG526603	9/23/2027	Department Inspection By: <u>WKP</u> Barometric Pressure Gauge: <u>1008</u> ID#: <u>34420</u> Barometric Pressure ID#: <u>28427</u> Gauge: <u>1009</u> Instrument: <u>1010</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>09/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>09/22/2027</u>	
Simulator	Serial #	Lot #	Expiration																												
0.000	MP6294	N/A	N/A																												
0.040	MP6295	25090	3/11/2027																												
0.100	MP6296	24390	10/29/2026																												
0.200	MP6297	25020	1/14/2027																												
0.300	MP6298	25150	5/6/2027																												
0.080 DGS	N/A	AG526603	9/23/2027																												
☒ Post Optical Bench Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 40%;">Expiration</th> </tr> <tr> <td>0.050</td> <td>MP6291</td> <td>202603A</td> <td>03/03/2028</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202603B</td> <td>03/03/2028</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202603D</td> <td>03/04/2028</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>23825080A1</td> <td>10/5/2027</td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202603A	03/03/2028	0.080	MP6292	202603B	03/03/2028	0.200	MP6293	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	Gauge ID #: <u>34419</u> Gauge: <u>1008</u> Instrument: <u>1008</u>									
Simulator	Serial #	Lot #	Expiration																												
0.050	MP6291	202603A	03/03/2028																												
0.080	MP6292	202603B	03/03/2028																												
0.200	MP6293	202603D	03/04/2028																												
0.080 DGS	N/A	23825080A1	10/5/2027																												
Notes/Suggested Service:		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Optical Bench Adjustment																													
		☒ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other:																													
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																													
		Digitally signed by Taylor Gutschow Date: 2026.08.02 20:49:24 -04'00'																													
		Digitally signed by Kaitlyn Spearin Date: 2026.08.03 12:47:42 -04'00'																													
		Tech Review Admin Review																													

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																															
WALTON COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 07/23/2026 Software: 8100.27 SN 80-001312 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr> <tr><td>Control Test</td><td>0.046</td><td>12:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>0.047</td><td>12:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0473</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2198</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Went</i>	Test	g/210L	Time	Air Blank	0.000	12:02	Control Test	0.046	12:03	Air Blank	0.000	12:03	Control Test	0.047	12:04	Air Blank	0.000	12:04	Control Test	0.047	12:05	Air Blank	0.000	12:05	Control Test Stats			Average	0.0473		Std Dev	0.0006		Rel Std Dev(%)	1.2198		WALTON COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 07/23/2026 Software: 8100.27 SN 80-001312 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr> <tr><td>Control Test</td><td>0.076</td><td>12:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0767</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7531</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Went</i>	Test	g/210L	Time	Air Blank	0.000	12:13	Control Test	0.076	12:14	Air Blank	0.000	12:14	Control Test	0.077	12:15	Air Blank	0.000	12:15	Control Test	0.077	12:16	Air Blank	0.000	12:17	Control Test Stats			Average	0.0767		Std Dev	0.0006		Rel Std Dev(%)	0.7531		WALTON COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 07/23/2026 Software: 8100.27 SN 80-001312 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:32</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.195</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Went</i>	Test	g/210L	Time	Air Blank	0.000	12:32	Control Test	0.196	12:32	Air Blank	0.000	12:33	Control Test	0.196	12:33	Air Blank	0.000	12:34	Control Test	0.195	12:35	Air Blank	0.000	12:35	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		WALTON COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 07/23/2026 Software: 8100.27 SN 80-001312 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature: <i>Went</i>	Test	g/210L	Time	Air Blank	0.000	10:09	Control Test	0.079	10:10	Air Blank	0.000	10:10	Control Test	0.078	10:11	Air Blank	0.000	10:11	Control Test	0.078	10:11	Air Blank	0.000	10:12	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test	0.046	12:03																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.047	12:04																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	0.047	12:05																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0473																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2198																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:13																																																																																																																																																	
Control Test	0.076	12:14																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.077	12:15																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.077	12:16																																																																																																																																																	
Air Blank	0.000	12:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0767																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7531																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:32																																																																																																																																																	
Control Test	0.196	12:32																																																																																																																																																	
Air Blank	0.000	12:33																																																																																																																																																	
Control Test	0.196	12:33																																																																																																																																																	
Air Blank	0.000	12:34																																																																																																																																																	
Control Test	0.195	12:35																																																																																																																																																	
Air Blank	0.000	12:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.079	10:10																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.078	10:11																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.078	10:11																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312 ✓
07/28/2026 09:01:09 *Sud*

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.0910 (-0.0200) 0.1130 (-0.0050)
Sample #2 0.1160 (-0.0050) 0.0930 (0.0130)
Sample #3 0.1120 (0.0150) 0.1090 (0.0240)
Sample #4 0.0980 (0.0290) 0.0950 (0.0310)
Avg % Abs 0.1087 (0.0130) 0.0990 (0.0227)
STD DEV 0.0095 (0.0171) 0.0087 (0.0091)
REL STD DEV 8.698 (131.446) 8.806 (40.031)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.8280 (-0.0160) 1.5480 (-0.0060)
Sample #2 0.8320 (0.0170) 1.5310 (0.0160)
Sample #3 0.8290 (0.0180) 1.5270 (0.0070)
Sample #4 0.8210 (0.0500) 1.5320 (0.0250)
Avg % Abs 0.8273 (0.0283) 1.5300 (0.0160)
STD DEV 0.0057 (0.0188) 0.0026 (0.0090)
REL STD DEV 0.687 (66.249) 0.173 (56.250)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 1.8730 (-0.0110) 3.6310 (-0.0380)
Sample #2 1.8590 (-0.0100) 3.5900 (-0.0120)
Sample #3 1.8680 (0.0230) 3.6040 (0.0040)
Sample #4 1.8830 (0.0220) 3.5960 (0.0000)
Avg % Abs 1.8700 (0.0117) 3.5967 (-0.0027)
STD DEV 0.0121 (0.0188) 0.0070 (0.0083)
REL STD DEV 0.648 (160.890) 0.195 (312.250)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 3.5820 (-0.0200) 6.8930 (0.0020)
Sample #2 3.5850 (-0.0040) 6.8840 (0.0110)
Sample #3 3.5830 (0.0140) 6.8910 (0.0080)
Sample #4 3.5880 (0.0010) 6.8850 (0.0050)
Avg % Abs 3.5853 (0.0037) 6.8867 (0.0080)
STD DEV 0.0025 (0.0093) 0.0038 (0.0030)
REL STD DEV 0.070 (253.407) 0.055 (37.500)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 5.2270 (-0.0090) 9.9730 (-0.0180)
Sample #2 5.2410 (0.0010) 9.9880 (-0.0080)
Sample #3 5.2640 (0.0020) 10.0140 (-0.0230)
Sample #4 5.2460 (0.0200) 10.0130 (-0.0080)
Avg % Abs 5.2503 (0.0077) 10.0050 (-0.0130)
STD DEV 0.0121 (0.0107) 0.0147 (0.0087)
REL STD DEV 0.230 (139.470) 0.147 (66.617)

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312
07/28/2026 09:01:09

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -295.54
First Order Coef 2657.11
Second Order Coef 22.87

<<<< 9um >>>>

Zero Order Coef -132.99
First Order Coef 1315.48
Second Order Coef 12.56

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	-0.0003
0.100	0.100	0.0002
0.200	0.200	-0.0000
0.300	0.300	0.0000

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0001
0.040	0.040	-0.0001
0.100	0.100	0.0000
0.200	0.200	0.0000
0.300	0.300	0.0000

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample

Sample #1	3140.00	3435.00
Sample #2	3148.00	3442.00
Sample #3	3034.00	3429.00
Sample #4	3112.00	3440.00
Avg	3098.0000	3437.0000
STD DEV	58.2752	7.0000
REL STD DEV	1.881	0.204
H2O adjust (mg/l*10k)	711	372

Barometric Pressure = 1008

*****CALIBRATION SUCCESSFUL*****

✓
SW

Post Optical Bench Adjustment Stability Checks 80-001312


SUH

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																			
WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001312 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </table>	Air Blank	0.000	10:41	Control Test	0.050	10:42	Air Blank	0.000	10:42	Control Test	0.050	10:43	Air Blank	0.000	10:43	Control Test	0.051	10:44	Air Blank	0.000	10:45	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001312 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </table>	Air Blank	0.000	10:31	Control Test	0.080	10:32	Air Blank	0.000	10:32	Control Test	0.078	10:33	Air Blank	0.000	10:34	Control Test	0.079	10:34	Air Blank	0.000	10:35	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		WALTON COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/28/2026 Software: 8100.27 SN 80-001312 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>		Air Blank	0.000	10:36	Control Test	0.080	10:37	Air Blank	0.000	10:37	Control Test	0.080	10:37	Air Blank	0.000	10:38	Control Test	0.080	10:38	Air Blank	0.000	10:39	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	10:41																																																																																																				
Control Test	0.050	10:42																																																																																																				
Air Blank	0.000	10:42																																																																																																				
Control Test	0.050	10:43																																																																																																				
Air Blank	0.000	10:43																																																																																																				
Control Test	0.051	10:44																																																																																																				
Air Blank	0.000	10:45																																																																																																				
Control Test Stats																																																																																																						
Average	0.0503																																																																																																					
Std Dev	0.0006																																																																																																					
Rel Std Dev(%)	1.1471																																																																																																					
Air Blank	0.000	10:31																																																																																																				
Control Test	0.080	10:32																																																																																																				
Air Blank	0.000	10:32																																																																																																				
Control Test	0.078	10:33																																																																																																				
Air Blank	0.000	10:34																																																																																																				
Control Test	0.079	10:34																																																																																																				
Air Blank	0.000	10:35																																																																																																				
Control Test Stats																																																																																																						
Average	0.0790																																																																																																					
Std Dev	0.0010																																																																																																					
Rel Std Dev(%)	1.2658																																																																																																					
Air Blank	0.000	10:36																																																																																																				
Control Test	0.080	10:37																																																																																																				
Air Blank	0.000	10:37																																																																																																				
Control Test	0.080	10:37																																																																																																				
Air Blank	0.000	10:38																																																																																																				
Control Test	0.080	10:38																																																																																																				
Air Blank	0.000	10:39																																																																																																				
Control Test Stats																																																																																																						
Average	0.0800																																																																																																					
Std Dev	0.0000																																																																																																					
Rel Std Dev(%)	0.0000																																																																																																					
Operator's Signature <i>Miguel</i>	Operator's Signature <i>Miguel</i>	Operator's Signature <i>Miguel</i>																																																																																																				

see
following
page
Suh

WALTON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001312 ✓
07/28/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.199	10:26
Air Blank	0.000	10:27
Control Test	0.198	10:27
Air Blank	0.000	10:28
Control Test	0.199	10:28
Air Blank	0.000	10:29
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 11:56

Date of Inspection: 07/30/2026

Serial Number: 80-001312
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23725080A1 Exp: 10/05/2027
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.080
0.000	0.051	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081
0.000	0.051	0.078	0.198	0.081
0.000	0.051	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.051	0.079	0.198	0.081
0.000	0.051	0.079	0.198	0.081
0.000	0.051	0.079	0.198	0.081
0.000	0.050	0.079	0.198	0.081

Standard Deviations	0.0005	0.0003	0.0000	0.0009
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WEN-CHI K PIERSON

Signature and Printed Name

07/30/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001312, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001312</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>WALTON COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/30/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:56</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi

Pierson

Digitally signed by Wen-Chi Pierson
Date: 2026.07.31 09:07:44
-04'00'

Date

07/30/2026

WEN-CHI K PIERSON,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality