

Alcohol Testing Program - Instrument Processing Sheet

Agency: Lee County Sheriff's Office Instrument Serial Number: 80-000939

Date In: 3/25/2026 DI Completion Date: 4/2/2026 ☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: <u>TDG</u> Date: <u>3/25/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>4/2/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: No box.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>215</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.152 (.139-.169) 36 mm 0.160 (.156-.190) 53 mm 0.250 (.228-.278) 103 mm 0.531 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>26932</u> Gauge: <u>1020</u> Instrument: <u>1014</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> <th style="width:45%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202406K</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>6/19/2026</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202406L</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>6/19/2026</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202406N</td> <td>☐ Other: _____</td> </tr> <tr> <td>6/20/2026</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG429602</td> <td></td> </tr> <tr> <td>10/22/2026</td> <td></td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K	☐ Battery Replacement	6/19/2026	☐ Dry Gas Regulator Replacement	0.080	MP6287	202406L	☐ Tank Sensor Tare	6/19/2026	☐ Breath Tube Replacement	0.200	MP6288	202406N	☐ Other: _____	6/20/2026		0.080 DGS	N/A	AG429602		10/22/2026	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP6286	202406K	☐ Battery Replacement																										
								6/19/2026	☐ Dry Gas Regulator Replacement																										
						0.080	MP6287	202406L	☐ Tank Sensor Tare																										
								6/19/2026	☐ Breath Tube Replacement																										
						0.200	MP6288	202406N	☐ Other: _____																										
								6/20/2026																											
						0.080 DGS	N/A	AG429602																											
10/22/2026																																			
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:45%;">Expiration</th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A								
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:45%;">Expiration</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A																
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Gauge ID #: _____ Gauge: _____ Instrument: _____																																			
Notes/Suggested Service:																																			
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td colspan="2"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC </td> </tr> <tr> <td colspan="2"> ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use </td> </tr> <tr> <td colspan="2"> ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td style="width:50%;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.04.06 16:14:22 -04'00' </td> <td style="width:50%; text-align: right;"> Shayla Platt Digitally signed by Shayla Platt Date: 2026.04.09 08:42:03 -04'00' </td> </tr> </table>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC		☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use		☒ Conduct an Agency Inspection Before Evidentiary Use		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.04.06 16:14:22 -04'00'	Shayla Platt Digitally signed by Shayla Platt Date: 2026.04.09 08:42:03 -04'00'																										
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																			
☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																			
☒ Conduct an Agency Inspection Before Evidentiary Use																																			
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.04.06 16:14:22 -04'00'	Shayla Platt Digitally signed by Shayla Platt Date: 2026.04.09 08:42:03 -04'00'																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%;">Tech Review</td> <td style="width:50%;">Admin Review</td> </tr> </table>		Tech Review	Admin Review																																
Tech Review	Admin Review																																		

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 04/02/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	10:16	Control Test	0.049	10:17	Air Blank	0.000	10:17	Control Test	0.049	10:18	Air Blank	0.000	10:19	Control Test	0.049	10:19	Air Blank	0.000	10:20	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 04/02/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	10:24	Control Test	0.079	10:24	Air Blank	0.000	10:25	Control Test	0.079	10:26	Air Blank	0.000	10:26	Control Test	0.079	10:27	Air Blank	0.000	10:27	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 04/02/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.196</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.197</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5076</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	10:10	Control Test	0.198	10:11	Air Blank	0.000	10:11	Control Test	0.196	10:12	Air Blank	0.000	10:12	Control Test	0.197	10:13	Air Blank	0.000	10:14	Control Test Stats			Average	0.1970		Std Dev	0.0010		Rel Std Dev(%)	0.5076		LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 04/02/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>MC</i>	Test	g/210L	Time	Air Blank	0.000	10:07	Control Test	0.078	10:07	Air Blank	0.000	10:08	Control Test	0.078	10:08	Air Blank	0.000	10:08	Control Test	0.078	10:09	Air Blank	0.000	10:09	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:16																																																																																																																																																	
Control Test	0.049	10:17																																																																																																																																																	
Air Blank	0.000	10:17																																																																																																																																																	
Control Test	0.049	10:18																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test	0.049	10:19																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test	0.079	10:24																																																																																																																																																	
Air Blank	0.000	10:25																																																																																																																																																	
Control Test	0.079	10:26																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.079	10:27																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.198	10:11																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.196	10:12																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test	0.197	10:13																																																																																																																																																	
Air Blank	0.000	10:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5076																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.078	10:07																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.078	10:08																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.078	10:09																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 12:44

Date of Inspection: 04/02/2026

Serial Number: 80-000939
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.080	0.198	0.078
0.000	0.049	0.079	0.198	0.078

Standard Deviations	0.0004	0.0003	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/02/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000939, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000939</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/02/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:44</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/02/2026

Date

**Taylor
Gutschow**

**TAYLOR D GUTSCHOW,
Department Inspector**

Digitally signed by Taylor
Gutschow
Date: 2026.04.02 15:08:51
-04'00'