



INSTRUMENT PROCESSING SHEET

Agency Lafayette County SO

S/N 80-000776

Florida Department of Law Enforcement

Date In 12/30/2024 DI Completion Date 1/30/2025

Ship P/U H/D CMI EE


Intake By ALL Date 12/30/2024 Quality Checks By PN Date 1/29/2025 Flow Calibration By Date
Annual Registration Return from CMI / EE
Visual Inspection: Case Handle Keyboard Dry Gas Shelf Feet Breath Tube Ports Screws Tight
Other Equipment/ Accessories: Power cord Printer Cable Static Bag 12V DC Cable
Notes: dropoff, no box
Breath Tube Screen Replace External O-Rings Instrument Set Up Verified R-Value 163 Flow Verification (L/s)
Flow Column # ATP 102
32 mm .156 (.139 - .169)
36 mm .164 (.156 - .190)
53 mm .234 (.228 - .278)
103 mm .492 (.447 - .547)
Barometric Pressure Check Gauge ID # 30793
Stability Checks
Simulator Serial # Lot #/Exp
0.050 MP6291 202406K 06/19/2026
0.080 MP6292 202406L 06/19/2026
0.200 MP6293 202406N 06/20/2026
0.080 DGS N/A AG325603 09/13/2025
Flow Column # 5L/min - 17mm 15L/min - 53mm 30L/min - 103mm
R-Value Post Calibration Verification (L/s)
Flow Column # 32 mm 36 mm 53 mm 103 mm
Maintenance By Date
Battery Replacement Dry Gas Regulator Replacement Breath Tube Replacement Other

Calibration Adjustment By Department Inspection By PN
Barometric Pressure Gauge ID #
Simulator Serial # Lot # Expiration
0.000 N/A N/A
0.040
0.100
0.200
0.300
0.080 DGS N/A
Post Calibration Adjustment Stability Checks
Simulator Serial # Lot # Expiration
0.050
0.080
0.200
0.080 DGS N/A
Barometric Pressure ID# 30793
Gauge 1018 Instrument 1012
Mouth Alcohol Solution Lot # 2024-A
Acetone Stock Solution Lot # 2023-B
Simulator Serial Number
0.000 MP6289
Interferent MP6290
0.050 MP6291
0.080 MP6292
0.200 MP6293
Attachments
Form 41 Stability Checks Calibration Certificate Calibration Adjustment Post-Stability Checks Flow Calibration Form 40 Other Stability Water
Instrument Complies with Chapter 11D-8, FAC
Instrument Does Not Comply with Chapter 11D-8, FAC
Return to/Place into Evidentiary Use
Remain Out of Evidentiary Use
Conduct an Agency Inspection Before Evidentiary Use

Notes/Suggested Service:
Miscommunication resulted in agency name being changed to "Troop H" during Quality Checks. Agency name was correctly changed back after the inspection was completed. A post-inspection stability check was performed with water to ensure correct agency name was retained in instrument. (see Stability Water attachment) PN 1/30/205
Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.02.03 11:50:50 -05'00'
Shayla Platt Digitally signed by Shayla Platt Date: 2025.02.03 13:28:27 -05'00'
Tech Review / Date Admin Review / Date

TROOP H
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
01/29/2025
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:19 |
| Control Test | 0.048 | 11:19 |
| Air Blank | 0.000 | 11:20 |
| Control Test | 0.048 | 11:20 |
| Air Blank | 0.000 | 11:21 |
| Control Test | 0.048 | 11:22 |
| Air Blank | 0.000 | 11:22 |
| Control Test Stats | | |
| Average | 0.0480 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |


Operator's Signature

TROOP H
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
01/29/2025
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:24 |
| Control Test | 0.080 | 11:25 |
| Air Blank | 0.000 | 11:25 |
| Control Test | 0.080 | 11:26 |
| Air Blank | 0.000 | 11:26 |
| Control Test | 0.079 | 11:27 |
| Air Blank | 0.000 | 11:27 |
| Control Test Stats | | |
| Average | 0.0797 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.7247 | |


Operator's Signature

TROOP H
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
01/29/2025
Software: 8100.27


| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:31 |
| Control Test | 0.201 | 11:32 |
| Air Blank | 0.000 | 11:32 |
| Control Test | 0.200 | 11:33 |
| Air Blank | 0.000 | 11:33 |
| Control Test | 0.200 | 11:34 |
| Air Blank | 0.000 | 11:35 |
| Control Test Stats | | |
| Average | 0.2003 | |
| Std Dev | 0.0006 | |
| Rel Std Dev(%) | 0.2882 | |


Operator's Signature

TROOP H
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
01/29/2025
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 11:38 |
| Control Test | 0.081 | 11:38 |
| Air Blank | 0.000 | 11:38 |
| Control Test | 0.082 | 11:39 |
| Air Blank | 0.000 | 11:39 |
| Control Test | 0.080 | 11:40 |
| Air Blank | 0.000 | 11:40 |
| Control Test Stats | | |
| Average | 0.0813 | |
| Std Dev | 0.0010 | |
| Rel Std Dev(%) | 1.2346 | |

DGS


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ~~TROOP H~~ LAFAYETTE COUNTY SO

Serial Number: 80-000776

Time of Inspection: 12:59 1/31/2025 Date of Inspection: 01/30/2025
PN

Software: 8100.27

| Check or Test | YES | NO | Check or Test | YES | NO |
|---|-----|----|--|-----|----|
| Diagnostic Check (Pre-Inspection): OK | Yes | | Date and/or Time Adjusted | | No |
| Minimum Sample Volume Check: OK | Yes | | Barometric Pressure Sensor Check: OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | | Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | | Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026 | 0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026 | 0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025 |
|----------------------------|---|---|---|---|
| 0.000 | 0.047 | 0.080 | 0.202 | 0.081 |
| 0.000 | 0.047 | 0.080 | 0.203 | 0.081 |
| 0.000 | 0.047 | 0.080 | 0.203 | 0.081 |
| 0.000 | 0.047 | 0.082 | 0.203 | 0.081 |
| 0.000 | 0.047 | 0.080 | 0.204 | 0.081 |
| 0.000 | 0.048 | 0.080 | 0.205 | 0.081 |
| 0.000 | 0.048 | 0.080 | 0.203 | 0.081 |
| 0.000 | 0.046 | 0.081 | 0.203 | 0.080 |
| 0.000 | 0.047 | 0.081 | 0.204 | 0.080 |
| 0.000 | 0.047 | 0.081 | 0.203 | 0.081 |


| | | | | |
|---------------------|--------|--------|--------|--------|
| Standard Deviations | 0.0005 | 0.0007 | 0.0008 | 0.0004 |
|---------------------|--------|--------|--------|--------|

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:


The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO
Signature and Printed Name

01/30/2025
Date



Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000776, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

| | | | |
|-------------------|-------------------|-------------------------------|-------|
| Serial Number: | <u>80-000776</u> | UNCERTAINTY* ± | |
| Owning Agency: | <u>TROOPH</u> | 0.050 g/210 L | 0.004 |
| Calibration Date: | <u>01/30/2025</u> | 0.080 g/210 L | 0.004 |
| Calibration Time: | <u>12:59</u> | 0.200 g/210 L | 0.007 |
| | | 0.080 g/210 L Dry Gas Control | 0.005 |

LAFAYETTE COUNTY SO
1/31/2025 PN

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



PHIL NICODEMO,
Department Inspector

01/30/2025

Date

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

STABILITY (WATER)

PN 1/30/2025



LAFAYETTE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000776
01/30/2025
Software: 8100.27

| Test | g/210L | Time |
|--------------------|--------|-------|
| Air Blank | 0.000 | 14:06 |
| Control Test | 0.000 | 14:07 |
| Air Blank | 0.000 | 14:07 |
| Control Test | 0.000 | 14:08 |
| Air Blank | 0.000 | 14:09 |
| Control Test | 0.000 | 14:09 |
| Air Blank | 0.000 | 14:10 |
| Control Test Stats | | |
| Average | 0.0000 | |
| Std Dev | 0.0000 | |
| Rel Std Dev(%) | 0.0000 | |

Operator's Signature