

Stability Checks

80-000770 subd
8/13/25

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet
<p>Performed Root Case Analysis</p> <p>HAMILTON COUNTY SO Intoxilyzer - Alcolho Analyzer Model 8000 SN: 80-000770 08/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.001 11:23 Control Test 0.049 11:24 Air Blank 0.100 11:25 Control Test 0.150 11:25 Air Blank 0.100 11:26 Control Test 0.049 11:27 Air Blank 0.100 11:27 Control Test Stats Average 1.0493 Std Dev 1.0006 Rel Std Dev(%) 1.1703</p> <p>Operator's Signature <i>Missbold</i></p>	<p>Performed Root Case Analysis</p> <p>HAMILTON COUNTY SO Intoxilyzer - Alcolho Analyzer Model 8000 SN: 80-000770 08/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:29 Control Test 0.080 11:30 Air Blank 0.000 11:30 Control Test 0.080 11:31 Air Blank 0.000 11:32 Control Test 0.079 11:32 Air Blank 0.000 11:33 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247</p> <p>Operator's Signature <i>Missbold</i></p>	<p>Performed Root Case Analysis</p> <p>HAMILTON COUNTY SO Intoxilyzer - Alcolho Analyzer Model 8000 SN: 80-000770 08/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:17 Control Test 0.202 11:18 Air Blank 0.000 11:18 Control Test 0.202 11:19 Air Blank 0.000 11:20 Control Test 0.202 11:20 Air Blank 0.000 11:21 Control Test Stats Average 0.2120 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <i>Missbold</i></p>	<p>Performed Root Case Analysis</p> <p>HAMILTON COUNTY SO Intoxilyzer - Alcolho Analyzer Model 8000 SN: 80-000770 08/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:35 Control Test 0.078 11:35 Air Blank 0.000 11:36 Control Test 0.079 11:36 Air Blank 0.000 11:37 Control Test 0.079 11:37 Air Blank 0.000 11:37 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339</p> <p>Operator's Signature <i>Missbold</i></p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HAMILTON COUNTY SO
Time of Inspection: 11:53

Date of Inspection: 08/14/2025

Serial Number: 80-000770
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.050	0.079	0.202	0.079
0.000	0.050	0.079	0.201	0.078
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079
0.000	0.051	0.080	0.201	0.079
0.000	0.050	0.080	0.202	0.079


Standard Deviations	0.0006	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5



Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

08/14/2025
 Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 2331 Phillips Road
 Tallahassee, FL 32308

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000770, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000770</u>	UNCERTAINTY* ±	
Owning Agency:	<u>HAMILTON COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/14/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:53</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


 LEANDRA HIGGINBOTHAM,
 Department Inspector

08/14/2025
 Date

FDLE/ATP Form 69 October 2024
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality