

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: **MANATEE COUNTY SO**
Time of Inspection: **09:19**

Date of Inspection: **06/12/2023**

Serial Number: **80-006565**
Software: **8100.27**

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/14/2024 <i>06/15/2024</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402357111 Exp: 02/10/2025
0.000	0.050	0.080	0.200	0.082
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.080	0.200	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



WILLIAM D COLEMAN

Signature and Printed Name

06/12/2023
Date



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: Manatee CSO

Instrument Serial Number: 80-006565

AGENCY INSPECTION DISCREPANCY

Agency Inspector: William Coleman

Inspection Date: 05/15/2023

- Agency inspection discrepancy options including test repeats, calendar month, expired solutions, and lot numbers.

REQUESTED CORRECTIVE ACTION

- Requested corrective action options such as recording amendments, providing explanations, and removing instruments.

Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.06.15 14:22:19 -04'00' Signature of Alcohol Testing Program Staff Member

06/15/2023

Date

Ammended
**Florida Department of Law Enforcement
 Alcohol Testing Program**

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
 Time of Inspection: 09:50

Date of Inspection: 05/15/2023

Serial Number: 80-006565
 Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202112A Exp: 12/07/2023	0.08g/210L Test (g/210L) Lot#:202108D Exp: 08/17/2023	0.20g/210L Test (g/210L) Lot#:2021106B <i>WDC</i> Exp: 06/22/2023 <i>202106B</i>	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402357111 Exp: 02/10/2025
0.000	0.051	0.081	0.200	0.081
0.000	0.051	0.081	0.200	0.081
0.000	0.051	0.081	0.201	0.080


Number of Simulators Used: 5

Remarks:

Wrong .20 Solution Lot # entered. (WDC)

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

WILLIAM COLEMAN

05/15/2023
Date