

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083	≤0.003 of Wet																																																																																																																																																															
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																																
PORT ORANGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005402 09/25/2025 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr><tr><td>Control Test</td><td>0.048</td><td>14:35</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:35</td></tr><tr><td>Control Test</td><td>0.048</td><td>14:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:36</td></tr><tr><td>Control Test</td><td>0.047</td><td>14:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:37</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0477</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr></table> menpi Operator's Signature					Test	g/210L	Time	Air Blank	0.000	14:34	Control Test	0.048	14:35	Air Blank	0.000	14:35	Control Test	0.048	14:36	Air Blank	0.000	14:36	Control Test	0.047	14:37	Air Blank	0.000	14:37	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		PORT ORANGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005402 09/25/2025 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:47</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:48</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:49</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr></table> menpi Operator's Signature					Test	g/210L	Time	Air Blank	0.000	14:46	Control Test	0.077	14:46	Air Blank	0.000	14:47	Control Test	0.077	14:47	Air Blank	0.000	14:48	Control Test	0.078	14:49	Air Blank	0.000	14:49	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466		PORT ORANGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005402 09/25/2025 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:55</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.1963</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr></table> menpi Operator's Signature					Test	g/210L	Time	Air Blank	0.000	14:51	Control Test	0.197	14:52	Air Blank	0.000	14:52	Control Test	0.196	14:53	Air Blank	0.000	14:54	Control Test	0.196	14:54	Air Blank	0.000	14:55	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		PORT ORANGE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005402 09/25/2025 Software: 8100.27 <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr><tr><td>Control Test</td><td>0.078</td><td>14:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:58</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr><tr><td>Control Test</td><td>0.077</td><td>14:59</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr><tr><td colspan="3">Control Test Stats</td></tr><tr><td>Average</td><td>0.0773</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr></table> menpi Operator's Signature					Test	g/210L	Time	Air Blank	0.000	14:57	Control Test	0.078	14:57	Air Blank	0.000	14:58	Control Test	0.077	14:58	Air Blank	0.000	14:58	Control Test	0.077	14:59	Air Blank	0.000	14:59	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																																	
Air Blank	0.000	14:34																																																																																																																																																																	
Control Test	0.048	14:35																																																																																																																																																																	
Air Blank	0.000	14:35																																																																																																																																																																	
Control Test	0.048	14:36																																																																																																																																																																	
Air Blank	0.000	14:36																																																																																																																																																																	
Control Test	0.047	14:37																																																																																																																																																																	
Air Blank	0.000	14:37																																																																																																																																																																	
Control Test Stats																																																																																																																																																																			
Average	0.0477																																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																																	
Control Test	0.077	14:46																																																																																																																																																																	
Air Blank	0.000	14:47																																																																																																																																																																	
Control Test	0.077	14:47																																																																																																																																																																	
Air Blank	0.000	14:48																																																																																																																																																																	
Control Test	0.078	14:49																																																																																																																																																																	
Air Blank	0.000	14:49																																																																																																																																																																	
Control Test Stats																																																																																																																																																																			
Average	0.0773																																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																																	
Control Test	0.197	14:52																																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																																	
Control Test	0.196	14:53																																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																																	
Control Test	0.196	14:54																																																																																																																																																																	
Air Blank	0.000	14:55																																																																																																																																																																	
Control Test Stats																																																																																																																																																																			
Average	0.1963																																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																																		
Rel Std Dev(%)	0.2941																																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																																	
Air Blank	0.000	14:57																																																																																																																																																																	
Control Test	0.078	14:57																																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																																	
Control Test	0.077	14:58																																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																																	
Control Test	0.077	14:59																																																																																																																																																																	
Air Blank	0.000	14:59																																																																																																																																																																	
Control Test Stats																																																																																																																																																																			
Average	0.0773																																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																																		

0-08g/210L
WET

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 60-005402
09/25/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.078	14:41
Air Blank	0.000	14:41
Control Test	0.076	14:42
Air Blank	0.000	14:42
Control Test	0.077	14:43
Air Blank	0.000	14:44
Control Test Stats		
Average	0.0770	
Std Dev	0.0010	
Rel Std Dev(%)	1.2987	


Operator's Signature

second control Test
outside of nominal
range. Performed
root cause analysis
and corrected simulator
connection to instrument,
overan stability.

NKp 09/25/2025

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ORANGE PD
Time of Inspection: 10:39

Date of Inspection: 09/25/2025

Serial Number: 80-005402
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi K Pierson 09/25/2025

WEN-CHI K PIERSON
Signature and Printed Name

09/25/2025
Date

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ORANGE PD
Time of Inspection: 13:49

Date of Inspection: 10/02/2025

Serial Number: 80-005402
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AC510701 Exp: 04/17/2027
0.000	0.048	0.079	0.196	0.077
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.048	0.079	0.197	0.076
0.000	0.049	0.079	0.197	0.077
0.000	0.049	0.079	0.197	0.076
0.000	0.050	0.078	0.197	0.076
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.078	0.196	0.076

Standard Deviations	0.0005	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:



The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM
Signature and Printed Name

10/02/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005402, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005402</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT ORANGE PD</u>	0.050 g/210 L	0.004
Calibration Date:	<u>10/02/2025</u>	0.080 g/210 L	0.004
Calibration Time:	<u>13:49</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

10/02/2025 Date
LEANDRA HIGGINBOTHAM,
Department Inspector

Service • Integrity • Respect • Quality