

# Instrument Processing Sheet

Agency: St. Petersburg Police Department Instrument Serial Number: 80-001151  
 Date In: 10/30/2025 DI Completion Date: 12/10/2025  Ship  P/U  H/D  CMI  EE

<b>Intake</b> By: <u>TDG</u> Date: <u>10/30/25</u>	<b>Quality Checks</b> By: <u>TDG</u> Date: <u>12/9/25</u>	<b>Flow Adjustment</b> By: _____ Date: _____															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Dropped off by Agency Inspector. No box.	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>246</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP101</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.515</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1016</u> Instrument: <u>1016</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG526603 9/23/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP6286	202406K 6/19/2026	0.080	MP6287	202406L 6/19/2026	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603 9/23/2027	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
Simulator	Serial #	Lot#/Exp															
0.050	MP6286	202406K 6/19/2026															
0.080	MP6287	202406L 6/19/2026															
0.200	MP6288	202406N 6/20/2026															
0.080 DGS	N/A	AG526603 9/23/2027															
		<b>Maintenance</b> By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other: _____															

<b>Optical Bench Adjustment</b> By: _____	<b>Department Inspection</b> By: <u>TDG</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>33364</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1017</u> Instrument: <u>1017</u> Mouth Alcohol Solution Lot #: <u>2025-C</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6284																																								
Interferent	MP6285																																								
0.050	MP6286																																								
0.080	MP6287																																								
0.200	MP6288																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Barometric Pressure Gauge: _____ ID#: _____	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other: _____																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								

Notes/Suggested Service: Returning to the agency in a loaner box. (TDG 12/10/25)	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Digitally signed by <table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center;"> <b>Shayla Platt</b>                      Tech Review                      Date: 2025.12.14 13:46:19 -05'00'                 </td> <td style="width: 50%; text-align: center;"> <b>LeAndra Higginbotham</b>                      Admin Review                      Date: 2025.12.15 11:40:55 -05'00'                 </td> </tr> </table>	<b>Shayla Platt</b> Tech Review Date: 2025.12.14 13:46:19 -05'00'	<b>LeAndra Higginbotham</b> Admin Review Date: 2025.12.15 11:40:55 -05'00'
<b>Shayla Platt</b> Tech Review Date: 2025.12.14 13:46:19 -05'00'	<b>LeAndra Higginbotham</b> Admin Review Date: 2025.12.15 11:40:55 -05'00'		

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ST. PETERSBURG PD  
Time of Inspection: 13:02

Date of Inspection: 12/09/2025

Serial Number: 80-001151  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/09/2025  
Date

# Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
<p>ST. PETERSBURG PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001151 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:24</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:22	Control Test	0.049	13:22	Air Blank	0.000	13:23	Control Test	0.049	13:23	Air Blank	0.000	13:24	Control Test	0.050	13:25	Air Blank	0.000	13:25	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel. Std Dev(%)	1.1703		<p>ST. PETERSBURG PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001151 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.081</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:39	Control Test	0.081	13:40	Air Blank	0.000	13:40	Control Test	0.081	13:41	Air Blank	0.000	13:41	Control Test	0.080	13:42	Air Blank	0.000	13:43	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel. Std Dev(%)	0.7157		<p>ST. PETERSBURG PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001151 12/05/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.201</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:17</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.5000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:15	Control Test	0.201	13:16	Air Blank	0.000	13:16	Control Test	0.199	13:17	Air Blank	0.000	13:17	Control Test	0.200	13:16	Air Blank	0.000	13:19	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel. Std Dev(%)	0.5000		<p>ST. PETERSBURG PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001151 12/09/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:29</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:29	Control Test	0.080	13:29	Air Blank	0.000	13:30	Control Test	0.080	13:30	Air Blank	0.000	13:30	Control Test	0.079	13:31	Air Blank	0.000	13:31	Control Test Stats			Average	0.0797		Std Dev	0.0016		Rel. Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test	0.049	13:22																																																																																																																																																	
Air Blank	0.000	13:23																																																																																																																																																	
Control Test	0.049	13:23																																																																																																																																																	
Air Blank	0.000	13:24																																																																																																																																																	
Control Test	0.050	13:25																																																																																																																																																	
Air Blank	0.000	13:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.081	13:40																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test	0.081	13:41																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.080	13:42																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7157																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.201	13:16																																																																																																																																																	
Air Blank	0.000	13:16																																																																																																																																																	
Control Test	0.199	13:17																																																																																																																																																	
Air Blank	0.000	13:17																																																																																																																																																	
Control Test	0.200	13:16																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel. Std Dev(%)	0.5000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:29																																																																																																																																																	
Control Test	0.080	13:29																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.080	13:30																																																																																																																																																	
Air Blank	0.000	13:30																																																																																																																																																	
Control Test	0.079	13:31																																																																																																																																																	
Air Blank	0.000	13:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel. Std Dev(%)	0.7247																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST PETERSBURG PD  
Time of Inspection: 12:54

Date of Inspection: 12/10/2025

Serial Number: 80-001151  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.199	0.080


Standard Deviations	0.0004	0.0005	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/10/2025  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001151, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001151</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ST.PETERSBURG PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/10/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:54</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor  
Gutschow  
Digitally signed by Taylor  
Gutschow  
Date: 2025.12.10 14:57:17  
-05'00'

12/10/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality