

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> THOMAS, CHRISTOPHER, L	<b>Time</b> 16:51:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/12/2026 16:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> THOMAS, CHRISTOPHER, L	<b>Time</b> 22:05:21	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/04/2026 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 03/05/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> THOMAS, CHRISTOPHER, L	<b>Time</b> 19:08:29	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/05/2026 19:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

**Agency** MARTIN COUNTY SO                      **Date** 04/02/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D                      **Time** 13:02:06                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 10/22/2026

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 04/02/2026 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 02/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GUTSCHOW,TAYLOR,D	<b>Time</b> 11:07:04	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 03/06/2026 13:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

**Agency** MARTIN COUNTY SO                      **Date** 03/09/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D                      **Time** 12:17:22                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

Data Download Date/Time 03/09/2026 12:18

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

<b>Agency</b> MARTIN COUNTY SO	<b>Date</b> 04/07/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> THOMAS, CHRISTOPHER, L	<b>Time</b> 00:37:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/07/2026 00:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

<b>Agency</b> HOMESTEAD PD	<b>Date</b> 01/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ACKERMAN,ELVIS,A	<b>Time</b> 19:11:39	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402806340
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 01/22/2026 19:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

<b>Agency</b> HOMESTEAD PD	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ACKERMAN,ELVIS,A	<b>Time</b> 18:11:26	<b>Software</b> 8100.27
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F
<b>0.08 Gas Lot #</b> 402806340		
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026
		<b>Expiration</b> 07/26/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 18:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

<b>Agency</b> HOMESTEAD PD	<b>Date</b> 03/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ACKERMAN,ELVIS,A	<b>Time</b> 19:56:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402806340
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 07/26/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 03/23/2026 19:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

<b>Agency</b> OCALA PD	<b>Date</b> 03/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> JOEDICKE, WILLIAM, R	<b>Time</b> 00:09:47	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

00: Interferent Detect HOOKED UP WRONG SIM.

**Data Download Date/Time** 03/04/2026 19:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

<b>Agency</b> OCALA PD	<b>Date</b> 04/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> JOEDICKE, WILLIAM, R	<b>Time</b> 00:27:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 31124080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.197
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.000
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance - HOOKED UP WROMG SIM.

**Data Download Date/Time** 04/23/2026 17:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006221

<b>Agency</b> FHP TROOP B	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> MURPHY,JAVON,D	<b>Time</b> 01:05:44	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402909409
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/29/2026
<b>Number of Simulators</b> 5	<b>Compliance</b> Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: NOT ENOUGH MA

**Data Download Date/Time** 03/25/2026 01:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO      **Date** 02/11/2026      **Type** Agency Inspection  
**Inspector** HIGGINBOTHAM, LEANDRA,      **Time** 10:56:22      **Software** 8100.27

**0.05 Lot #**      **0.08 Lot #**      **0.20 Lot #**      **0.08 Gas Lot #**  
**Expiration**      **Expiration**      **Expiration**      **Expiration**

**Number of Simulators**      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

**Data Download Date/Time** 02/12/2026 17:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

**Agency** BREVARD COUNTY SO      **Date** 02/12/2026      **Type** Department Inspection  
**Inspector** HIGGINBOTHAM, LEANDRA,      **Time** 17:06:57      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0006      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.047
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.202
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.082
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.047
0.08 g/210L Test # 10	0.083
0.20 g/210L Test # 10	0.202
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0013
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

**Remarks:**

**Data Download Date/Time** 02/12/2026 17:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

<b>Agency</b> BREVARD COUNTY SO	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WAGNER,ROBERT,F	<b>Time</b> 11:37:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2025	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/25/2026 11:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

<b>Agency</b> BREVARD COUNTY SO	<b>Date</b> 04/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LEME,SANTIAGO,	<b>Time</b> 09:26:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 25925080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/27/2026 09:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

<b>Agency</b> HENDRY COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> LINK,CHRIS,D	<b>Time</b> 07:11:24	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.085
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.085
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.085
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO      **Date** 02/13/2026      **Type** Agency Inspection  
**Inspector** LINK,CHRIS,D      **Time** 07:54:42      **Software** 8100.27

**0.05 Lot #** 202506E      **0.08 Lot #** 202410E      **0.20 Lot #** 202410F      **0.08 Gas Lot #** 402872586  
**Expiration** 06/24/2027      **Expiration** 10/22/2026      **Expiration** 10/30/2026      **Expiration** 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 10:16

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

<b>Agency</b> HENDRY COUNTY SO	<b>Date</b> 03/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LINK,CHRIS,D	<b>Time</b> 07:13:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.085
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2026 10:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

**Agency** HENDRY COUNTY SO                      **Date** 04/14/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D                      **Time** 15:37:22                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.082
		0.20 g/210L Test # 5	0.203
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.203
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.082
0.20 g/210L Test # 9	0.203
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.203
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

**Data Download Date/Time** 04/14/2026 15:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 10:05:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/12/2026 10:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 03/05/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> SPEARIN,KATIE,T	<b>Time</b> 13:23:10	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 28424080A3
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 11/05/2026
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0002	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.047
		0.08 g/210L Test # 4	0.077
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.047
		0.08 g/210L Test # 5	0.077
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.077
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.077
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.077
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

**Data Download Date/Time** 03/05/2026 13:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 04/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> NOVAK, CHRISTOPHER,	<b>Time</b> 10:25:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403411042
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/02/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/22/2026 10:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

<b>Agency</b> CLAY COUNTY SO	<b>Date</b> 01/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> JOHNSON, MATTHEW, P	<b>Time</b> 12:54:35	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 01524080A1
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 02/03/2026 18:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 12:56:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 12:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 12:41:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 12:40

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

**Agency** PALM BEACH CSO                      **Date** 03/09/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D              **Time** 13:03:01                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

---

Data Download Date/Time 03/09/2026 13:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 13:21:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006238

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 15:26:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2026 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 11:58:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 11:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

**Agency** PALM BEACH COUNTY SO      **Date** 02/09/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 13:49:53      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 09/23/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 02/09/2026 13:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 13:20:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 13:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 14:02:38	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 14:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 15:10:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance DRY GAS TUBE NOT CONNECTED.

**Data Download Date/Time** 04/17/2026 15:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 16:39:01	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 16:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 15:42:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:40

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

**Agency** PALM BEACH CSO                      **Date** 03/09/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D              **Time** 13:11:37                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

Data Download Date/Time 03/09/2026 14:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 13:59:13	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 12:47:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2026 12:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAKE, DAVID, A	<b>Time</b> 00:45:23	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 3024028434
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 01/22/2026 01:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/18/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAKE,DAVID,A	<b>Time</b> 23:00:16	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 3024028434
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 02/19/2026 01:54

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAKE,DAVID,A	<b>Time</b> 20:56:57	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 3024028434
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

08: DID NOT CONNECT HOSE.

**Data Download Date/Time** 03/16/2026 12:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

**Agency** FL HIGHWAY PATROL      **Date** 03/19/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 13:35:14      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

**Data Download Date/Time** 03/19/2026 14:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/02/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BLAKE,DAVID,A	<b>Time</b> 01:37:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 3024028434
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/08/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

INTOX FREE OF CRACKS OR DAMAGE

**Data Download Date/Time** 04/03/2026 18:31

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SEARS, COLBY, D	<b>Time</b> 18:28:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.204
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/09/2026 18:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SEARS, COLBY, D	<b>Time</b> 17:42:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/06/2026 17:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SEARS, COLBY, D	<b>Time</b> 17:56:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

08: RFI Detect.

**Data Download Date/Time** 03/06/2026 18:02

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006261

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SEARS, COLBY, D	<b>Time</b> 19:50:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/04/2026 19:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO,NICHOLAS,R	<b>Time</b> 16:06:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.000
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.000
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.000
		0.08 g/210L Test # (Repeat) 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside TOL. SOL CH LOT202510E 10/28/2027

**Data Download Date/Time** 02/02/2026 19:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 02/27/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO, NICHOLAS, R	<b>Time</b> 16:14:39	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/27/2026 16:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 03/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO, NICHOLAS, R	<b>Time</b> 16:34:34	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/29/2026 16:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

<b>Agency</b> VOLUSIA COUNTY SO	<b>Date</b> 04/29/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MALETTO,NICHOLAS,R	<b>Time</b> 18:21:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 403135737
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 08/30/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/29/2026 18:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 01/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 20:02:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/25/2026 19:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 02/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 20:07:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/25/2026 20:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 03/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 21:42:56	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/30/2026 21:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

<b>Agency</b> GAINESVILLE PD	<b>Date</b> 04/24/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> STEMEN, AARON, M	<b>Time</b> 01:49:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/24/2026 01:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

<b>Agency</b> HENDRY COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LINK,CHRIS,D	<b>Time</b> 07:43:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 09:56

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

<b>Agency</b> HENDRY COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> LINK,CHRIS,D	<b>Time</b> 08:25:45	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 10:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

<b>Agency</b> HENDRY COUNTY SO	<b>Date</b> 03/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> LINK,CHRIS,D	<b>Time</b> 07:43:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/06/2026 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

**Agency** HENDRY COUNTY SO                      **Date** 04/14/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D                      **Time** 15:05:58                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 10/22/2026

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

#### Remarks:

Data Download Date/Time 04/14/2026 15:04

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 01/14/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SHELLEY,SAMUEL,L	<b>Time</b> 09:25:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.195
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.196
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.048
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.197
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

TESTED INCORRECT SIM.

**Data Download Date/Time** 01/14/2026 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 02/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SHELLEY,SAMUEL,L	<b>Time</b> 12:38:45	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/09/2026 13:46

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 03/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SHELLEY,SAMUEL,L	<b>Time</b> 11:01:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/17/2026 12:05

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 01/05/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> PIERSON, WEN-CHI, K	<b>Time</b> 09:00:48	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 01/05/2026 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

**Agency** ESCAMBIA COUNTY SO      **Date** 01/05/2026      **Type** Department Inspection  
**Inspector** PIERSON, WEN-CHI, K      **Time** 13:36:14      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.194
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.194
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.193
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.194
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.077
0.20 g/210L Test # 8	0.194
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.194
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.077
0.20 g/210L Test # 10	0.194
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

---

Data Download Date/Time 01/05/2026 14:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 02/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SHELLEY,SAMUEL,L	<b>Time</b> 11:45:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.053
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.053
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 02/09/2026 12:50

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 03/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SHELLEY,SAMUEL,L	<b>Time</b> 11:50:59	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 03/17/2026 14:28

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

<b>Agency</b> ESCAMBIA COUNTY SO	<b>Date</b> 04/21/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GRUS,TRENTON,L	<b>Time</b> 10:33:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402987895
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/06/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

IN COMPLIANCE

**Data Download Date/Time** 04/21/2026 12:16

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

<b>Agency</b> MIAMI PD	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FERNANDEZ,ANGEL,	<b>Time</b> 11:45:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402732434
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/28/2026

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

<b>Agency</b> MIAMI PD	<b>Date</b> 03/23/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FERNANDEZ,ANGEL,	<b>Time</b> 18:06:34	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403318190
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/11/2028

  

<b>Number of Simulators</b> 5	<b>Compliance</b> Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/23/2026 18:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

<b>Agency</b> MIAMI PD	<b>Date</b> 04/24/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> FERNANDEZ, ANGEL,	<b>Time</b> 10:25:18	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403318190
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 04/11/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.182
Mouth Alcohol Test: Slope Not Met	Yes	0.20 g/210L Test # (Repeat) 1	0.200
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.045
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.20 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.08 g/210L Dry Gas Std Test: OK	Yes	0.20 g/210L Test # 2	0.182
		0.20 g/210L Test # (Repeat) 2	0.201
		0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.183
		0.20 g/210L Test # (Repeat) 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance. 20: CHANGE SOLUTIONS e.

**Data Download Date/Time** 04/24/2026 10:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

<b>Agency</b> MONROE COUNTY SO	<b>Date</b> 01/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> GUTSCHOW,TAYLOR,D	<b>Time</b> 11:05:06	<b>Software</b> 8100.27

<b>0.05 Lot #</b>	<b>0.08 Lot #</b>	<b>0.20 Lot #</b>	<b>0.08 Gas Lot #</b>
<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>	<b>Expiration</b>

**Number of Simulators** **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

**Remarks:**

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

**Data Download Date/Time** 01/22/2026 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

**Agency** MONROE COUNTY SO      **Date** 01/23/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 13:02:33      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 09/23/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.077
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.196
		0.08 g/210L Dry Gas Std Test # 5	0.077
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.077
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.077
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Data Download Date/Time 01/23/2026 13:01

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

<b>Agency</b> MONROE COUNTY SO	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ZITO, ANTHONY, P	<b>Time</b> 12:41:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207508
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/02/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

PK

**Data Download Date/Time** 02/16/2026 12:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

<b>Agency</b> MONROE COUNTY SO	<b>Date</b> 03/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ZITO, ANTHONY, P	<b>Time</b> 14:41:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207508
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/02/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

PK MAIN

**Data Download Date/Time** 03/12/2026 14:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

<b>Agency</b> MONROE COUNTY SO	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ZITO, ANTHONY, P	<b>Time</b> 14:30:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202410F	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403411042
<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/02/2028

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.191
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

POST EVIDENTIARY AL. GOING TO FDLE

**Data Download Date/Time** 03/25/2026 14:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

<b>Agency</b> MONROE COUNTY SO	<b>Date</b> 04/28/2026	<b>Type</b> Department Inspection
<b>Inspector</b> GUTSCHOW, TAYLOR, D	<b>Time</b> 11:50:40	<b>Software</b> 8100.27

<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202603B	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> AG526603
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 03/03/2028	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0004                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.193
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.191
		0.20 g/210L Test # (Repeat) 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.189
		0.20 g/210L Test # (Repeat) 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

0.08 g/210L Test # 5	0.078
0.20 g/210L Test # 5	0.193
0.20 g/210L Test # (Repeat) 5	0.196
0.08 g/210L Dry Gas Std Test # 5	0.079
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.195
0.20 g/210L Test # (Repeat) 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.20 g/210L Test # (Repeat) 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.20 g/210L Test # (Repeat) 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.20 g/210L Test # (Repeat) 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.20 g/210L Test # (Repeat) 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006471

0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0026
0.20 g/210L Test (Std Dev) (Repeat)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

20: Control Outside Tolerance. SIM POWER CORD CAME LOOSE. .

---

**Data Download Date/Time** 04/28/2026 11:49

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL,NEIL,F	<b>Time</b> 11:55:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403374313
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/25/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.084
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/09/2026 12:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL      **Date** 01/20/2026      **Type** Department Inspection  
**Inspector** PIERSON,WEN-CHI,K      **Time** 12:13:43      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.086
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.085
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.087
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.084
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.087
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.083
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.087
		0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

0.08 g/210L Test # 5	0.081
0.20 g/210L Test # 5	0.200
0.08 g/210L Dry Gas Std Test # 5	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.082
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.087
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.081
Interferent Detect Test # 10	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0013

**Remarks:**

08: Control Outside Tolerance.

---

**Data Download Date/Time** 01/26/2026 10:13

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

**Agency** FL HIGHWAY PATROL      **Date** 01/29/2026      **Type** Department Inspection  
**Inspector** PIERSON, WEN-CHI, K      **Time** 10:10:13      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement  
Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

**Remarks:**

---

**Data Download Date/Time** 01/29/2026 14:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/06/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL,NEIL,F	<b>Time</b> 13:02:43	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402946336
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/17/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 02/06/2026 14:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> RUBEL, NEIL, F	<b>Time</b> 12:08:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402946336
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/17/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 03/12/2026 14:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> MCDOWELL, MICHAEL, J	<b>Time</b> 09:01:58	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402946336
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/17/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 04/04/2026 10:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

<b>Agency</b> PINECREST PD	<b>Date</b> 01/15/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WADDELL,NATHANIEL,J	<b>Time</b> 09:10:19	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202410E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/15/2026 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

<b>Agency</b> PINECREST PD	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> VEGA,OBEL,	<b>Time</b> 15:40:39	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:43



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.196
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.047
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

Data Download Date/Time 03/06/2026 15:04

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

<b>Agency</b> BROWARD COUNTY SO	<b>Date</b> 01/14/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER, ANAYA, S	<b>Time</b> 10:54:03	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402783528
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 01/14/2026

  

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

S TILMAN IN TRAINING

**Data Download Date/Time** 01/14/2026 11:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

<b>Agency</b> BROWARD COUNTY SO	<b>Date</b> 02/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER, ANAYA, S	<b>Time</b> 09:38:15	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402783528
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/27/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

S. TILLMAN IN TRAINING.

**Data Download Date/Time** 02/04/2026 09:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

<b>Agency</b> BROWARD COUNTY SO	<b>Date</b> 03/04/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER, ANAYA, S	<b>Time</b> 08:09:31	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402783528
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/27/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

S. TILLAN IN TRAINING.

**Data Download Date/Time** 03/04/2026 08:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

<b>Agency</b> BROWARD COUNTY SO	<b>Date</b> 04/01/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FRAZIER, ANAYA, S	<b>Time</b> 08:24:41	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402783528
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 06/27/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

S. TILLMAN IN TRAINING

**Data Download Date/Time** 04/01/2026 08:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 15:57:11	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 15:57

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

**Agency** PALM BEACH COUNTY SO      **Date** 02/09/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 13:32:43      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0001      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

**Remarks:**

**Data Download Date/Time** 02/09/2026 13:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 15:07:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 16:22:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 13:04:13	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2026 13:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 13:28:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 13:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO      **Date** 02/09/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 13:14:49      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0004      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Data Download Date/Time 02/09/2026 13:12

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 13:05:55	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 13:15

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 13:43:29	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 13:42

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

<b>Agency</b> PALM BEACH COUNTY SO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 12:42:34	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2026 12:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

**Agency** PALM BEACH COUNTY SO      **Date** 04/23/2026      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D      **Time** 15:02:55      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG429602  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 10/22/2026

**Number of Simulators** 5      **Average Standard Deviation** 0.0003      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.196
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.196
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 04/23/2026 15:00

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 13:23:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/16/2026 13:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 02/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 15:24:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/13/2026 15:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

**Agency** PALM BEACH CSO                      **Date** 03/09/2026                      **Type** Department Inspection  
**Inspector** GUTSCHOW,TAYLOR,D              **Time** 12:24:23                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG526603  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 09/23/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0001                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

#### Remarks:

Data Download Date/Time 03/09/2026 12:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 03/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE,JASON,C	<b>Time</b> 15:54:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 19825080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/13/2026 15:52

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006478

<b>Agency</b> PALM BEACH CSO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KARLECKE, JASON, C	<b>Time</b> 14:37:07	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 32324080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/17/2026 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 09:32:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

08: Control Outside Tolerance.

**Data Download Date/Time** 01/12/2026 09:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 13:53:00	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 13:51

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

Agency MANATEE COUNTY SO	Date 03/16/2026	Type Agency Inspection	
Inspector COLEMAN, WILLIAM,	Time 13:54:04	Software 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

Data Download Date/Time 03/16/2026 13:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 04/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> NOVAK, CHRISTOPHER,	<b>Time</b> 09:28:14	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2026	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: .

**Data Download Date/Time** 04/22/2026 09:26

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 09:38:39	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/12/2026 09:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 14:00:32	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 13:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 03/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 09:10:04	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/16/2026 09:06

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 04/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> NOVAK, CHRISTOPHER,	<b>Time</b> 09:13:51	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403141809
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 09/10/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/22/2026 09:12

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

<b>Agency</b> HILLSBOROUGH CSO	<b>Date</b> 01/15/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN, KEVIN, C	<b>Time</b> 08:23:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 29425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 12/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 01/15/2026 09:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

**Agency** HILLSBOROUGH CSO                      **Date** 01/28/2026                      **Type** Department Inspection  
**Inspector** SPEARIN,KATIE,T                      **Time** 11:51:21                      **Software** 8100.27

**0.05 Lot #** 202406K                      **0.08 Lot #** 202406L                      **0.20 Lot #** 202406N                      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026                      **Expiration** 06/19/2026                      **Expiration** 06/20/2026                      **Expiration** 04/17/2027

**Number of Simulators** 5                      **Average Standard Deviation** 0.0003                      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 01/28/2026 12:06

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

<b>Agency</b> HILLSBOROUGH CSO	<b>Date</b> 03/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN,KEVIN,C	<b>Time</b> 10:19:37	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A6
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 03/12/2026 11:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

<b>Agency</b> HILLSBOROUGH CSO	<b>Date</b> 04/09/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> BROWN,KEVIN,C	<b>Time</b> 08:40:26	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 04125080A6
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 03/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

AI B KROME #263389 PRESENT FOR INSPECTION

**Data Download Date/Time** 04/09/2026 09:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

<b>Agency</b> FLAGLER COUNTY SO	<b>Date</b> 01/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WEAVER,DANIEL,	<b>Time</b> 13:48:42	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 01/16/2026 13:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

<b>Agency</b> FLAGLER COUNTY SO	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> WEAVER,DANIEL,	<b>Time</b> 14:24:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	
		Interferent Detect Test # (Repeat) 3	INT

**Remarks:**

00: Ambient Fail. Int Det: RFI Detect - RADIO. MONTHLY INSP

**Data Download Date/Time** 02/28/2026 14:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

<b>Agency</b> FLAGLER COUNTY SO	<b>Date</b> 03/18/2026	<b>Type</b> Department Inspection	
<b>Inspector</b> HIGGINBOTHAM, LEANDRA,	<b>Time</b> 14:36:25	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 23825080A1
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 10/05/2027
<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0004	<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.083
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.083
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.083
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.083
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

---

Data Download Date/Time 03/18/2026 14:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

<b>Agency</b> FLAGLER COUNTY SO	<b>Date</b> 04/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> FAIRCLOTH, EDWARD, W	<b>Time</b> 10:15:24	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202510F	<b>0.08 Gas Lot #</b> 23725080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/29/2027	<b>Expiration</b> 10/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

MONTHLY INSPECTION

**Data Download Date/Time** 04/17/2026 10:12



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 01/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ETTMAN, KRISTOFER, C	<b>Time</b> 04:17:40	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

A F / M A: No Sample Provided.

**Data Download Date/Time** 01/28/2026 11:41

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 02/24/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SAVARD, CHARLES, M	<b>Time</b> 11:55:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/24/2026 12:46

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 03/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> ETTMAN, KRISTOFER, C	<b>Time</b> 02:45:28	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/05/2026 23:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006621

<b>Agency</b> KENNEDY SPACE CENTER	<b>Date</b> 04/25/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SAVARD, CHARLES, M	<b>Time</b> 07:54:17	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A1
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/27/2026 23:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

<b>Agency</b> BOCA RATON PD	<b>Date</b> 01/13/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CRAWFORD, ANDREW,	<b>Time</b> 16:11:22	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

**Remarks:**

08: RFI Detect IN CAR RADIO SALLYPORT. PASS ALL OKAY

**Data Download Date/Time** 01/13/2026 16:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

<b>Agency</b> BOCA RATON PD	<b>Date</b> 02/17/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CRAWFORD, ANDREW,	<b>Time</b> 10:10:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

PASS

**Data Download Date/Time** 02/17/2026 10:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

<b>Agency</b> BOCA RATON PD	<b>Date</b> 03/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CRAWFORD, ANDREW,	<b>Time</b> 15:18:58	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

PASS

**Data Download Date/Time** 03/11/2026 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

<b>Agency</b> BOCA RATON PD	<b>Date</b> 04/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> CRAWFORD, ANDREW,	<b>Time</b> 10:48:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 08425080A2
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/05/2027

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

PASS

**Data Download Date/Time** 04/16/2026 10:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 17:00:18	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135736
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/29/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/30/2026 18:13

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/22/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 16:39:23	<b>Software</b> 8100.27	

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135736
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/29/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 02/22/2026 17:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 13:18:48	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135736
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/29/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/23/2026 16:12

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/26/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 19:27:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403135736
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/29/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

NA

**Data Download Date/Time** 04/26/2026 19:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

<b>Agency</b> FHP	<b>Date</b> 01/22/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> SOUZA,MAYKOLL,G	<b>Time</b> 10:28:23	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403342588
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/09/2028
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/22/2026 10:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> SOUZA,MAYKOLL,G	<b>Time</b> 10:22:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403342588
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 05/09/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/22/2026 10:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 17:36:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 01/30/2026 18:04

**Printed: 30 April 2026**

COBRA ©1997-2014 CMI, Inc.

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 16:51:27	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

N/A

**Data Download Date/Time** 02/22/2026 17:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/20/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 13:33:06	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 03/23/2026 16:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/26/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> KEMP,TREVOR,D	<b>Time</b> 19:15:52	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403268835
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 02/13/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

NA

**Data Download Date/Time** 04/26/2026 19:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 15:53:01	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403419188
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/13/2028

**Number of Simulators** 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/12/2026 15:52

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 09:41:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403419188
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 09:37

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 03/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 10:20:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403419188
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/16/2026 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 04/22/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> NOVAK, CHRISTOPHER,	<b>Time</b> 13:47:02	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403419188
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/13/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 04/22/2026 13:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/11/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> PALESE, ANTHONY, W	<b>Time</b> 20:21:54	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

M/A.

**Data Download Date/Time** 02/14/2026 17:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/26/2026	<b>Type</b> Agency Inspection	
<b>Inspector</b> PALESE, ANTHONY, W	<b>Time</b> 17:35:26	<b>Software</b> 8100.27	
<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026
<b>Number of Simulators</b> 5		<b>Compliance</b> Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/26/2026 16:36

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/15/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> PALESE, ANTHONY, W	<b>Time</b> 15:17:08	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 402872586
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 10/12/2026

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

**Remarks:**

Time-Date changed.

**Data Download Date/Time** 04/15/2026 15:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 01/12/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 15:59:33	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202402E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403411042
<b>Expiration</b> 02/27/2026	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/02/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 01/12/2026 15:55

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 02/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 09:47:21	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403411042
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/02/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/16/2026 09:43

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

<b>Agency</b> MANATEE COUNTY SO	<b>Date</b> 03/16/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> COLEMAN, WILLIAM,	<b>Time</b> 08:50:05	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202510E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403411042
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/28/2027	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 08/02/2028

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/16/2026 08:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL      **Date** 02/12/2026      **Type** Department Inspection  
**Inspector** HIGGINBOTHAM, LEANDRA,      **Time** 17:23:46      **Software** 8100.27

**0.05 Lot #** 202406K      **0.08 Lot #** 202406L      **0.20 Lot #** 202406N      **0.08 Gas Lot #** AG510701  
**Expiration** 06/19/2026      **Expiration** 06/19/2026      **Expiration** 06/20/2026      **Expiration** 04/17/2027

**Number of Simulators** 5      **Average Standard Deviation** 0.0005      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.202
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Inspection Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.202
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.053
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.053
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.052
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.053
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0009
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

**Remarks:**

**Data Download Date/Time** 02/12/2026 17:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 23:16:30	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/31/2026 23:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

**Agency** FL HIGHWAY PATROL      **Date** 04/29/2026      **Type** Agency Inspection  
**Inspector** HITCHCOCK, WILLIAM, D      **Time** 21:24:07      **Software** 8100.27

**0.05 Lot #** 202506E      **0.08 Lot #** 202510E      **0.20 Lot #** 202510F      **0.08 Gas Lot #** 403207507  
**Expiration** 06/24/2027      **Expiration** 10/28/2027      **Expiration** 10/29/2027      **Expiration** 11/26/2027

**Number of Simulators** 5      **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.047
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.198
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.047
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.05 g/210L Test # (Repeat) 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

05: Control Outside Tolerance.REPEATED OK

**Data Download Date/Time** 04/29/2026 21:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 01/30/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 00:04:20	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/02/2026 10:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 02/28/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 13:05:12	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 02/28/2026 20:15

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 03/31/2026	<b>Type</b> Agency Inspection
<b>Inspector</b> HITCHCOCK, WILLIAM, D	<b>Time</b> 15:59:49	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202506E	<b>0.08 Lot #</b> 202410E	<b>0.20 Lot #</b> 202410F	<b>0.08 Gas Lot #</b> 403207507
<b>Expiration</b> 06/24/2027	<b>Expiration</b> 10/22/2026	<b>Expiration</b> 10/30/2026	<b>Expiration</b> 11/26/2027

**Number of Simulators** 5

**Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

**Remarks:**

**Data Download Date/Time** 03/31/2026 23:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

<b>Agency</b> FL HIGHWAY PATROL	<b>Date</b> 04/07/2026	<b>Type</b> Department Inspection
<b>Inspector</b> HIGGINBOTHAM, LEANDRA,	<b>Time</b> 17:11:09	<b>Software</b> 8100.27

  

<b>0.05 Lot #</b> 202406K	<b>0.08 Lot #</b> 202406L	<b>0.20 Lot #</b> 202406N	<b>0.08 Gas Lot #</b> 23825080A1
<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/19/2026	<b>Expiration</b> 06/20/2026	<b>Expiration</b> 10/05/2027

  

<b>Number of Simulators</b> 5	<b>Average Standard Deviation</b> 0.0003	<b>Compliance</b> Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.080
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.080
		Interferent Detect Test # 5	INT

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

**Remarks:**

Data Download Date/Time 04/07/2026 17:08

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.