




Alcohol Testing Program - Instrument Processing Sheet

Agency: SARASOTA COUNTY SO Instrument Serial Number: 80-005936  
 Date In: 1/13/2026 DI Completion Date: 1/15/26  Ship  P/U  H/D  CMI  EE

Intake By: <u>KTS</u> Date: <u>1/13/2026</u>		Quality Checks By: <u>KTS</u> Date: <u>1/14/2026</u>		Flow Adjustment By: _____																																									
<input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: AI indicated instrument needs battery change. KTS 1/13/26  Fraying on breath hose- taped up. KTS 1/14/26		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>199</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.492</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1010</u> Instrument: <u>1012</u> <input checked="" type="checkbox"/> Stability Checks		Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																									
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 4/17/2027</td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/2026	0.080	MP5089	202406L 6/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	AG510701 4/17/2027	Maintenance By: <u>KTS</u> Date: <u>1/13/2026</u> <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:																										
Simulator	Serial #	Lot#/Exp																																											
0.050	MP5088	202406K 6/19/2026																																											
0.080	MP5089	202406L 6/19/2026																																											
0.200	MP5090	202406N 6/20/2026																																											
0.080 DGS	N/A	AG510701 4/17/2027																																											
Optical Bench Adjustment By: _____ Barometric Pressure Gauge: _____ ID#: _____		Department Inspection By: <u>KTS</u> Barometric Pressure ID#: <u>28427</u> Gauge: <u>1011</u> Instrument: <u>1013</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/27</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/27</u>																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP5088																																												
0.080	MP5089																																												
0.200	MP5090																																												
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks		<input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																																											
Gauge ID #: _____ Gauge: _____ Instrument: _____		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC  <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use  <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																											
Notes/Suggested Service: Flow verification, barometric pressure check and stabilities conducted on 1/14/26 KTS 1/15/26 Tech Review: Added checkmark to annual box. KTS 1/23/26		Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2026.01.23 12:33:43 -05'00'</small>		LeAndra Higginbotham <small>Digitally signed by LeAndra Higginbotham Date: 2026.01.28 12:09:22 -05'00'</small>																																									
		Tech Review		Admin Review																																									

# Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005936 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:48</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:51</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <p style="text-align: right;">             -----            Operator's Signature         </p>	Test	g/210L	Time	Air Blank	0.000	08:48	Control Test	0.050	08:49	Air Blank	0.000	08:50	Control Test	0.049	08:50	Air Blank	0.000	08:51	Control Test	0.049	08:51	Air Blank	0.000	08:52	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005936 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p style="text-align: right;">             -----            Operator's Signature         </p>	Test	g/210L	Time	Air Blank	0.000	09:03	Control Test	0.080	09:04	Air Blank	0.000	09:04	Control Test	0.080	09:05	Air Blank	0.000	09:06	Control Test	0.080	09:06	Air Blank	0.000	09:07	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005936 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.200</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr> <tr><td>Control Test</td><td>0.201</td><td>09:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:12</td></tr> <tr><td>Control Test</td><td>0.201</td><td>09:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2007</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr> </tbody> </table> <p style="text-align: right;">             -----            Operator's Signature         </p>	Test	g/210L	Time	Air Blank	0.000	09:09	Control Test	0.200	09:10	Air Blank	0.000	09:11	Control Test	0.201	09:11	Air Blank	0.000	09:12	Control Test	0.201	09:13	Air Blank	0.000	09:13	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005935 01/14/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:53</td></tr> <tr><td>Control Test</td><td>0.082</td><td>08:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:54</td></tr> <tr><td>Control Test</td><td>0.082</td><td>08:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test</td><td>0.081</td><td>08:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </tbody> </table> <p style="text-align: right;">             -----            Operator's Signature         </p>	Test	g/210L	Time	Air Blank	0.000	08:53	Control Test	0.082	08:54	Air Blank	0.000	08:54	Control Test	0.082	08:55	Air Blank	0.000	08:55	Control Test	0.081	08:55	Air Blank	0.000	08:56	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:48																																																																																																																																																	
Control Test	0.050	08:49																																																																																																																																																	
Air Blank	0.000	08:50																																																																																																																																																	
Control Test	0.049	08:50																																																																																																																																																	
Air Blank	0.000	08:51																																																																																																																																																	
Control Test	0.049	08:51																																																																																																																																																	
Air Blank	0.000	08:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test	0.080	09:04																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.080	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test	0.080	09:06																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.200	09:10																																																																																																																																																	
Air Blank	0.000	09:11																																																																																																																																																	
Control Test	0.201	09:11																																																																																																																																																	
Air Blank	0.000	09:12																																																																																																																																																	
Control Test	0.201	09:13																																																																																																																																																	
Air Blank	0.000	09:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2877																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:53																																																																																																																																																	
Control Test	0.082	08:54																																																																																																																																																	
Air Blank	0.000	08:54																																																																																																																																																	
Control Test	0.082	08:55																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.081	08:55																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0817																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7070																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO  
Time of Inspection: 14:32

Date of Inspection: 01/15/2026

Serial Number: 80-005936  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.081	0.201	0.082
0.000	0.050	0.080	0.202	0.082
0.000	0.050	0.080	0.201	0.082
0.000	0.051	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.081
0.000	0.051	0.081	0.201	0.080
0.000	0.050	0.081	0.202	0.080
0.000	0.052	0.080	0.202	0.080
0.000	0.050	0.081	0.201	0.080

Standard Deviations	0.0006	0.0005	0.0006	0.0009
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5


**Remarks:**

Int Det: .

Changed simulator o-rings to ensure tight connection to instrument. Repeated interferent detect test. KTS 1/15/26


The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN  
\_\_\_\_\_  
Signature and Printed Name

01/15/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005936, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005936</u>	UNCERTAINTY* ±
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L
Calibration Date:	<u>01/15/2026</u>	0.080 g/ 210 L
Calibration Time:	<u>14:32</u>	0.200 g/ 210 L
		0.080 g/ 210 L Dry Gas Control
		0.005

All results are reported in g/ 210 L.  
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.  
\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).  
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.  
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/15/2026

Date

KATIE SPEARIN,

Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality