

Alcohol Testing Program - Instrument Processing Sheet

Agency: Collier County Sheriff's Office

Instrument Serial Number: 80-000941

Date In: 6/26/2026

DI Completion Date: 7/15/2026

☐ Ship

☒ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>7/14/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>7/15/2026</u>		Flow Adjustment By: _____																															
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: No box.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>249</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm 0.160 (.139-.169) 36 mm 0.179 (.156-.190) 53 mm 0.250 (.228-.278) 103 mm 0.500 (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1020</u> Instrument: <u>1016</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																															
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 20%;">Serial #</th> <th style="width: 25%;">Lot#/Exp</th> <th style="width: 40%;">Maintenance By: _____ Date: _____</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202603A</td> <td>☐ Battery Replacement</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Dry Gas Regulator Replacement</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202603B</td> <td>☐ Tank Sensor Tare</td> </tr> <tr> <td>3/3/2028</td> <td>☐ Breath Tube Replacement</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202603D</td> <td>☐ Other: _____</td> </tr> <tr> <td>3/4/2028</td> <td></td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>28424080A3</td> <td></td> </tr> <tr> <td>11/5/2026</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202603A	☐ Battery Replacement	3/3/2028	☐ Dry Gas Regulator Replacement	0.080	MP6287	202603B	☐ Tank Sensor Tare	3/3/2028	☐ Breath Tube Replacement	0.200	MP6288	202603D	☐ Other: _____	3/4/2028		0.080 DGS	N/A	28424080A3		11/5/2026	
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																										
						0.050	MP6286	202603A	☐ Battery Replacement																										
								3/3/2028	☐ Dry Gas Regulator Replacement																										
0.080	MP6287	202603B	☐ Tank Sensor Tare																																
		3/3/2028	☐ Breath Tube Replacement																																
0.200	MP6288	202603D	☐ Other: _____																																
		3/4/2028																																	
0.080 DGS	N/A	28424080A3																																	
		11/5/2026																																	
Optical Bench Adjustment By: _____		Department Inspection By: <u>TDG</u>																																	
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>33363</u>																																	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1020</u> Instrument: <u>1016</u>																															
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																															
0.040				Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>																															
0.100				Simulator	Serial Number																														
0.200				0.000	MP6284																														
0.300				Interferent	MP6285																														
0.080 DGS	N/A			0.050	MP6286																														
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP6287																														
Simulator	Serial #	Lot #	Expiration	0.200	MP6288																														
0.050				Attachments																															
0.080				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment ☐ Post-Stability Checks ☐ Flow Adjustment ☐ Form 40 ☐ Other: _____																															
0.200																																			
0.080 DGS	N/A																																		
Gauge ID #: _____ Gauge: _____ Instrument: _____																																			
Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																															
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																															
				☒ Conduct an Agency Inspection Before Evidentiary Use																															
Wen-Chi Pierson				LeAndra Higginbotham																															
Digitally signed by Wen-Chi Pierson Date: 2026.07.21 12:56:34 -04'00'				Digitally signed by LeAndra Higginbotham Date: 2026.07.21 17:49:15 -04'00'																															
Tech Review				Admin Review																															

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
COLLIER COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 07/15/2026 Software: 8100.27 SN 80-000941 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1713</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:34	Control Test	0.049	11:34	Air Blank	0.000	11:35	Control Test	0.050	11:36	Air Blank	0.000	11:36	Control Test	0.049	11:37	Air Blank	0.000	11:37	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1713		COLLIER COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 07/15/2026 Software: 8100.27 SN 80-000941 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:42	Control Test	0.078	11:42	Air Blank	0.000	11:43	Control Test	0.077	11:44	Air Blank	0.000	11:44	Control Test	0.078	11:45	Air Blank	0.000	11:45	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		COLLIER COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 07/15/2026 Software: 8100.27 SN 80-000941 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:26	Control Test	0.198	11:27	Air Blank	0.000	11:28	Control Test	0.198	11:28	Air Blank	0.000	11:29	Control Test	0.198	11:30	Air Blank	0.000	11:30	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		COLLIER COUNTY SO Intoxilyzer - Alconol Analyzer Model: 8000 07/15/2026 Software: 8100.27 SN 80-000941 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	11:20	Control Test	0.079	11:21	Air Blank	0.000	11:21	Control Test	0.079	11:21	Air Blank	0.000	11:22	Control Test	0.079	11:22	Air Blank	0.000	11:22	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test	0.049	11:34																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.050	11:36																																																																																																																																																	
Air Blank	0.000	11:36																																																																																																																																																	
Control Test	0.049	11:37																																																																																																																																																	
Air Blank	0.000	11:37																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1713																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test	0.078	11:42																																																																																																																																																	
Air Blank	0.000	11:43																																																																																																																																																	
Control Test	0.077	11:44																																																																																																																																																	
Air Blank	0.000	11:44																																																																																																																																																	
Control Test	0.078	11:45																																																																																																																																																	
Air Blank	0.000	11:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:26																																																																																																																																																	
Control Test	0.198	11:27																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.198	11:28																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.198	11:30																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test	0.079	11:21																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test	0.079	11:21																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test	0.079	11:22																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 15:27

Date of Inspection: 07/15/2026

Serial Number: 80-000941
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.049	0.077	0.198	0.079
0.000	0.049	0.077	0.198	0.079
0.000	0.049	0.077	0.198	0.078
0.000	0.049	0.077	0.198	0.078
0.000	0.049	0.077	0.198	0.078
0.000	0.050	0.077	0.198	0.078
0.000	0.050	0.077	0.198	0.078
0.000	0.049	0.076	0.199	0.078
0.000	0.049	0.075	0.198	0.078
0.000	0.049	0.075	0.197	0.078


Standard Deviations	0.0004	0.0008	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/15/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000941, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000941</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/15/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:27</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.07.16 10:12:19
-04'00'

07/15/2026

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality