Alcohol Testing Program - Instrument Processing Sheet

Agency: FFWCC **Instrument Serial Number:** 80-007496

Date In: 2/11/2026 **DI Completion Date:** 3/2/2026 ☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: WKP Date: 2/11/2026

☒ Annual *
☐ Dropped Off *
☐ Registration *
☐ Return from CMI / EE *
☐ Training Instrument *

Visual Inspection

☒ Case ☒ Handle
☒ Keyboard ☒ Dry Gas Shelf
☒ Feet ☒ Breath Tube
☒ Ports ☒ Screws Tight

Other Equipment/Accessories *

☐ Power Cord ☐ Printer Cable
☐ Static Bag ☐ 12V DC Cable

Notes:

Quality Checks By: KTS Date: 2/25/2026

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value: 208
☒ Flow Verification (L/s)
Flow Column #: ATP103

32 mm	0.152	(.139-.169)
36 mm	0.164	(.156-.190)
53 mm	0.238	(.228-.278)
103 mm	0.500	(.447-.547)

☒ Barometric Pressure Check
Gauge ID #: 28421
Gauge: 1017 Instrument: 1018
☒ Stability Checks

Simulator	Serial #	Lot#/Exp
0.050	MP5088	202406K
		6/19/2026
0.080	MP5089	202406L
		6/19/2026
0.200	MP5090	202406N
		6/20/2026
0.080 DGS	N/A	28424080A3
		11/5/2026

Flow Adjustment By:

Flow Column #:

☐ 5L/min – 17mm
☐ 15L/min – 53mm
☐ 30L/min – 103mm

☐ R-Value:
☐ Post Adjustment Verification (L/S)
Flow Column #:

32 mm	(.139-.169)
36 mm	(.156-.190)
53 mm	(.228-.278)
103 mm	(.447-.547)

Maintenance By: Date:

☐ Battery Replacement
☐ Dry Gas Regulator Replacement
☐ Tank Sensor Tare
☐ Breath Tube Replacement
☐ Other:

Optical Bench Adjustment By:

Barometric Pressure Gauge: ID#:

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Optical Bench Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Gauge ID #:
Gauge: Instrument:

Department Inspection By: WKP

Barometric Pressure ID#: 34419
Gauge: 1021 Instrument: 1022
Mouth Alcohol Solution Lot #: 2025-D Exp: 9/25/2027
Acetone Stock Solution Lot #: 2025-B Exp: 9/22/2027

Simulator	Serial Number
0.000	MP5086
Interferent	MP5087
0.050	MP5088
0.080	MP5089
0.200	MP5090

Attachments

☒ Form 41
☒ Stability Checks
☒ Calibration Certificate
☐ Optical Bench Adjustment
☐ Post-Stability Checks
☐ Flow Adjustment
☒ Form 40
☐ Other:

Notes/Suggested Service: * Completed by KTS on 2/25/26. 2/26/26

☒ **Instrument Complies with Chapter 11D-8, FAC**
☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**

☒ **Return to/Place into Evidentiary Use**
☐ **Remain Out of Evidentiary Use**

☒ **Conduct an Agency Inspection Before Evidentiary Use**

Shayla Platt Digitally signed by Shayla Platt Date: 2026.03.06 14:49:45 -05'00'

Tech Review

Shayla Platt Digitally signed by Shayla Platt Date: 2026.03.06 14:49:58 -05'00'

Admin Review

Stability Checks

0.05g/210L

0.047 to 0.053 ✓

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007496
03/02/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.049	10:31
Air Blank	0.000	10:32
Control Test	0.049	10:32
Air Blank	0.000	10:33
Control Test	0.049	10:34
Air Blank	0.000	10:34
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

0.08g/210L

0.077 to 0.083 ✓

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007496
03/02/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:37
Control Test	0.079	10:37
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:40
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

0.20g/210L

0.194 to 0.206 ✓

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007496
03/02/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:42
Control Test	0.198	10:43
Air Blank	0.000	10:43
Control Test	0.197	10:44
Air Blank	0.000	10:45
Control Test	0.197	10:45
Air Blank	0.000	10:46
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Operator's Signature

DGS 0.08g/210L

0.077 to 0.083 ✓ *≤0.003 of Wet* ✓

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007496
03/02/2026
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:55
Control Test	0.081	10:55
Air Blank	0.000	10:55
Control Test	0.081	10:56
Air Blank	0.000	10:56
Control Test	0.081	10:57
Air Blank	0.000	10:57
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC
Time of Inspection: 09:44 Date of Inspection: 03/02/2026
Serial Number: 80-007496
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: ____

Remarks:
BYPASS AI TO OPERATE INSTRUMENT. COMPLIANCE NOT DETERMINED.

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WEN-CHI K PIERSON
Signature and Printed Name

03/02/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC
Time of Inspection: 12:51 Date of Inspection: 03/02/2026
Serial Number: 80-007496
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28424080A3 Exp: 11/05/2026
0.000	0.048	0.079	0.198	0.082
0.000	0.048	0.078	0.198	0.081
0.000	0.049	0.078	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.048	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.082
0.000	0.049	0.079	0.198	0.081
Standard Deviations	0.0004	0.0004	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

WEN-CHI K PIERSON
Signature and Printed Name

03/02/2026
Date

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007496 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007496
Owning Agency:	FFWCC
Calibration Date:	03/02/2026
Calibration Time:	12:51

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.008
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

03/02/2026
Date

Wen-Chi Pierson
Digitally signed by Wen-Chi Pierson
Date: 2026.03.03 08:56:12 -05'00'

WEN-CHI K PIERSON,
Department Inspector