

Alcohol Testing Program - Instrument Processing Sheet

Agency: Florida Highway Patrol

Instrument Serial Number: 80-003408

Date In: 5/7/2026

DI Completion Date: 6/9/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>TDG</u> Date: <u>5/18/2026</u>		Quality Checks By: <u>TDG</u> Date: <u>6/8/2026</u>		Flow Adjustment By: _____																																																																																
☒ Annual ☐ Dropped Off ☐ Registration ☒ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>202</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.148</u> (.139-.169) 36 mm <u>0.171</u> (.156-.190) 53 mm <u>0.238</u> (.228-.278) 103 mm <u>0.515</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>33363</u> Gauge: <u>1018</u> Instrument: <u>1020</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm (.139-.169) 36 mm (.156-.190) 53 mm (.228-.278) 103 mm (.447-.547)																																																																																
						<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot#/Exp</th> <th style="width:45%;">Maintenance By: _____ Date: _____</th> </tr> <tr> <td>0.050</td> <td>MP6286</td> <td>202406K 6/19/2026</td> <td rowspan="5"> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other: </td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202603B 3/3/2028</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202406N 6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG526603</td> </tr> <tr> <td>9/23/2027</td> </tr> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____	0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:	0.080	MP6287	202603B 3/3/2028	0.200	MP6288	202406N 6/20/2026	0.080 DGS	N/A	AG526603	9/23/2027																																																											
						Simulator	Serial #	Lot#/Exp	Maintenance By: _____ Date: _____																																																																											
						0.050	MP6286	202406K 6/19/2026	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Tank Sensor Tare ☐ Breath Tube Replacement ☐ Other:																																																																											
						0.080	MP6287	202603B 3/3/2028																																																																												
0.200	MP6288	202406N 6/20/2026																																																																																		
0.080 DGS	N/A	AG526603																																																																																		
		9/23/2027																																																																																		
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">Gauge: <u>1020</u> Instrument: <u>1021</u></th> </tr> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> <td>Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u></td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> <td>Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> <td style="background-color: #d3d3d3;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table> </td> </tr> <tr> <td colspan="4" style="background-color: #d3d3d3;"> ☐ Post Optical Bench Adjustment Stability Checks </td> <td>0.080 MP6287</td> </tr> <tr> <td colspan="4" style="background-color: #d3d3d3;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200 MP6288</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td rowspan="4" style="background-color: #d3d3d3;"> Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table> </td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table> </td> </tr> <tr> <td colspan="4"> Gauge ID #: _____ Gauge: _____ Instrument: _____ </td> <td colspan="2" rowspan="3" style="background-color: #d3d3d3; vertical-align: top;"> Notes/Suggested Service: </td> </tr> <tr> <td colspan="4" rowspan="2"> Notes/Suggested Service: </td> </tr> <tr> </tr> <tr> <td colspan="4" rowspan="2" style="background-color: #d3d3d3; vertical-align: top;"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use </td> <td colspan="2" rowspan="2" style="background-color: #d3d3d3; vertical-align: top;"> <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00' </td> <td style="width:50%; vertical-align: top;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table> </td> </tr> </table>		Simulator	Serial #	Lot #	Expiration	Gauge: <u>1020</u> Instrument: <u>1021</u>	0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>	0.040				Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>	0.100				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table>	Simulator	Serial Number	0.000	MP6284	0.200	MP6285	0.300	MP6286	0.080 DGS	N/A	☐ Post Optical Bench Adjustment Stability Checks				0.080 MP6287	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200 MP6288</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td rowspan="4" style="background-color: #d3d3d3;"> Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table> </td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>				Simulator	Serial #	Lot #	Expiration	0.200 MP6288	0.050				Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table>	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:	0.080				0.200				0.080 DGS	N/A			Gauge ID #: _____ Gauge: _____ Instrument: _____				Notes/Suggested Service:		Notes/Suggested Service:				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00' </td> <td style="width:50%; vertical-align: top;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table>		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00'	Tech Review	Admin Review
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1020</u> Instrument: <u>1021</u>																																																																																
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>																																																																																
0.040				Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>																																																																																
0.100				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:50%;">Simulator</th> <th style="width:50%;">Serial Number</th> </tr> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>0.200</td> <td>MP6285</td> </tr> <tr> <td>0.300</td> <td>MP6286</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> </tr> </table>	Simulator	Serial Number	0.000	MP6284	0.200	MP6285	0.300	MP6286	0.080 DGS	N/A																																																																						
Simulator	Serial Number																																																																																			
0.000	MP6284																																																																																			
0.200	MP6285																																																																																			
0.300	MP6286																																																																																			
0.080 DGS	N/A																																																																																			
☐ Post Optical Bench Adjustment Stability Checks				0.080 MP6287																																																																																
<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <th style="width:15%;">Simulator</th> <th style="width:20%;">Serial #</th> <th style="width:20%;">Lot #</th> <th style="width:20%;">Expiration</th> <th style="width:25%;">0.200 MP6288</th> </tr> <tr> <td>0.050</td> <td></td> <td></td> <td></td> <td rowspan="4" style="background-color: #d3d3d3;"> Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table> </td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </table>				Simulator	Serial #	Lot #	Expiration	0.200 MP6288	0.050				Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table>	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:	0.080				0.200				0.080 DGS	N/A																																																											
Simulator	Serial #	Lot #	Expiration	0.200 MP6288																																																																																
0.050				Attachments <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment </td> <td style="width:50%; vertical-align: top;"> ☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other: </td> </tr> </table>	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																																																																														
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment	☐ Post-Stability Checks ☐ Flow Adjustment ☒ Form 40 ☐ Other:																																																																																			
0.080																																																																																				
0.200																																																																																				
0.080 DGS	N/A																																																																																			
Gauge ID #: _____ Gauge: _____ Instrument: _____				Notes/Suggested Service:																																																																																
Notes/Suggested Service:																																																																																				
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%; vertical-align: top;"> LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00' </td> <td style="width:50%; vertical-align: top;"> Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00' </td> </tr> <tr> <td style="text-align: center;">Tech Review</td> <td style="text-align: center;">Admin Review</td> </tr> </table>		LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00'	Tech Review	Admin Review																																																																											
						LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.06.10 08:54:56 -04'00'	Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.06.15 07:51:14 -04'00'																																																																													
Tech Review	Admin Review																																																																																			

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 11:49

Date of Inspection: 06/08/2026

Serial Number: 80-003408
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:


Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/08/2026
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																					
FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.000</td><td>13:41</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:38	Control Test	0.048	13:38	Air Blank	0.000	13:39	Control Test	0.048	13:40	Air Blank	0.000	13:40	Control Test	0.048	13:41	Air Blank	0.000	13:41	Control Test	0.000	13:41	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:33</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.000</td><td>13:34</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:31	Control Test	0.078	13:31	Air Blank	0.000	13:32	Control Test	0.077	13:32	Air Blank	0.000	13:33	Control Test	0.077	13:33	Air Blank	0.000	13:34	Control Test	0.000	13:34	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 06/08/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.000</td><td>13:12</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2906</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	13:08	Control Test	0.198	13:09	Air Blank	0.000	13:10	Control Test	0.199	13:10	Air Blank	0.000	13:11	Control Test	0.199	13:12	Air Blank	0.000	13:12	Control Test	0.000	13:12	Control Test Stats			Average	0.1987		Std Dev	0.0006		Rel Std Dev(%)	0.2906		Printed to external printer. Will attach. ML 6/8/24
Test	g/210L	Time																																																																																																																						
Air Blank	0.000	13:38																																																																																																																						
Control Test	0.048	13:38																																																																																																																						
Air Blank	0.000	13:39																																																																																																																						
Control Test	0.048	13:40																																																																																																																						
Air Blank	0.000	13:40																																																																																																																						
Control Test	0.048	13:41																																																																																																																						
Air Blank	0.000	13:41																																																																																																																						
Control Test	0.000	13:41																																																																																																																						
Control Test Stats																																																																																																																								
Average	0.0480																																																																																																																							
Std Dev	0.0000																																																																																																																							
Rel Std Dev(%)	0.0000																																																																																																																							
Test	g/210L	Time																																																																																																																						
Air Blank	0.000	13:31																																																																																																																						
Control Test	0.078	13:31																																																																																																																						
Air Blank	0.000	13:32																																																																																																																						
Control Test	0.077	13:32																																																																																																																						
Air Blank	0.000	13:33																																																																																																																						
Control Test	0.077	13:33																																																																																																																						
Air Blank	0.000	13:34																																																																																																																						
Control Test	0.000	13:34																																																																																																																						
Control Test Stats																																																																																																																								
Average	0.0773																																																																																																																							
Std Dev	0.0006																																																																																																																							
Rel Std Dev(%)	0.7466																																																																																																																							
Test	g/210L	Time																																																																																																																						
Air Blank	0.000	13:08																																																																																																																						
Control Test	0.198	13:09																																																																																																																						
Air Blank	0.000	13:10																																																																																																																						
Control Test	0.199	13:10																																																																																																																						
Air Blank	0.000	13:11																																																																																																																						
Control Test	0.199	13:12																																																																																																																						
Air Blank	0.000	13:12																																																																																																																						
Control Test	0.000	13:12																																																																																																																						
Control Test Stats																																																																																																																								
Average	0.1987																																																																																																																							
Std Dev	0.0006																																																																																																																							
Rel Std Dev(%)	0.2906																																																																																																																							

FHP TROOP E MIAMI
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003408
06/08/2026
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	12:16
Control Test	0.079	12:16
Air Blank	0.000	12:17
Control Test	0.079	12:17
Air Blank	0.000	12:17
Control Test	0.079	12:18
Air Blank	0.000	12:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MC

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 11:08

Date of Inspection: 06/09/2026

Serial Number: 80-003408
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG526603 Exp: 09/23/2027
0.000	0.048	0.076	0.199	0.080
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.049	0.077	0.199	0.079

Standard Deviations	0.0003	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/09/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003408, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003408</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/09/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:08</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/09/2026

Date

**Taylor
Gutschow**

**TAYLOR D GUTSCHOW,
Department Inspector**

Digitally signed by Taylor
Gutschow
Date: 2026.06.09 11:36:10
-04'00'