

Alcohol Testing Program - Instrument Processing Sheet

Agency: Boynton Beach Police Department Instrument Serial Number: 80-001190  
 Date In: 11/14/2025 DI Completion Date: 2/5/2026  Ship  P/U  H/D  CMI  EE

Intake By: <u>TDG</u> Date: <u>1/16/2026</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	Quality Checks By: <u>TDG</u> Date: <u>2/4/2026</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>194</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.164</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.531</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1022</u> Instrument: <u>1021</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)
--	---	---

Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:
0.050	MP6286	202406K	<input type="checkbox"/> Battery Replacement	
		6/19/2026		
0.080	MP6287	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement	
		6/19/2026		
0.200	MP6288	202406N	<input type="checkbox"/> Tank Sensor Tare	
		6/20/2026		
0.080 DGS	N/A	AG429602	<input type="checkbox"/> Breath Tube Replacement	
		10/22/2026		

**Optical Bench Adjustment** By: \_\_\_\_\_ **Department Inspection** By: TDG

Barometric Pressure Gauge: _____ ID#: _____				Barometric Pressure ID#: <u>33364</u>	
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1015</u>	Instrument: <u>1014</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2024-B</u>	Exp: <u>7/19/2026</u>
0.100				Simulator	Serial Number
0.200				0.000	MP6284
0.300				Interferent	MP6285
0.080 DGS	N/A			0.050	MP6286
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.080	MP6287
Simulator	Serial #	Lot #	Expiration	0.200	MP6288
0.050				<b>Attachments</b>	
0.080				<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks
0.200				<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Adjustment
0.080 DGS	N/A			<input checked="" type="checkbox"/> Calibration Certificate	<input checked="" type="checkbox"/> Form 40
Gauge ID #: _____				<input type="checkbox"/> Optical Bench Adjustment	<input type="checkbox"/> Other:
Gauge: _____ Instrument: _____					

Notes/Suggested Service:

- Instrument Complies with Chapter 11D-8, FAC
- Instrument Does Not Comply with Chapter 11D-8, FAC
- Return to/Place into Evidentiary Use
- Remain Out of Evidentiary Use
- Conduct an Agency Inspection Before Evidentiary Use

LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.02.10 13:23:52 -05'00' Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.02.11 09:30:11 -05'00'

**Tech Review** **Admin Review**

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BOYNTON BEACH PD  
Time of Inspection: 12:30

Date of Inspection: 02/04/2026

Serial Number: 80-001190  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: \_\_\_\_\_

Remarks:  
AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


TAYLOR D GUTSCHOW

---

Signature and Printed Name

02/04/2026  
Date

# Stability Checks

<b>0.05g/210L</b> 0.047 to 0.053	<b>0.08g/210L</b> 0.077 to 0.083	<b>0.20g/210L</b> 0.194 to 0.206	<b>DGS 0.08g/210L</b> 0.077 to 0.083    ✓    ≤0.003 of Wet    ✓																																																																																																																																																
<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000    SN 80-001190 02/04/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:55</td></tr> <tr><td>Average</td><td>0.0453</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	12:51	Control Test	0.050	12:52	Air Blank	0.000	12:53	Control Test	0.049	12:53	Air Blank	0.000	12:54	Control Test	0.049	12:54	Air Blank	0.000	12:55	Control Test	0.049	12:55	Average	0.0453		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000    SN 80-001190 02/04/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:02</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:05</td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:01	Control Test	0.080	13:02	Air Blank	0.000	13:02	Control Test	0.080	13:03	Air Blank	0.000	13:04	Control Test	0.080	13:04	Air Blank	0.000	13:05	Control Test	0.080	13:05	Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000    SN 80-001190 02/04/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.202</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.201</td><td>13:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.202</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test</td><td>0.202</td><td>13:12</td></tr> <tr><td>Average</td><td>0.2017</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2863</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:09	Control Test	0.202	13:09	Air Blank	0.000	13:10	Control Test	0.201	13:10	Air Blank	0.000	13:11	Control Test	0.202	13:12	Air Blank	0.000	13:12	Control Test	0.202	13:12	Average	0.2017		Std Dev	0.0006		Rel Std Dev(%)	0.2863		<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000    SN 80-001190 02/04/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:14</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:16</td></tr> <tr><td>Control Test</td><td>0.0783</td><td>13:16</td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	13:14	Control Test	0.078	13:14	Air Blank	0.000	13:15	Control Test	0.078	13:15	Air Blank	0.000	13:15	Control Test	0.079	13:16	Air Blank	0.000	13:16	Control Test	0.0783	13:16	Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.050	12:52																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.049	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.049	12:54																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.049	12:55																																																																																																																																																	
Average	0.0453																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:01																																																																																																																																																	
Control Test	0.080	13:02																																																																																																																																																	
Air Blank	0.000	13:02																																																																																																																																																	
Control Test	0.080	13:03																																																																																																																																																	
Air Blank	0.000	13:04																																																																																																																																																	
Control Test	0.080	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.080	13:05																																																																																																																																																	
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:09																																																																																																																																																	
Control Test	0.202	13:09																																																																																																																																																	
Air Blank	0.000	13:10																																																																																																																																																	
Control Test	0.201	13:10																																																																																																																																																	
Air Blank	0.000	13:11																																																																																																																																																	
Control Test	0.202	13:12																																																																																																																																																	
Air Blank	0.000	13:12																																																																																																																																																	
Control Test	0.202	13:12																																																																																																																																																	
Average	0.2017																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:14																																																																																																																																																	
Control Test	0.078	13:14																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.078	13:15																																																																																																																																																	
Air Blank	0.000	13:15																																																																																																																																																	
Control Test	0.079	13:16																																																																																																																																																	
Air Blank	0.000	13:16																																																																																																																																																	
Control Test	0.0783	13:16																																																																																																																																																	
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BOYNTON BEACH PD  
Time of Inspection: 12:40

Date of Inspection: 02/05/2026

Serial Number: 80-001190  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.201	0.077
0.000	0.048	0.078	0.201	0.077
0.000	0.048	0.079	0.202	0.077
0.000	0.049	0.078	0.201	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.080	0.201	0.077
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.079	0.201	0.076


Standard Deviations	0.0004	0.0007	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/05/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001190 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001190</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BOYNTON BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/05/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:40</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor  
Gutschow  
Digitally signed by Taylor Gutschow  
Date: 2026.02.05 16:05:44 -05'00'

02/05/2026

Date

TAYLOR D GUTSCHOW,  
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality