



AGENCY INSPECTION AND OTHER ELECTRONIC DATA REVIEW

Agency: FHP Troop F

Instrument Serial Number: 80-006758

AGENCY INSPECTION DISCREPANCY

Agency Inspector: Taylor Ledford

Inspection Date: 07/20/2023

A test was repeated without reason for repeat or the corrective action given

Alcohol Free/Mouth Alcohol Alcohol Free Test Interferent Detect

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Agency Inspection not conducted for calendar month

Expired Alcohol Reference Solution or Dry Gas Standard was used for inspection

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Lot Number and/or Expiration Date for Standards entered incorrectly

0.050 g/210L 0.080 g/210L 0.200 g/210L 0.080 g/210L Dry Gas

Other Electronic Data Review/Comments: The 0.05 lot number is mistyped. A remark describing the corrective action (RFI Detect) on the Alcohol Free Test must be included.

REQUESTED CORRECTIVE ACTION

Record hand-written amendments on FDLE/ATP Form 40, initial and date amendments, mark report "AMENDED", and forward copy to Department Inspector

Provide written explanation regarding referenced item(s) to Department Inspector

Remove instrument from evidentiary use until otherwise directed by the Department

Other:

Taylor Gutschow

Digitally signed by Taylor Gutschow Date: 2023.10.30 13:18:15 -04'00'

10/30/2023

Signature of Alcohol Testing Program Staff Member

Date

"Amended"

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP F
Time of Inspection: 10:00

Date of Inspection: 07/20/2023

Serial Number: 80-006758
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:2023J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402357111 Exp: 02/10/2025
0.000 / 0.000	0.049	0.078	0.198	0.082
RFI / 0.000	0.049	0.078	0.198	0.082
/ 0.000	0.049	0.078	0.198	0.082

Number of Simulators Used: 5

Remarks:

00: RFI Detect. 0.05 lot # mistyped. Correct lot # 202303J. (TL)

Re tested Alcohol free test, after RFI occurred. (TL) 11/6/23
Was repeated.

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name TAYLOR M LEDFORD

07/20/2023
Date