

Instrument Processing Sheet


Agency: DAYTONA BEACH PD Instrument Serial Number: 80-001135
 Date In: 12/1/2025 DI Completion Date: 12/3/2025 Ship P/U H/D CMI EE

Intake By: <u>KTS</u> Date: <u>12/1/25</u>	Quality Checks By: <u>KTS</u> Date: <u>12/3/25</u>	Flow Adjustment By: _____ Date: _____															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>230</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.156</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.515</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1014</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)															
	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 4/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/2026	0.080	MP5089	202406L 6/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	AG510701 4/17/2027	Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:
Simulator	Serial #	Lot#/Exp															
0.050	MP5088	202406K 6/19/2026															
0.080	MP5089	202406L 6/19/2026															
0.200	MP5090	202406N 6/20/2026															
0.080 DGS	N/A	AG510701 4/17/2027															

Optical Bench Adjustment By: _____	Department Inspection By: <u>KTS</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28427</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1015</u> Instrument: <u>1015</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6289																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Barometric Pressure Gauge: _____ ID#: _____	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																				
Simulator	Serial #	Lot #	Expiration																																						
0.050																																									
0.080																																									
0.200																																									
0.080 DGS	N/A																																								

Notes/Suggested Service: Tech Review: Added check mark to return method. KTS 12/12/25	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Shayla Platt Date: 2025.12.14 13:55:18 -05'00' Digitally signed by Wen-Chi Pierson Pierson Date: 2025.12.18 11:03:45 -05'00' Tech Review Admin Review
---	--


Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																																																				
<p>DAYTONA BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001135 12/03/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>07:19</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:20</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:21</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>07:23</td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>	Air Blank	0.000	07:19	Control Test	0.049	07:20	Air Blank	0.000	07:20	Control Test	0.049	07:21	Air Blank	0.000	07:21	Control Test	0.049	07:22	Air Blank	0.000	07:23	Control Test	0.049	07:23	Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>DAYTONA BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001135 12/03/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>07:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:28</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:28</td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Control Test Stats</p> <p>Average 0.2013</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2868</p> <p>Operator's Signature </p>	Air Blank	0.000	07:24	Control Test	0.079	07:24	Air Blank	0.000	07:25	Control Test	0.079	07:26	Air Blank	0.000	07:26	Control Test	0.079	07:27	Air Blank	0.000	07:28	Control Test	0.079	07:28	Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>DAYTONA BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001135 12/03/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>07:29</td></tr> <tr><td>Control Test</td><td>0.202</td><td>07:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:30</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:31</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:32</td></tr> <tr><td>Control Test</td><td>0.201</td><td>07:32</td></tr> <tr><td>Average</td><td>0.2013</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2868</td><td></td></tr> </table> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>	Air Blank	0.000	07:29	Control Test	0.202	07:29	Air Blank	0.000	07:30	Control Test	0.201	07:31	Air Blank	0.000	07:31	Control Test	0.201	07:32	Air Blank	0.000	07:32	Control Test	0.201	07:32	Average	0.2013		Std Dev	0.0006		Rel Std Dev(%)	0.2868		<p>DAYTONA BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-001135 12/03/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>07:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:16</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>07:17</td></tr> <tr><td>Control Test</td><td>0.080</td><td>07:17</td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </table> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>	Air Blank	0.000	07:15	Control Test	0.080	07:15	Air Blank	0.000	07:15	Control Test	0.079	07:16	Air Blank	0.000	07:16	Control Test	0.080	07:16	Air Blank	0.000	07:17	Control Test	0.080	07:17	Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Air Blank	0.000	07:19																																																																																																																																					
Control Test	0.049	07:20																																																																																																																																					
Air Blank	0.000	07:20																																																																																																																																					
Control Test	0.049	07:21																																																																																																																																					
Air Blank	0.000	07:21																																																																																																																																					
Control Test	0.049	07:22																																																																																																																																					
Air Blank	0.000	07:23																																																																																																																																					
Control Test	0.049	07:23																																																																																																																																					
Average	0.0490																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	07:24																																																																																																																																					
Control Test	0.079	07:24																																																																																																																																					
Air Blank	0.000	07:25																																																																																																																																					
Control Test	0.079	07:26																																																																																																																																					
Air Blank	0.000	07:26																																																																																																																																					
Control Test	0.079	07:27																																																																																																																																					
Air Blank	0.000	07:28																																																																																																																																					
Control Test	0.079	07:28																																																																																																																																					
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	07:29																																																																																																																																					
Control Test	0.202	07:29																																																																																																																																					
Air Blank	0.000	07:30																																																																																																																																					
Control Test	0.201	07:31																																																																																																																																					
Air Blank	0.000	07:31																																																																																																																																					
Control Test	0.201	07:32																																																																																																																																					
Air Blank	0.000	07:32																																																																																																																																					
Control Test	0.201	07:32																																																																																																																																					
Average	0.2013																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2868																																																																																																																																						
Air Blank	0.000	07:15																																																																																																																																					
Control Test	0.080	07:15																																																																																																																																					
Air Blank	0.000	07:15																																																																																																																																					
Control Test	0.079	07:16																																																																																																																																					
Air Blank	0.000	07:16																																																																																																																																					
Control Test	0.080	07:16																																																																																																																																					
Air Blank	0.000	07:17																																																																																																																																					
Control Test	0.080	07:17																																																																																																																																					
Average	0.0797																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7247																																																																																																																																						
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																				

Stability checks: Root Cause Analysis performed. Dry Gas Standard wasn't attached during first measurement. KTS 12/3/25

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001135
12/03/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:11
Control Test	0.000	07:12
Air Blank	0.000	07:12
Control Test	0.000	07:12
Air Blank	0.000	07:13
Control Test	0.079	07:13
Air Blank	0.000	07:14
Control Test Stats		
Average	0.0530	
Std Dev	0.0459	
Rel Std Dev(%)	86.6077	



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH PD
Time of Inspection: 09:57

Date of Inspection: 12/03/2025

Serial Number: 80-001135
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.050	0.079	0.201	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.079	0.202	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.079	0.202	0.079

Standard Deviations	0.0004	0.0004	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

12/03/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001135, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001135</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>DAYTONA BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/03/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>09:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


Date 12/03/2025
KATIE TSPEARIN,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality