

Alcohol Testing Program - Instrument Processing Sheet

Agency: FFWCC Instrument Serial Number: 80-000902  
 Date In: 1/29/2026 DI Completion Date: 1/30/2026  Ship  P/U  H/D  CMI  EE





Intake By: <u>KTS</u> Date: <u>1/29/2026</u>	Quality Checks By: <u>KTS</u> Date: <u>1/30/2026</u>	Flow Adjustment By: _____
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes:	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.164</u> (.156-.190) 53 mm <u>0.234</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1015</u> Instrument: <u>1014</u> <input checked="" type="checkbox"/> Stability Checks	Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)

Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:
0.050	MP6291	202406K	<input type="checkbox"/> Battery Replacement	
		6/19/2026		
0.080	MP6292	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement	
		6/19/2026		
0.200	MP6293	202406N	<input type="checkbox"/> Tank Sensor Tare	
		6/20/2026		
0.080 DGS	N/A	AG510701	<input type="checkbox"/> Breath Tube Replacement	
		4/17/2027		

Optical Bench Adjustment	By: _____	Department Inspection	By: <u>KTS</u>																												
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>28427</u>																													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A				Gauge: <u>1014</u> Instrument: <u>1014</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u>	
Simulator	Serial #	Lot #	Expiration																												
0.000		N/A	N/A																												
0.040																															
0.100																															
0.200																															
0.300																															
0.080 DGS	N/A																														
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6294</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																	
Simulator	Serial Number																														
0.000	MP6294																														
Interferent	MP5087																														
0.050	MP5088																														
0.080	MP5089																														
0.200	MP5090																														
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:									
Simulator	Serial #	Lot #	Expiration																												
0.050																															
0.080																															
0.200																															
0.080 DGS	N/A																														
Gauge ID #: _____		Gauge: _____ Instrument: _____																													

Notes/Suggested Service:	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use
	Digitally signed by LeAndra Higginbotham Date: 2026.02.05 10:11:47 <b>LeAndra Higginbotham</b> Tech Review
	Digitally signed by Shayla Platt Date: 2026.02.12 13:41:28 <b>Shayla Platt</b> Admin Review

# Stability Checks

<p>0.050 g/210L 0.047 to 0.053 g/210L</p>	<p>0.080 g/210L 0.077 to 0.083 g/210L</p>	<p>0.200 g/210L 0.194 to 0.206 g/210L</p>	<p>DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet</p>																																																																																																																																																
<p>FFWC Intoxilyzer - Alcohol Analyzer Model 8100 SN 81-005912 01/20/2026 Software: 8101.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:45</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:46</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>06:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:48</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:48</td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	06:45	Control Test	0.048	06:45	Air Blank	0.000	06:46	Control Test	0.048	06:47	Air Blank	0.000	06:47	Control Test	0.049	06:48	Air Blank	0.000	06:48	Control Test	0.048	06:48	Average	0.0483		Std Dev	0.0006		Rel. Std Dev(%)	1.1945		<p>FFWC Intoxilyzer - Alcohol Analyzer Model 8100 SN 81-005912 01/20/2026 Software: 8101.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:52</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:53</td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	06:49	Control Test	0.079	06:50	Air Blank	0.000	06:51	Control Test	0.079	06:51	Air Blank	0.000	06:52	Control Test	0.079	06:53	Air Blank	0.000	06:53	Control Test	0.079	06:53	Average	0.0790		Std Dev	0.0006		Rel. Std Dev(%)	0.0000		<p>FFWC Intoxilyzer - Alcohol Analyzer Model 8100 SN 81-005912 01/20/2026 Software: 8101.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:55</td></tr> <tr><td>Control Test</td><td>0.202</td><td>06:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:56</td></tr> <tr><td>Control Test</td><td>0.201</td><td>06:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:57</td></tr> <tr><td>Control Test</td><td>0.202</td><td>06:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:58</td></tr> <tr><td>Control Test</td><td>0.202</td><td>06:58</td></tr> <tr><td>Average</td><td>0.2017</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2863</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	06:55	Control Test	0.202	06:55	Air Blank	0.000	06:56	Control Test	0.201	06:56	Air Blank	0.000	06:57	Control Test	0.202	06:57	Air Blank	0.000	06:58	Control Test	0.202	06:58	Average	0.2017		Std Dev	0.0006		Rel. Std Dev(%)	0.2863		<p>FFWC Intoxilyzer - Alcohol Analyzer Model 8100 SN 81-005912 01/20/2026 Software: 8101.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:41</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:43</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:44</td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> <p>Operator's Signature: </p>	Test	g/210L	Time	Air Blank	0.000	06:41	Control Test	0.080	06:41	Air Blank	0.000	06:42	Control Test	0.080	06:42	Air Blank	0.000	06:43	Control Test	0.079	06:43	Air Blank	0.000	06:44	Control Test	0.079	06:44	Average	0.0797		Std Dev	0.0006		Rel. Std Dev(%)	0.7247	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:45																																																																																																																																																	
Control Test	0.048	06:45																																																																																																																																																	
Air Blank	0.000	06:46																																																																																																																																																	
Control Test	0.048	06:47																																																																																																																																																	
Air Blank	0.000	06:47																																																																																																																																																	
Control Test	0.049	06:48																																																																																																																																																	
Air Blank	0.000	06:48																																																																																																																																																	
Control Test	0.048	06:48																																																																																																																																																	
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:49																																																																																																																																																	
Control Test	0.079	06:50																																																																																																																																																	
Air Blank	0.000	06:51																																																																																																																																																	
Control Test	0.079	06:51																																																																																																																																																	
Air Blank	0.000	06:52																																																																																																																																																	
Control Test	0.079	06:53																																																																																																																																																	
Air Blank	0.000	06:53																																																																																																																																																	
Control Test	0.079	06:53																																																																																																																																																	
Average	0.0790																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:55																																																																																																																																																	
Control Test	0.202	06:55																																																																																																																																																	
Air Blank	0.000	06:56																																																																																																																																																	
Control Test	0.201	06:56																																																																																																																																																	
Air Blank	0.000	06:57																																																																																																																																																	
Control Test	0.202	06:57																																																																																																																																																	
Air Blank	0.000	06:58																																																																																																																																																	
Control Test	0.202	06:58																																																																																																																																																	
Average	0.2017																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:41																																																																																																																																																	
Control Test	0.080	06:41																																																																																																																																																	
Air Blank	0.000	06:42																																																																																																																																																	
Control Test	0.080	06:42																																																																																																																																																	
Air Blank	0.000	06:43																																																																																																																																																	
Control Test	0.079	06:43																																																																																																																																																	
Air Blank	0.000	06:44																																																																																																																																																	
Control Test	0.079	06:44																																																																																																																																																	
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7247																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FWCC

Time of Inspection: 08:56

Date of Inspection: 01/30/2026

Serial Number: 80-000902

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG510701 Exp: 04/17/2027
0.000	0.048	0.078	0.202	0.079
0.000	0.047	0.079	0.202	0.079
0.000	0.048	0.079	0.201	0.079
0.000	0.048	0.080	0.201	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.202	0.079
0.000	0.048	0.079	0.201	0.080
0.000	0.048	0.079	0.202	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.079	0.201	0.079


Standard Deviations	0.0005	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (  ) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

01/30/2026  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000902, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000902</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/30/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:56</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

\_\_\_\_\_  
Date 01/30/2026  
KATIE T SPEARIN,  
Department Inspector

FDLE/ATP Form 69 January 2026  
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality