

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO	Date 01/11/2026	Type Agency Inspection
Inspector BRENCKLE,TIMOTHY,M	Time 21:25:08	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 02/27/2026	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

INSPECTED AT N JAIL

Data Download Date/Time 01/11/2026 21:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO	Date 02/11/2026	Type Agency Inspection
Inspector SANDERS,JOSHUA,P	Time 08:57:34	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 10124080A2
Expiration 02/27/2026	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 06/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 02/11/2026 09:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO	Date 03/30/2026	Type Agency Inspection
Inspector SANDERS,JOSHUA,P	Time 08:47:13	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 10124080A2
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PASSED INSPECTION AT JAIL

Data Download Date/Time 03/30/2026 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Agency SARASOTA COUNTY SO	Date 04/22/2026	Type Department Inspection
Inspector SPEARIN,KATIE,T	Time 15:31:09	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202603B	0.20 Lot # 202406N	0.08 Gas Lot # 23825080A1
Expiration 06/19/2026	Expiration 03/03/2028	Expiration 06/20/2026	Expiration 10/05/2027

Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.202
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.202
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 04/23/2026 06:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 01/15/2026	Type Department Inspection	
Inspector SPEARIN,KATIE,T	Time 12:27:36	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # AG510701
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 04/17/2027
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.198
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.079
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 01/15/2026 12:29

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 02/08/2026	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 08:24:16	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 231603
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2026
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.037
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.037
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.037
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.

Data Download Date/Time 02/08/2026 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 03/25/2026	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 09:58:17	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 231603
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 07/05/2026
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2026 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

Agency BARTOW PD	Date 04/08/2026	Type Agency Inspection	
Inspector SILLS,MARK,A	Time 14:29:04	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 231603
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 07/05/2026
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/08/2026 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK	Date 01/21/2026	Type Agency Inspection
Inspector SAAVEDRA, OSCAR,	Time 17:43:54	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403056709
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 05/28/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: No Sample Provided.

Data Download Date/Time 01/21/2026 17:42

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK	Date 02/12/2026	Type Agency Inspection
Inspector SAAVEDRA, OSCAR,	Time 17:28:49	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202506E	0.20 Lot # 202410F	0.08 Gas Lot # 403056709
Expiration 06/24/2027	Expiration 06/24/2027	Expiration 10/30/2026	Expiration 05/28/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.200
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	INT
		Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: RFI Detect. 00: RFI Detect.

Data Download Date/Time 02/12/2026 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK	Date 03/19/2026	Type Agency Inspection
Inspector GORNEY, WILLIAM, M	Time 11:40:33	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402987895
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/19/2026 11:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001360

Agency EVERGLADES NAT PARK	Date 04/21/2026	Type Agency Inspection	
Inspector GORNEY, WILLIAM, M	Time 12:24:48	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 40351665
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/16/2028
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2026 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL **Date** 01/30/2026 **Type** Agency Inspection
Inspector HITCHCOCK, WILLIAM, D **Time** 07:15:17 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202510E **0.20 Lot #** 202410F **0.08 Gas Lot #** 403207507
Expiration 06/24/2027 **Expiration** 10/22/2026 **Expiration** 10/30/2026 **Expiration** 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.041
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.050
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.208
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.050
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.05 g/210L Test # (Repeat) 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance.REPEATED OK

Data Download Date/Time 02/02/2026 11:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL	Date 02/28/2026	Type Agency Inspection
Inspector HITCHCOCK, WILLIAM, D	Time 20:13:20	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403207507
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2026 20:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL	Date 03/31/2026	Type Agency Inspection
Inspector HITCHCOCK, WILLIAM, D	Time 23:46:46	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403207507
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/26/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/31/2026 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

Agency FL HIGHWAY PATROL	Date 04/29/2026	Type Agency Inspection
Inspector HITCHCOCK, WILLIAM, D	Time 21:28:18	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 403207507
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 11/26/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2026 21:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO	Date 01/18/2026	Type Agency Inspection
Inspector LAVANCE,ALEYA,D	Time 22:47:59	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/18/2026 22:43

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO	Date 02/22/2026	Type Agency Inspection
Inspector CHANDLER,VANESSA,N	Time 22:53:16	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/22/2026 22:51

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

Agency CHARLOTTE COUNTY SO	Date 03/29/2026	Type Agency Inspection
Inspector LAVANCE,ALEYA,D	Time 18:26:57	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 03/29/2026 18:32

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD	Date 01/28/2026	Type Agency Inspection
Inspector STEMEN, AARON, M	Time 20:02:30	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 01/28/2026 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD	Date 02/25/2026	Type Agency Inspection
Inspector STEMEN, AARON, M	Time 20:07:08	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A1
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/25/2026 20:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD	Date 03/30/2026	Type Agency Inspection
Inspector STEMEN, AARON, M	Time 21:42:33	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A1
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/30/2026 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

Agency GAINESVILLE PD	Date 04/24/2026	Type Agency Inspection
Inspector STEMEN, AARON, M	Time 01:45:21	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A1
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/24/2026 01:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO	Date 03/06/2026	Type Department Inspection	
Inspector PIERSON, WEN-CHI, K	Time 10:53:26	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # 28424080A3
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 11/05/2026
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.082
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0006

Remarks:

Data Download Date/Time 03/06/2026 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

Agency LEVY COUNTY SO	Date 03/12/2026	Type Agency Inspection
Inspector LEVY, JAYSON, I	Time 20:04:32	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 03326080A2
Expiration 06/19/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.077
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.053
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.077
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.077
		0.08 g/210L Test # (Repeat) 3	0.077
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance HOSE CONNECTION.

Data Download Date/Time 03/18/2026 20:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO	Date 01/07/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 10:53:39	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG500903
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 01/09/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.204
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2026 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO **Date** 01/16/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 10:00:37 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0004 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.083
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.206
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.084
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.052
		0.08 g/210L Test # 4	0.084
		0.20 g/210L Test # 4	0.206
		0.08 g/210L Dry Gas Std Test # 4	0.082
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.052
		0.08 g/210L Test # 5	0.084
		0.20 g/210L Test # 5	0.206
		0.08 g/210L Dry Gas Std Test # 5	0.082
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.052
0.08 g/210L Test # 6	0.084
0.20 g/210L Test # 6	0.207
0.08 g/210L Dry Gas Std Test # 6	0.082
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.052
0.08 g/210L Test # 7	0.084
0.20 g/210L Test # 7	0.206
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.053
0.08 g/210L Test # 8	0.084
0.20 g/210L Test # 8	0.206
0.08 g/210L Dry Gas Std Test # 8	0.082
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.053
0.08 g/210L Test # 9	0.085
0.20 g/210L Test # 9	0.206
0.08 g/210L Dry Gas Std Test # 9	0.082
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.053
0.08 g/210L Test # 10	0.085
0.20 g/210L Test # 10	0.206
0.08 g/210L Dry Gas Std Test # 10	0.082
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0006
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/16/2026 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO	Date 03/04/2026	Type Agency Inspection
Inspector DEANE, M,	Time 11:44:34	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG500903
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 01/09/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.082
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2026 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

Agency PINELLAS COUNTY SO	Date 04/01/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 11:03:33	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG500903
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 01/09/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.083
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.205
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.083
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/03/2026 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT	Date 01/21/2026	Type Agency Inspection
Inspector POTTER, MICHAEL, R	Time 14:00:11	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NS MAYPORT PD

Data Download Date/Time 01/21/2026 13:59

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT	Date 02/27/2026	Type Agency Inspection
Inspector POTTER, MICHAEL, R	Time 14:13:00	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.053
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.085
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ MAYPORT PD

Data Download Date/Time 02/27/2026 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT	Date 03/25/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 07:00:01	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT NAS MAYPORT

Data Download Date/Time 03/25/2026 06:58

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

Agency NAS MAYPORT	Date 04/08/2026	Type Agency Inspection
Inspector POTTER, MICHAEL, R	Time 11:44:07	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NS MAYPORT PD

Data Download Date/Time 04/08/2026 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO	Date 01/07/2026	Type Agency Inspection
Inspector DEANE, M,	Time 11:32:59	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG409204
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 04/01/2026

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/07/2026 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO **Date** 01/16/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 09:46:22 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0006 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.050
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.201
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.201
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.052
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0006
0.08 g/210L Test (Std Dev)	0.0008
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0007

Remarks:

Data Download Date/Time 01/16/2026 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO	Date 02/04/2026	Type Agency Inspection
Inspector DEANE, M,	Time 15:08:04	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG409204
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 04/01/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO	Date 03/04/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 13:52:12	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG409204
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 04/01/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2026 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001415

Agency PINELLAS COUNTY SO	Date 04/01/2026	Type Agency Inspection	
Inspector DEANE, M,	Time 14:38:01	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # AG525301
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 09/10/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/03/2026 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO	Date 01/19/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 10:12:46	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 17725080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/19/2026 10:16

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO	Date 02/16/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 10:11:09	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 17725080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2026 15:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO	Date 03/18/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 12:16:51	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 17725080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.202
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/18/2026 12:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001416

Agency ORANGE COUNTY SO	Date 04/08/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 10:38:10	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 17725080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/08/2026 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO	Date 01/19/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 10:07:51	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 00925080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/19/2026 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO	Date 02/16/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 10:15:50	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 00925080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2026 15:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO	Date 03/18/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 12:47:42	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 00925080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/18/2026 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

Agency ORANGE COUNTY SO	Date 04/08/2026	Type Agency Inspection
Inspector MELVILLE,KELLY,M	Time 11:59:32	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 00925080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/08/2026 13:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO	Date 01/12/2026	Type Agency Inspection
Inspector THOMAS,CHRISTOPHER,L	Time 17:24:42	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2026 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO	Date 02/04/2026	Type Agency Inspection
Inspector THOMAS, CHRISTOPHER, L	Time 23:25:54	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/04/2026 23:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO	Date 03/05/2026	Type Agency Inspection
Inspector THOMAS, CHRISTOPHER, L	Time 23:47:04	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/05/2026 23:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001439

Agency MARTIN COUNTY SO	Date 04/06/2026	Type Agency Inspection
Inspector THOMAS, CHRISTOPHER, L	Time 19:46:04	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/06/2026 19:46

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK	Date 02/12/2026	Type Agency Inspection
Inspector SAAVEDRA, OSCAR,	Time 18:10:44	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402987895
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.202
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/12/2026 18:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK	Date 03/19/2026	Type Agency Inspection
Inspector GORNEY, WILLIAM, M	Time 11:55:37	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402987895
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/19/2026 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001442

Agency BISCAYNE NAT PARK	Date 04/21/2026	Type Agency Inspection
Inspector GORNEY, WILLIAM, M	Time 12:29:13	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402987895
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.082
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2026 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO	Date 01/27/2026	Type Agency Inspection	
Inspector SPEARIN,KATIE,T	Time 14:17:59	Software 8100.27	

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 01/30/2026 08:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO **Date** 01/30/2026 **Type** Department Inspection
Inspector SPEARIN,KATIE,T **Time** 08:02:08 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG510701
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 04/17/2027

Number of Simulators 5 **Average Standard Deviation** 0.0005 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.200
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.081
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0007
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 01/30/2026 08:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO	Date 02/20/2026	Type Agency Inspection	
Inspector CAUDLE,PAUL,A	Time 11:53:13	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 251683
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/20/2026 12:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO	Date 03/28/2026	Type Agency Inspection
Inspector HAYES, CLINTON, R	Time 19:41:09	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 251683
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.075
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

08:CONTROL OUTSIDE.

Data Download Date/Time 03/28/2026 19:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

Agency SUMTER COUNTY SO	Date 04/28/2026	Type Agency Inspection
Inspector HOLLIS, TERENCE, L	Time 10:43:26	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 251683
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.075
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.075
		0.08 g/210L Test # (Repeat) 3	0.075
		0.20 g/210L Test # 3	0.193
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 04/28/2026 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD	Date 01/10/2026	Type Agency Inspection
Inspector BARKLEY,JEFFREY,G	Time 14:34:15	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/10/2026 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD	Date 02/16/2026	Type Agency Inspection
Inspector BARKLEY,JEFFREY,G	Time 19:20:04	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2026 19:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD	Date 03/03/2026	Type Agency Inspection
Inspector BARKLEY,JEFFREY,G	Time 17:50:21	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/03/2026 18:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

Agency CAPE CORAL PD	Date 04/11/2026	Type Agency Inspection
Inspector BARKLEY,JEFFREY,G	Time 12:39:27	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/11/2026 13:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO	Date 01/30/2026	Type Agency Inspection
Inspector BISHOP,TREVOR,C	Time 20:36:15	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 25925080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/30/2026 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO	Date 02/13/2026	Type Agency Inspection
Inspector BISHOP,TREVOR,C	Time 19:19:45	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 25925080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2026 19:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO	Date 03/27/2026	Type Agency Inspection	
Inspector BISHOP,TREVOR,C	Time 17:25:11	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 25925080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/27/2026 17:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

Agency PUTNAM COUNTY SO	Date 04/28/2026	Type Agency Inspection
Inspector BISHOP,TREVOR,C	Time 17:29:22	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 25925080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/28/2026 17:27

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency TITUSVILLE PD	Date 03/08/2026	Type Department Inspection
Inspector HIGGINBOTHAM, LEANDRA,	Time 14:00:41	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # 28424080A3
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 11/05/2026

Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.076
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.076
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.075
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.076
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.045
		0.08 g/210L Test # 4	0.075
		0.20 g/210L Test # 4	0.203
		0.08 g/210L Dry Gas Std Test # 4	0.076
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.046
		0.08 g/210L Test # 5	0.075
		0.20 g/210L Test # 5	0.204
		0.08 g/210L Dry Gas Std Test # 5	0.076
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.046
0.08 g/210L Test # 6	0.076
0.20 g/210L Test # 6	0.203
0.08 g/210L Dry Gas Std Test # 6	0.076
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.045
0.08 g/210L Test # 7	0.075
0.20 g/210L Test # 7	0.203
0.08 g/210L Dry Gas Std Test # 7	0.076
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.046
0.08 g/210L Test # 8	0.076
0.20 g/210L Test # 8	0.204
0.08 g/210L Dry Gas Std Test # 8	0.076
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.045
0.08 g/210L Test # 9	0.075
0.20 g/210L Test # 9	0.204
0.08 g/210L Dry Gas Std Test # 9	0.076
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.045
0.08 g/210L Test # 10	0.075
0.20 g/210L Test # 10	0.204
0.08 g/210L Dry Gas Std Test # 10	0.076
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 03/08/2026 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Agency TITUSVILLE PD	Date 03/18/2026	Type Department Inspection	
Inspector HIGGINBOTHAM, LEANDRA,	Time 15:08:33	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # 23825080A1
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 10/05/2027
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.051
		0.08 g/210L Test # 4	0.081
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.051
		0.08 g/210L Test # 5	0.081
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.051
0.08 g/210L Test # 6	0.081
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.051
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.051
0.08 g/210L Test # 8	0.081
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.051
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.051
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/18/2026 15:07

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO	Date 01/23/2026	Type Agency Inspection
Inspector LANDAS,CARTER,	Time 09:34:01	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 01/23/2026 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO	Date 02/22/2026	Type Agency Inspection
Inspector LANDAS, CARTER,	Time 07:36:49	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.199
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.080
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.200
		0.08 g/210L Test # (Repeat) 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: WRONG SIMULATOR ATTACHED.

Data Download Date/Time 02/22/2026 08:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO	Date 03/27/2026	Type Agency Inspection
Inspector LANDAS, CARTER,	Time 19:02:28	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

TESTED AT NJC

Data Download Date/Time 03/27/2026 22:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001651

Agency COLLIER COUNTY SO	Date 04/25/2026	Type Agency Inspection	
Inspector LANDAS, CARTER,	Time 07:46:22	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/25/2026 10:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001652

Agency WAUCHULA PD	Date 03/06/2026	Type Agency Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 13:42:23	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/06/2026 13:39

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001652

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.197
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0005

Remarks:

Data Download Date/Time 03/09/2026 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD	Date 01/21/2026	Type Agency Inspection
Inspector CHOUINARD, MITCHELL, J	Time 08:53:06	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: INSUFFICIENT SAMPLE

Data Download Date/Time 01/21/2026 09:05

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD	Date 02/23/2026	Type Agency Inspection
Inspector CHOUINARD, MITCHELL, J	Time 08:44:47	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	Yes	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Time-Date changed. IN COMPLIANCE.

Data Download Date/Time 02/26/2026 08:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD	Date 03/26/2026	Type Agency Inspection
Inspector WILLMES,AUSTIN,J	Time 14:34:02	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 03/26/2026 14:31

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

Agency ST PETERSBURG PD	Date 04/27/2026	Type Agency Inspection
Inspector WILLMES,AUSTIN,J	Time 13:54:54	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 04125080A5
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 04/27/2026 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO	Date 01/29/2026	Type Agency Inspection	
Inspector MILLS,S,T	Time 19:25:39	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403374313
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/25/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.084
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/29/2026 19:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO	Date 02/27/2026	Type Agency Inspection	
Inspector MILLS,SCOTT,T	Time 20:35:09	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 403374313
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 06/25/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.083
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # (Repeat) 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # (Repeat) 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

00: RFI Detect.

Data Download Date/Time 02/27/2026 20:41

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO	Date 03/25/2026	Type Agency Inspection	
Inspector MILLS,SCOTT,T	Time 20:38:24	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 403374313
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 06/25/2028
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.082
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/25/2026 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

Agency ST JOHNS COUNTY SO	Date 04/29/2026	Type Agency Inspection	
Inspector MILLS,SCOTT,T	Time 20:25:08	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 403374313
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 06/25/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.052
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.052
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/29/2026 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD	Date 01/15/2026	Type Agency Inspection
Inspector DE JESUS, RAY,	Time 18:16:10	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.192
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/15/2026 18:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD	Date 03/05/2026	Type Agency Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 09:53:50	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/05/2026 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD **Date** 03/05/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:35:14 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0003 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.079
		0.20 g/210L Test # 4	0.199
		0.08 g/210L Dry Gas Std Test # 4	0.078
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.079
		0.20 g/210L Test # 5	0.200
		0.08 g/210L Dry Gas Std Test # 5	0.078
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.200
0.08 g/210L Dry Gas Std Test # 6	0.078
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.200
0.08 g/210L Dry Gas Std Test # 7	0.078
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.199
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.200
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.200
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 03/05/2026 13:34

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD	Date 03/15/2026	Type Agency Inspection
Inspector DE JESUS, RAY,	Time 23:15:42	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.199
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.074
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.074
		0.08 g/210L Test # (Repeat) 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance-SIMULATOR WAS COLD.

Data Download Date/Time 03/15/2026 23:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

Agency NORTH MIAMI PD	Date 04/21/2026	Type Agency Inspection	
Inspector DE JESUS, RAY,	Time 19:34:38	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 26824080A1
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 11/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.046
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.046
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/21/2026 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 01/16/2026	Type Agency Inspection	
Inspector LEVY, JAYSON, I	Time 07:11:39	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 01524080A
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2026
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.069
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.048
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.082
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.206
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.079
		0.08 g/210L Test # (Repeat) 3	0.081
		0.20 g/210L Test # 3	0.206
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance SIM LOOSE.

Data Download Date/Time 01/16/2026 07:14

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 02/16/2026	Type Agency Inspection
Inspector LEVY, JAYSON, I	Time 09:44:33	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 10124080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.073
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.083
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.202
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.047
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.084
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # 3	0.082
		0.08 g/210L Test # (Repeat) 3	0.084
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance LOOSE HOSE.

Data Download Date/Time 02/16/2026 09:44

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 03/27/2026	Type Agency Inspection
Inspector LEVY, JAYSON, I	Time 12:25:58	Software 8100.27

0.05 Lot # 202406K	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 03326080A2
Expiration 06/19/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/05/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.071
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.081
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.051
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.061
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.083
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.203
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.089
		0.08 g/210L Test # (Repeat) 3	0.084
		0.20 g/210L Test # 3	0.203
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance SIM ISSUE.

Data Download Date/Time 03/27/2026 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001720

Agency ALACHUA PD	Date 04/01/2026	Type Agency Inspection	
Inspector LEVY, JAYSON, I	Time 10:53:32	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 10124080A2
Expiration 06/19/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2026
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.055
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.051
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.082
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.207
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.055
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.207
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.041
		0.05 g/210L Test # (Repeat) 3	0.055
		0.08 g/210L Test # 3	0.085
		0.20 g/210L Test # 3	0.207
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

05: Control Outside Tolerance UNK WHY LOW READING.

Data Download Date/Time 04/01/2026 12:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD	Date 01/23/2026	Type Agency Inspection
Inspector HODGDON, TODD, S	Time 15:10:15	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 32324080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.074
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.076
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.193
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.046
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.075
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.076
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.08 g/210L Test # (Repeat) 3	0.076
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/23/2026 15:08

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD	Date 02/26/2026	Type Agency Inspection
Inspector LUTZ, EDWIN, J	Time 17:06:12	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 32324080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/26/2026 17:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD	Date 03/28/2026	Type Agency Inspection
Inspector BARNETT,JAMES,C	Time 19:57:45	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 29425080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/28/2026 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001721

Agency PALM BAY PD	Date 04/27/2026	Type Agency Inspection	
Inspector LUTZ, EDWIN, J	Time 00:58:56	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 29425080A3
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # (Repeat) 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # (Repeat) 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

20: RFI Detect WEARING APPLE WATCH, REMOVED.

Data Download Date/Time 04/27/2026 00:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Agency LEE COUNTY SO	Date 01/21/2026	Type Agency Inspection
Inspector BONSALL, MATTHEW, A	Time 14:39:35	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 250284
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.192
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/22/2026 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.198
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 01/23/2026 12:35

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001737

Agency SOUTH MIAMI PD	Date 02/27/2026	Type Agency Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 16:01:20	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/27/2026 15:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO	Date 01/18/2026	Type Agency Inspection
Inspector LAVANCE,ALEYA,D	Time 23:24:53	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 01/18/2026 23:29

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO	Date 02/22/2026	Type Agency Inspection
Inspector CHANDLER,VANESSA,N	Time 21:36:23	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 15025080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 07/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.045
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.194
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.194
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

CLEANED AND INSPECTED

Data Download Date/Time 02/22/2026 21:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

Agency CHARLOTTE COUNTY SO **Date** 03/29/2026 **Type** Agency Inspection
Inspector LAVANCE,ALEYA,D **Time** 19:09:32 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** **0.20 Lot #** **0.08 Gas Lot #**
Expiration 06/24/2027 **Expiration** **Expiration** **Expiration**

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.044
Minimum Sample Volume Check: OK		Interferent Detect Test # 1	INT
Barometric Pressure Sensor Check: OK		Alcohol Free Test # 2	0.000
Alcohol Free Subject Test: 0.000	Yes	0.05 g/210L Test # 2	0.045
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 2	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 3	0.000
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # 3	0.044
Alcohol Free Test: OK	Yes	Interferent Detect Test # 3	INT
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

05: Control Outside Tolerance. Non-compliance:.05 SIMULATOR OUT OF RANGE.

Data Download Date/Time 03/29/2026 19:13

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD	Date 02/04/2026	Type Agency Inspection
Inspector SPEARIN,KATIE,T	Time 08:45:55	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

BYPASSED TO OPERATE INSTRUMENT.COMPLIANCE NOT DETERMINED.

Data Download Date/Time 02/05/2026 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.050
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.080
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.080
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.200
0.08 g/210L Dry Gas Std Test # 8	0.080
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.080
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.080
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0003
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0003

Remarks:

Data Download Date/Time 02/05/2026 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD	Date 03/05/2026	Type Agency Inspection
Inspector YEATES,SCOTT,W	Time 19:32:19	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.052
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.055
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: Interferent Detect.

Data Download Date/Time 03/05/2026 19:49

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD	Date 04/17/2026	Type Agency Inspection
Inspector YEATES,SCOTT,W	Time 18:19:39	Software 8100.27

0.05 Lot # 202506E	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration 06/24/2027	Expiration	Expiration	Expiration

Number of Simulators 5 **Compliance** No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	INT
Date and/or Time Adjusted	No	Alcohol Free Test # (Repeat) 1	0.000
Minimum Sample Volume Check: OK		0.05 g/210L Test # 1	0.042
Barometric Pressure Sensor Check: OK		0.05 g/210L Test # (Repeat) 1	0.048
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.042
Diagnostic Check (Post-Inspection): OK	No	0.05 g/210L Test # (Repeat) 2	INT
Alcohol Free Test: OK	Yes	Interferent Detect Test # 2	INT
0.05 g/210L Test: OK	No	Alcohol Free Test # (Repeat) 3	0.000
0.08 g/210L Test: OK	No	0.05 g/210L Test # 3	0.044
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

00: Interferent Detect. 05: Control Outside Tolerance, Interferent Detect. Non-compliance: .

Data Download Date/Time 04/18/2026 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

Agency CLEARWATER PD	Date 04/17/2026	Type Agency Inspection
Inspector YEATES,SCOTT,W	Time 19:08:32	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 29425080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	INT
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.079
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.196
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.084
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.081
Diagnostic Check (Post-Inspection): OK	Yes	Interferent Detect Test # 1	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 2	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 2	0.052
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.079
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.088
		0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.082
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.045
		0.08 g/210L Test # (Repeat) 3	0.083
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.084
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Interferent Detect. 08: Control Outside Tolerance.

Data Download Date/Time 04/18/2026 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO	Date 01/16/2026	Type Agency Inspection
Inspector KARLECKE,JASON,C	Time 16:19:57	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.082
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/16/2026 16:21

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO **Date** 02/09/2026 **Type** Department Inspection
Inspector GUTSCHOW,TAYLOR,D **Time** 13:59:20 **Software** 8100.27

0.05 Lot # 202406K **0.08 Lot #** 202406L **0.20 Lot #** 202406N **0.08 Gas Lot #** AG526603
Expiration 06/19/2026 **Expiration** 06/19/2026 **Expiration** 06/20/2026 **Expiration** 09/23/2027

Number of Simulators 5 **Average Standard Deviation** 0.0002 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.080
		0.20 g/210L Test # 5	0.199
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.199
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.080
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.080
0.20 g/210L Test # 9	0.199
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.199
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 02/09/2026 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO	Date 02/13/2026	Type Agency Inspection
Inspector KARLECKE,JASON,C	Time 11:54:05	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 32324080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.081
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.082
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/13/2026 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO	Date 03/13/2026	Type Agency Inspection
Inspector KARLECKE, JASON, C	Time 16:10:47	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/13/2026 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

Agency PALM BEACH COUNTY SO	Date 04/17/2026	Type Agency Inspection
Inspector KARLECKE,JASON,C	Time 12:06:19	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 19825080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 09/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.051
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.051
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/17/2026 12:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO	Date 01/12/2026	Type Agency Inspection
Inspector COLEMAN, WILLIAM,	Time 09:25:50	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403419188
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/13/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/12/2026 09:23

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO	Date 02/16/2026	Type Agency Inspection	
Inspector COLEMAN, WILLIAM,	Time 13:54:26	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 403419188
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 08/13/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/16/2026 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO	Date 03/16/2026	Type Agency Inspection
Inspector COLEMAN, WILLIAM,	Time 09:12:40	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 403419188
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 08/13/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/16/2026 09:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Agency MANATEE COUNTY SO	Date 03/30/2026	Type Department Inspection	
Inspector SPEARIN,KATIE,T	Time 12:47:47	Software 8100.27	

0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # 23825080A1
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 10/05/2027

Number of Simulators 5	Average Standard Deviation 0.0001	Compliance Yes
-------------------------------	--	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.196
		0.08 g/210L Dry Gas Std Test # 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.048
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.079
		Interferent Detect Test # 5	INT

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.197
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.078
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.197
0.08 g/210L Dry Gas Std Test # 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.196
0.08 g/210L Dry Gas Std Test # 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.196
0.08 g/210L Dry Gas Std Test # 10	0.079
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0000
0.08 g/210L Test (Std Dev)	0.0000
0.20 g/210L Test (Std Dev)	0.0005
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 03/30/2026 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 01/14/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 13:04:37	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.081
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/14/2026 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 02/19/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 14:09:20	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 23725080A1
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 10/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/19/2026 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 03/24/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 14:09:51	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 36525080A4
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.203
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.201
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT PTFD

Data Download Date/Time 03/24/2026 14:15

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 04/10/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 13:12:48	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 36525080A4
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

IN STANDBY @ PTFD

Data Download Date/Time 04/10/2026 13:24

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 04/13/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 18:01:03	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 36525080A4
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2028

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.081
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.201
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

PLACED INTO EVID USE WITH CLAY COUNTY S/O

Data Download Date/Time 04/13/2026 18:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

Agency JACKSONVILLE SO	Date 04/15/2026	Type Agency Inspection
Inspector TURNER,JOSEPH,B	Time 12:41:25	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 36525080A4
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 02/05/2028

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.200
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.050
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.201
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

ADDED ACETONE SOLUTION.

Data Download Date/Time 04/15/2026 12:47

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

Agency PORT ST LUCIE PD	Date 01/30/2026	Type Department Inspection	
Inspector SPEARIN,KATIE,T	Time 08:22:14	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # AG510701
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 04/17/2027
Number of Simulators 5	Average Standard Deviation 0.0004	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 1	0.079
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.000
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # (Repeat) 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 3	0.079
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.049
		0.08 g/210L Test # 4	0.080
		0.20 g/210L Test # 4	0.198
		0.08 g/210L Dry Gas Std Test # 4	0.000
		0.08 g/210L Dry Gas Std Test # (Repeat) 4	0.079
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.050

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

0.08 g/210L Test # 5	0.080
0.20 g/210L Test # 5	0.198
0.08 g/210L Dry Gas Std Test # 5	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 5	0.079
Interferent Detect Test # 5	INT
Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.080
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.050
0.08 g/210L Test # 7	0.081
0.20 g/210L Test # 7	0.199
0.08 g/210L Dry Gas Std Test # 7	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.050
0.08 g/210L Test # 8	0.080
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 8	0.079
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.050
0.08 g/210L Test # 9	0.081
0.20 g/210L Test # 9	0.198
0.08 g/210L Dry Gas Std Test # 9	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 9	0.079
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.050
0.08 g/210L Test # 10	0.080
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.000
0.08 g/210L Dry Gas Std Test # (Repeat) 10	0.079
Interferent Detect Test # 10	INT

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

**Florida Department of Law Enforcement
Alcohol Testing Program**

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001962

0.05 g/210L Test (Std Dev)	0.0005
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0004
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000
0.08 g/210L Dry Gas Std Test (Std Dev) (Repeat)	0.0003

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/30/2026 08:23

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD	Date 01/06/2026	Type Agency Inspection
Inspector POTTER, MICHAEL, R	Time 12:32:28	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.200
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

IN SERV @ NAS JAX PD

Data Download Date/Time 01/06/2026 13:06

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD	Date 02/26/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 13:33:14	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.193
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT NAS JAX

Data Download Date/Time 02/26/2026 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002097

Agency NAS JACKSONVILLE PD	Date 03/24/2026	Type Agency Inspection
Inspector LIGHTSEY,ROBERT,D	Time 07:02:31	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 13225080A2
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 06/05/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.202
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.200
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

EVID USE AT NAS JAX

Data Download Date/Time 03/24/2026 07:04

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST	Date 01/20/2026	Type Agency Inspection
Inspector ANDERSON,LEE,B	Time 14:01:47	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot #	0.08 Gas Lot #
Expiration 06/24/2027	Expiration 10/22/2026	Expiration	Expiration

Number of Simulators 5 Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		Interferent Detect Test # 1	INT
Alcohol Free Subject Test: 0.000	Yes	Alcohol Free Test # 2	0.000
Mouth Alcohol Test: Slope Not Met	Yes	0.05 g/210L Test # 2	0.048
Interferent Detect Test: Interferent Detect	Yes	0.08 g/210L Test # 2	0.073
Diagnostic Check (Post-Inspection): OK	No	Interferent Detect Test # 2	INT
Alcohol Free Test: OK	Yes	Alcohol Free Test # 3	0.000
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # 3	0.049
0.08 g/210L Test: OK	No	0.08 g/210L Test # 3	0.073
0.20 g/210L Test: OK	No	Interferent Detect Test # 3	INT
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/20/2026 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002170

Agency NAS KEY WEST	Date 01/20/2026	Type Agency Inspection	
Inspector ANDERSON,LEE,B	Time 15:24:58	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403411042
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 08/02/2028
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

TEST 1 EXTERNAL DISRUPTION ELECTRIC/CELL SECOND TEST SUCCES

Data Download Date/Time 01/20/2026 15:42

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD	Date 01/27/2026	Type Agency Inspection
Inspector HUNNEFELD,RALPH,A	Time 08:54:24	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 23725080A1
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 10/05/2027

Number of Simulators 4 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.080
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.195
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.071
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.081
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.071
		0.08 g/210L Test # (Repeat) 3	0.081
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

08: Control Outside Tolerance.

Data Download Date/Time 01/27/2026 08:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD	Date 02/11/2026	Type Agency Inspection
Inspector HUNNEFELD,RALPH,A	Time 07:59:54	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 23725080A1
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 10/05/2027

Number of Simulators 4

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.196
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/12/2026 07:02

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

Agency NEW SMYRNA BEACH PD **Date** 03/28/2026 **Type** Agency Inspection
Inspector HUNNEFELD,RALPH,A **Time** 17:36:30 **Software** 8100.27

0.05 Lot # 202506E **0.08 Lot #** 202510E **0.20 Lot #** **0.08 Gas Lot #**
Expiration 06/24/2027 **Expiration** 10/28/2027 **Expiration** **Expiration**

Number of Simulators 4

Compliance No

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # (Repeat) 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.070
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # (Repeat) 1	0.075
Alcohol Free Subject Test: 0.000	Yes	Interferent Detect Test # 1	INT
Mouth Alcohol Test: Slope Not Met	Yes	Alcohol Free Test # (Repeat) 2	0.000
Interferent Detect Test: Interferent Detect	Yes	0.05 g/210L Test # 2	0.048
Diagnostic Check (Post-Inspection): OK	No	0.08 g/210L Test # 2	0.070
Alcohol Free Test: OK	Yes	0.08 g/210L Test # (Repeat) 2	0.074
0.05 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Test: OK	No	Alcohol Free Test # (Repeat) 3	0.000
0.20 g/210L Test: OK	No	0.05 g/210L Test # 3	0.048
0.08 g/210L Dry Gas Std Test: OK	No	0.08 g/210L Test # 3	0.071
		0.08 g/210L Test # (Repeat) 3	0.073
		Interferent Detect Test # 3	INT

Remarks:

00: Ambient Fail-CLEARED AIR IN ROOM. 08: Control Outside Tolerance - ADJUSTED SIMULATOR SEAL, Co. Non-compliance: CONTROL OUTSIDE TOLERANCE X2 0.08.

Data Download Date/Time 03/29/2026 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD	Date 01/15/2026	Type Agency Inspection
Inspector SOCARRAS,PEDRO,A	Time 11:46:32	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402872586
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 10/12/2026

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI PRICE #958 PRESENT

Data Download Date/Time 01/15/2026 13:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD	Date 02/12/2026	Type Agency Inspection
Inspector SOCARRAS,PEDRO,A	Time 10:48:11	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 402872586
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 10/12/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	INT

Remarks:

AI GUZMAN #846 PRESENT

Data Download Date/Time 02/12/2026 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD	Date 03/26/2026	Type Agency Inspection
Inspector SOCARRAS,PEDRO,A	Time 11:55:40	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 402872586
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 10/12/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.050
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.051
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI GUZMAN #846 PRESENT

Data Download Date/Time 03/26/2026 14:19

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

Agency MIAMI BEACH PD	Date 04/28/2026	Type Agency Inspection
Inspector SOCARRAS,PEDRO,A	Time 11:02:22	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202510F	0.08 Gas Lot # 402872586
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/29/2027	Expiration 10/12/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

AI GUZMAN #846 PRESENT

Data Download Date/Time 04/28/2026 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.048
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.079
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.048
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.079
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.048
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.078
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0003
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 03/20/2026 16:00

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD	Date 01/07/2026	Type Agency Inspection	
Inspector MONTERO,JOSE,	Time 11:17:27	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403255643
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 01/29/2028
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	0.000
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # (Repeat) 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.20 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.08 g/210L Dry Gas Std Test: OK	Yes	Interferent Detect Test # 2	0.000
		Interferent Detect Test # (Repeat) 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.078
		Interferent Detect Test # 3	0.000
		Interferent Detect Test # (Repeat) 3	INT

Remarks:

REDID THE SOLUTION. FDLE TAYLOR ADVISED..

Data Download Date/Time 01/07/2026 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD	Date 02/06/2026	Type Agency Inspection	
Inspector MONTERO,JOSE,	Time 09:51:59	Software 8100.27	
0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403255643
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 01/29/2028
Number of Simulators 5	Compliance Yes		

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.199
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.079
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.080
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.199
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.080
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

A F / M A: TOO MUCH MOUTH ALCOHOL..

Data Download Date/Time 02/06/2026 09:55

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD	Date 03/04/2026	Type Agency Inspection
Inspector MONTERO,JOSE,	Time 09:54:47	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403255643
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 01/29/2028

Number of Simulators 5	Compliance Yes
-------------------------------	-----------------------

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.198
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.079
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.199
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 03/04/2026 09:57

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

Agency HIALEAH PD	Date 04/08/2026	Type Agency Inspection	
Inspector MONTERO,JOSE,	Time 10:07:54	Software 8100.27	
0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403255643
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 01/29/2028
Number of Simulators 5		Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.079
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.197
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.077
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.050
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.198
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.077
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.050
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.198
		0.08 g/210L Dry Gas Std Test # 3	0.077
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 04/08/2026 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Agency HIALEAH PD	Date 02/27/2026	Type Agency Inspection
Inspector GUTSCHOW,TAYLOR,D	Time 09:45:25	Software 8100.27

0.05 Lot #	0.08 Lot #	0.20 Lot #	0.08 Gas Lot #
Expiration	Expiration	Expiration	Expiration

Number of Simulators **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	No		
Date and/or Time Adjusted	No		
Minimum Sample Volume Check: OK			
Barometric Pressure Sensor Check: OK			
Alcohol Free Subject Test: 0.000	No		
Mouth Alcohol Test: Slope Not Met	No		
Interferent Detect Test: Interferent Detect	No		
Diagnostic Check (Post-Inspection): OK	No		
Alcohol Free Test: OK	No		
0.05 g/210L Test: OK	No		
0.08 g/210L Test: OK	No		
0.20 g/210L Test: OK	No		
0.08 g/210L Dry Gas Std Test: OK	No		

Remarks:

AI NOT CONDUCTED. COMPLIANCE NOT DETERMINED.

Data Download Date/Time 03/30/2026 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.047
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.201
0.08 g/210L Dry Gas Std Test # 6	0.077
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.047
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.201
0.08 g/210L Dry Gas Std Test # 7	0.077
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.047
0.08 g/210L Test # 8	0.079
0.20 g/210L Test # 8	0.201
0.08 g/210L Dry Gas Std Test # 8	0.078
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.048
0.08 g/210L Test # 9	0.078
0.20 g/210L Test # 9	0.201
0.08 g/210L Dry Gas Std Test # 9	0.077
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.048
0.08 g/210L Test # 10	0.078
0.20 g/210L Test # 10	0.201
0.08 g/210L Dry Gas Std Test # 10	0.078
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0005
0.20 g/210L Test (Std Dev)	0.0003
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0004

Remarks:

Data Download Date/Time 04/23/2026 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD	Date 01/29/2026	Type Agency Inspection
Inspector OMALLEY,BRIAN,P	Time 11:40:46	Software 8100.27

0.05 Lot # 202402E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403005538
Expiration 02/27/2026	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/25/2027

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.082
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.083
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.083
		Interferent Detect Test # 3	INT

Remarks:

IN COMPLIANCE

Data Download Date/Time 02/02/2026 10:38

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD	Date 02/28/2026	Type Agency Inspection
Inspector HALL,RASHAUN,K	Time 17:23:41	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403005538
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.047
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/28/2026 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002987

Agency MIAMI GARDENS PD	Date 03/30/2026	Type Agency Inspection
Inspector HALL,RASHAUN,K	Time 16:10:43	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202410E	0.20 Lot # 202410F	0.08 Gas Lot # 403005538
Expiration 06/24/2027	Expiration 10/22/2026	Expiration 10/30/2026	Expiration 03/25/2027

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.080
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.195
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.079
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.048
		0.08 g/210L Test # 3	0.079
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT

Remarks:

INCOMPLIANCE

Data Download Date/Time 03/30/2026 16:10

Printed: 30 April 2026

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD	Date 01/21/2026	Type Agency Inspection
Inspector ROY,GARY,A	Time 17:14:01	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.077
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.194
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.049
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 01/21/2026 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD	Date 02/12/2026	Type Agency Inspection
Inspector ROY,GARY,A	Time 15:55:36	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.046
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.075
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.078
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.047
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.076
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.196
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.078
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.047
		0.08 g/210L Test # 3	0.076
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.079
		Interferent Detect Test # 3	INT

Remarks:

Data Download Date/Time 02/12/2026 15:52

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD	Date 03/19/2026	Type Agency Inspection
Inspector ROY,GARY,A	Time 15:06:44	Software 8100.27

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5

Compliance Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK		0.05 g/210L Test # (Repeat) 1	0.049
Barometric Pressure Sensor Check: OK		0.08 g/210L Test # 1	0.078
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # 1	0.194
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.05 g/210L Test # (Repeat) 2	0.049
0.08 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.195
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.05 g/210L Test # (Repeat) 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.195
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

05: RFI Detect.

Data Download Date/Time 03/19/2026 15:03

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD	Date 04/02/2026	Type Department Inspection	
Inspector PIERSON, WEN-CHI, K	Time 12:29:36	Software 8100.27	
0.05 Lot # 202406K	0.08 Lot # 202406L	0.20 Lot # 202406N	0.08 Gas Lot # 28325080A1
Expiration 06/19/2026	Expiration 06/19/2026	Expiration 06/20/2026	Expiration 10/05/2027
Number of Simulators 5	Average Standard Deviation 0.0003	Compliance Yes	

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.048
Minimum Sample Volume Check: OK	Yes	0.08 g/210L Test # 1	0.078
Barometric Pressure Sensor Check: OK	Yes	0.20 g/210L Test # 1	0.196
Alcohol Free Subject Test: 0.000	Yes	0.08 g/210L Dry Gas Std Test # 1	0.081
Mouth Alcohol Test: Slope Not Met	Yes	Interferent Detect Test # 1	INT
Interferent Detect Test: Interferent Detect	Yes	Alcohol Free Test # 2	0.000
Diagnostic Check (Post-Inspection): OK	Yes	0.05 g/210L Test # 2	0.048
Alcohol Free Test: OK	Yes	0.08 g/210L Test # 2	0.078
0.05 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.197
0.08 g/210L Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.081
0.20 g/210L Test: OK	Yes	Interferent Detect Test # 2	INT
0.08 g/210L Dry Gas Std Test: OK	Yes	Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.078
		0.20 g/210L Test # 3	0.197
		0.08 g/210L Dry Gas Std Test # 3	0.081
		Interferent Detect Test # 3	INT
		Alcohol Free Test # 4	0.000
		0.05 g/210L Test # 4	0.048
		0.08 g/210L Test # 4	0.078
		0.20 g/210L Test # 4	0.197
		0.08 g/210L Dry Gas Std Test # 4	0.081
		Interferent Detect Test # 4	INT
		Alcohol Free Test # 5	0.000
		0.05 g/210L Test # 5	0.049
		0.08 g/210L Test # 5	0.078
		0.20 g/210L Test # 5	0.197
		0.08 g/210L Dry Gas Std Test # 5	0.081
		Interferent Detect Test # 5	INT

Florida Department of Law Enforcement
Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Alcohol Free Test # 6	0.000
0.05 g/210L Test # 6	0.049
0.08 g/210L Test # 6	0.078
0.20 g/210L Test # 6	0.198
0.08 g/210L Dry Gas Std Test # 6	0.081
Interferent Detect Test # 6	INT
Alcohol Free Test # 7	0.000
0.05 g/210L Test # 7	0.049
0.08 g/210L Test # 7	0.079
0.20 g/210L Test # 7	0.197
0.08 g/210L Dry Gas Std Test # 7	0.081
Interferent Detect Test # 7	INT
Alcohol Free Test # 8	0.000
0.05 g/210L Test # 8	0.049
0.08 g/210L Test # 8	0.078
0.20 g/210L Test # 8	0.198
0.08 g/210L Dry Gas Std Test # 8	0.081
Interferent Detect Test # 8	INT
Alcohol Free Test # 9	0.000
0.05 g/210L Test # 9	0.049
0.08 g/210L Test # 9	0.079
0.20 g/210L Test # 9	0.197
0.08 g/210L Dry Gas Std Test # 9	0.081
Interferent Detect Test # 9	INT
Alcohol Free Test # 10	0.000
0.05 g/210L Test # 10	0.049
0.08 g/210L Test # 10	0.079
0.20 g/210L Test # 10	0.198
0.08 g/210L Dry Gas Std Test # 10	0.081
Interferent Detect Test # 10	INT
0.05 g/210L Test (Std Dev)	0.0004
0.08 g/210L Test (Std Dev)	0.0004
0.20 g/210L Test (Std Dev)	0.0006
0.08 g/210L Dry Gas Std Test (Std Dev)	0.0000

Remarks:

Data Download Date/Time 04/02/2026 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Inspection Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

Agency TALLAHASSEE PD	Date 04/09/2026	Type Agency Inspection	
Inspector ROY,GARY,A	Time 14:14:09	Software 8100.27	

0.05 Lot # 202506E	0.08 Lot # 202510E	0.20 Lot # 202410F	0.08 Gas Lot # 31124080A2
Expiration 06/24/2027	Expiration 10/28/2027	Expiration 10/30/2026	Expiration 12/05/2026

Number of Simulators 5 **Compliance** Yes

Check or Test	Result	Test	Result
Diagnostic Check (Pre-Inspection): OK	Yes	Alcohol Free Test # 1	0.000
Date and/or Time Adjusted	No	0.05 g/210L Test # 1	0.049
Minimum Sample Volume Check: OK		0.08 g/210L Test # 1	0.076
Barometric Pressure Sensor Check: OK		0.20 g/210L Test # 1	0.212
Alcohol Free Subject Test: 0.000	Yes	0.20 g/210L Test # (Repeat) 1	0.206
Mouth Alcohol Test: Slope Not Met	Yes	0.08 g/210L Dry Gas Std Test # 1	0.080
Interferent Detect Test: Interferent Detect	Yes	Interferent Detect Test # 1	INT
Diagnostic Check (Post-Inspection): OK	Yes	Alcohol Free Test # 2	0.000
Alcohol Free Test: OK	Yes	0.05 g/210L Test # 2	0.049
0.05 g/210L Test: OK	Yes	0.08 g/210L Test # 2	0.077
0.08 g/210L Test: OK	Yes	0.20 g/210L Test # 2	0.210
0.20 g/210L Test: OK	Yes	0.20 g/210L Test # (Repeat) 2	0.205
0.08 g/210L Dry Gas Std Test: OK	Yes	0.08 g/210L Dry Gas Std Test # 2	0.080
		Interferent Detect Test # 2	INT
		Alcohol Free Test # 3	0.000
		0.05 g/210L Test # 3	0.049
		0.08 g/210L Test # 3	0.077
		0.20 g/210L Test # 3	0.208
		0.20 g/210L Test # (Repeat) 3	0.204
		0.08 g/210L Dry Gas Std Test # 3	0.080
		Interferent Detect Test # 3	INT

Remarks:

20: Control Outside Tolerance.

Data Download Date/Time 04/09/2026 14:10