



INSTRUMENT PROCESSING SHEET

Agency Osceola County Corrections

S/N 80-000968

Florida Department of Law Enforcement

Date In 01/24/2025

DI Completion Date 02/04/2025

Ship P/U H/D CMI EE

Intake By ALL Date 01/24/2025 Quality Checks By DA Date 01/29/2025 Flow Calibration By Date. Includes sections for Annual/Registration, Visual Inspection, Other Equipment, Notes, Quality Checks (Breath Tube Screen, R-Value, Flow Verification), Flow Calibration (5L/min, 15L/min, 30L/min), and Maintenance (Battery, Dry Gas Regulator, Breath Tube).

Calibration Adjustment and Department Inspection. Calibration Adjustment includes Barometric Pressure Gauge and Post Calibration Adjustment Stability Checks. Department Inspection includes Barometric Pressure ID# 30793, Gauge 1020, Mouth Alcohol Solution Lot # 2024-A, Acetone Stock Solution Lot # 2024-B, and Attachments (Form 41, Stability Checks, Calibration Certificate, Calibration Adjustment, Post-Stability Checks, Flow Calibration, Form 40, Other).

Notes/Suggested Service and Compliance. Notes/Suggested Service section for handwritten input. Compliance section with checkboxes: Instrument Complies with Chapter 11D-8, FAC; Instrument Does Not Comply with Chapter 11D-8, FAC; Return to/Place into Evidentiary Use; Remain Out of Evidentiary Use; Conduct an Agency Inspection Before Evidentiary Use. Includes digital signatures for Taylor Gutschow and Shayla Platt.

# Stability Checks 80-000968 1/29/25 DA

OSCEOLA CORRECTIONS  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000968  
 01/29/2025  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:40
Control Test	0.049	14:40
Air Blank	0.000	14:41
Control Test	0.048	14:41
Air Blank	0.000	14:42
Control Test	0.048	14:43
Air Blank	0.000	14:43
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	



Operator's Signature

OSCEOLA CORRECTIONS  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000968  
 01/29/2025  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:56
Control Test	0.080	14:57
Air Blank	0.000	14:58
Control Test	0.078	14:58
Air Blank	0.000	14:59
Control Test	0.078	15:00
Air Blank	0.000	15:00
Control Test Stats		
Average	0.0787	
Std Dev	0.0012	
Rel Std Dev(%)	1.4678	

Wet



Operator's Signature

OSCEOLA CORRECTIONS  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000968  
 01/29/2025  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:06
Control Test	0.196	15:07
Air Blank	0.000	15:07
Control Test	0.195	15:08
Air Blank	0.000	15:09
Control Test	0.194	15:09
Air Blank	0.000	15:10
Control Test Stats		
Average	0.1950	
Std Dev	0.0010	
Rel Std Dev(%)	0.5128	



Operator's Signature

OSCEOLA CORRECTIONS  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-000968  
 01/29/2025  
 Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:44
Control Test	0.077	14:45
Air Blank	0.000	14:45
Control Test	0.077	14:46
Air Blank	0.000	14:46
Control Test	0.078	14:46
Air Blank	0.000	14:47
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

DGS



Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA CORRECTIONS

Serial Number: 80-000968

Time of Inspection: 10:07

Date of Inspection: 02/04/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.048	0.077	0.198	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.079	0.197	0.078
0.000	0.048	0.079	0.196	0.078
0.000	0.047	0.079	0.197	0.078

Standard Deviations	0.0005	0.0008	0.0006	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004      Number of Simulators Used: 5

Remarks:

The above instrument complies (  ) does not comply (    ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




DESTINEE N ARMSTRONG

Signature and Printed Name

02/04/2025

Date





Florida Department of Law Enforcement  
 Alcohol Testing Program  
 2331 Phillips Road  
 Tallahassee, FL 32308

# Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000968, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000968</u>	UNCERTAINTY* ±	
Owning Agency:	<u>OSCEOLA CORRECTIONS</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/04/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:07</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/04/2025

Date

DESTINEE N ARMSTRONG,  
 Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality