

Instrument Processing Sheet

Agency: OKALOOSA COUNTY SO Instrument Serial Number: 80-000739
 Date In: 12/1/2025 DI Completion Date: 12/3/2025 Ship P/U H/D CMI EE

Intake By: <u>KTS</u> Date: <u>12/1/25</u>	Quality Checks By: <u>KTS</u> Date: <u>12/3/25</u>	Flow Adjustment By: <u>KTS</u> Date: <u>12/3/25</u>															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Return unworked <input type="checkbox"/> Training Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Shipped in FDLE loaner box. KTS 12/1/2025	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>165</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP103</u> 32 mm <u>0.136</u> (.139-.169) 36 mm <u>0.152</u> (.156-.190) 53 mm <u>0.222</u> (.228-.278) 103 mm <u>0.480</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1014</u> Instrument: <u>1015</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202406K 6/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202406L 6/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> <td>202406N 6/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG510701 4/17/2027</td> </tr> </tbody> </table>	Simulator	Serial #	Lot#/Exp	0.050	MP5088	202406K 6/19/2026	0.080	MP5089	202406L 6/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	AG510701 4/17/2027	Flow Column #: <u>ATP105</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value: <u>167</u> <input checked="" type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: <u>ATP103</u> 32 mm <u>0.144</u> (.139-.169) 36 mm <u>0.160</u> (.156-.190) 53 mm <u>0.242</u> (.228-.278) 103 mm <u>0.484</u> (.447-.547)
Simulator	Serial #	Lot#/Exp															
0.050	MP5088	202406K 6/19/2026															
0.080	MP5089	202406L 6/19/2026															
0.200	MP5090	202406N 6/20/2026															
0.080 DGS	N/A	AG510701 4/17/2027															
		Maintenance By: _____ Date: _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement and Tank Sensor Tare <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other:															

Optical Bench Adjustment By: _____	Department Inspection By: <u>KTS</u>																																								
Barometric Pressure Gauge: _____ ID#: _____	Barometric Pressure ID#: <u>28427</u>																																								
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Optical Bench Adjustment Stability Checks	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge: <u>1015</u> Instrument: <u>1016</u> Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u> Acetone Stock Solution Lot #: <u>2025-B</u> Exp: <u>9/22/2027</u> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5090</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6289	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000		N/A	N/A																																						
0.040																																									
0.100																																									
0.200																																									
0.300																																									
0.080 DGS	N/A																																								
Simulator	Serial Number																																								
0.000	MP6289																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
Barometric Pressure Gauge: _____ ID#: _____	Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Adjustment <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Other:																																								

Notes/Suggested Service: Performed Root Cause Analysis. User error and/or equipment error not found. KTS 12/3/25 Re-printed Department Inspection due to printer issue. KTS 12/3/25 Tech Review: Added check mark to return method. KTS 12/12/25	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <hr/> <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <hr/> <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Shayla Platt Date: 2025.12.14 13:53:08 -05'00' Digitally signed by Wen-Chi Pierson Date: 2025.12.18 10:55:50 -05'00'
	Shayla Platt Tech Review
	Wen-Chi Pierson Admin Review

Stability Checks

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>OKALOOSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000739 12/03/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:12</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:14</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:15</td></tr> <tr><td>Control Test</td><td>0.048</td><td>06:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.000</td><td>06:16</td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Kattyn S</i></p>	Test	g/210L	Time	Air Blank	0.000	06:12	Control Test	0.048	06:13	Air Blank	0.000	06:14	Control Test	0.048	06:14	Air Blank	0.000	06:15	Control Test	0.048	06:16	Air Blank	0.000	06:16	Control Test	0.000	06:16	Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>OKALOOSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000739 12/03/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:18</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:19</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:20</td></tr> <tr><td>Control Test</td><td>0.078</td><td>06:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:21</td></tr> <tr><td>Control Test</td><td>0.000</td><td>06:21</td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Kattyn S</i></p>	Test	g/210L	Time	Air Blank	0.000	06:18	Control Test	0.078	06:18	Air Blank	0.000	06:19	Control Test	0.078	06:20	Air Blank	0.000	06:20	Control Test	0.078	06:21	Air Blank	0.000	06:21	Control Test	0.000	06:21	Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>OKALOOSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000739 12/03/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:23</td></tr> <tr><td>Control Test</td><td>0.196</td><td>06:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:25</td></tr> <tr><td>Control Test</td><td>0.195</td><td>06:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:26</td></tr> <tr><td>Control Test</td><td>0.195</td><td>06:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:27</td></tr> <tr><td>Control Test</td><td>0.000</td><td>06:27</td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2956</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Kattyn S</i></p>	Test	g/210L	Time	Air Blank	0.000	06:23	Control Test	0.196	06:24	Air Blank	0.000	06:25	Control Test	0.195	06:25	Air Blank	0.000	06:26	Control Test	0.195	06:26	Air Blank	0.000	06:27	Control Test	0.000	06:27	Average	0.1953		Std Dev	0.0006		Rel Std Dev(%)	0.2956		<p>OKALOOSA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000739 12/03/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>06:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>06:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:17</td></tr> <tr><td>Control Test</td><td>0.079</td><td>06:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>06:17</td></tr> <tr><td>Control Test</td><td>0.000</td><td>06:17</td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>Kattyn S</i></p>	Test	g/210L	Time	Air Blank	0.000	06:15	Control Test	0.080	06:15	Air Blank	0.000	06:16	Control Test	0.079	06:16	Air Blank	0.000	06:17	Control Test	0.079	06:17	Air Blank	0.000	06:17	Control Test	0.000	06:17	Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:12																																																																																																																																																	
Control Test	0.048	06:13																																																																																																																																																	
Air Blank	0.000	06:14																																																																																																																																																	
Control Test	0.048	06:14																																																																																																																																																	
Air Blank	0.000	06:15																																																																																																																																																	
Control Test	0.048	06:16																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.000	06:16																																																																																																																																																	
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:18																																																																																																																																																	
Control Test	0.078	06:18																																																																																																																																																	
Air Blank	0.000	06:19																																																																																																																																																	
Control Test	0.078	06:20																																																																																																																																																	
Air Blank	0.000	06:20																																																																																																																																																	
Control Test	0.078	06:21																																																																																																																																																	
Air Blank	0.000	06:21																																																																																																																																																	
Control Test	0.000	06:21																																																																																																																																																	
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:23																																																																																																																																																	
Control Test	0.196	06:24																																																																																																																																																	
Air Blank	0.000	06:25																																																																																																																																																	
Control Test	0.195	06:25																																																																																																																																																	
Air Blank	0.000	06:26																																																																																																																																																	
Control Test	0.195	06:26																																																																																																																																																	
Air Blank	0.000	06:27																																																																																																																																																	
Control Test	0.000	06:27																																																																																																																																																	
Average	0.1953																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2956																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	06:15																																																																																																																																																	
Control Test	0.080	06:15																																																																																																																																																	
Air Blank	0.000	06:16																																																																																																																																																	
Control Test	0.079	06:16																																																																																																																																																	
Air Blank	0.000	06:17																																																																																																																																																	
Control Test	0.079	06:17																																																																																																																																																	
Air Blank	0.000	06:17																																																																																																																																																	
Control Test	0.000	06:17																																																																																																																																																	
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Performed Root Case Analysis																																																																																																																																																			

Flow Adjustment. KTS 12/3/25

OKALOOSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000739
12/03/2025
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
 SQRT(Diff)) = 6.855
2: Rate (Liters/min) = 15
 SQRT(Diff)) = 12.000
3: Rate (Liters/min) = 30
 SQRT(Diff)) = 21.562
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 658
Rounded Intercept = -601154
Correlation = 0.99839

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKALOOSA COUNTY SO

Serial Number: 80-000739

Time of Inspection: 08:32

Date of Inspection: 12/03/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG510701 Exp: 04/17/2027
0.000	0.047	0.078	0.194	0.079
0.000	0.047	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.047	0.077	0.196	0.079
0.000	0.047	0.077	0.195	0.078
0.000	0.047	0.077	0.195	0.078
0.000	0.047	0.078	0.195	0.079
0.000	0.047	0.078	0.195	0.079
0.000	0.047	0.077	0.195	0.079

Standard Deviations	0.0004	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




KATIE T SPEARIN

Signature and Printed Name

12/03/2025

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000739, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000739</u>	UNCERTAINTY* ±	
Owning Agency:	<u>OKALOOSA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/03/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>08:32</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/03/2025

Date

KATIE T SPEARIN,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality