

Gutschow, Taylor

From: Gutschow, Taylor
Sent: Friday, July 7, 2023 11:42 AM
To: Pastor, Jose; 'Fernandez, Angel'
Subject: Missed March AI 80-007516
Attachments: Blank Form 40.pdf; Affected Breath Tests 80-007516.pdf

Good morning,

It looks like instrument 7516 was used to conduct breath tests back in March (see Affected Breath Tests attachment) but never received an Agency Inspection for that month. If an Agency Inspection was conducted, please send me the Form 40. If one was not conducted, please fill out the Blank Form 40 that has been attached, include a remark about why the inspection did not occur, and then scan it back to me.

If the inspection did not occur, you should also contact your local State Attorney's Office, provide them with the list of affected breath tests, and let them know that these tests were run on a non-compliant instrument. They can contact me if they have any additional questions.

Taylor Gutschow

Department Inspector for the Alcohol Testing Program
Florida Department of Law Enforcement
Fort Myers Regional Operations Center
4700 Terminal Drive, Suite 3
Fort Myers, FL 33907

Desk #239-335-7500
Cell #850-728-7022

Internal x467500

PLEASE NOTE : Florida has a very broad public records law. Most written communications to or from state officials regarding state business are considered to be public records and will be made available to the public and the media upon request. Your e-mail messages may, therefore, be subject to public disclosure.