

Alcohol Testing Program - Instrument Processing Sheet

Agency: St. Johns County SO

Instrument Serial Number: 80-007366

Date In: 7/29/2026

DI Completion Date: 08/04/2026

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By: SLH Date: 7/30/2026		Quality Checks By: SLH Date: 7/30/2026		Flow Adjustment By:																																					
☒ Annual ☒ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Keyboard ☒ Feet ☒ Ports ☒ Handle ☒ Dry Gas Shelf ☒ Breath Tube ☒ Screws Tight Other Equipment/Accessories ☐ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: 210 ☒ Flow Verification (L/s) Flow Column #: ATP 102 32 mm0.152(.139-.169) 36 mm0.171(.156-.190) 53 mm0.242(.228-.278) 103 mm0.511(.447-.547) ☒ Barometric Pressure Check Gauge ID #: 28427 Gauge: 1008Instrument: 1008 ☒ Stability Checks <table><tr><th>Simulator</th><th>Serial #</th><th>Lot#/Exp</th></tr><tr><td>0.050</td><td>MP5088</td><td>202603A 03/03/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B 03/03/2028</td></tr><tr><td>0.200</td><td>MP5089</td><td>202603D 03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1 10/5/2027</td></tr></table>		Simulator	Serial #	Lot#/Exp	0.050	MP5088	202603A 03/03/2028	0.080	MP5089	202603B 03/03/2028	0.200	MP5089	202603D 03/04/2028	0.080 DGS	N/A	23825080A1 10/5/2027	Flow Column #: ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm R-Value: Post Adjustment Verification (L/S) Flow Column #: 32 mm(.139-.169) 36 mm(.156-.190) 53 mm(.228-.278) 103 mm(.447-.547) <table><tr><th>Maintenance By:</th><th>Date:</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Tank Sensor Tare</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other:</td><td></td></tr></table>		Maintenance By:	Date:	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Tank Sensor Tare		☐ Breath Tube Replacement		☐ Other:										
Simulator	Serial #	Lot#/Exp																																							
0.050	MP5088	202603A 03/03/2028																																							
0.080	MP5089	202603B 03/03/2028																																							
0.200	MP5089	202603D 03/04/2028																																							
0.080 DGS	N/A	23825080A1 10/5/2027																																							
Maintenance By:	Date:																																								
☐ Battery Replacement																																									
☐ Dry Gas Regulator Replacement																																									
☐ Tank Sensor Tare																																									
☐ Breath Tube Replacement																																									
☐ Other:																																									
Optical Bench AdjustmentBy: SLH		Department InspectionBy: SLH																																							
Barometric Pressure Gauge: 1008ID#: 34419		Barometric Pressure ID#: 28427																																							
<table><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>25090</td><td>03/11/2027</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>25020</td><td>01/14/2027</td></tr><tr><td>0.300</td><td>MP6298</td><td>25150</td><td>05/06/2027</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG526603</td><td>09/23/2027</td></tr></table>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	25090	03/11/2027	0.100	MP6296	24390	10/29/2026	0.200	MP6297	25020	01/14/2027	0.300	MP6298	25150	05/06/2027	0.080 DGS	N/A	AG526603	09/23/2027	<table><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></table>	Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																						
0.000	MP6294	N/A	N/A																																						
0.040	MP6295	25090	03/11/2027																																						
0.100	MP6296	24390	10/29/2026																																						
0.200	MP6297	25020	01/14/2027																																						
0.300	MP6298	25150	05/06/2027																																						
0.080 DGS	N/A	AG526603	09/23/2027																																						
Simulator	Serial Number																																								
0.000	MP5086																																								
Interferent	MP5087																																								
0.050	MP5088																																								
0.080	MP5089																																								
0.200	MP5090																																								
☒ Post Optical Bench Adjustment Stability Checks		☒ Form 41																																							
<table><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td>MP5088</td><td>202603A</td><td>03/03/2028</td></tr><tr><td>0.080</td><td>MP5089</td><td>202603B</td><td>03/03/2028</td></tr><tr><td>0.200</td><td>MP5090</td><td>202603D</td><td>03/04/2028</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>23825080A1</td><td>10/5/2027</td></tr></table>	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202603A	03/03/2028	0.080	MP5089	202603B	03/03/2028	0.200	MP5090	202603D	03/04/2028	0.080 DGS	N/A	23825080A1	10/5/2027	☒ Stability Checks	☒ Post-Stability Checks																			
Simulator	Serial #	Lot #	Expiration																																						
0.050	MP5088	202603A	03/03/2028																																						
0.080	MP5089	202603B	03/03/2028																																						
0.200	MP5090	202603D	03/04/2028																																						
0.080 DGS	N/A	23825080A1	10/5/2027																																						
Gauge ID #: 28427		☒ Calibration Certificate	☐ Flow Adjustment																																						
Gauge: 1009Instrument: 1012		☒ Optical Bench Adjustment	☐ Form 40																																						
Notes/Suggested Service:		☒ Other: Calculation Form																																							
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																							
		Wen-Chi Pierson Digitally signed by Wen-Chi Pierson Date: 2026.08.06 15:37:50 -04'00' Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.08.07 06:19:21 -04'00'																																							
		Tech ReviewAdmin Review																																							

Stability Checks

80-00 **7366**

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ±0.003 g/210L of Wet																																																																																																																																																
ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:47	Control Test	0.049	12:47	Air Blank	0.000	12:48	Control Test	0.049	12:48	Air Blank	0.000	12:49	Control Test	0.049	12:50	Air Blank	0.000	12:50	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783 X</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:13	Control Test	0.078	11:14	Air Blank	0.000	11:14	Control Test	0.079	11:15	Air Blank	0.000	11:16	Control Test	0.078	11:16	Air Blank	0.000	11:17	Control Test Stats			Average	0.0783 X		Std Dev	0.0006		Rel Std Dev(%)	0.7370		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:19	Control Test	0.200	11:19	Air Blank	0.000	11:20	Control Test	0.199	11:21	Air Blank	0.000	11:21	Control Test	0.200	11:22	Air Blank	0.000	11:22	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel Std Dev(%)	0.2892		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 07/30/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817 X</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	11:28	Control Test	0.082	11:28	Air Blank	0.000	11:29	Control Test	0.081	11:29	Air Blank	0.000	11:29	Control Test	0.082	11:30	Air Blank	0.000	11:30	Control Test Stats			Average	0.0817 X		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:47																																																																																																																																																	
Control Test	0.049	12:47																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.049	12:48																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.049	12:50																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:13																																																																																																																																																	
Control Test	0.078	11:14																																																																																																																																																	
Air Blank	0.000	11:14																																																																																																																																																	
Control Test	0.079	11:15																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.078	11:16																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783 X																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:19																																																																																																																																																	
Control Test	0.200	11:19																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test	0.199	11:21																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test	0.200	11:22																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2892																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.082	11:28																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.081	11:29																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.082	11:30																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0817 X																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7070																																																																																																																																																		

will need optical bench
adjustment

Verified instrument was
able to read gas pressure. ☒

ST JOHNS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
08/03/2026 08:55:40

SW

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.0710	(0.0000)	0.1410	(-0.0160)	
Sample #2	0.0720	(-0.0170)	0.1330	(-0.0150)	
Sample #3	0.0910	(-0.0120)	0.1110	(0.0040)	
Sample #4	0.0710	(0.0000)	0.1250	(0.0000)	
Avg % Abs	0.0780	(-0.0097)	0.1230	(-0.0037)	
STD DEV	0.0113	(0.0087)	0.0111	(0.0100)	
REL STD DEV	14.448	(90.382)	9.053	(273.181)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.7810	(-0.0020)	1.4340	(0.0140)	
Sample #2	0.7780	(0.0170)	1.4370	(0.0210)	
Sample #3	0.7510	(0.0390)	1.4220	(0.0220)	
Sample #4	0.7640	(0.0420)	1.4560	(0.0220)	
Avg % Abs	0.7643	(0.0327)	1.4383	(0.0217)	
STD DEV	0.0135	(0.0137)	0.0170	(0.0006)	
REL STD DEV	1.767	(41.787)	1.185	(2.665)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.7780	(0.0000)	3.3470	(0.0120)	
Sample #2	1.7810	(0.0000)	3.3480	(0.0110)	
Sample #3	1.7760	(0.0150)	3.3370	(0.0250)	
Sample #4	1.7790	(0.0150)	3.3330	(0.0220)	
Avg % Abs	1.7787	(0.0100)	3.3393	(0.0193)	
STD DEV	0.0025	(0.0087)	0.0078	(0.0074)	
REL STD DEV	0.141	(86.603)	0.233	(38.126)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.4270	(-0.0080)	6.4310	(-0.0010)	
Sample #2	3.4490	(-0.0220)	6.4040	(0.0210)	
Sample #3	3.4370	(-0.0020)	6.4130	(0.0120)	
Sample #4	3.4540	(-0.0030)	6.4310	(0.0110)	
Avg % Abs	3.4467	(-0.0090)	6.4160	(0.0147)	
STD DEV	0.0087	(0.0113)	0.0137	(0.0055)	
REL STD DEV	0.253	(125.216)	0.214	(37.552)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.0420	(-0.0070)	9.3120	(-0.0110)	
Sample #2	5.0090	(0.0160)	9.2770	(0.0370)	
Sample #3	5.0430	(0.0220)	9.2840	(0.0370)	
Sample #4	5.0220	(0.0350)	9.2950	(0.0410)	
Avg % Abs	5.0247	(0.0243)	9.2853	(0.0383)	
STD DEV	0.0172	(0.0097)	0.0091	(0.0023)	
REL STD DEV	0.341	(39.915)	0.098	(6.025)	

ST JOHNS COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-007366
 08/03/2026 08:55:40

Auto Calibration

pg 2 of 2

```

<<<<<      3um      >>>>>
-----
Zero Order Coef   -199.08
First Order Coef  2717.84
Second Order Coef  32.39
  
```

```

<<<<<      9um      >>>>>
-----
                -168.63
                1421.45
                14.46
  
```

```

-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0003
0.040    0.040    0.0002
0.100    0.099    0.0005
0.200    0.201    -0.0006
0.300    0.300    0.0002
-----
  
```

```

-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0001
0.040    0.040    -0.0000
0.100    0.100    0.0005
0.200    0.200    -0.0005
0.300    0.300    0.0002
-----
  
```

```

<<<<<      3um      >>>>>
-----
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1      3539.00
Sample #2      3543.00
Sample #3      3539.00
Sample #4      3529.00
Avg            3537.0000
STD DEV        7.2111
REL STD DEV     0.204
H2O adjust (mg/l*10k) 272

<<<<<      9um      >>>>>
-----
                3290.00
                3254.00
                3262.00
                3238.00
                3251.3333
                12.2202
                0.376
                558
  
```

Barometric Pressure = 1009

*****CALIBRATION SUCCESSFUL*****

SWA

Post Optical Bench Adjustment Stability Checks 80-007366

* 0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	* 0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L ≤0.003 g/210L of Wet																																																																																																												
ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	12:46	Control Test	0.050	12:47	Air Blank	0.000	12:47	Control Test	0.049	12:48	Air Blank	0.000	12:48	Control Test	0.050	12:49	Air Blank	0.000	12:50	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 08/03/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:06	Control Test	0.079	13:07	Air Blank	0.000	13:08	Control Test	0.079	13:08	Air Blank	0.000	13:09	Control Test	0.079	13:09	Air Blank	0.000	13:10	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ST JOHNS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007366 08/04/2026 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:55</td></tr> <tr><td>Control Test</td><td>0.197</td><td>13:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:56</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:57</td></tr> <tr><td>Control Test</td><td>0.196</td><td>13:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2941</td><td></td></tr> </tbody> </table> <i>Miguel</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	13:55	Control Test	0.197	13:56	Air Blank	0.000	13:56	Control Test	0.196	13:57	Air Blank	0.000	13:57	Control Test	0.196	13:58	Air Blank	0.000	13:59	Control Test Stats			Average	0.1963		Std Dev	0.0006		Rel Std Dev(%)	0.2941		see next page
Test	g/210L	Time																																																																																																													
Air Blank	0.000	12:46																																																																																																													
Control Test	0.050	12:47																																																																																																													
Air Blank	0.000	12:47																																																																																																													
Control Test	0.049	12:48																																																																																																													
Air Blank	0.000	12:48																																																																																																													
Control Test	0.050	12:49																																																																																																													
Air Blank	0.000	12:50																																																																																																													
Control Test Stats																																																																																																															
Average	0.0497																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1625																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	13:06																																																																																																													
Control Test	0.079	13:07																																																																																																													
Air Blank	0.000	13:08																																																																																																													
Control Test	0.079	13:08																																																																																																													
Air Blank	0.000	13:09																																																																																																													
Control Test	0.079	13:09																																																																																																													
Air Blank	0.000	13:10																																																																																																													
Control Test Stats																																																																																																															
Average	0.0790																																																																																																														
Std Dev	0.0000																																																																																																														
Rel Std Dev(%)	0.0000																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	13:55																																																																																																													
Control Test	0.197	13:56																																																																																																													
Air Blank	0.000	13:56																																																																																																													
Control Test	0.196	13:57																																																																																																													
Air Blank	0.000	13:57																																																																																																													
Control Test	0.196	13:58																																																																																																													
Air Blank	0.000	13:59																																																																																																													
Control Test Stats																																																																																																															
Average	0.1963																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	0.2941																																																																																																														

* performed 8/4/2026 sut

ST JOHNS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007366
08/03/2026
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	13:02
Control Test	0.081	13:02
Air Blank	0.000	13:03
Control Test	0.081	13:03
Air Blank	0.000	13:03
Control Test	0.081	13:04
Air Blank	0.000	13:04
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ST JOHNS COUNTY SO
Time of Inspection: 16:10

Date of Inspection: 08/04/2026

Serial Number: 80-007366
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.197	0.081
0.000	0.050	0.078	0.197	0.081
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.078	0.197	0.081
0.000	0.049	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.081
0.000	0.049	0.079	0.197	0.081
Standard Deviations	0.0005	0.0004	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

08/04/2026
Date

Florida Department of Law Enforcement
Alcohol Testing Program - Average Calculations for Calibration Certificate

Calculations By: SLH

Serial Number: 80-007366

Date: 8/4/2026

Solution:

0.05 g/210 L Test: 0.494 / 10 = 0.049 g/210L
(SUM) (AVERAGE)

0.08 g/210L Test: 0.788 / 10 = 0.079 g/210L
(SUM) (AVERAGE)

0.20 g/210L Test: 1.970 / 10 = 0.197 g/210L
(SUM) (AVERAGE)

0.08 g/210L Dry Gas Std Test: 0.807 / 10 = 0.081 g/210L
(SUM) (AVERAGE)

Comments:

Average = (Sum of observations) / (Total number of observations)

All results are reported in g/210L



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number **80-007366**, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-007366	Average of $y \pm$ Uncertainty*
Owning Agency:	ST JOHNS COUNTY SO	0.049 g/ 210 L 0.004
Calibration Date:	08/04/2026	0.079 g/ 210 L 0.004
Calibration Time:	16:10	0.197 g/ 210 L 0.008
		0.081 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/04/2026

Date

LeAndra
Higginbotham

Digitally signed by LeAndra
Higginbotham
Date: 2026.08.05 14:42:50 -04'00'

LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 August 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1