

Alcohol Testing Program - Instrument Processing Sheet

Agency: BAKER COUNTY SO

Instrument Serial Number: 80-007439

Date In: 6/25/2026

DI Completion Date: 6/26/2026

☒ Ship

☐ P/U

☐ H/D

☐ CMI

☐ EE

Intake By: <u>WKP</u> Date: <u>6/25/2026</u>		Quality Checks By: <u>WKP</u> Date: <u>6/26/2026</u>		Flow Adjustment By: _____																																						
☒ Annual ☐ Dropped Off ☐ Registration ☐ Return from CMI / EE ☐ Training Instrument Visual Inspection ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/Accessories ☒ Power Cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes:		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value: <u>161</u> ☒ Flow Verification (L/s) Flow Column #: <u>ATP 105</u> 32 mm <u>0.160</u> (.139-.169) 36 mm <u>0.175</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.503</u> (.447-.547) ☒ Barometric Pressure Check Gauge ID #: <u>28427</u> Gauge: <u>1016</u> Instrument: <u>1015</u> ☒ Stability Checks		Flow Column #: _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value: _____ ☐ Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																																						
						<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> <th>Maintenance By:</th> <th>Date:</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5088</td> <td>202603A</td> <td rowspan="2">☐ Battery Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5089</td> <td>202603B</td> <td rowspan="2">☐ Dry Gas Regulator Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>3/3/2028</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5090</td> <td>202603D</td> <td rowspan="2">☐ Tank Sensor Tare</td> <td rowspan="2"></td> </tr> <tr> <td>3/4/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>23825080A1</td> <td rowspan="2">☐ Breath Tube Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>10/5/2027</td> </tr> <tr> <td colspan="2"></td> <td colspan="2"></td> <td>☐ Other:</td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:	0.050	MP5088	202603A	☐ Battery Replacement		3/3/2028	0.080	MP5089	202603B	☐ Dry Gas Regulator Replacement		3/3/2028	0.200	MP5090	202603D	☐ Tank Sensor Tare		3/4/2026	0.080 DGS	N/A	23825080A1	☐ Breath Tube Replacement		10/5/2027					☐ Other:	
						Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:																																
						0.050	MP5088	202603A	☐ Battery Replacement																																	
								3/3/2028																																		
0.080	MP5089	202603B	☐ Dry Gas Regulator Replacement																																							
		3/3/2028																																								
0.200	MP5090	202603D	☐ Tank Sensor Tare																																							
		3/4/2026																																								
0.080 DGS	N/A	23825080A1	☐ Breath Tube Replacement																																							
		10/5/2027																																								
				☐ Other:																																						
Optical Bench Adjustment By: _____		Department Inspection By: <u>WKP</u>																																								
Barometric Pressure Gauge: _____ ID#: _____		Barometric Pressure ID#: <u>34419</u>																																								
Simulator	Serial #	Lot #	Expiration	Gauge: <u>1018</u> Instrument: <u>1016</u>																																						
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u>	Exp: <u>9/25/2027</u>																																					
0.040				Acetone Stock Solution Lot #: <u>2025-B</u>	Exp: <u>9/22/2027</u>																																					
0.100				Simulator	Serial Number																																					
0.200				0.000	MP5086																																					
0.300				Interferent	MP5087																																					
0.080 DGS	N/A			0.050	MP5088																																					
☐ Post Optical Bench Adjustment Stability Checks				0.080	MP5089																																					
Simulator	Serial #	Lot #	Expiration	0.200	MP5090																																					
0.050				Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Optical Bench Adjustment																																						
0.080																																										
0.200																																										
0.080 DGS	N/A																																									
Gauge ID #: _____				☐ Post-Stability Checks																																						
Gauge: _____ Instrument: _____				☐ Flow Adjustment																																						
Notes/Suggested Service: Tech Correction: Added signature to Department Inspection Report. WKP 7/7/2026				☐ Form 40																																						
				☐ Other:																																						
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																						
				☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																						
				☒ Conduct an Agency Inspection Before Evidentiary Use																																						
LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2026.07.07 08:58:10 -04'00'				Kaitlyn Spearin Digitally signed by Kaitlyn Spearin Date: 2026.07.09 10:14:42 -04'00'																																						
				Tech Review Admin Review																																						

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤ 0.003 of Wet ☒
BRKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/26/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:13 Control Test 0.050 10:13 Air Blank 0.000 10:20 Control Test 0.050 10:21 Air Blank 0.000 10:21 Control Test 0.049 10:22 Air Blank 0.000 10:22 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 <i>Wentz</i> Operator's Signature	BRKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/26/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:25 Control Test 0.079 10:25 Air Blank 0.000 10:26 Control Test 0.079 10:27 Air Blank 0.000 10:27 Control Test 0.079 10:28 Air Blank 0.000 10:28 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Wentz</i> Operator's Signature	BRKER COUNTY SC Intoxilyzer - Alcohol Analyzer Model 8000 06/26/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:29 Control Test 0.200 10:30 Air Blank 0.000 10:30 Control Test 0.198 10:31 Air Blank 0.000 10:32 Control Test 0.199 10:32 Air Blank 0.000 10:33 Control Test Stats Average 0.1990 Std Dev 0.0010 Rel Std Dev(%) 0.5025 <i>Wentz</i> Operator's Signature	BRKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/26/2026 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:36 Control Test 0.080 10:36 Air Blank 0.000 10:37 Control Test 0.080 10:37 Air Blank 0.000 10:37 Control Test 0.079 10:37 Air Blank 0.000 10:38 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 <i>Wentz</i> Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 14:42

Date of Inspection: 06/26/2026

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202603A Exp: 03/03/2028	0.08g/210L Test (g/210L) Lot#:202603B Exp: 03/03/2028	0.20g/210L Test (g/210L) Lot#:202603D Exp: 03/04/2028	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23825080A1 Exp: 10/05/2027
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.078	0.198	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.078	0.198	0.078
0.000	0.050	0.078	0.198	0.077
0.000	0.050	0.079	0.198	0.077

Standard Deviations	0.0005	0.0005	0.0004	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. *Timed out, repeated. WCP 6/26/2026*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Wen-Chi Pierson

WEN-CHI K PIERSON

Signature and Printed Name

06/26/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/26/2026</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:42</u>	0.200 g/ 210 L 0.008
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Wen-Chi
Pierson

Digitally signed by Wen-Chi Pierson
Date: 2026.06.30 09:28:05 -04'00'

06/26/2026

Date

WEN-CHI K PIERSON,
Department Inspector

FDLE/ATP Form 69 January 2026

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality