

Alcohol Testing Program - Instrument Processing Sheet

Agency: Collier County Sheriff's Office Instrument Serial Number: 80-001202
 Date In: 2/2/2026 DI Completion Date: 2/27/2026 Ship P/U H/D CMI EE

Intake By: <u>TDG</u> Date: <u>2/3/2026</u> <input checked="" type="checkbox"/> Annual <input checked="" type="checkbox"/> Dropped Off <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE <input type="checkbox"/> Training Instrument Visual Inspection <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/Accessories <input type="checkbox"/> Power Cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: No box	Quality Checks By: <u>TDG</u> Date: <u>2/27/2026</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value: <u>222</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column #: <u>ATP106</u> 32 mm <u>0.152</u> (.139-.169) 36 mm <u>0.167</u> (.156-.190) 53 mm <u>0.246</u> (.228-.278) 103 mm <u>0.500</u> (.447-.547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID #: <u>33364</u> Gauge: <u>1017</u> Instrument: <u>1014</u> <input checked="" type="checkbox"/> Stability Checks	Flow Adjustment By: _____ Flow Column #: _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value: _____ <input type="checkbox"/> Post Adjustment Verification (L/S) Flow Column #: _____ 32 mm _____ (.139-.169) 36 mm _____ (.156-.190) 53 mm _____ (.228-.278) 103 mm _____ (.447-.547)																													
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot#/Exp</th> <th>Maintenance By:</th> <th>Date:</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6286</td> <td>202406K</td> <td rowspan="2"><input type="checkbox"/> Battery Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>6/19/2026</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6287</td> <td>202406L</td> <td rowspan="2"><input type="checkbox"/> Dry Gas Regulator Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>6/19/2026</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6288</td> <td>202406N</td> <td rowspan="2"><input type="checkbox"/> Tank Sensor Tare</td> <td rowspan="2"></td> </tr> <tr> <td>6/20/2026</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG429602</td> <td rowspan="2"><input type="checkbox"/> Breath Tube Replacement</td> <td rowspan="2"></td> </tr> <tr> <td>10/22/2026</td> </tr> </tbody> </table>			Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:	0.050	MP6286	202406K	<input type="checkbox"/> Battery Replacement		6/19/2026	0.080	MP6287	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement		6/19/2026	0.200	MP6288	202406N	<input type="checkbox"/> Tank Sensor Tare		6/20/2026	0.080 DGS	N/A	AG429602	<input type="checkbox"/> Breath Tube Replacement		10/22/2026
Simulator	Serial #	Lot#/Exp	Maintenance By:	Date:																											
0.050	MP6286	202406K	<input type="checkbox"/> Battery Replacement																												
		6/19/2026																													
0.080	MP6287	202406L	<input type="checkbox"/> Dry Gas Regulator Replacement																												
		6/19/2026																													
0.200	MP6288	202406N	<input type="checkbox"/> Tank Sensor Tare																												
		6/20/2026																													
0.080 DGS	N/A	AG429602	<input type="checkbox"/> Breath Tube Replacement																												
		10/22/2026																													

Optical Bench Adjustment By: _____ **Department Inspection** By: TDG
 Barometric Pressure Gauge: _____ ID#: _____ Barometric Pressure ID#: 33364

Simulator	Serial #	Lot #	Expiration	Gauge: <u>1017</u> Instrument: <u>1014</u>
0.000		N/A	N/A	Mouth Alcohol Solution Lot #: <u>2025-D</u> Exp: <u>9/25/2027</u>
0.040				Acetone Stock Solution Lot #: <u>2024-B</u> Exp: <u>7/19/2026</u>
0.100				Simulator
0.200				Serial Number
0.300				0.000 MP6284
0.080 DGS	N/A			Interferent MP6285
<input type="checkbox"/> Post Optical Bench Adjustment Stability Checks				0.050 MP6286
Simulator	Serial #	Lot #	Expiration	0.080 MP6287
0.050				0.200 MP6288

Simulator Serial # Lot # Expiration 0.050 0.080 0.200 0.080 DGS N/A	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Optical Bench Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Adjustment <input type="checkbox"/> Form 40 <input type="checkbox"/> Other:
---	--

Gauge ID #: _____
 Gauge: _____ Instrument: _____

Notes/Suggested Service:

<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use	Digitally signed by Shayla Platt Date: 2026.03.06 14:40:30 -05'00' Digitally signed by Shayla Platt Date: 2026.03.06 14:40:43 -05'00'
---	--

Shayla Platt Shayla Platt
 Tech Review Admin Review

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	≤0.003 of Wet																																																																																																																																															
<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001202 02/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	10:18	Control Test	0.048	10:18	Air Blank	0.000	10:19	Control Test	0.048	10:20	Air Blank	0.000	10:20	Control Test	0.048	10:21	Air Blank	0.000	10:21	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001202 02/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	10:26	Control Test	0.079	10:27	Air Blank	0.000	10:28	Control Test	0.078	10:28	Air Blank	0.000	10:29	Control Test	0.078	10:30	Air Blank	0.000	10:30	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001202 02/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.197</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1977</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2921</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	10:09	Control Test	0.198	10:10	Air Blank	0.000	10:10	Control Test	0.198	10:11	Air Blank	0.000	10:11	Control Test	0.197	10:12	Air Blank	0.000	10:13	Control Test Stats			Average	0.1977		Std Dev	0.0006		Rel Std Dev(%)	0.2921		<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001202 02/27/2026 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	10:33	Control Test	0.078	10:33	Air Blank	0.000	10:34	Control Test	0.078	10:34	Air Blank	0.000	10:35	Control Test	0.079	10:35	Air Blank	0.000	10:35	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:18																																																																																																																																																	
Control Test	0.048	10:18																																																																																																																																																	
Air Blank	0.000	10:19																																																																																																																																																	
Control Test	0.048	10:20																																																																																																																																																	
Air Blank	0.000	10:20																																																																																																																																																	
Control Test	0.048	10:21																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.079	10:27																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.078	10:28																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test	0.078	10:30																																																																																																																																																	
Air Blank	0.000	10:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.198	10:10																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.198	10:11																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.197	10:12																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1977																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2921																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.078	10:33																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test	0.078	10:34																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test	0.079	10:35																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 12:30

Date of Inspection: 02/27/2026

Serial Number: 80-001202
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.079	0.198	0.079
0.000	0.048	0.078	0.197	0.079

Standard Deviations	0.0000	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/27/2026
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001202, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001202</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/27/2026</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:30</u>	0.200 g/ 210 L	0.008
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2026.03.04 11:10:18
-05'00'

02/27/2026 Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 January 2026
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality